



Universiteit
Leiden
The Netherlands

Detection of single cell microbial life artificially inoculated in mudstone analogue material using a miniature LIMS system

Riedo, A.; Koning, C. de; Stevens, A.H.; López, A.C.; Tulej, M.; Wurz, P.; ... ; Ehrenfreund, P.

Citation

Riedo, A., Koning, C. de, Stevens, A. H., López, A. C., Tulej, M., Wurz, P., ... Ehrenfreund, P. (2019). Detection of single cell microbial life artificially inoculated in mudstone analogue material using a miniature LIMS system. *Egu General Assembly Conference Abstracts*, 16763. Retrieved from <https://hdl.handle.net/1887/85374>

Version: Publisher's Version

License: [Creative Commons CC BY 4.0 license](https://creativecommons.org/licenses/by/4.0/)

Downloaded from: <https://hdl.handle.net/1887/85374>

Note: To cite this publication please use the final published version (if applicable).



Detection of single cell microbial life artificially inoculated in mudstone analogue material using a miniature LIMS system

Andreas Riedo (1,2), Coen de Koning (1), Adam H. Stevens (3), Alena Cedeño López (4), Marek Tulej (2), Peter Wurz (2), Charles S. Cockell (3), and Pascale Ehrenfreund (1)

(1) Sackler Laboratory for Astrophysics, Leiden Observatory, Leiden University, Leiden, Netherlands (riedo@strw.leidenuniv.nl), (2) Space Research and Planetary Sciences, Physics Institute, University of Bern, Bern, Switzerland, (3) School of Physics and Astronomy, UK Centre for Astrobiology, University of Edinburgh, Edinburgh, United Kingdom, (4) Department of Chemistry and Biochemistry, University of Bern, Bern, Switzerland

In situ detection and identification of signatures of life on Mars, past or present, is one of the most challenging and important tasks in current space science and astrobiology. However, the detection is extremely challenging and depends on numerous parameters, starting from the selection of field site for material sampling to the application of appropriate instrumentation for analysis. From previous and ongoing space exploration missions on Mars we know that for a positive localisation and identification of potentially available signatures of life a combined and complementary instrumentation effort is required. In this contribution, we discuss the latest results towards detection and identification of the biogenic chemical fingerprint of single cell microbial life, artificially inoculated in Mars analogue mudstone material using our miniature high-sensitivity Laser Ablation Ionisation Mass Spectrometer (LIMS).

The LIMS system employed in this study consists of a miniature reflectron-type time-of-flight mass analyser (160 mm x Ø 60 mm) that is coupled to a femtosecond laser system ($\lambda = 775$ nm, $\tau \sim 190$ fs) used to induce the ablation and ionisation of sample material [1]. For the localisation of microbial life structures, the mudstone material was spot-wise chemically analysed in depth and monitoring biogenic elements, such as carbon, phosphorus, sulphur [2, 3].

The host matrix was artificially produced using commercially available minerals that resemble the chemical composition of Martian mudstone [4]. A total of 14 mudstone samples were chemically investigated of which only one half contained a low number density of artificially inoculated microbes. The individual samples were treated differently, from anaerobic to aerobic, wet to dry, irradiated to un-irradiated using a gamma-ray source. The measurements described in this contribution show that the different sample treatments did not result in the detection of significant elemental differences between samples. However, the monitoring of major and minor biogenic elements, such as e.g., carbon, sulphur allows for localisation of the biogenic element fingerprint of single cell microbes. The presented studies are of high interest for future high-resolution instrumentation concepts for in situ detection of signs of life.

References

- [1] A. Riedo, M. Neuland, S. Meyer, M. Tulej and P. Wurz, "Coupling of LMS with a fs-laser ablation ion source: elemental and isotope composition measurements", *J. Anal. At. Spectrom.*, 28, 2013, 1256 – 1269.
- [2] M. Tulej, A. Neubeck, M. Ivarsson, A. Riedo, M.B. Neuland, S. Meyer and P. Wurz, "Chemical composition of micrometer-sized filaments in an aragonite host by a miniature laser ablation/ionization mass spectrometer", *Astrobiol.*, 15, 2015, 669 – 682.
- [3] V. Grimaudo, P. Moreno-García, A. Riedo, M.B. Neuland, M. Tulej, P. Broekmann and P. Wurz, "High-resolution chemical depth profiling of solid material using a miniature laser ablation/ionization mass spectrometer", *Anal. Chem.*, 87, 2015, 2037 - 2041.
- [4] A.H. Stevens, E. Steer, A. McDonald, E.S. Amador, and C.S. Cockell, "Y-Mars: An Astrobiological Analogue of Martian Mudstone", *Earth Space Sci.*, 5, 2018. 163-174.