

Measuring local pH in electrochemistry

Monteiro M.C., Koper M.T.M.

Citation

Monteiro M.C., K. M. T. M. (2020). Measuring local pH in electrochemistry. *Current Opinion In Electrochemistry*. doi:10.1016/j.coelec.2020.100649

| Version: | Publisher's Version |
|------------------|------------------------------------|
| License: | Creative Commons CC BY 4.0 license |
| Downloaded from: | https://hdl.handle.net/1887/138242 |

Note: To cite this publication please use the final published version (if applicable).

Journal Pre-proof

Measuring local pH in electrochemistry

Mariana C.O. Monteiro, Marc T.M. Koper

PII: S2451-9103(20)30190-3

DOI: https://doi.org/10.1016/j.coelec.2020.100649

Reference: COELEC 100649

To appear in: Current Opinion in Electrochemistry

Received Date: 21 September 2020

Accepted Date: 23 October 2020

Please cite this article as: Monteiro MCO, Koper MTM, Measuring local pH in electrochemistry, *Current Opinion in Electrochemistry*, https://doi.org/10.1016/j.coelec.2020.100649.

This is a PDF file of an article that has undergone enhancements after acceptance, such as the addition of a cover page and metadata, and formatting for readability, but it is not yet the definitive version of record. This version will undergo additional copyediting, typesetting and review before it is published in its final form, but we are providing this version to give early visibility of the article. Please note that, during the production process, errors may be discovered which could affect the content, and all legal disclaimers that apply to the journal pertain.

© 2020 Elsevier B.V.



Measuring local pH in electrochemistry

Mariana C. O. Monteiro, Marc T. M. Koper*

Leiden Institute of Chemistry, Leiden University, P.O. Box 9502, 2300 RA Leiden, The Netherlands *Corresponding author: m.koper@lic.leidenuniv.nl

Abstract

Localized pH measurements are important in various areas of electrochemistry, from corrosion to bio-electrochemistry and electrocatalysis. Different techniques are available to perform these measurements and offer numerous possibilities in terms of spatial and temporal resolution, sensibility, and precision. In this brief review we present the recent progress made and summarize the main techniques available for localized pH measurements in electrochemistry such as scanning probe techniques (SECM, SICM, SIET), laser (confocal) fluorescence microscopy, rotating ring-disc electrode (RRDE) voltammetry, and infra-red spectroscopy, among others.

Keywords: local pH, electrochemistry, SECM, fluorescence, RRDE

1. Introduction

The concept of pH as we know it today dates back to 1909 when Sørensen introduced the pH scale as a way to express the concentration of hydrogen ions.[1] In electrochemistry, the local concentration of protons in solution plays a significant role in reactions of different fields and is a function of the substrate geometry, current density, mass transport, and the buffer capacity of the electrolyte. Corrosion via anodic oxidation is usually associated with metal dissolution and hydrolysis, thus locally producing protons.[2] Various biological processes cause intra- or extracellular pH changes and the quantification of these pH changes can help to understand the associated mechanisms .[3] Numerous electrocatalytic processes such as O2 evolution or CO2 reduction, that consume or produce H⁺ or OH⁻, generate a pH gradient between the working electrode (substrate) and the bulk of the electrolyte.[4] In all these applications, measuring the pH locally and with high sensitivity is desired to better understand the electrochemical processes involved.

Shortly after the invention of the glass pH electrode, scientists started putting efforts towards its miniaturization to perform what one could consider the first localized pH measurement ever reported: pH of arterial blood flowing through a small cannula in the cortex of a monkey.[5] It was already concluded at that time, that producing miniaturized pH probes "is quite possible, with practice". Since then, electrochemists have invested significant efforts to develop not only new probes but also new techniques to measure pH locally with high spatial and temporal resolution. Here, we give an overview of those developments, together with a brief discussion of their advantages, drawbacks, and applicability. The techniques can be divided into direct and indirect methods to measure pH, namely techniques in which the signal monitored is the proton concentration by measuring its electrochemical potential (scanning probe microscopy, rotating ring-disc electrode) or techniques that probe species, whose signal is a function of the proton concentration (fluorescence and infrared intensities). Localized pH measurements can help to elucidate reaction mechanisms and assist in the deconvolution of pH effects from other electrolyte or structural surface effects. The rate at which the pH changes can also help understanding the stability of a certain substrate, to probe mass transport phenomena or reactivity. Spatial resolution, temporal resolution, and sensitivity

will dictate which of these system properties can be captured by the different methods.

2. Scanning probe techniques: SECM, SICM, and SIET

Scanning probe microscopy (SPM) techniques used for pH measurements usually consist of monitoring the electrochemical signal of a miniaturized tip (ultramicroelectrode or nanoelectrode) while the tip is moved above a substrate. High spatial resolution can be achieved in three dimensions (XYZ) and is limited by the size and geometry of the tip, taking into account the radial diffusion profile towards ultramicroelectrodes (UMEs). The spatial resolution can be improved and the shielding can be minimized by miniaturizing the pH sensor and the insulating layer around it. Temporal resolution, precision, selectivity, and stability, on the other hand, are highly dependent on the nature of the pH sensor. In the next sections we discuss the main SPM techniques used to probe pH locally in electrochemistry, namely: Scanning Electrochemical Microscopy (SECM), Scanning Ionselective Electrode Technique (SIET) and Scanning Ion Conductive Microscopy (SICM).

2.1. SECM

Among the different SPM techniques, SECM is the most employed to probe local pH. Normally, the tip is made of an inert or metallic material with a pH-sensitive electrochemical response, surrounded by an insulator. A schematic representation of an SECM setup is shown in Figure 1a. pH measurements have been performed using either potentiometric or voltammetric sensors. Potentiometric pH probes consist of substrates whose open circuit potential shows a Nernstian (59 mV/pH) or super-Nernstian (> 59 mV/pH) shift as a function of pH. They are mainly based on metals, metal/metal oxides (TiO₂, RuO₂, RhO₂, Ta₂O₅, IrO₂, PtO₂, and ZrO₂)[6] or electrodeposited polymer films.[7] Among these, IrO₂ is the most popular substrate. Tefashe et al. used a Pt/IrO₂ probe to measure pH during localized corrosion of AZ31B magnesium alloy electrochemically coated with poly(3,4ethylenedioxythiophene) (PEDOT).[8] The pH map obtained after 30 min immersion in 0.01 M NaCl can be seen in Figure 1b, recorded with the pH probe positioned 10 µm from the surface. Other metal supports can also be used, as seen in the work of Santos and co-workers[9], where pH above a copper surface during nitrate reduction was measured with a Au/IrO₂ probe. Polymer films have also been used as potentiometric sensors. The most popular is polyaniline (PANI) due to its high conductivity and ease of synthesis. Recently Song et al. reported the use of a PANI coated Pt ultramicroelectrode to monitor the extracellular pH of MCF-7 cells under electrical stimulation.[10]

The main advantage of potentiometric pH sensors, in general, is their easy manufacturing, which facilitates their miniaturization. However, the fact that the pH response comes from the interaction of protons in solution with a solid-state film, means that potentiometric pH sensors have a relatively long response time of usually a few seconds, which can change as a function of pH.[11] The overall performance is also highly affected by the quality and thickness of the film. The synthesis of potentiometric pH sensors usually does not involve complicated steps and is often done via electrodeposition. Although there is an extensive list of synthesis protocols reported in literature[12], the



Figure 1. a) Schematic representation of a scanning probe setup, b) pH map of uncoated and PEDOT-coated AZ31B Mg alloy after 30 minutes in 0.01 M NaCl, measured with a Pt/IrOx microprobe operated in constant height potentiometric mode of SECM. Adapted from [8]. c) pH map during oxygen reduction on a 10 μ m diameter Pt disc recorded with a voltammetric pH nanosensor made of a syringaldazine polymeric film. Adapted with permission from [15]. Copyright (2015) American Chemical Society. d) Image of the shifts in the reduction peak potential of a Pt nanoelectrode during oxygen reduction at a Ag-based gas diffusion electrode. Each grid point was measured at 1 μ m tip-surface distance utilizing shear force constant distance mode SECM. Adapted with permission from [14]. Copyright (2018) Wiley- VCH Verlag GmbH & Co. KGaA. e) Buffering effect during CO₂ reduction measured with a voltammetric pH sensor based on a gold UME functionalized with a 4-NTP self-assembled monolayer.(M.C.O. Monteiro et al. unpublished). f) Z-directional pH profile recorded with a glass capillary filled with an ion-selective cocktail, from 50 μ m above the S. mutans biofilm to 1000 μ m in the bulk solution after addition of 30 mM sucrose in artificial saliva (pH 6.0). Adapted with permission from [23]. Copyright (2017) American Chemical Society.

success rate, even when following a well-established recipe, is known to be low. Potential drift can also prevent direct pH determination from the initial calibration curve and often requires recalibration.[13] Besides, the potential gradient across the interface being studied can potentially affect the open circuit potential recorded at the tip. The mentioned drawbacks can lead to strong deviations in the measurements (especially when pH maps are being constructed) and require a systematic data processing protocol.

Alternatively, SECM voltammetric pH sensors can be employed. They are based on the current-potential response of a certain redox reaction (upon cycling) and the pH is determined by the Nernstian shift of the reduction or oxidation midpeak potential. Voltammetric probes have the major advantage that the response time is only dependent on the time required to record a cyclic voltammogram and can easily be tuned. Botz et al. used the PtO reduction voltammetry to successfully probe pH during oxygen reduction (ORR) over a silver gas diffusion electrode as can be seen in Figure 1d.[14] Although a Pt-UME is suitable to probe the high alkalinity developed during ORR, unfortunately, it cannot be easily applied to other systems, as the Pt response is not exclusively selective to pH. Michalak et al. have also probed pH during ORR using a voltammetric pH nanosensor based on a syringaldazine film. The pH map obtained above a Pt-UME is shown in Figure 1c.[15] The same type of sensor was also successfully applied to monitor the extracellular pH of adherently mammalian cells.[16] As syringaldazine growing is electropolymerized on a substrate, the influence of film thickness and stability on the pH response is still a concern. We have recently presented a pH sensor that overcomes film limitations, based on a self-assembled monolayer of 4-nitrothiophenol (4-NTP) on a Au-UME.[17] The sensor is extremely sensitive, capturing differences as small as 0.1 pH units and is stable and selective in diverse electrolytes and under different gaseous atmospheres. This enables, for instance, pH measurements during CO₂ reduction on gold using SECM, as shown in Figure 1e (M.C.O. Monteiro, unpublished).

For mapping pH with high resolution above a substrate, pH sensors must be positioned accurately in solution at a controlled distance from the surface of interest. Commonly used SECM approach techniques need a mediator or a diffusion-limited reaction taking place at the tip to determine the tip-to-sample distance. However, not all pH sensors allow for diffusioncontrolled feedback, as not always current can be drawn without destabilizing the tip response. The development of doublebarrelled UMEs or dual probes overcomes positioning issues allowing for diffusion-controlled feedback, using the amperometric side of the tip, while carrying out pH measurements with the other side.[18,19] Other (more refined) distance control feedback systems employ shear-force[20], introduced by the group of Schuhmann, or alternating current-SECM (AC-SECM)[21]. We have also recently presented the application of a capacitive approach, performed in air, to determine the tip-tosurface distance without destabilizing the pH sensor.[17] The intricacies and applicability of commonly used distance control methods are well discussed in the book of Bard and Mirkin.[22]

2.2. Scanning Ion-selective Electrode (SIET)

SIET is a technique very similar to SECM, albeit employing potentiometric pH sensors that consist of a liquid ion-selective membrane enclosed in a (pulled) micro/nanopipette. Different proton selective cocktails have been reported and many are commercially available (Fluka, Selectophore[®]). Joshi et al. reported the use of an H⁺ selective cocktail to map the pH of the microenvironment produced by the lactate-producing S. mutans biofilm. Figure 1f shows a profile in z-direction of the pH monitored 950 µm across the diffusion layer.[23] The dynamic pH range of these probes is usually limited so they are mostly used for measurements in the physiologically relevant pH range, or in corrosion processes where the pH changes are not too drastic.[24–26] The main drawbacks here are similar to those discussed for the potentiometric SECM probes, i.e. slow time response (limiting the scanning speed that can be used when



Figure 2. a) Schematic representation of the operation of double-barrel nanopipettes for simultaneous SICM imaging and pH measurement and b) fluorescence and SICM imaging of a group of low-buffered CD44^{GFP-high} breast cancer MCF7 cells in estradiol-deprived medium (–E2), together with a high resolution pH map, all simultaneously obtained from a single scan. Adapted from [28]. c) Optical micrograph of a calcite microcrystal compared to a simultaneously recorded SICM topography image and pH map measured 100 nm from the microcrystal (bulk pH 6.85). Adapted with permission from [29]. Copyright (2013) American Chemical Society.

imaging pH), fouling, and damage to the ion-selective electrode. Additionally, manufacturing requires expertise. New fabrication procedures have been reported to improve the response time by decreasing the electrode resistance. However, to date, no pH measurements have been reported with these improved probes.[27]

2.3. Scanning Ion Conductive Microscopy (SICM)

SICM is another powerful technique that has been used to perform pH measurements. In standard SICM, the scanning probe is a pulled nanopipette that is filled with an electrolyte containing a quasi-reference counter electrode (e.g. Ag/AgCl). Distance control is obtained by applying a potential between the tip and an external quasi-reference counter electrode placed in the bulk electrolyte which generates an ionic conductance current (feedback signal). To perform pH measurements with SICM, double-barrel SICM-SECM pipettes are used, which allow for simultaneously controlling the position of the electrode with high spatial resolution and measuring the pH locally.

A schematic representation of a SICM experimental setup is shown in Figure 2a. Here, a zwitterionic label-free nanoprobe was combined with SICM in a double-barrelled nanopipette to probe the topography and pH of single living cancer cells, with 50 nm spatial resolution.[28] Figure 2b shows the simultaneously obtained fluorescence image, SICM topography, and 3D pH distributions of a group of low-buffered breast cancer MCF7 cells in an estradiol-deprived medium. Despite the remarkable results, the zwitterionic pH probe response is only linear in a narrow pH range between 5 and 8, making it mainly applicable to biological systems. Other (more robust) pH sensing substrates, on the other hand, allow for pH measurements in a wider pH range. Nadappuram et al.[29] used nanoscale double-barrelled pipettes with IrO₂ for pH measurements and a SICM barrel for distance control. The probe was used to map the topography and pH of a calcite microcrystal. The maps obtained with a 10 µm lateral resolution can be seen in Figure 2c, where they are compared to an optical image of the microcrystal. Polymeric pH sensors have also been used in combination with SICM, as reported in the work of Morris et al.[30], who used a Au/PANi barrel to measure pH in porous membranes. While SECM pH probes are employed with

or without positional feedback (depending on the probe's size and the nature of the substrate), positioning control is intrinsic of SICM, which makes it, combined with other techniques as SECM and fluorescence, a powerful tool for obtaining detailed information of the electrochemical interface. Similar to SIET, fabrication of the probes requires expertise.

3. Rotating Ring-Disc Electrode

Measuring pH under well-defined mass transport conditions using a Rotating Ring-Disc Electrode (RRDE) system was introduced by Albery and Calvo[31] and since then mainly employed to probe electrocatalytic reactions. The operation principle is similar to SECM in the sense that an electrode is used to measure pH. In the RRDE system, a shaft is connected to a rotator and the crosssection of the shaft contains two working electrodes, a disc electrode surrounded by a thin ring electrode. The ring is used to probe the flux of species coming from the disc with a certain collection efficiency that will depend on the electrode's dimensions (radius, spacing) and experimental conditions. Initially, mainly bare Pt was used as ring material, however, that limits the number of systems that can be investigated as Pt is not a very selective H⁺ sensor. In later work, the ring was modified with a potentiometric pH sensing layer, as IrO₂, which allows the use of the technique to probe a larger variety of systems, as shown in Figure 3a. With the electrode shown in Figure 3a, Zimer et al.[32] investigated the influence of rotation rate on the steadystate water reduction on copper. Results from Figure 3b show the pH measured by the ring as a function of the rotation rate and indicate that a steady-state is achieved at rotation rates over 1200 rpm. RRDE has also been employed to study other reactions as CO₂ hydrogen oxidation[33], reduction[34], oxvgen reduction[35], and ethanol oxidation[36]. The main advantage of using Rotating Ring-Disc Electrode (RRDE) is that it is one of the few techniques that allows for the measurement of pH under defined mass transport conditions. On the other hand, the measurements are not spatially resolved in XY (as it probes the flux of species coming from the whole disc surface). The resolution in Z is limited and is a function of the rotation speed. It can be obtained by modelling of the system. Additionally, the



Figure 3. a) Schematic representation of an RRDE pH sensor and b) measurements performed with it during water reduction on copper at different rotation rates. Adapted with permission from [32]. Copyright (2015 Elsevier B.V. c) Schematic representation of a single compartment electrochemical cell coupled with a laser scanning confocal microscope, d) chemical structure of the LysoSensor Green DND-189 dye together with its normalized fluorescence response in phosphate buffer from pH 4 to 7, e) pH distribution at the anode and cathode during electrocoagulation, after a current density of 2 mA cm⁻² is applied for 60 s. Figures c, d and e are adapted from [46]. Copyright (2019) Wiley VCH Verlag GmbH & Co. KGaA.

time resolution to fully capture the process taking place at the disc is dependent on the time response of the pH sensors used.

4. Optical techniques

Various methods can be used to probe proton concentration gradients locally by using the optical properties of (electro)generated or consumed species, fluorescence being the most popular. In general, the measured optical property can be absorption or fluorescence intensity, decay time, reflectance, refractive index, light scattering, or light polarization.[37] We have selected fluorescence, infrared and Raman spectroscopy to briefly discuss here.

4.1. Fluorescence microscopy

Fluorescence has been widely used to indirectly measure pH in electrochemical systems. The most employed strategy is to add a pH-sensitive fluorophore to the electrolyte. The protons or OH produced/consumed by an electrochemical process can modulate the signal of the species in solution, which is captured using a fluorescence microscope. The workable pH range is determined by the pK_a of the fluorophores used. It is a versatile technique, that can be adapted to numerous electrochemical systems. The classic wide-field illumination setup allows for in situ mappings of pH in X- and Y-directions.[38,39] Although it has been used to record 2D pH images of different electrochemical systems, its well-known major drawback is the large signal contribution of out-of-focus light. The use of confocal laser scanning microscopy (CLSM) can overcome that by using a spatial pinhole to eliminate the out-of-focus light contribution to the image formation. It allows for mapping pH in 3D, by stacking 2D images acquired sequentially at different positions on the Z-axis. Pioneer contributions were made by the group of Unwin, demonstrating that 3D fluorescence maps can provide quantitative imaging of pH profiles during water reduction on Pt-UMEs.[40,41] CLSM has been successfully applied to measure pH in the fields of bioelectrochemistry[42,43], electrosynthesis[44], electrocatalysis[45,46], and the development of new dyes for expanding the workable pH range and stability can broaden its use even more.[47] Recently, the work of Fuladpanjeh-Hojaghan et al.[46] presented an electrochemical cell coupled with a laser scanning confocal microscope able to perform quantitative pH mapping under operando conditions during electrocoagulation

(Figure 3c). Two aluminium plates are used as cathode and anode and a combination of pH-sensitive fluorescent dyes with different pK_a allows for pH detection in a range from 1.5 to 8.5. An example of the dye response upon a change in pH is shown in Figure 3d for the LysoSensor Green DND-189 (LSG, pK_a 5.2). pH maps under operando conditions were obtained from the cathode and anode side and an example is shown in Figure 3e, which is an image taken after 60 seconds of polarization. The anodic reaction causes aluminium dissolution and consequently aluminium cation hydrolysis, leading to acidification near the electrode surface. On the cathode side, the water reduction reaction generates a more alkaline pH near the surface. A setup for performing time-resolved CLSM pH measurements has also been recently introduced by Pande et al.[48] Fluorescein was used as a pH-sensitive fluorophore, to study the diffusion layer during ORR on platinum and the effect of sulfate buffering on the pH profiles. The results of a time-dependent numerical model give good agreement with the experimental data.

One of the main advantages of CLSM is that, differently from SPM techniques, it enables one to probe an entire macroscopic sample in real-time with high spatial resolution. Although pH maps can be obtained relatively quickly, the need of adding a fluorophore to the electrolyte is a drawback as it may affect the electrochemical process being studied. In general, the resolution limit of pH measurements using fluorescence falls into the diffraction limit of conventional light microscopes (~ 250 nm, considering the wavelength of green light and a numerical aperture of 1), unless a super-resolution microscopy technique is used, e.g. Stimulated emission depletion (STED), where pH measurements in a living cell have been performed with a resolution down to 20 nm.[49] This and other super-resolution techniques have been mainly used to probe biological substrates. Their application to other electrochemical systems is not straightforward and sometimes not possible due to the lack of stability of the pH sensing molecules.

4.2. Other optical techniques

pH measurements at the electrochemical interface during CO_2 reduction have recently been reported using surface-enhanced infrared absorption spectroscopy (SEIRAS).[50–52] The signal intensity of species that compose a buffer system, such as CO_2/HCO_3^- or $H_2PO_4^{-/}HPO_4^{-2}/PO_4^{-3-}$, is monitored. Using the equilibrium equations, the average proton concentration at the

these species. Valuable information can be obtained regarding surface in XY. These measurements do not provide spatial the species in solution only a few nanometers from the electrode

interface can be indirectly derived from the ratio of the signal of surface in the Z direction, averaged over a large portion of the

Table 1. Overview of techniques discussed in this work to measure local pH in electrochemistry

| Technique | pH sensor | Range | Application | Туре | Reference |
|----------------|---|--------------|---|----------------------|--------------|
| SECM | | | | | |
| Potentiometric | IrOx/Au | 2-11 | Nitrate reduction | direct | [9] |
| | IrOx/Pt | 1-13 | Corrosion of 316L stainless steel, Corrosion of AZ31B Magnesium Allov | direct | [8,55,56] |
| | Sb/Sb ₂ O ₃ | 4-11 | Corrosion of AZ63 magnesium alloy | direct | [19] |
| | Polyaniline/ Pt | 4-8 | extracellular pH | direct | [10] |
| | Pt | 1-8 | Methanol oxidation, enzymatic processes | direct | [57] |
| Voltammetric | syringaldazine | 2-12 | Oxygen reduction, extracellular pH | direct | [15,16] |
| | 4-NTP monolayer on Au | 1-12 | Hydrogen evolution, CO ₂ reduction on gold | direct | [17] |
| SICM | zwitterion nano-membrane | 4-9 | Extracellular pH | direct | [28] |
| | IrOx/C | 2-10 | Dissolution of a calcite microcrystal | direct | [29] |
| | Polyaniline/Au | 2.5-12 | Nano channels of a membrane | direct | [30] |
| SIET | H ⁺ -selective cocktail | 4-10 | Microbial metabolism | direct | [23] |
| | Hydrogen ionophore I - cocktail B (Selectophore TM) | 5-12.5 | Corrosion of MA8 Mg alloy | direct | [26] |
| | LSG + 5(6)- | | | | |
| FLUORESCENCE | carboxynaphtho-fluorescein | 1.5-8.5 | Electrocoagulation | indirect | [46] |
| | fluorescein | 5.5-7 | H_2O and O_2 reduction, enzymatic O_2 reduction, coaxial flow micro-reactor | indirect | [41,43,44] |
| RRDE | Pt | 5-9 | H ₂ evolution and oxidation | direct | [33] |
| | IrO ₂ | 2.5-12.5 | H ₂ O reduction | direct | [32] |
| | Pt, IrO_2 | n.a. | CO_2 reduction on gold | direct | [34] |
| INFRARED | $H_2PO_4^{-}/HPO_4^{2-}/PO_4^{3-}$ CO ₂ / HCO ³⁻ | 5-13 n.a. | CO ₂ reduction on copper CO ₂ reduction on gold | indirect indirect | [52] [50] |
| RAMAN | 4-mercaptobenzoic acid/Au nanoparticle | 2-12 | Aerosol microdroplets | indirect | [54] |

resolution and require IR active species in the electrolyte. Similarly to the dyes used in fluorescence, the pK_a of these species in equilibrium will dictate the pH range that can be measured. Raman Spectroscopy has also been used to probe pH. Wei et al. developed

pH nanoprobes based on plasmonic gold nanoparticles functionalized with a pH-sensitive amine or carboxylic acid. The nanoprobes were used to detect the intracellular pH of PC-3 cancer cells[53] and the pH gradients within phosphate-buffered aerosol microdroplets with a lateral resolution of 5 μ m.[54] The technique is powerful to investigate micro-environments, however, can only be applied to specific systems where the addition of the signal enhancing pH-sensitive particles is feasible. Differently than in SPM, the resolution that can be achieved in pH measurements performed with optical techniques is defined by the operating system and not by the size of the pH probe.

Final considerations 5.

Having discussed the main techniques that can be used to measure pH in electrochemistry, we have grouped a few examples in Table 1 to provide a general overview of the methods available, the pH range that can be measured, and the kind of systems that can be investigated with each technique. To conclude, we highlight the following:

• The main advantages of SPM for performing localized pH measurements are the high spatial resolution in X-Y-Z that can be achieved, and the versatility in terms of the sample, probes, and modes available, to be chosen according to the application. A wide pH range can be studied with a series of established potentiometric and voltammetric sensors. Voltammetric sensors are especially interesting, due to the high temporal resolution that can be achieved. The main disadvantage of SPM is shielding caused by the probe.[58] However, this can be significantly minimized by the use of nanopipettes. Additionally, these contributions can be simulated and accounted for with finite element models.

- Optical methods have the great advantage of being noninvasive techniques, and the use of a confocal laser microscope allows for achieving spatial resolutions comparable to SPM techniques. However, altering the electrolyte composition by the addition of a fluorophore is a major drawback. Additionally, the operational pH range is usually limited by the pK_a of the pH-sensitive species.
- Infrared and Raman spectroscopy can be powerful tools to probe small volumes of the electrochemical interface, although the spatial resolution in Z cannot be tuned and the XY resolution is limited by the beam size.
- Combining SPM and optical techniques can be highly effective to obtain detailed information about complex electrochemical systems, see e.g. SECM-SICM[29], SECMfluorescence[59], SECM-infrared[60].
- RRDE pH measurements are especially interesting for mechanistic studies as the defined mass transport control of the system allows for precise modeling of the reaction interface. On the other hand, the measurements are not spatially resolved and represent an average response of the whole working electrode.

There has been intensive development of methods for measuring pH in electrochemistry lately. Knowing precisely the research questions that need answers is key for defining the most suitable technique to investigate a given system. We believe the combined knowledge obtained through the different techniques discussed here, can enable a deeper understanding of complex electrochemical systems.

Conflict of interest statement

Nothing declared.

Acknowledgments

This work was supported by the European Commission under contract 722614 (Innovative training network Elcorel).

References

[1] S.P.L. Sörensen, Über die Messung und Bedeutung der Wasserstoffionen-konzentration bei biologischen Prozessen, Ergebnisse Der Physiol. (1909) 131–200.

[2] N. Jadhav, V.J. Gelling, Review—The Use of Localized Electrochemical Techniques for Corrosion Studies, J. Electrochem. Soc. 166 (2019) C3461–C3476.

[3] R. Orij, S. Brul, G.J. Smits, Intracellular pH is a tightly controlled signal in yeast, Biochim. Biophys. Acta - Gen. Subj. 1810 (2011) 933–944.

[4] H. Ooka, M.C. Figueiredo, M.T.M. Koper, Competition between Hydrogen Evolution and Carbon Dioxide Reduction on Copper Electrodes in Mildly Acidic Media, Langmuir. 33 (2017) 9307–9313.

[5] L.F. Nims, Glass Electrodes and Apparatus for Direct Recording of pH in Vivo, Yale J. Biol. Med. 10 (1938) 241–246.

[6] L. Manjakkal, D. Szwagierczak, R. Dahiya, Metal oxides based electrochemical pH sensors: Current progress and future perspectives, Prog. Mater. Sci. 109 (2020) 100635.

[7] O. Korostynska, K. Arshak, E. Gill, A. Arshak, Review on state-of-the-art in polymer based pH sensors, Sensors. 7 (2007) 3027–3042.

[8] U.M. Tefashe, P. Dauphin-Ducharme, M. Danaie, Z.P. Cano, J.R. Kish, G.A. Botton, J. Mauzeroll, Localized Corrosion Behavior of AZ31B Magnesium Alloy with an Electrodeposited Poly(3,4-Ethylenedioxythiophene) Coating, J. Electrochem. Soc. 162 (2015) C536–C544.

[9] C.S. Santos, A.S. Lima, D. Battistel, S. Daniele, M. Bertotti, Fabrication and Use of Dual-function Iridium Oxide Coated Gold SECM Tips. An Application to pH Monitoring above a Copper Electrode Surface during Nitrate Reduction., Electroanalysis. 28 (2016) 1441–1447.

[10] R. Song, Q. Xiong, T. Wu, X. Ning, F. Zhang, Q. Wang, P. He, Real-time monitoring of extracellular pH using a pH-potentiometric sensing SECM dual-microelectrode, Anal. Bioanal. Chem. 412 (2020) 3737–3743.

[11] X. Zhang, B. Ogorevc, J. Wang, Solid-state pH nanoelectrode based on polyaniline thin film electrodeposited onto ion-beam etched carbon fiber, Anal. Chim. Acta. 452 (2002) 1–10.

[12] H. Jang, J. Lee, Iridium oxide fabrication and application: A review, J. Energy Chem. 46 (2020) 152–172.

[13] M. Etienne, P. Dierkes, T. Erichsen, W. Schuhmann, I. Fritsch, Constant-distance mode scanning potentiometry. High resolution pH measurements in three-dimensions, Electroanalysis. 19 (2007) 318–323.

**[14] A. Botz, J. Clausmeyer, D. Öhl, T. Tarnev, D. Franzen, T. Turek, W. Schuhmann, Local Activities of Hydroxide and Water Determine the Operation of Silver-Based Oxygen Depolarized Cathodes, Angew. Chemie - Int. Ed. 57 (2018) 12285–12289.

pH maps above a silver gas diffusion electrode were constructed during oxygen reduction using SECM. A platinum nanoelectrode is used as pH sensor and shear-force is used for distance control.

*[15] M. Michalak, M. Kurel, J. Jedraszko, D. Toczydlowska, G. Wittstock, M. Opallo, W. Nogala, **Voltammetric pH** Nanosensor, Anal. Chem. 87 (2015) 11641–11645.

A polymer film of syringaldazine is used as voltammetric pH sensor. pH changes during water reduction are measured with high spatial resolution.

[16] R.-E. Munteanu, L. Stănică, M. Gheorghiu, S. Gáspár, Measurement of the Extracellular pH of Adherently Growing Mammalian Cells with High Spatial Resolution Using a Voltammetric pH Microsensor, Anal. Chem. 90 (2018) 6899– 6905.

**[17] M.C.O. Monteiro, L. Jacobse, T. Touzalin, M.T.M. Koper, Mediator-Free SECM for Probing the Diffusion Layer pH with Functionalized Gold Ultramicroelectrodes, Anal. Chem. 92 (2020) 2237–2243.

A voltammetric sensor based on a self-assembled monolayer on gold is presented to measure pH with SECM, with high time resolution. The probe is stable in different gaseous environments and extremely sensitive.

[18] Z. Zhu, Z. Ye, Q. Zhang, J. Zhang, F. Cao, Novel dual Pt-Pt/IrOx ultramicroelectrode for pH imaging using SECM in both potentiometric and amperometric modes, Electrochem. Commun. 88 (2018) 47–51.

**[19] D. Filotás, B.M. Fernández-Pérez, L. Nagy, G. Nagy, R.M. Souto, Multi-barrel electrodes containing an internal micro-reference for the improved visualization of galvanic corrosion processes in magnesium-based materials using potentiometric scanning electrochemical microscopy, Sensors Actuators, B Chem. 296 (2019) 126625.

A triple-barrel microelectrode containing an internal reference is used for simultaneous monitoring Mg2+ and pH during corrosion of a Mg alloy using potentiometric SECM. The use of an internal reference eliminate electric field effects and improves the chemical resolution.

[20] B. Ballesteros Katemann, A. Schulte, W. Schuhmann, Constant-distance mode scanning electrochemical microscopy (SECM) - Part I: Adaptation of a non-optical shear-force-based positioning mode for SECM tips, Chem. - A Eur. J. 9 (2003) 2025–2033.

[21] K. Eckhard, W. Schuhmann, Alternating current techniques in scanning electrochemical microscopy (AC-SECM), Analyst. 133 (2008) 1486–1497.

[22] A.J. Bard, M. V. Mirkin, Scanning Electrochemical Microscopy, Second, CRC Press, 2012.

[23] V.S. Joshi, P.S. Sheet, N. Cullin, J. Kreth, D. Koley, Real-Time Metabolic Interactions between Two Bacterial Species Using a Carbon-Based pH Microsensor as a Scanning Electrochemical Microscopy Probe, Anal. Chem. 89 (2017) 11044–11052.

[24] S. V. Lamaka, M. Taryba, M.F. Montemor, H.S. Isaacs, M.G.S. Ferreira, Quasi-simultaneous measurements of ionic currents by vibrating probe and pH distribution by ion-selective microelectrode, Electrochem. Commun. 13 (2011) 20–23.

[25] R. Oltra, Mass transport control of localised corrosion processes: in situ local probing and modelling, Corros. Eng. Sci. Technol. 53 (2018) 2–8.

[26] A.S. Gnedenkov, D. Mei, S. V. Lamaka, S.L. Sinebryukhov, D. V. Mashtalyar, I.E. Vyaliy, M.L. Zheludkevich, S. V. Gnedenkov, Localized currents and pH distribution studied during corrosion of MA8 Mg alloy in the cell culture medium, Corros. Sci. 170 (2020) 108689.

[27] Á. Varga, L. Nagy, J. Izquierdo, I. Bitter, R.M. Souto, G. Nagy, Development of Solid Contact Micropipette Zn-Ion Selective Electrode for Corrosion Studies, Anal. Lett. 44 (2011) 2876–2886.

**[28] Y. Zhang, Y. Takahashi, S.P. Hong, F. Liu, J. Bednarska, P.S. Goff, P. Novak, A. Shevchuk, S. Gopal, I. Barozzi, L. Magnani, H. Sakai, Y. Suguru, T. Fujii, A. Erofeev, P. Gorelkin, A. Majouga, D.J. Weiss, C. Edwards, A.P. Ivanov, D. Klenerman, E. V. Sviderskaya, J.B. Edel, Y. Korchev, High-resolution label-free 3D mapping of extracellular pH of single living cells, Nat. Commun. 10 (2019) 1–9.

A zwitterionic label-free pH nanoprobe is used in combination with SICM. The topography and pH of single living cancer cells is measured with 50 nm spatial resolution.

[29] B.P. Nadappuram, K. McKelvey, R. Al Botros, A.W. Colburn, P.R. Unwin, Fabrication and characterization of dual function nanoscale pH-scanning ion conductance microscopy (SICM) probes for high resolution pH mapping, Anal. Chem. 85 (2013) 8070–8074.

[30] C.A. Morris, C.C. Chen, T. Ito, L.A. Baker, Local pH measurement with scanning ion conductance microscopy, J. Electrochem. Soc. 160 (2013) H430–H435.

[31] W.J. Albery, E.J. Calvo, Ring-disc electrodes. Part 21. pH measurement with the ring, J. Chem. Soc. Faraday Trans. 1 Phys. Chem. Condens. Phases. 79 (1983) 2583–2596.

*[32] A.M. Zimer, M. Medina da Silva, E.G. Machado, H. Varela, L.H. Mascaro, E.C. Pereira, Development of a versatile rotating ring-disc electrode for in situ pH measurements, Anal. Chim. Acta. 897 (2015) 17–23.

RRDE is used for pH measurements during water reduction. The ring is functionalized with IrOx for pH sensing and shows fast response due to the small thickness of the ring layer.

[33] Y. Yokoyama, K. Miyazaki, Y. Miyahara, T. Fukutsuka, T. Abe, In Situ Measurement of Local pH at Working Electrodes in Neutral pH Solutions by the Rotating Ring Disk Electrode Technique, ChemElectroChem. (2019) 4750–4756.

[34] F. Zhang, A.C. Co, Direct Evidence of Local pH Change and the Role of Alkali Cation during CO_2 Electroreduction in Aqueous Media, Angew. Chemie - Int. Ed. 59 (2020) 1674–1681.

[35] P. Steegstra, E. Ahlberg, In situ pH measurements with hydrous iridium oxide in a rotating ring disc configuration, J. Electroanal. Chem. 685 (2012) 1–7.

[36] M.C. Figueiredo, R.M. Arán-Ais, V. Climent, T. Kallio, J.M. Feliu, Evidence of Local pH Changes during Ethanol Oxidation at Pt Electrodes in Alkaline Media, ChemElectroChem. 2 (2015) 1254–1258.

[37] D. Wencel, T. Abel, C. McDonagh, Optical chemical pH sensors, Anal. Chem. 86 (2014) 15–29.

[38] W.J. Bowyer, J. Xie, R.C. Engstrom, Fluorescence imaging of the heterogeneous reduction of oxygen, Anal. Chem. 68 (1996) 2005–2009.

[39] J.C. O'Brien, J. Shumaker-Parry, R.C. Engstrom, Microelectrode Control of Surface-Bound Enzymatic Activity, Anal. Chem. 70 (1998) 1307–1308. [40] S. Cannan, I. Douglas Macklam, P.R. Unwin, Threedimensional imaging of proton gradients at microelectrode surfaces using confocal laser scanning microscopy, Electrochem. Commun. 4 (2002) 886–892.

*[41] N.C. Rudd, S. Cannan, E. Bitziou, I. Ciani, A.L. Whitworth, P.R. Unwin, Fluorescence confocal laser scanning microscopy as a probe of pH gradients in electrode reactions and surface activity, Anal. Chem. 77 (2005) 6205–6217.

The authors present the use of fluorescence confocal laser scanning microscopy to quantify three-dimensional pH gradients near electrode surfaces. The method was applied to study oxygen and water reduction and to detect defects in microelectrode arrays.

[42] J.T. Hou, W.X. Ren, K. Li, J. Seo, A. Sharma, X.Q. Yu, J.S. Kim, Fluorescent bioimaging of pH: From design to applications, Chem. Soc. Rev. 46 (2017) 2076–2090.

[43] B. Tassy, A.L. Dauphin, H.M. Man, H. Le Guenno, E. Lojou, L. Bouffier, A. De Poulpiquet, In Situ Fluorescence Tomography Enables a 3D Mapping of Enzymatic O_2 Reduction at the Electrochemical Interface, Anal. Chem. 92 (2020) 7249–7256.

[44] A. Abou-Hassan, J.F. Dufrêchfer, O. Sandre, G. Mériguet, O. Bernard, V. Cabuil, Fluorescence confocal laser scanning microscopy for pH mapping in a coaxial flow microreactor: Application in the synthesis of superparamagnetic nanoparticles, J. Phys. Chem. C. 113 (2009) 18097–18105.

[45] A.J. Leenheer, H.A. Atwater, Imaging Water-Splitting Electrocatalysts with pH-Sensing Confocal Fluorescence Microscopy, J. Electrochem. Soc. 159 (2012) H752–H757.

[46] B. Fuladpanjeh-Hojaghan, M.M. Elsutohy, V. Kabanov, B. Heyne, M. Trifkovic, E.P.L. Roberts, **In-Operando Mapping of pH Distribution in Electrochemical Processes, Angew. Chemie Int. Ed. 58 (2019) 16815–16819.

An experimental setup is presented to measure pH during electrocoagulation using confocal lazer microscopy. Simultaneous pH maps for the anode and cathode reaction are constructed under operando conditions.

[47] A. Bigdeli, F. Ghasemi, S. Abbasi-Moayed, M. Shahrajabian, N. Fahimi-Kashani, S. Jafarinejad, M.A. Farahmand Nejad, M.R. Hormozi-Nezhad, Ratiometric fluorescent nanoprobes for visual detection: Design principles and recent advances - A review, Anal. Chim. Acta. 1079 (2019) 30–58.

**[48] N. Pande, S.K. Chandrasekar, D. Lohse, G. Mul, J.A. Wood, B.T. Mei, D. Krug, Electrochemically Induced pH Change: Time Resolved Confocal Fluorescence Microscopic Measurements and Comparison with Numerical Model, J. Phys. Chem. Lett. (2020).

Confocal fluorescence microscopy is used to measure timeresolved pH profiles, which are compared to numerical models. Additionally crucial factors affecting the pH measurements as attenuation of light and the role of dye migration are discussed.

[49] D.S. Richardson, C. Gregor, F.R. Winter, N.T. Urban, S.J. Sahl, K.I. Willig, S.W. Hell, SRpHi ratiometric pH biosensors for super-resolution microscopy, Nat. Commun. 8 (2017).

[50] O. Ayemoba, A. Cuesta, Spectroscopic Evidence of Size-Dependent Buffering of Interfacial pH by Cation Hydrolysis during CO_2 Electroreduction, ACS Appl. Mater. Interfaces. 9 (2017) 27377–27382.

[51] J. Zhou, H. Ma, Design principles of spectroscopic probes for biological applications, Chem. Sci. 7 (2016) 6309–6315.

*[52] K. Yang, R. Kas, W.A. Smith, In Situ Infrared Spectroscopy Reveals Persistent Alkalinity near Electrode Surfaces during CO₂ Electroreduction, J. Am. Chem. Soc. 141 (2019) 15891–15900.

Infrared spectroscopy is used to probe pH during CO_2 reduction on copper films, based on the ratio of phosphate buffer species. Measurements at high current densities were performed and correlated with a model.

[53] H. Wei, M.R. Willner, L.C. Marr, P.J. Vikesland, Highly stable SERS pH nanoprobes produced by co-solvent controlled AuNP aggregation, Analyst. 141 (2016) 5159–5169.

[54] H. Wei, E.P. Vejerano, W. Leng, Q. Huang, M.R. Willner, L.C. Marr, P.J. Vikesland, Aerosol microdroplets exhibit a stable pH gradient, Proc. Natl. Acad. Sci. U. S. A. 115 (2018) 7272–7277.

[55] Z. Zhu, Q. Zhang, P. Liu, J. Zhang, F. Cao, Quasisimultaneous electrochemical/chemical imaging of local Fe2+ and pH distributions on 316 L stainless steel surface, J. Electroanal. Chem. (2020) 114107.

[56] Z. Zhu, X. Liu, Z. Ye, J. Zhang, F. Cao, J. Zhang, A fabrication of iridium oxide film pH micro-sensor on Pt ultramicroelectrode and its application on in-situ pH distribution of 316L stainless steel corrosion at open circuit potential, Sensors Actuators, B Chem. 255 (2018) 1974–1982.

[57] D.J. Chen, R.W. Penhallurick, Y.Y.J. Tong, A versatile and robust surface-poison-resisting Scanning Amperometric Proton Microscopy, J. Electroanal. Chem. (2020) 113918.

*[58] R.A.J. Critelli, M. Bertotti, R.M. Torresi, **Probe effects on concentration profiles in the diffusion layer : Computational modeling and near-surface pH measurements using microelectrodes**, Electrochim. Acta. 292 (2018) 511–521.

A finite element model is used to account for the physical blocking effect an SECM tip has on the pH profiles measured during hydrogen evolution.

[59] F.M. Boldt, J. Heinze, M. Diez, J. Petersen, M. Börsch, Real-time pH microscopy down to the molecular level by combined scanning electrochemical microscopy/single-molecule fluorescence spectroscopy, Anal. Chem. 76 (2004) 3473–3481.

[60] L. Wang, J. Kowalik, B. Mizaikoff, C. Kranz, Combining scanning electrochemical microscopy with infrared attenuated total reflection spectroscopy for in situ studies of electrochemically induced processes, Anal. Chem. 82 (2010) 3139–3145.

Declaration of interest

In this paper, we review different methodologies for measuring local (surface) pH in electrochemistry. This is an important topic in various areas of (surface) electrochemistry.