



Universiteit  
Leiden

The Netherlands

## Uncovering the value of autonomic signs and seizure detection in epilepsy care

Westrhenen, A. van

### Citation

Westrhenen, A. van. (2023, September 12). *Uncovering the value of autonomic signs and seizure detection in epilepsy care*. Retrieved from <https://hdl.handle.net/1887/3640064>

Version: Publisher's Version

License: [Licence agreement concerning inclusion of doctoral thesis in the Institutional Repository of the University of Leiden](#)

Downloaded from: <https://hdl.handle.net/1887/3640064>

**Note:** To cite this publication please use the final published version (if applicable).



## CHAPTER 2

# **Ictal autonomic changes as a tool for seizure detection: a systematic review**

**van Westrhenen A, De Cooman T, Lazon RHC,  
Van Huffel S, Thijs RD**

# ABSTRACT

## Introduction

Adequate epileptic seizure detection may have the potential to minimize seizure-related complications and improve treatment evaluation. Autonomic changes often precede ictal electroencephalographic discharges and therefore provide a promising tool for timely seizure detection. We reviewed the literature for seizure detection algorithms using autonomic nervous system parameters.

## Methods

The PubMed and Embase databases were systematically searched for original human studies that validate an algorithm for automatic seizure detection based on autonomic function alterations. Studies on neonates only and pilot studies without performance data were excluded. Algorithm performance was compared for studies with a similar design (retrospective vs. prospective) reporting both sensitivity and false alarm rate (FAR). Quality assessment was performed using QUADAS-2 and recently reported quality standards on reporting seizure detection algorithms.

## Results

Twenty-one out of 638 studies were included in the analysis. Fifteen studies presented a single-modality algorithm based on heart rate variability ( $n = 10$ ), heart rate ( $n = 4$ ), or QRS morphology ( $n = 1$ ), while six studies assessed multimodal algorithms using various combinations of HR, corrected QT interval, oxygen saturation, electrodermal activity, and accelerometry. Most studies had small sample sizes and a short follow-up period. Only two studies performed a prospective validation. A tendency for a lower FAR was found for retrospectively validated algorithms using multimodal autonomic parameters compared to those using single modalities (mean sensitivity per participant 71-100% vs. 64-96% and mean FAR per participant 0.0-2.4/h vs. 0.7-5.4/h).

## Conclusions

The overall quality of studies on seizure detection using autonomic parameters is low. Unimodal autonomic algorithms cannot reach acceptable performance as false alarm rates are still too high. Larger prospective studies are needed to validate multimodal automatic seizure detection.


## INTRODUCTION

Epileptic seizures are potentially dangerous as they can lead to complications, including injury, status epilepticus, and sudden unexpected death in epilepsy (SUDEP).<sup>1</sup> Adequate seizure detection may have the potential to minimize these complications and to ameliorate treatment evaluation, as seizures — particularly those at night — are often underreported.<sup>2-5</sup> Detection devices may also help to improve the independence and quality of life of people with epilepsy and their caregivers.<sup>3,6</sup>

Several parameters, including movement, sound, and autonomic nervous system changes, can be used to detect seizures. This review focuses on changes in autonomic function, including cardiovascular, respiratory, and transpiration changes.<sup>7</sup> Seizures can alter autonomic function, particularly if the central autonomic network is involved. The most common expression is a sudden increase in sympathetic tone.<sup>7,8</sup> Ictal tachycardia (IT) is a very frequent sign, with prevalence rates ranging from 80 to 100%.<sup>9,10</sup> IT is a hallmark of convulsive seizures (i.e., focal to bilateral tonic-clonic as well as generalized tonic-clonic seizures), and more common in temporal lobe vs. extratemporal lobe seizures.<sup>9</sup> Changes in autonomic function can precede ictal electroencephalographic (EEG) discharges by several seconds.<sup>10-12</sup> Preictal tachycardia has an incidence rate of approximately one-third of seizures.<sup>13</sup> Autonomic alterations may therefore provide an adequate tool for early seizure detection and facilitate timely interventions. Ictal arrhythmias and desaturations are more common but are thought to be self-limiting, while postictal arrhythmias and apneas may lead to SUDEP.<sup>14-17</sup> SUDEP usually occurs several minutes after a convulsive seizure (mean 10 min, range 2-17 min).<sup>18</sup> Raising an alarm at seizure onset may be sufficient to allow timely intervention. We aimed to systematically review different seizure detection algorithms based on autonomic function changes.

## METHODS

This systematic review was conducted in accordance with the preferred reporting items for systematic reviews and meta-analyses (PRISMA) guideline.<sup>19</sup> The PubMed and Embase databases were systematically searched through May 2018 for original studies validating an algorithm for automatic seizure detection based on heart rate (HR), heart rate variability (HRV), oxygen saturation (SpO<sub>2</sub>), electrodermal activity (EDA, reflecting changes in transpiration), or a



**Figure 1** Flowchart of the search for applicable studies

combination of the aforementioned. A sequence of synonyms for ‘autonomic variables,’ ‘seizures,’ and ‘detection’ were used as search terms. Studies were included if they met the following criteria: (1) human studies; (2) written in English; (3) reporting on children or adults with any type of epilepsy; (4) validating an algorithm for automatic seizure detection using autonomic parameters; (5) reporting at least one performance measure [sensitivity, positive predictive value (PPV), false alarm rate (FAR), or detection latency (DL)]. Studies on neonates only were excluded, because both seizure and autonomic

function characteristics differ greatly at this age compared to older age. Pilot studies lacking performance data, as well as conference abstracts and reviews were also excluded (Fig. 1).

One author (AvW) screened all titles and abstracts, as well as the full texts of the remaining studies. For each article included, the following parameters were recorded: method of automatic seizure detection, type of autonomic variable, individual characteristics, number and types of seizures analyzed, prospective or retrospective validation, total recording time and performance of the algorithm (including sensitivity, PPV, FAR, and DL). We compared algorithm performance using multimodal autonomic parameters versus those using single modalities, provided that the studies (1) had a similar design (prospective vs. retrospective) and (2) reported both sensitivity and FAR.

The quality of the included studies was evaluated using the QUADAS-2.<sup>20</sup> This tool consists of four domains (patient selection, index test, reference standard, and flow and timing) and different signaling questions to assist in judgments of the risk of bias and applicability. Additionally, we assessed all included studies according to the recently proposed standards for clinical validation of seizure detection devices (SDDs).<sup>21</sup>

## RESULTS

Out of the 638 articles identified, 86 studies were selected based on title and abstract. After full-text screening, 21 studies were included for further analysis. Most of the excluded articles lacked the validation of a seizure detection algorithm (Fig. 1). The characteristics of the included studies are summarized in Table 1. Most of the studies ( $n = 15$ ) focused on ictal cardiac changes as a tool for seizure detection algorithms, including HRV ( $n = 10$ ),<sup>8, 22-30</sup> HR ( $n = 4$ ),<sup>31-34</sup> and changes in QRS morphology ( $n = 1$ ).<sup>35</sup> Six studies used multimodal algorithms, including combinations of HR, corrected QT interval (QTc), SpO<sub>2</sub>, EDA, and accelerometry (ACC).<sup>2, 36-40</sup> None of the included studies validated an algorithm based on oxygen saturation or EDA alone. Most studies were conducted in adults, but two studies included a pediatric population,<sup>23, 40</sup> and six studies included both children and adults.<sup>22, 25, 35-37, 39</sup> Fourteen studies prospectively enrolled their participants,<sup>8, 22, 23, 26, 28, 30-33, 36-40</sup> but only two studies prospectively validated their algorithm.<sup>31, 33</sup> Most studies had small sample sizes (median population size 14, IQR 7-26). The number of seizures analyzed per patient tended to be low (median number of seizures per participant 3, IQR 2-7). The total recording time used to validate the algorithm varied from 7 min to 158

h per person (median recording time per participant 34 h, IQR 3-86 h), but was not specified in two studies. Seizure onset was mostly focal ( $n = 14$ ),<sup>8, 22, 24-26, 28, 30, 31, 33, 34, 37, 39, 40, 42</sup> but was focal and generalized in some ( $n = 4$ )<sup>2, 23, 35, 42</sup> or not specified in others ( $n = 3$ ).<sup>32, 36, 38</sup> All four performance measures (sensitivity, PPV, FAR, and DL) were only reported in three out of 21 studies;<sup>22, 33, 39</sup> eight studies reported three,<sup>2, 23-25, 28, 30, 31, 42</sup> eight studies reported two,<sup>8, 26, 34, 36-38, 40, 43</sup> one study reported one,<sup>41</sup> and one study only reported sensitivity and PPV data for some of the subjects.<sup>32</sup>

## Heart rate analysis

Heart rate was monitored using single or multiple lead electrocardiography (ECG) in 14 of 18 studies,<sup>8, 22-26, 28, 32, 34-37, 42, 43</sup> Alternative methods included photoplethysmography (PPG) in a wearable sensor ( $n = 2$ )<sup>2, 30</sup> and an implanted heart rate sensor (AspireSR) ( $n = 2$ ).<sup>31, 33</sup>

Heart rate measurement was done using various methods of R-peak detection, including those proposed by Pan and Tompkins,<sup>30, 41</sup> Kohler,<sup>28</sup> Yeh and Wang,<sup>22-24</sup> or unspecified methods.<sup>8, 25, 26, 31-34, 42</sup> Some studies applied noise filtering techniques to diminish false R-peak detection, including high- and low-pass noise filters<sup>8, 22-24, 26, 30</sup> or a specific algorithm (baseline estimation and denoising with sparsity).<sup>42</sup> One case study prospectively assessed a HR algorithm using a vagal nerve stimulation (VNS) device with a fixed HR sensitivity threshold.<sup>33</sup> Alarms were generated when the HR augmentation exceeded 50% of the baseline HR. Eleven out of twelve seizures were detected (sensitivity 92%), together with 128 false alarms (FAR 1.88/h; 68 h recordings). A second prospective validation study of the same VNS device compared different HR thresholds ( $\geq 20\%$ ,  $\geq 40\%$ , and  $\geq 60\%$  increases from baseline) in 16 adults with refractory epilepsy.<sup>31</sup> Lower thresholds resulted in higher sensitivity and higher FAR than higher thresholds (e.g., sensitivity 59.3% and FAR 7.2/h for threshold  $\geq 20\%$  vs. sensitivity 18.8% and FAR 0.5/h for thresholds  $\geq 60\%$ ). Similar effects of varying the thresholds (for both the relative HR increase and the duration of HR increase) were reported in two studies on retrospectively validated HR algorithms.<sup>32, 34</sup> A follow-up using the same dataset examined different factors that may influence the probability of seizure detection.<sup>44</sup> The best regression model was created with variables including age, gender, etiology, seizure class, and years with epilepsy.



**Table 1** Characteristics of included studies

Study	Autonomic parameter	Measurement	Device	Algorithm	Prospective/retrospective validation	Population (n)	No. of seizures/TRT <sup>a</sup>	Type of seizures (n)	Mean age (y) [range]	Performance (mean of values per patient)
De Cooman et al. <sup>22</sup>	Cardiac	HRV	Single-lead ECG	Noise filtering; High- and low-pass Butterworth filters. HRI-extract algorithm and HRI feature extraction with patient-independant SVM classifier; LOPO CT; Fastened HRI-SVM seizure detection	Retrospective	Refractory temporal lobe epilepsy (17)	127/918 h	FOIA, FOBTC	33.5 [9-54]	Sens: 83.2% [50-100%] (overall: 81.9%) PPV: 7.9% [0.4-21%] (overall: 5.4%) FAR: 2.01/h [0.88-3.52/h] (overall: 1.97/h) DL: 13.3 s [-18.2-54.3] (overall: 17.8 s)
De Cooman et al. <sup>23</sup>	Cardiac	HRV	Single-lead ECG	See De Cooman <sup>22</sup> ; Patient-specific heuristic adaptive classifier	Retrospective	1) Children (14) 2) Other group of children (14)	107/695 h	GOS (30), FOS (77)	NA	Patient-independent: Sens: (overall: 81.3%) PPV: NA FAR (overall: 0.75/h) DL: NA Patient-specific: Sens: (overall: 77.6%) PPV: (overall: 30.7%) FAR (overall: 0.33/h) DL: 19.1 s
De Cooman et al. <sup>24</sup>	Cardiac	HRV	Single-lead ECG	See De Cooman <sup>22</sup> ; Patient-specific heuristic adaptive classifier and real-time adaptive classifier	Retrospective	Temporal lobe epilepsy (19)	153/2833 h	FOS, FOIA, FOBTC, U, sub-clinical	NA	Patient-independent: Sens: (overall: 78.4%) PPV: (overall: 2.4%) FAR: (overall: 1.73/h) DL: NA Patient-specific: Sens: (overall: 76.5%) PPV: (overall: 3.7%) FAR: (overall: 1.09/h) DL: NA Adaptive: Sens: (overall: 77.1%) PPV: (overall: 3.3%) FAR: (overall: 1.24/h) DL: NA

Table 1 (Continued)

Study	Autonomic parameter	Measurement	Device	Algorithm	Prospective/retrospective validation	Population (n)	No. of seizures/TRT <sup>a</sup>	Type of seizures (n)	Mean age (y) [range]	Performance (mean of values per patient)
Fujiwara et al. <sup>23</sup>	Cardiac	HRV	ECG	<i>Time domain analysis</i> : mean NN, SDNN, RMSSD, TP, NN50. <i>Frequency domain analysis</i> : LF, HF, LF/HF. Analysis over 2-5 min. <i>Algorithm</i> ≥ T <sub>i</sub> and Q statistics exceed limit > 10 s continuously	Retrospective	Refractory focal epilepsy (14)	11/ 69 h	FOS, awakening seizures (11)	30.6 [14-63]	T <sub>i</sub> (Overall): Sens: 55% PPV: NA FAR: 1.2/h DL: - 52.4 ± 216 s Q (Overall): Sens: 91% PPV: NA FAR: 0.7/h DL: - 49.4 ± 262 s
Jeppesen et al. <sup>24</sup>	Cardiac	HRV	Single-lead ECG	<i>Noise filtering</i> : High-pass filter + manual edit. Automatic R-peak detection. <i>Lorenz plot<sup>b</sup> analysis</i> : SD1, SD2, CSI, mCSI, CVI	Retrospective	Temporal lobe epilepsy (19)	11/ 13 h	FOIA (11)	NA	Sens: 88% (CSI-30), (overall: 73%, CSI-30, mCSI-50) PPV: NA FAR: NA DL: - 5-60 s
Jeppesen et al. <sup>8</sup>	Cardiac	HRV	Single-lead ECG	<i>Noise filtering</i> : High-pass filter + manual edit. Automatic R-peak detection. <i>Frequency domain analysis</i> : HF-power (using FFT) HR-diff. <i>Lorenz plot<sup>b</sup> analysis</i> : CSI, mCSI	Retrospective	Focal epilepsy (17)	47/ ±27 h	FOS (44), FOBTC (3)	39 [20-55]	Sens: 81% (mCSI-100) (overall: 74%, mCSI-100) PPV: NA FAR: NA DL: 16 s [6-50]

Table 1 (Continued)

Study	Autonomic parameter	Measurement	Device	Algorithm	Prospective/retrospective validation	Population (n)	No. of seizures/TRTs	Type of seizures (n)	Mean age (y) [range]	Performance (mean of values per patient)
Moridani et al. <sup>27</sup>	Cardiac	HRV	ECG	R-peak detection by Pan & Tompkins' algorithm <i>Time domain analysis:</i> SDNN, RMSSD, NN50, pNN50 <i>Frequency domain analysis:</i> LF, HF, VLF, LF/HF <i>Poincaré plot analysis:</i> SD1, SD2, SD2/SD1	Retrospective	Focal epilepsy (7)	11/±6 h	NA	NA	Sens: (overall: 88.3%) PPV: NA FAR: NA DL: NA
Pavei et al. <sup>28</sup>	Cardiac	HRV	ECG	<i>Noise filtering:</i> Visual artifact inspection, high- and low-pass Butter-worth filters. QRS detection algorithm by Kohler. <i>Time domain analysis:</i> SDNN, RMSSD <i>Frequency domain analysis:</i> LF, HF (with FFT) <i>SampEn:</i> Entropy changes. Lorenz plot analysis: CSI, CVI	Retrospective	Temporal lobe epilepsy (12)	34/171 h	FOIA (34)	34.5 SD 7.5	Sens: (overall: 94.1%) PPV: (overall: 95.6%) FAR: (overall: 0.49/h) DL: NA
Qarqae et al. <sup>24</sup>	Cardiac	HRV	Single-lead ECG	<i>Noise filtering:</i> Baseline estimation and de-noising with sparsity. QRS detection algorithm. Outlier removal, linear interpolation. <i>Time Frequency analysis:</i> MP-WVD algorithm. SVM classifier for EEG features.	Retrospective	Focal epilepsy (7)	68/NA	FOA, FOIA, FOBTC	43.6 [26-65]	ECC: Sens: 96.4% [75-100%] PPV: NA FAR: 5.4/h [1.5-9.5/h] DL: 13.1 s [8-20.5] ECG + EEG: Sens: 100% PPV: NA FAR: 1.6/h [0-3.5/h] DL: 12.3 s [3-26]

Table 1 (Continued)

Study	Autonomic parameter	Measurement	Device	Algorithm	Prospective/retrospective validation	Population (n)	No. of seizures/TRT <sup>a</sup> -h	Type of seizures (n)	Mean age (y) [range]	Performance (mean of values per patient)
Vandecasteele et al. <sup>36</sup>	Cardiac	HRV/PRV	180° eMotion Faros and Empatica E4 smart-watch	R-peak detection by Pan-Tompkins' algorithm. HRV analysis: Empatica Method of Varon. PRV analysis: Method of Lázaro. Seizure detection algorithm with SVM classifier by de Cooman. <i>Feature extraction:</i> HR <sub>peak</sub> , HR <sub>base</sub> , and STDHR <sub>base</sub>	Retrospective	Temporal lobe epilepsy (11)	47/701 h	NA	42.7 [19-67]	Wearable ECG: Sens: 64% (overall: 70%) PPV: 2.03% (overall: 2.15%) FAR: 2.35/h (overall 2.11/h) DL: NA Hospital ECG: Sens: 57% (overall: 57%) PPV: 2.22% (overall: 1.93%) FAR: 2.05/h (overall: 1.92/h) DL: NA PPG: Sens: 33% (overall:32%) PPV: 1.43% (overall: 1.12%) FAR: 1.88/h (overall:1.80/h) DL: NA
Varon et al. <sup>43</sup>	Cardiac	QRS morphology	Single-lead ECG	R-peak detection via Pan-Tompkins' algorithm. <i>Algorithm 1:</i> principal component analysis for changes in QRS morphology. <i>Algorithm 2:</i> ictal acceleration of HR quantified by using phase-rectified signal averaging (PRSA)	Retrospective	1) Children with refractory epilepsy (37) 2) Women with epilepsy (5)	1) 98 2) 10/ ± 5 h-	1) FOS (48) (28 frontal, 20 temporal) GOS (50) (29 T/TC, 11 MC, 10 absences) 2) FOS (10)	1) 9.2 [3-16] 2) [31-48]	Algorithm 1: Sens: 89.5% (F1), 86% (G1), 100% (F2) PPV: 85.7% (F1), 57.3% (G1), 52.6% (F2) FAR: NA DL: NA Algorithm 2: Sens: 100% (F1), 90% (G1), 100% (F2) PPV: 90.5% (F1), 77.5% (G1), 71.4% (F2) FAR: NA DL: NA

**Table 1 (Continued)**

Study	Autonomic parameter	Measurement	Device	Algorithm	Prospective/retrospective validation	Population (n)	No. of seizures/TRT <sup>a</sup>	Type of seizures (n)	Mean age (y) [range]	Performance (mean of values per patient)
Elmpt van et al. <sup>32</sup>	Cardiac	HR	2-lead ECG	R identification (increase signal > 250 $\mu$ V in 10 ms). <i>HR</i> : baseline (60 s before increase), pericardial period (120 s after), SD, min and max. <i>Detection threshold</i> : *zSD from baseline for $\geq$ 5 QRS complexes. Curve-fitting algorithm and onset detection algorithm	Retrospective	Severe epilepsy (10)	10/9 h	T, TC, MC, atypical absences	34.1 [21-50]	Sens: NA PPV: NA FAR: NA DL: NA HR increase in 48.1% of seizures. Great variability in sensitivity and PPV. Better performance when combined with ACC
Osorio et al. <sup>34</sup>	Cardiac	HR	ECG	True beat range determined (30-180 bpm). <i>5 s moving window</i> : RRI determination. <i>Time of beat sequence</i> (TOBS): RHR, 4 threshold values (T) and 3 duration values (D)	Retrospective	Focal onset epilepsy (81) Dataset 1 (41) 2 (40)	241/ 6935 h	FOS	NA	Lowest settings T, D Sens: 98.8% PPV: NA FAR: 9.5/h (1), 7.2/h (2) DL: NA Highest settings T, D Sens: 85.5% PPV: NA FAR: 1.1/h (1), 0.7/h (2) DL: NA

Table 1 (Continued)

Study	Autonomic parameter	Measurement	Device	Algorithm	Prospective/retrospective validation	Population (n)	No. of seizures/TRT <sup>a</sup>	Type of seizures (n)	Mean age (y) [range]	Performance (mean of values per patient)
Boon et al. <sup>31</sup>	Cardiac	HR	VNS-AspireSR	Relative HR increase > 1 s above threshold (≥ 20%, ≥ 40%, ≥ 60% above baseline HR)	Prospective	Refractory epilepsy (16)	66/NA	FOS (8), FOA (26), FOIA (31), FOBTC e (17), U (5)	39.6 SD 13.4 [19-66]	Threshold > 20%: Sens: 16/27 = 59.3% PPV: NA FAR: 7.2/h [95%CI 5.31-9.94] DL: 6.0 s [- 112-105] Threshold > 40%: Sens: 8/23 = 34.8% PPV: NA FAR: 2.7/h [95%CI 1.70-3.91] DL: 27.5 s [0-57] Threshold > 60%: Sens: 3/16 = 18.8% PPV: NA FAR: 0.5/h [95%CI 0.20-0.96] DL: 35.0 s [4-40]
Hampel et al. <sup>32</sup>	Cardiac	HR	VNS-AspireSR	Heartbeat sensitivity threshold 50% compared to baseline	Prospective	Refractory epilepsy (1)	12/66 h	FOS with hyperkinetic movements	29	Sens: 92% PPV: 8% FAR: 1.88/h (n = 126) DL: 7.4 s (±5)
Andel, van et al. <sup>36</sup>	Combined	HR, ACC	Shimmer sensor (chest) ECG+3D ACC	Algorithm 1: No. of s in which summed waveform lengths-fixed threshold within a fixed window. Detection if no. > window length/4 Algorithm 2: HR > threshold. Algorithm 3: combination of summed waveform length OR HR	Retrospective	Epilepsy (43)	86/402 hr	Major motor (86) (18 TC, 41 T, 18 HM, 9 Cluster)	Median 15 [2-65]	All seizures: Sens: 60% (A1), 56% (A2), 71% (A3) PPV: NA FAR: 0.5/h (A1), 0.3/h (A2), 0.7/h (A3). DL: NA Clinically urgent seizures: Sens: 74% (A1), 71% (A2), 87% (A3) PPV: NA FAR: 0.6/h (A1), 0.3/h (A2), 0.8 (A3). DL: NA

Table 1 (Continued)

Study	Autonomic parameter	Measurement	Device	Algorithm	Prospective/retrospective validation	Population (n)	No. of seizures/ TRT <sup>2</sup>	Type of seizures (n)	Mean age (y) [range]	Performance (mean of values per patient)
Cogan et al. <sup>2</sup>	Combined	HR, SpO <sub>2</sub> , EDA	Affective Q-curve & Nonin WristOx2 sensor (P)	Seizure pattern analysis (HR <sup>+</sup> , SpO <sub>2</sub> <sup>+</sup> , EDA <sup>+</sup> ). Biosignal algorithm. Personalized parameters (P)	Retrospective	Focal epilepsy (10)	26/ 340 h	FOIA (23), FOBTC (2), GTCS (1)	41.8 [21-64]	3 Sensors (n = 6): Sens: 100%, 100% (P) PPV: 86%, 100% (P) FAR: 0.015/h, 0.000/h (P) DL: NA
Goldenholz et al. <sup>37</sup>	Combined	HR, QTC, SpO <sub>2</sub>	Single-lead ECG, Radical-7	SpO <sub>2</sub> : ictal drop, provided it remained > 50%. Optimal balance 80-86%. Calculation of HR and QTC by Bazette method	Retrospective	Refractory epilepsy (45)	15/ 7104 h	FOS (119), FOBTC (32)	40 [14-68]	Sens: (overall: 81-94% (FOBTC), 25-36%(FOS)) PPV: NA FAR: (overall: 0.4-2.4/h) DL: NA
Heldberg et al. <sup>38</sup>	Combined	EDA, ACC	Empatica E3 wrist-band	EDA: low-pass filter, cutoff frequency of 1.5 Hz. Time windows: 10 s 50% overlap and 5 min 80% overlap. Decomposing signal (Ledalab algorithm). Feature extraction (56) of EDA and ACC. kNN (11 features) and random forest (26 features) classifiers	Retrospective	Epilepsy (8)	55/ 540 h	Motor seizures (21), nonmotor (34)	NA	kNN classifier: Sens: 76.2% (M), 97.1% (n-M) PPV: 4.6% (M), 9.7% (n-M) FAR: NA DL: NA Random forest: Sens: 90.5% (M), 85.3% (n-M) PPV: 5.6% (M), 12.3% (n-M) FAR: NA DL: NA
Onorati et al. <sup>39</sup>	Combined	EDA, ACC	Empatica E3, E4, iCALM	10 s sliding epochs (75% overlap), feature extraction, classifier, decision thresholds. (3 sets: Poh's (19), larger (46), reduced (25))	Retrospective	Epilepsy (69) /24 children, 45 adults)	55/ 5928 h	FOBTC (49), FOTC (6)	Median [4-60]	Sens: 83.6% (C1), 92.7% (C2), 94.6 (C3) PPV: 39% (C1), 50% (C2), 51% (C3) FAR: 0.29/day (C1), 0.21 (C2), 0.20 (C3) DL: 31.2 s (C1), 29.3 s (C2), 29.3 s (C3)



Table 1 (Continued)

Study	Autonomic parameter	Measurement	Device	Algorithm	Prospective/retrospective validation	Population (n)	No. of seizures/TRT <sup>a</sup>	Type of seizures (n)	Mean age (y) [range]	Performance (mean of values per patient)
Poh et al. <sup>46</sup>	Combined	EDA, ACC	Custom-built wrist-worn biosensors	10 s epochs, sliding window with 75% overlap, preprocessing, 19 time, frequency, and nonlinear features extracted to form feature vectors. SVM to classify vectors as (non) seizure. Cross-validation	Retrospective	Focal epilepsy (7)	16/688 bp	FOBTC (16)	10 SD 4.6	Nonpatient-specific: Sens: 14/16 = 88% PPV: NA FAR: 0.04/h (n = 28) DL: NA Semi-patient specific: Sens: 15/16 = 94% PPV: NA FAR: 0.04/h DL: NA

ACC accelerometry, CS/cardiac sympathetic index (SD2/SD1), CVI cardiac vagal index (log10(SD2xSD1)), DL detection latency, ECG electrocardiogram, EDA electrodermal activity, EEG electroencephalography, FAR rate false alarm rate, FFT fast-Fourier transformation, FV false negative, FOA focal onset aware seizures, FOBTC focal onset to bilateral tonic-clonic, FOIA focal onset with impaired awareness, FOS focal onset seizures, FOTC focal onset tonic-clonic, GOS general onset seizures, GTC generalized tonic-clonic, HF high frequency (0.15–0.4 Hz), HR heart rate, HR<sub>base</sub> average HR over the 60 s before the start of the HR increase, HR<sub>diff</sub> heart rate differentiation, HR<sub>peak</sub> peak HR at the end of the HR increase, HRV heart rate variability, IHR instantaneous heart rate (inverse of RRI), LFI low frequency (0.04–0.15 Hz), LO/PO-C7 leave-one-(patient)-out cross-testing, mCSI/modified CSI (SD2<sup>2</sup>/SD1), MC myoclonic, MP-IVD algorithm matching pursuit and Wigner-Ville distribution algorithm, NA not applicable, No. number, MPS non-patient-specific, PPG photoplethysmography, PRV pulse rate variability, RHR relative heart rate, RRR-R interval, RMSSD root mean square of difference in adjacent RRI, SampEn sample entropy, s seconds, SD standard deviation, Sens sensitivity, SPS semi-patient-specific, STDHR<sub>base</sub> standard deviation of the HR over the 60 s before the start of the HR increase, SVM support vector machine, TP total power, TRT total recording time, T/TC tonic or tonic-clonic, U unknown, V/HF very high frequency (0.4–0.5 Hz), VLF very low frequency (0.0001–0.04 Hz), VNS vagus nerve stimulation, Y years

<sup>a</sup>Data used for validation

<sup>b</sup>Lorenz plot=Poincaré plot

<sup>c</sup>Data used for validation

<sup>d</sup>F1 focal seizures in children, G1 generalized seizures in children (F1+G1=training set), F2 focal seizures adult, used for validation

<sup>e</sup>Specified for all seizures, only 66 analyzed

<sup>f</sup>Training and test data combined

<sup>g</sup>Percentage of evaluable data

<sup>h</sup>3525 h without seizures were also tested for false positives



## Heart rate variability (HRV)

All the HRV-focused studies performed retrospective validations.<sup>8, 22-26, 28, 30, 41, 42</sup>

Different HRV features were selected and specific feature thresholds were classified as 'ictal' or 'interictal.' Nine out of ten HRV studies applied linear analysis<sup>8, 22-25, 28, 30, 41, 42</sup> using time domain<sup>22-25, 28, 30, 41, 42</sup> and frequency domain<sup>8, 25, 28, 41, 42</sup> features. Time domain analysis focuses on the instantaneous HR; the interval between two normal QRS complexes, abbreviated to 'NN.' Different time domain features, such as the mean NN interval or the distribution of NN have been used for seizure detection. Four studies extracted and classified these time domain features using a support vector machine (SVM) classifier and validated the same HRV algorithm in different populations.<sup>22-24, 30</sup> The first retrospective study of seventeen people with temporal lobe epilepsy found a mean sensitivity of 83.2% with a FAR of 2.01/h.<sup>22</sup> The second study extracted ECG or PPG data from three different heart rate sensors worn by eleven adults with temporal lobe epilepsy.<sup>30</sup> The best performance was obtained using a wearable ECG device, with a sensitivity of 64% and a FAR of 2.35/h. A third study tested the algorithm in 28 children and showed a higher overall sensitivity (81.3%) and a lower FAR (0.75/h).<sup>23</sup> Performance, particularly FAR, improved when applying a patient-specific heuristic classifier. The latter was confirmed in the fourth study of data from nineteen people with temporal lobe epilepsy from a pre-existing epilepsy database.<sup>24</sup> The authors also proposed an adaptive seizure detection algorithm, and showed that similar results were obtained with simulated 'real-time' user feedback.

Frequency domain analysis is used to extract the frequency components of the HR signal, each with its own physiological footprint: low frequency (LF 0.04-0.15 Hz), high frequency (HF 0.15-0.40 Hz), very low frequency (VLF 0.0001-0.04 Hz), and very high frequency (VHF 0.4-0.5 Hz). Different frequencies were identified by power spectral density analysis of HRV in four studies,<sup>8, 25, 28, 41</sup> and two studies sped up this process by applying an efficiency algorithm [fast Fourier transform (FFT)].<sup>8, 28</sup> The LF/HF ratio, reflecting the balance of sympathetic and parasympathetic function, was examined in two studies.<sup>25, 41</sup>

One of these studies tested a seizure detection algorithm combining both time and frequency domain features on eleven focal seizures upon awakening.<sup>25</sup> Ten of the eleven seizures were detected prior to seizure onset (sensitivity 91%, DL – 494 ± 262 s). Another study of seven adults with focal epilepsy that used time-frequency analysis of HRV based on a combination of the matching-pursuit and Wigner-Ville distribution algorithms reported a sensitivity of 96.4% with high

FAR (5.4/h).<sup>42</sup> Combining ECG and EEG algorithms yielded better performance (sensitivity 100%, FAR 1.6/h).

To assess the dynamic properties of ictal HR changes, nonlinear analysis can be applied, such as a Lorenz (or Poincaré) plot. This method plots the current R-R interval against the next R-R value. Standard deviations in the transverse (SD1) and longitudinal (SD2) directions of these plots can be calculated, and higher ratios of SD2/SD1 reflect increased sympathetic tone. These ratios can be used in seizure detection algorithms, since an increase in sympathetic tone is often seen during the preictal and early ictal phases. One small retrospective study proposed the modified cardio sympathetic index (mCSI) as a new measure in seizure detection that reflects the sympathetic tone.<sup>26</sup> A seizure detection algorithm based on changes in mCSI yielded a sensitivity of 88% in five people with temporal lobe epilepsy (FAR not reported). A larger follow-up study of adults with focal epilepsy compared frequency domain analysis with Lorenz plot analysis.<sup>8</sup> mCSI appeared more sensitive, but FARs were not reported.

The two remaining studies of HRV combined linear and nonlinear analysis.<sup>28, 41</sup> The first retrospective study of seven people with focal epilepsy reported an overall sensitivity of 88.3% with a specificity of 86.2% after selecting an optimal performance threshold for each patient.<sup>41</sup> The second study combined time-frequency and Lorenz plot analysis with a second nonlinear analysis of 'sample entropy'.<sup>28</sup> This parameter quantifies the regularity and complexity of a time series, and entropy decreases can be seen during the ictal phase. Applying all these methods together to ECG data from twelve temporal lobe epilepsy patients resulted in overall sensitivity of 94.1% with a FAR of 0.49/h.

Another retrospective study reported two different seizure detection algorithms based on changes in QRS morphology (algorithm 1) and cardiorespiratory interactions (algorithm 2).<sup>35</sup> The first algorithm captured five consecutive QRS complexes, aligned them with respect to the R peak, and assembled them into one QRS matrix. Principal component analysis was used to select different features from this QRS matrix. This process was repeated for every heart beat, which resulted in a sensitivity of 89.5-100% for detecting focal onset seizures and 86% for generalized onset seizures. The second algorithm was based on the well-known modulatory effects of respiration on HRV. These cardiorespiratory changes were quantified using phase-rectified signal averaging — a methodology used to detect quasi-periodicities in nonstationary signals such as the resampled RR interval time series — and were used for

seizure detection. Slightly better performance was achieved by the second algorithm, which yielded a sensitivity of 100% for focal onset seizures and 90% for generalized onset seizures. In this study, 10.4-90% of the generated alarms were false, and this percentage was lower for the second algorithm.

## **Combining autonomic parameters**

All multimodal autonomic algorithms were retrospectively validated. A combination of three biosignals, measured by two different devices, was used for seizure detection in a study of ten subjects with focal epilepsy.<sup>2</sup> An algorithm based on a specific seizure pattern of increased HR, decreased SpO<sub>2</sub>, and increased EDA was able to detect all seizures in six out of ten patients with a low FAR of 0.015/h. Specific thresholds of HR, QTC, and SpO<sub>2</sub> were combined in an algorithm tested on a larger study population of 45 people with refractory epilepsy.<sup>37</sup> Only half of the collected data was used for analysis, and a sensitivity of 81-94% was found for focal to bilateral tonic-clonic seizures, while focal seizures without bilateral spreading showed worse performance, with a sensitivity of 25-36%. Overall FAR ranged from 0.4-2.4/h.

Three other retrospective validation studies combined EDA and accelerometry (ACC), measured with one device.<sup>28-40</sup> Different classifiers were used to select features of EDA and ACC. The first study tested two machine learning algorithms, the k-nearest neighbor (kNN) and random forest classifiers. The kNN classifier achieved the best results with eleven features and was most sensitive for nonmotor seizures (sensitivity 97.1%, FAR not reported). The random forest classifier selected 26 features and showed its best performance with motor seizures (sensitivity 90.5%, FAR not reported). A second study used a SVM classifier to extract 19 features (16 ACC and 3 EDA).<sup>40</sup> Fourteen out of sixteen focal onset seizures with bilateral spreading were detected (sensitivity 88%) and FAR was 0.04/h. The same feature set was used in the third study and compared to a larger (40 ACC and 6 EDA) and a reduced (22 ACC and 3 EDA) feature set.<sup>39</sup> Retrospectively tested on 24 children and 45 adults with focal epilepsy, the reduced set showed the best performance (sensitivity 94.6%, FAR 0.20/ day).

A multicenter study combined HR and ACC measures in 95 people with nocturnal major motor seizures.<sup>36</sup> Data from only 23 patients could be used to retrospectively validate three different algorithms based on changes in HR, ACC, and 'HR or ACC.' Clinically urgent seizures were detected well (sensitivity 71-87%), but FAR was relatively high (2.3-6.3/night), with wide variation between subjects.

**Table 2** Quality of the included studies according to QUADAS-2

Study	Risk of Bias				Concerns regarding applicability		
	Patient selection	Index tests	Reference standard	Flow and timing	Patient selection	Index tests	Reference standard
Van Andel et al. <sup>36</sup>	●	●	●	●	●	●	●
Boon et al. <sup>31</sup>	●	●	●	●	●	●	●
Cogan et al. <sup>2</sup>	●	●	●	●	●	●	●
De Cooman et al. <sup>22</sup>	●	●	●	●	●	●	●
De Cooman et al. <sup>23</sup>	●	●	●	●	●	●	●
De Cooman et al. <sup>24</sup>	●	●	●	●	●	●	●
Elmpt, van et al. <sup>32</sup>	●	●	●	●	●	●	●
Fujiwara et al. <sup>25</sup>	●	●	●	●	●	●	●
Goldenholz et al. <sup>37</sup>	●	●	●	●	●	●	●
Hampel et al. <sup>33</sup>	●	●	●	●	●	●	●
Heldberg et al. <sup>38</sup>	●	●	●	●	●	●	●
Jeppesen et al. <sup>26</sup>	●	●	●	●	●	●	●
Jeppesen et al. <sup>8</sup>	●	●	●	●	●	●	●
Moridani et al. <sup>27</sup>	●	●	●	●	●	●	●
Onorati et al. <sup>39</sup>	●	●	●	●	●	●	●
Osorio et al. <sup>34</sup>	●	●	●	●	●	●	●
Pavei et al. <sup>28</sup>	●	●	●	●	●	●	●
Poh et al. <sup>40</sup>	●	●	●	●	●	●	●
Qaraq et al. <sup>29</sup>	●	●	●	●	●	●	●
Vandecasteele et al. <sup>30</sup>	●	●	●	●	●	●	●
Varon et al. <sup>43</sup>	●	●	●	●	●	●	●

● low risk of bias ● unclear risk of bias ● high risk of bias.

## Quality of the included studies

According to the QUADAS-2 criteria, the overall quality of the included studies was medium-high (Table 2). Seventeen out of 21 studies were at risk of bias,

mainly due to an undefined patient selection process and fitting of the algorithm.<sup>2, 8, 22-26, 30, 32, 34, 37-43</sup> There was concern regarding the applicability of the selected patients in three studies, because the populations consisted of children only and/or were not well described.<sup>23, 25, 33</sup> Concerns about the applicability of the index test (i.e., the tested algorithm) arose in nine studies, mainly because the algorithm was fitted to one dataset.<sup>2, 8, 23, 25, 28, 30, 32, 36, 37</sup>

Based on the standards for the clinical validation of SDDs proposed by Beniczky and Ryvlin,<sup>21</sup> most studies were classified as phase 1 proof-of-principle studies, whereas three were classified as phase 0 initial studies,<sup>34, 41, 42</sup> and only one as a phase 2 study on a dedicated SDD<sup>31</sup> (Table 3). Seven other studies also tested a dedicated device but included small population sizes or did not address the safety of the device and were therefore classified as phase 1.<sup>2, 30, 33, 36, 38-40</sup> Ten studies trained and tested their algorithm on the same dataset,<sup>2, 8, 22, 26, 32, 34, 37, 40-42</sup> and only four used a predefined algorithm or cutoff values.<sup>30, 31, 33, 36</sup> Eighteen studies used video-EEG as reference standard; the remaining three used EEG or ECoG without video recordings.<sup>34, 41, 42</sup>

## DISCUSSION

The overall quality of studies on seizure detection using autonomic parameters is low. Small population sizes, short follow-up periods, and high study heterogeneity raise concerns about the applicability of the results. Available studies are mainly initial or proof-of-principle studies that lack long-term and real-time ambulatory monitoring, which is needed to obtain more reliable performance data and usability outcomes.

HR- or HRV-based algorithms are most frequently applied, but it is hard to compare the results of different studies due to wide variation in the detection techniques used and a lack of FAR data (Table 4). Additionally, FAR, when mentioned, is high for these studies and exceeds acceptable limits for daily practice. We could not compare the performance of HR- and HRV-based algorithms due to the wide variety of study designs employed. HRV-based algorithms seem attractive given their short detection latency, but they still require prospective validation. HRV is, however, situation dependent and affected by exercise, stress, respiration, and sleep stage.<sup>45-47</sup> These confounding factors make it more challenging to distinguish ictal patterns from non-ictal ones, resulting in lower accuracy.<sup>48</sup> Also, similar activation of the autonomic nervous system can occur before physiological arousal or other sleep-related movements.<sup>49</sup> Multimodal algorithms might help to lower FARs.

Table 3 Quality of validation studies of seizure detection, as assessed using standards proposed by Beniczky and Ryvlin

Study	Subjects		Recordings			Analysis & alarms			Reference standard						
	Simulation/ healthy subjects	No. of people with seizures	No. of seizures	Conventional methods	Dedicated device	Continuous	Multi-center	Offline/ retrospective	Training and testing using the dataset	Predefined algorithm & cutoff values	Real time	Blinded	Video or video- EEG recordings	Information from pt & care- givers	Study phase
Van Andel et al. <sup>16</sup>	-	20-50	≥75	-	+	+	+	+	-	+	-	-	+	-	1
Boon et al. <sup>17</sup>	-	10-20	30-75	-	+	+	+	-	-	+	+	-	+	+	2
Cogan et al. <sup>2</sup>	-	10-20	15-30	-	+	+	-	+	+	-	-	-	+	-	1
De Cooman et al. <sup>22</sup>	-	10-20	≥75	+	-	+	-	+	+	-	-	-	+	-	1
De Cooman et al. <sup>23</sup>	-	20-50	≥75	+	-	+	+	+	-	-	-	-	+	-	1
De Cooman et al. <sup>24</sup>	-	10-20	≥75	+	-	+	-	+	-	-	-	-	- <sub>b</sub>	- <sub>c</sub>	1
Van Elmpt et al. <sup>22</sup>	-	10-20	≥75	+	-	+	-	+	+	-	-	-	+	-	1
Fujiwara et al. <sup>25</sup>	-	10-20	1-15	+	-	-	+	+	-	-	-	-	+	-	1
Golden holz et al. <sup>17</sup>	-	20-50	≥75	+	-	+	-	+	+	-	-	-	+	-	1
Hampel et al. <sup>23</sup>	-	≥1	1-15	-	+	+	-	-	-	+	+	-	+	+	1
Heldberg et al. <sup>26</sup>	-	1-10	30-75	-	+	+	-	+	-	-	-	-	+	-	1
Jeppesen et al. <sup>26</sup>	-	1-10	1-15	+	-	-	-	+	+	-	-	-	+	-	1

**Table 3** (Continued)

Study	Subjects		Recordings			Analysis & alarms			Reference standard					
	Simulation/ healthy subjects	No. of people with seizures	Conventional methods	Dedicated device	Continuous	Multi-center	Offline/ retrospective	Training and testing using the dataset	Predefined algorithm & cutoff values	Real time	Blinded	Video or EEG recordings	Information from pt & caregivers	Study phase
Jeppesen et al. <sup>15</sup>	-	10-20 30-75	+	-	-	-	+	+	-	-	-	+	-	1
Moridani et al. <sup>27</sup>	-	1-10 1-15	+	-	+	-	+	+	-	-	-	- <sup>b</sup>	-	0
Onorati et al. <sup>38</sup>	-	≥50 30-75	-	+	+	+	+	-	-	-	-	+	-	1
Osorio et al. <sup>34</sup>	-	≥50 ≥75	+	-	+	+	+	+	-	-	-	- <sup>d</sup>	-	0
Pavei et al. <sup>28</sup>	-	10-20 30-75	+	-	-	-	+	-	-	-	-	+	-	1
Poh et al. <sup>40</sup>	-	1-10 15-30	-	+	-	-	+	+	-	-	-	+	-	1
Qarage et al. <sup>29</sup>	-	1-10 30-75	+	-	-	-	+	+	-	-	-	- <sup>b</sup>	-	0
Vandecasteele et al. <sup>36</sup>	-	10-20 30-75	+	+	+	-	+	-	+	-	-	+	-	1
Varon et al. <sup>39</sup>	-	20-50 ≥75	+	-	-	-	+	-	-	-	-	+	-	1
et al. <sup>43</sup>	-													

Phase 0: initial studies performed when initiating or developing a novel method. Phase 1: proof-of-principle studies. Phase 2: studies of a dedicated seizure detection device. Phase 3: studies allowing the final confirmation of safety and accuracy. Phase 4: in-field studies of seizure detection devices in the home environments of the patients, addressing aspects related to usability

No. number, *pt* patient

<sup>a</sup>Two different devices combined

<sup>b</sup>Available database with EEG recordings

<sup>c</sup>Simulated real-time feedback on detections

<sup>d</sup>ECoG without video

**Table 4** Performance of seizure detection algorithms grouped according to dataset size

Study	Validation of algorithm			Algorithm	Performance of algorithm			
	No. of sub-jects	No. of seizures/ TRT	Type of seizures		Sensitivity (%)	FAR	PPV (%)	DL (s) [range]
<b>Large datasets</b>								
Andel, van et al. <sup>36</sup>	23	86/402h <sup>a</sup>	All major motor <sup>b</sup>	Heart rate	60	0.5/h	NA	NA
				Movement	56	0.3/h	NA	NA
				Heart rate or movement	71	0.7/h	NA	NA
		59	Clinically urgent seizures <sup>c</sup>	Heart rate	74	0.6/h	NA	NA
				Movement	71	0.3/h	NA	NA
				Heart rate or movement	87	0.8/h	NA	NA
De Cooman et al. <sup>22</sup>	17	127/918h	FOS, including TCs		83.2 [50-100]	2.01/h [0.88-3.52/h]	7.9 [0.4-21]	13.3 [-18.2 to 54.3]
De Cooman et al. <sup>23</sup>	28	107/695h	Convulsive and clinical subtle seizures	Patient-independent	Overall: 81.3	Overall: 0.75/h	NA	NA
				Patient-specific	Overall: 77.6	Overall: 0.33/h	Overall: 30.7	19.1
De Cooman et al. <sup>24</sup>	19	153/2833h	FOS, including TCs (only clinical seizures)	Patient-independent	Overall: 78.4	Overall: 1.73/h	Overall: 2.4	NA
				Patient-specific	Overall: 76.5	Overall: 1.09/h	Overall: 3.7	NA
				Adaptive	Overall: 77.1	Overall: 1.24/h	Overall: 3.3	NA
Goldenholz et al. <sup>37</sup>	45	151/7104h	FOS, including TCs		Overall: 81-94 (FOBTC) 25-36(FOS) <sup>d</sup>	Overall: 0.4-2.4/h	NA	NA
Onorati et al. <sup>39</sup>	69	55/5928h	FOS, all TCs	Classifier 1	83.6	0.29/day	39	31.2
				Classifier 2	92.7	0.21/day	50	29.3
				Classifier 3	94.6	0.20/day	51	29.3
<b>Medium datasets</b>								
Boon et al. <sup>31</sup>	16	66/ NA	Different types of FOS, including TCs	Threshold >20%	59.3	7.2/h [95% CI 5.31-9.94]	NA	6.0 [-112 to 105]
				Threshold >40%	34.8	2.7/h [95%CI 1.70-3.91]	NA	27.5 [0- 57]
				Threshold >60%	18.8	0.5/h [95%CI 0.20-0.96]	NA	35.0 [4-40]
Heldberg et al. <sup>38</sup>	8	55/540h	Motor (M) and non-motor (nM) seizures	kNN classifier	76.2 (M) 97.1 (nM)	NA	4.6 (M) 9.7 (nM)	NA
				Random Forest	90.5 (M) 85.3 (nM)	NA	5.6 (M) 12.3(nM)	NA
Jeppesen et al. <sup>8</sup>	17	47/±27h	FOS, including TCs		81:(mCSI-100) (Overall: 74, mCSI-100)	NA	NA	16 [6-50]
Osorio et al. <sup>34</sup>	81	241/6935h	FOS	Lowest settings T,D Dataset (1) & (2)	98.8	9.5/h (1) 7.2/h (2)	NA	NA
				Highest settings T,D Dataset (1) & (2)	85.5	1.1/h (1) 0.7/h (2)	NA	NA



**Table 4** (Continued)

Study	Validation of algorithm			Algorithm	Performance of algorithm			
	No. of sub-jects	No. of seizures/ TRT	Type of seizures		Sensitivity (%)	FAR	PPV (%)	DL (s) [range]
Pavei et al. <sup>28</sup>	12	34/ 171h	FOIA		Overall: 94.1	Overall: 0.49/h	Overall: 95.6	NA
Poh et al. <sup>40</sup>	7	16/ 688h <sup>c</sup>	FOS, all TCs	Non-patient specific	88	0.04/h (n=28)	NA	NA
				Semi-patient specific	94	0.04/h	NA	NA
Qaraqe et al. <sup>29</sup>	7	68/ NA	FOS, including TCs	ECG	96.4 [75-100]	5.4/h [1.5- 9.5 /h]	NA	13.1 [8-20.5]
				ECG+EEG	100	1.6/h [0-3.5/h]	NA	12.3 [3-26]
Vandecasteele et al. <sup>30</sup>	11	47/ 701h	FOIA	Wearable ECG	64 (Overall: 70)	2.35/h (Overall: 2.11/h)	2.03 (Overall: 2.15)	NA
				Hospital ECG	57 (Overall: 57)	2.05/h (Overall: 1.92/h)	2.22 (Overall: 1.93)	NA
				PPG	33 (Overall:32)	1.88/h (Overall: 1.80/h)	1.43 (Overall: 1.12)	NA
<b>Small datasets</b>								
Cogan et al. <sup>2</sup>	6	10/ 340h	FOIA and TCs	3 Sensors Personalized	100	0.015/h	86	NA
Elmpt, van et al. <sup>32</sup>	10	104/ 9h	Motor seizures (T, TC, MC) & atypical absences		100	0.000/h	100	NA
Fujiwara et al. <sup>25</sup>	14	11/ 69h	FOS (awake)	T <sup>2</sup> statistics	Overall: 55	Overall: 1.2/h	NA	- 524 ± 216
				Q statistics	Overall: 91	Overall: 0.7/h	NA	- 494 ± 262
Hampel et al. <sup>33</sup>	1	12/ 68h	FOS with hyperkinetic movements		92	1.88/h	8	7.4 (±5)
Jeppesen et al. <sup>26</sup>	5	11/ 13h	FOIA		88 (CSI-30) (Overall: 73, CSI-30, mCSI-50)	NA	NA	-5 to 60
Moridani et al. <sup>27</sup>	7	11/ ±6h	FOS		Overall: 88.3	NA	NA	NA
Varon et al. <sup>43</sup>	42	108/ ±5h	FOS and GOS, including T, TC, MC and absences	Algorithm 1 <sup>§</sup>	89.5 (F1) 86 (G1) 100 (F2)	NA	85.7 (F1) 57.3 (G1) 52.6 (F2)	NA
				Algorithm 2 <sup>§</sup>	100(F1) 90(G1) 100(F2)	NA	90.5 (F1) 77.5 (G1) 71.4 (F2)	NA

CSI cardiac sympathetic index, DL detection latency, ECG electrocardiogram, EEG electroencephalography, FAR rate false alarm rate, FOBTc focal onset to bilateral tonic-clonic, FOIA focal onset with impaired awareness, FOS focal onset seizures, h hour, MC myoclonic, mCSI modified cardiac sympathetic index, NA not applicable, No. number, PPG photoplethysmography, s seconds, T tonic, TCs tonic-clonic seizures, TRT total recording time.

**Table 4** (Continued)

<sup>a</sup>Training and test set combined.

<sup>b</sup>Including tonic-clonic, tonic, hypermotor and cluster (series of at least five tonic or myoclonic spasms within 3 min).

<sup>c</sup>When attendance or intervention was deemed necessary, based on seizure severity, postictal arousal state, breathing difficulties, and distress.

<sup>d</sup>Percentage of evaluable data.

<sup>e</sup>Also 3525 hours without seizures tested for False positives.

<sup>f</sup>Great variability in sensitivity and PPV.

<sup>g</sup> F1: Focal seizures children, G1: generalized seizures children (F1 +G1 = training set), F2: focal seizures adult, used for validation.

A retrospective study of seven children with tonic-clonic seizures validated different unimodal and multimodal algorithms on the same dataset. All combinations of multimodal sensors, including ECG, EMG, and ACC, showed at least 75% lower FAR.<sup>50</sup> Studies differentiating outcome according to seizure type showed diverse results, indicating that that different seizure types may require different detection techniques. Multimodal techniques can provide a solution to this problem.<sup>51</sup> Another solution could be personalizing or tailoring the algorithm. One study group studied two different personalization strategies and calculated the number of seizures required for accurate tailoring.<sup>52</sup> The authors proposed an initialization phase to tailor an existing predefined algorithm to a patient-specific algorithm. Six to eight seizures seemed sufficient to set individual thresholds.<sup>52</sup> Another retrospective multicenter study proposed an automatic adaptive HRV algorithm and tested it on a database of 107 nocturnal seizures from 28 children.<sup>23</sup> After an initialization phase of five seizures, the personalized algorithm resulted in lower FARs compared to those obtained with the patient-independent algorithm. A follow-up study proposed an adaptive classifier with real-time user feedback that presented similar performance; this method might be better accepted in daily practice.<sup>24</sup>

## CONCLUSION

Autonomic function alterations seem to represent an attractive tool for timely seizure detection. Unimodal autonomic algorithms cannot, however, reach acceptable performance: while most algorithms are quite sensitive, false alarm rates are still too high. Multimodal algorithms and personalization of the algorithm are important strategies to improve performance. Larger, prospective, home-based studies with long-term follow-up are needed to validate these methods and to demonstrate the added value of SDDs in clinical care.

## References

1. Van Andel J, Thijs RD, de Weerd A, Arends J, Leijten F. Non-EEG based ambulatory seizure detection designed for home use: what is available and how will it influence epilepsy care? *Epilepsy Behav* 2016;57:82-89.
2. Cogan D, Birjandtalab J, Nourani M. Multi-biosignal analysis for epileptic seizure monitoring. *Int J Neural Syst* 2016;26:1-18.
3. Elger CE, Hoppe C. Diagnostic challenges in epilepsy: seizure under-reporting and seizure detection. *Lancet Neurol* 2018;17(3):279-288.
4. Van der Lende M, Cox FME, Visser GH, Sander JW, Thijs RD. Value of video monitoring for nocturnal seizure detection in a residential setting. *Epilepsia* 2016;57(11):1748-1753.
5. Van der Lende M, Hesdorffer DC, Sander JW, Thijs RD. Nocturnal supervision and SUDEP risk at different epilepsy care settings. *Neurology* 2018;91(16):e1508-e1518.
6. Van de Vel A, Cuppens K, Bonroy B et al. Non-EEG seizure detection systems and potential SUDEP prevention: state of the art: review and update. *Seizure* 2016;41:141-153.
7. Devinsky O. Effects of seizures on autonomic and cardiovascular function. *Epilepsy Curr* 2004;4(2):43-46.
8. Jeppesen J, Beniczky S, Johansen P, Sidenius P, Fuglsang-Frederiksen A. Detection of epileptic seizures with a modified heart rate variability algorithm based on Lorenz plot. *Seizure* 2015;24:1-7.
9. Sevcencu C, Struijk JJ. Autonomic alterations and cardiac changes in epilepsy. *Epilepsia* 2010;51(5):725-737.
10. Leutmezer F, Scherthaner C, Lurger S, Potzelberger K, Baumgartner C. Electrocardiographic changes at the onset of epileptic seizures. *Epilepsia* 2003;44(3):348-354.
11. Blumhardt LD, Smith PEM. Electrocardiographic accompaniments of temporal lobe epileptic seizures. *Lancet* 1986;1(8489):1051-1055.
12. Zijlmans M, Flanagan D, Gotman J. Heart rate changes and ECG abnormalities during epileptic seizures: prevalence and definition of an objective clinical sign. *Epilepsia* 2002;43(8):847-854.
13. Bruno E, Biondi A, Richardson MP. Pre-ictal heart rate changes : a systematic review and meta-analysis. *Seizure* 2018;55:48-56.
14. Bateman LM, Li C, Seyal M. Ictal hypoxemia in localization-related epilepsy: analysis of incidence, severity and risk factors. *Brain* 2008;131:3239-3245.
15. Lacuey N, Zonjy B, Hampson JP et al. The incidence and significance of periictal apnea in epileptic seizures. *Epilepsia* 2018;59(3):573-582.
16. Van der Lende M, Surges R, Sander JW, Thijs RD. Cardiac arrhythmias during or after epileptic seizures. *J Neurol Neurosurg Psychiatry* 2016;87(1):69-74.

## CHAPTER 2

17. Massey CA, Sowers LP, Dlouhy BJ, Richerson GB. SUDEP mechanisms: the pathway to prevention. *Nat Rev Neurol* 2014;10(5):271-282.
18. Ryvlin P, Nashef L, Lhatoo SD et al. Incidence and mechanisms of cardiorespiratory arrests in epilepsy monitoring units (MORTEMUS): a retrospective study. *Lancet Neurol* 2013;12:966-977.
19. Moher D, Shamseer L, Clarke M et al. Preferred reporting items for systematic review and meta-analysis protocols (PRISMA-P) 2015 statement. *Syst Rev* 2015;4(1):1-9.
20. Whiting PF, Rutjes AWS, Westwood ME et al. Research and reporting methods accuracy studies. *Ann Intern Med* 2011;155(4):529-536.
21. Beniczky S, Ryvlin P. Standards for testing and clinical validation of seizure detection devices. *Epilepsia* 2018;59(S1):9-13.
22. De Cooman T, Varon C, Hunyadi B et al. Online automated seizure detection in temporal lobe epilepsy patients using single-lead ECG. *Int J Neural Syst* 2017;27(7):1750022.
23. De Cooman T, Varon C, Van de Vel A et al. Adaptive nocturnal seizure detection using heart rate and low-complexity novelty detection. *Seizure* 2018;59:48-53.
24. De Cooman T, Kjaer TW, Van Huffel S, Sorensen HB. Adaptive heart rate-based epileptic seizure detection using real-time user feedback. *Physiol Meas* 2018;39:014005.
25. Fujiwara K, Miyajima M, Yamakawa T et al. Epileptic seizure prediction based on multivariate statistical process control of heart rate variability features. *IEEE Trans Biomed Eng* 2016;63(6):1321-1332.
26. Jeppesen J, Beniczky S, Johansen P, Sidenius P, Fuglsang-Frederiksen A. Using Lorenz plot and cardiac sympathetic index of heart rate variability for detecting seizures for patients with epilepsy. *IEEE Eng Med Biol Soc* 2014;4563-4566.
27. Moridani MK, Farhadi H. Heart rate variability as a biomarker for epilepsy seizure prediction. *Bratisl Lek Listy* 2017;118(1):3-8.
28. Pavei J, Heinzen RG, Novakova B et al. Early seizure detection based on cardiac autonomic regulation dynamics. *Front Physiol* 2017;8:1-12.
29. Qaraqe M, Ismail M, Serpedin E, Zulfi H. Epileptic seizure onset detection based on EEG and ECG data fusion. *Epilepsy Behav* 2016;58:48-60.
30. Vandecasteele K, De Cooman T, Gu Y et al. Automated epileptic seizure detection based on wearable ECG and PPG in a hospital environment. *Sensors* 2017;17(2338):1-12.
31. Boon P, Vonck K, van Rijckevorsel K et al. A prospective, multicenter study of cardiac-based seizure detection to activate vagus nerve stimulation. *Seizure* 2015;32:52-61.
32. Van Elmpt WJC, Nijsen TME, Griep PAM, Arends JBAM. A model of heart rate changes to detect seizures in severe epilepsy. *Seizure* 2006;15:366-375.

33. Hampel KG, Vatter H, Elger CE, Surges R. Cardiac-based vagus nerve stimulation reduced seizure duration in a patient with refractory epilepsy. *Seizure* 2015;26:81-85.
34. Osorio I. Automated seizure detection using EKG. *Int J Neural Syst* 2014;24(2):1450001.
35. Varon C, Jansen K, Lagae L, Van Huffel S. Can ECG monitoring identify seizures? *J Electrocardiol* 2015;48(6):1069-1074.
36. Van Andel J, Ungureanu C, Arends J et al. Multimodal, automated detection of nocturnal motor seizures at home: is a reliable seizure detector feasible? *Epilepsia Open* 2017;2(4):424-431.
37. Goldenholz DM, Kuhn A, Austermuehle A et al. Long-term monitoring of cardiorespiratory patterns in drug-resistant epilepsy. *Epilepsia* 2017;58(1):77-84.
38. Heldberg BE, Kautz T, Leutheuser H. Using wearable sensors for semiology-independent seizure detection—towards ambulatory monitoring of epilepsy. *IEEE Eng Med Biol Soc* 2015;5593-5596.
39. Onorati F, Regalia G, Caborni C et al. Multicenter clinical assessment of improved wearable multimodal convulsive seizure detectors. *Epilepsia* 2017;1:1-10.
40. Poh MZ, Loddenkemper T, Reinsberger C et al. Convulsive seizure detection using a wrist-worn electrodermal activity and accelerometry biosensor. *Epilepsia* 2012;53(5):93-97.
41. Moridani M, Farhadi H. Heart rate variability as a biomarker for epilepsy seizure prediction. *Bratisl Med J* 2017;118(1):3-8.
42. Qaraqe M, Ismail M, Serpedin E, Zul H. Epileptic seizure onset detection based on EEG and ECG data fusion. *Epilepsy Behav* 2016;58:48-60.
43. Varon C, Jansen K, Lagae L, Van Huffel S. Can ECG monitoring identify seizures? *J Electrocardiol* 2015;48(6):1069-1074.
44. Osorio I, Manly BFJ. Probability of detection of clinical seizures using heart rate changes. *Seizure* 2015;30:120-123.
45. Bernardi L, Wdowczyk-szulc J, Valenti C et al. Effects of controlled breathing, mental activity and mental stress with or without verbalization on heart rate variability. *J Am Coll Cardiol* 2000;35(6):1462-1469.
46. Niemeijer ND, Corssmit EP, Reijntjes RH et al. Sleep-mediated heart rate variability after bilateral carotid body tumor resection. *Sleep* 2015;38(4):633-639.
47. Karemaker JM. An introduction into autonomic nervous function. *Physiol Meas* 2017;38(5):R89-R118.
48. Delamont RS, Walker MC. Pre-ictal autonomic changes. *Epilepsy Res* 2011;97(3):267-272.

## CHAPTER 2

49. Gilfriche P, Arsac LM, Daviaux Y et al. Highly sensitive index of cardiac autonomic control based on time-varying respiration derived from ECG. *Am J Physiol Regul Integr Comp Physiol* 2018;315(3):R469-R478.
50. De Cooman T, Varon C, Van De Vel A et al. Comparison and combination of electrocardiogram, electromyogram and accelerometry for tonic-clonic seizure detection in children. In: *Proc 2018 Int Conf on Biomedical and Health Informatics*, Las Vegas, NV, USA, 4-7 March 2018.
51. Leijten FSS. Multimodal seizure detection: a review. *Epilepsia* 2018;59(S1):42-47.
52. Cogan D, Heydarzadeh M, Nourani M. Personalization of nonEEG-based seizure detection systems. *IEEE Eng Med Biol Soc* 2016;6349-6352.