



Universiteit
Leiden
The Netherlands

3D magnetic resonance fingerprinting on a low-field 50 mT point-of-care system prototype: evaluation of muscle and lipid relaxation time mapping and comparison with standard techniques

O'Reilly, T.; Börnert, P.; Liu, H.Y.; Webb, A.; Koolstra, K.

Citation

O'Reilly, T., Börnert, P., Liu, H. Y., Webb, A., & Koolstra, K. (2023). 3D magnetic resonance fingerprinting on a low-field 50 mT point-of-care system prototype: evaluation of muscle and lipid relaxation time mapping and comparison with standard techniques. *Magnetic Resonance Materials In Physics, Biology And Medicine*, 36(3), 499-512.
doi:10.1007/s10334-023-01092-0

Version: Publisher's Version

License: [Creative Commons CC BY 4.0 license](https://creativecommons.org/licenses/by/4.0/)

Downloaded from: <https://hdl.handle.net/1887/3720963>

Note: To cite this publication please use the final published version (if applicable).



3D magnetic resonance fingerprinting on a low-field 50 mT point-of-care system prototype: evaluation of muscle and lipid relaxation time mapping and comparison with standard techniques

Thomas O'Reilly¹ · Peter Börnert^{1,3} · Hongyan Liu⁴ · Andrew Webb¹ · Kirsten Koolstra²

Received: 9 March 2022 / Revised: 11 April 2023 / Accepted: 17 April 2023 / Published online: 18 May 2023
© The Author(s) 2023

Abstract

Objective To implement magnetic resonance fingerprinting (MRF) on a permanent magnet 50 mT low-field system deployable as a future point-of-care (POC) unit and explore the quality of the parameter maps.

Materials and methods 3D MRF was implemented on a custom-built Halbach array using a slab-selective spoiled steady-state free precession sequence with 3D Cartesian readout. Undersampled scans were acquired with different MRF flip angle patterns and reconstructed using matrix completion and matched to the simulated dictionary, taking excitation profile and coil ringing into account. MRF relaxation times were compared to that of inversion recovery (IR) and multi-echo spin echo (MESE) experiments in phantom and in vivo. Furthermore, B_0 inhomogeneities were encoded in the MRF sequence using an alternating TE pattern, and the estimated map was used to correct for image distortions in the MRF images using a model-based reconstruction.

Results Phantom relaxation times measured with an optimized MRF sequence for low field were in better agreement with reference techniques than for a standard MRF sequence. In vivo muscle relaxation times measured with MRF were longer than those obtained with an IR sequence (T_1 : 182 ± 21.5 vs 168 ± 9.89 ms) and with an MESE sequence (T_2 : 69.8 ± 19.7 vs 46.1 ± 9.65 ms). In vivo lipid MRF relaxation times were also longer compared with IR (T_1 : 165 ± 15.1 ms vs 127 ± 8.28 ms) and with MESE (T_2 : 160 ± 15.0 ms vs 124 ± 4.27 ms). Integrated ΔB_0 estimation and correction resulted in parameter maps with reduced distortions.

Discussion It is possible to measure volumetric relaxation times with MRF at $2.5 \times 2.5 \times 3.0$ mm³ resolution in a 13 min scan time on a 50 mT permanent magnet system. The measured MRF relaxation times are longer compared to those measured with reference techniques, especially for T_2 . This discrepancy can potentially be addressed by hardware, reconstruction and sequence design, but long-term reproducibility needs to be further improved.

Keywords Fingerprinting · Low field · Quantitative MRI · Halbach magnet · Matrix completion · ΔB_0 map estimation

✉ Kirsten Koolstra
K.Koolstra@lumc.nl

¹ Radiology, C.J. Gorter Center for MRI, Leiden University Medical Center, Albinusdreef 2, 2333 ZA Leiden, The Netherlands

² Radiology, Division of Image Processing, Leiden University Medical Center, Albinusdreef 2, 2333 ZA Leiden, The Netherlands

³ Philips Research, Röntgenstraße 24-26, 22335 Hamburg, Germany

⁴ Computational Imaging Group for MR Diagnostics & Therapy, Center for Imaging Sciences, University Medical Center Utrecht, Heidelberglaan 100, 3584 CX Utrecht, The Netherlands

Introduction

Low-field MRI (<0.1 Tesla) has seen a resurgence in interest in recent years as a means of making MRI more affordable and accessible [1]. An obvious disadvantage of low-field MRI systems is the reduction in signal-to-noise ratio (SNR) compared to their high-field counterparts due to the supra-linear relationship between B_0 field strength and SNR [2]. Some of the signal loss caused by the reduced magnetization and Larmor frequency can be recovered through efficient sequence design [3]; T_1 times are significantly reduced at lower field strength [4, 5], T_2 times are either the same or slightly longer [3, 6], and SAR is (ordinarily) not a limiting factor [7, 8]. One of the challenges with sequence

optimization for low field is that while relaxation times are reported in healthy tissue, knowledge of tissue properties in the diseased state is less well known.

The reduction in system cost of a low-field MRI scanner has the potential to change the role of MRI in the healthcare system. MRI has been shown to be an effective method for screening for (early) disease, but the high cost associated with an MR examination typically limits the availability to people with an elevated risk of disease development [9–15]. The lower cost of low-field MR combined with increased implant safety [8] and reduced implant-induced image artefacts [8, 16] means that screening can be provided to a much larger fraction of the population. Efficient quantitative acquisition strategies may potentially play an important role in this [17, 18]. An additional advantage of quantitative methods is that they allow post-acquisition image synthesis of many different image contrasts using a single data set. This can lead to greater scan automation [19] and could potentially reduce the need for trained operators [1, 20]. Furthermore, it is possible to design low-field systems that are light weight and portable, making them deployable as point-of-care (POC) units [21].

Magnetic resonance fingerprinting (MRF) is a quantitative technique that can rapidly acquire multiple tissue parameters simultaneously [22]. The use of a variable flip angle pattern in a steady-state sequence results in unique signal evolutions for different tissue types, which can be quantified by matching the measured signal to a pre-calculated dictionary. The technique has been used in many clinical applications and has also potential for synthetic MR [23, 24]. However, so far most of the work on MRF has been performed on clinical scanners. Recently, Sarracanie et al. have shown that MRF is feasible at 100 mT using a fixed permanent magnet-based system with sufficient B_0 homogeneity to enable balanced SSFP sequences to be run [25] and on a 6.5 mT system as well [26]. While susceptibility and chemical shift pose no significant issues at low field due to their scaling with B_0 field strength and B_1 distribution is unaffected by anatomy due to the much longer RF wavelength, other issues do arise. In particular, B_0 homogeneity is often degraded [27, 28], especially in POC systems that are also susceptible to B_0 field drift [29]. Furthermore, SNR is reduced because the lower B_0 field strength, gradient strength and slew rate are reduced compared to standard clinical systems, and parallel imaging is typically not possible as only a single receive coil is used since there is little SNR to be gained by using coil arrays in the coil noise dominated regime [30, 31].

In this work, we implement 3D MRF on an in-house developed 50 mT permanent magnet-based low-field MRI system and explore the quality of the parameter maps. As a first application, we focus on muscle and lipid measurements of the lower limb, which has potential future applications in nutritional assessment in underserved communities. We use

a Cartesian slab-selective spoiled steady-state with free precession (SSFP) sequence to reduce the sensitivity to B_0 inhomogeneity and demonstrate that B_0 induced image distortions can be corrected by adding ΔB_0 encoding to the MRF sequence. We account for RF coil ringing, which distorts the RF pulse shape and has a stronger effect at low fields (due to the low sample loading effects and Larmor frequency) in the matching process and use a flip angle train optimized for the much shorter T_1 times at 50 mT compared to 1.5 and 3 Tesla. We compare the matched relaxation times to those obtained with reference techniques in phantom and in vivo experiments. Finally, we use the obtained MRF relaxation times to synthesize several MRI contrasts retrospectively, to test the feasibility of simplifying the workflow of low-field MRI systems through a one-scan-only approach.

Methods

Hardware

All data were acquired on an in-house developed 50 mT Halbach-based MRI scanner described in detail previously [27, 32], shown in Fig. 1. The total weight of this system is about 100 kg (magnet: ~70 kg, gradients: ~7 kg, amplifiers: ~20 kg). Typical linewidths, measured as the full width at half maximum of a spectrum, were around 100 Hz on the lower leg after applying linear shims using the gradient coils. Heat introduced by the body causes an f_0 drift of around 1300 Hz per hour. Therefore, f_0 was determined prior to every scan to minimize RF off-resonance effects. The heating of the magnet is spatially homogenous, and the line width does not significantly broaden between the beginning and end of the scan; we therefore perform shimming only at the start of the scan session.

A 15 cm-long, 15 cm-diameter solenoid coil with 15 turns of 0.8 mm copper wire is used for RF transmit and receive. The loaded Q factor of the coil is around 70 giving a coil bandwidth of 30 kHz, sufficient to avoid any coil bandwidth-related image shading. The RF pulse from the spectrometer is amplified by a custom-built 1 kW RF amplifier described previously [29]. Power optimization is performed at the start of every imaging experiment by recording a series of FIDs with increasing power and integrating underneath the resulting spectra. A sinusoidal function is fitted to the integrated spectra to determine the power needed for a 90° and 180° flip angle. The power optimization is performed separately for the 100 μ s excitation and the 200 μ s inversion pulses due to the nonlinear scaling caused by ringing of the RF coil. The rise and decay time of the RF pulse was around 16 μ s and is considered in the slab profile correction in the matching process. An RF shielding blanket was placed over the subject to minimize external noise coupling into the RF coil.




Fig. 1 Experimental setup. **A** The Halbach magnet consists of 23 rings filled with neodymium boron iron magnets. **B** The gradient amplifiers (left), RF amplifier (middle) and the spectrometer (right). **C** An RF solenoid coil was used for transmitting and receiving the

signal. **D** The main magnetic field was measured with a robot. Values are shown as difference in Hz with respect to the center frequency. **E** The simulated B_1^+ profile along the main axis of the solenoid

Fingerprinting implementation


MRF data were acquired with a (spoiled) FISP sequence [33, 34], using an unbalanced gradient in the readout direction implemented on a Magritek Kea2 (Aachen, Germany) spectrometer. The other two phase encoding gradients were implemented in balanced mode, and no RF spoiling was used. An MRF flip angle train was designed using Cramér–Rao-bound optimization for optimal discrimination between four different (T_1, T_2) pairs: (247,85), (184,48), (122,85), (89,90) ms corresponding to the phantom described in the next section, and to five (T_1, T_2) pairs for the in vivo experiments: (20,20), (42,42), (120,80), (160,160), (440, 440) ms. In this process, the number of flip angles was fixed to 240, TR fixed to 12 ms, and the maximum flip angle was restricted to 60° . An unoptimized reference sequence based on the one used in Ref. [34] used the same number of shots, TR and maximum flip angle. The patterns are shown in Fig. 2A. TE was set to constant values of 6 ms. These two flip angle patterns and the corresponding dictionaries were analyzed with t-SNE, using the method described in Ref. [35], to confirm that the encoding capability was higher for the optimized sequence compared to the non-optimized sequence, as shown in Fig. 2B, C. The inversion pulse length was set to 100 μ s, to ensure sufficient inversion efficiency for the entire field of view (FOV). A 200 μ s block pulse

was used to excite a 3D imaging volume during the MRF train with a slab selective gradient of 100 kHz/m applied in the head/foot direction along the main axis of the magnet. The extended phase graph formalism was used to compute a three-dimensional dictionary [36, 37] for different $T_1/T_2/B_1^+$ combinations, containing a total of 145,500 elements. T_1 values ranged from 20 to 500 ms in steps of 5 ms, and T_2 values ranged from 10 to 500 ms in steps of 5 ms. A B_1^+ fraction (defined as the fraction of the measured value from the pulse calibration) ranging from 0.05 to 1.50 in steps of 0.05 was incorporated into the dictionary calculation to support slab excitation profile correction. The relatively short T_1 times at 50 mT allowed us to set the time between repetitions of the MRF train to 0.5 s. For the relatively longer T_1 values ($T_1 \sim 250$ ms), a steady state was reached after a single MRF train. This train contributed to less than 1% of the total signal for the adopted 3D Cartesian readout, such that any remaining steady-state effects on the matched parameter maps can be ignored.

Phantom construction

The phantom was constructed out of 110-mm-diameter, 80-mm-long plexiglass tube, with 3 additional 30-mm-diameter plexiglass tubes placed inside. The main tube structure was filled with a ‘reference liquid’ ($T_1 = T_2 = \sim 90$ ms);

Fig. 2 MRF flip angle pattern and k-space sampling pattern. **A** The optimized flip angle pattern (blue) and standard flip angle pattern (red) used for the experiments. **B** The three-dimensional embeddings of the corresponding dictionaries obtained with t-SNE (top) and the corresponding color-coded dictionary maps in T_1/T_2 space (bottom). **C** The similarity maps for fixed T_2 values (shown on the left) show a steeper diagonal structure for the optimized sequence compared to the standard sequence, confirming improved encoding capability. Similarity maps are shown for the T_1 ranges corresponding to each fixed T_2 , i.e., from top to bottom: $T_2 = 50$: $T_1 = 55\text{--}500$, $T_2 = 100$: $T_1 = 105\text{--}500$, $T_2 = 200$: $T_1 = 205\text{--}500$, $T_2 = 300$: $T_1 = 305\text{--}500$



the smaller tubes were filled with tissue mimicking liquids made from a mixture of agarose, copper sulfate and water with the following concentrations: (1) white matter: 1.6 mmol/L CuSO₄, 1% agarose by mass, (2) muscle: 2 mmol/L CuSO₄, 2% agarose by mass (3) lipid: 4 mmol/L CuSO₄, 0.5% agarose by mass, (4) background: 6 mmol/L CuSO₄, no agarose. This resulted in (T_1, T_2) combinations corresponding to (1) white matter: $T_1/T_2 = 247/85$ ms, (2) muscle: $T_1/T_2 = 184/48$ ms, (3) lipid: $T_1/T_2 = 122/85$ ms and (4) background: $T_1/T_2 = 89/90$ ms. These T_1/T_2 values were confirmed with spectroscopic measurements.

MR Data acquisition

Experiments were performed in a phantom and in 7 healthy volunteers after informed consent was obtained, conforming to the local ethical regulations. The data from one volunteer were corrupted and was not included in the analysis. MRF scans were acquired with a Cartesian sampling scheme in undersampled mode, acquiring a 4×4 fully sampled center region, while acquiring the rest of k-space in a random fashion, according to the sampling scheme shown in Fig. 3B.

Phantom scans used the following parameters: FOV = $150 \times 150 \times 100$ mm³, resolution = $2.5 \times 2.5 \times 5.0$ m m³, TE/TR = 6/12 ms, imaging bandwidth (BW) = 333 Hz/pixel, total undersampling factor = 7.5, scan time = 9 min. In vivo scans used the following scan parameters:

FOV = $170 \times 150 \times 99$ mm³, resolution = $2.5 \times 2.5 \times 3.0$ mm³, TE/TR = 6/12 ms, imaging bandwidth (BW) = 294 Hz/pixel, total undersampling factor = 8.6, scan time = 13 min and 6 s. In the phantom experiment, the MRF scan was acquired with the optimized flip angle pattern as well as with a more standard flip angle pattern used at high field in Ref. (38) for comparison. Spectroscopic reference measurements were obtained in each tube individually using six inversion times (25/50/75/100/150/300 ms) for T_1 values and Carr–Purcell–Meiboom–Gill (echo times 20/40/60/80/100/120/140/160/180/200 ms) for T_2 values. In one of the volunteers and in the phantom, standard quantitative (imaging) techniques were used to produce reference relaxation time maps on the same FOV as the MRF scans. Inversion recovery (IR) was used for T_1 mapping: resolution = $2.5 \times 2.5 \times 8.3$ mm³, TE/TR = 12/900 ms, echo train length (ETL) = 5, inversion times = 25/50/75/100/150/300 ms, imaging bandwidth (BW) = 294 Hz/pixel, scan time = 12 min and 58 s. Multi-echo-spin-echo (MESE) was used for T_2 mapping: resolution = $2.5 \times 2.5 \times 8.3$ mm³, TR = 1250 ms, TEs = 20/40/60/80/100/120/140/160/180/200 ms, imaging bandwidth (BW) = 294 Hz/pixel, scan time = 15 min. In the phantom experiment, reference techniques were acquired on the same resolution as the MRF scans. Finally, in one of the volunteers, an alternating TE pattern ($\Delta TE = 150$ μ s) was implemented along the MRF train to support MRF-integrated ΔB_0 estimation. A TSE-based B_0 map [39]




Fig. 3 **A** Schematic overview of the MRF pulse sequence. This includes the time from the start of the inversion pulse until the end of the first TR. This is a spoiled slab-selective 3D SSFP sequence, with unbalanced gradients in the readout direction. **B** 4D sampling

scheme. For each MRF frame, the $c \times c \times n_{\text{readout}}$ center region of k-space was always acquired and used to estimate the rank of the MRF data. Note that the readout dimension (not shown) was always fully sampled. **C** Processing pipeline for the MRF data

was acquired on the same FOV for comparison using the following scan parameters: resolution = $2.5 \times 2.2 \times 3.0$ mm³, TE/TR = 12/300 ms, readout gradient shift = 150 μ s, echo train length (ETL) = 5, imaging bandwidth (BW) = 294 Hz/pixel, scan time = 9 min.

Reconstruction of MRF images

Full MRF k-space data were reconstructed from undersampled k-space data using a matrix completion-based reconstruction [38, 40]. The fully sampled lines in the 4×4 center of the 3D k-space matrix were used as calibration data. The 5 most significant singular values obtained with the singular value decomposition (SVD) were used to estimate the rank of the MRF data and to form a projection matrix; 50 iterations were used to ensure convergence. After reconstruction, the 3D k-space data were filtered using a 3D sine-bell squared filter to reduce noise.

For one of the volunteers, a ΔB_0 map was estimated from the alternating TE pattern (TE₁ = 6 ms, TE₂ = 6.15 ms) in the MRF sequence, as shown in Online Resource 3A. This was done by first reconstructing the time series corresponding to each TE with matrix completion. After this, two averaged

images were computed for TE₁ and TE₂, containing a phase difference introduced by B_0 , from which ΔB_0 was reconstructed using total variation regularization and spherical harmonic decomposition. This ΔB_0 map was used in a model-based reconstruction framework to correct for image distortions [39] after which the combined MRF images were matched to the original dictionary. This processing pipeline is schematically shown in Online Resource 3B, where the gray boxes indicate the additional processing steps that are performed for the MRF data acquired with the alternating TE pattern.

Matching process

The shape of the RF pulse in the time domain was measured using an oscilloscope connected to a pickup coil inside the transmit coil. The corresponding excitation profile was calculated using the Shinnar–Le–Roux (SLR) algorithm, resulting in an inhomogeneous excitation profile along the bore, as plotted in Fig. 1E. This excitation profile was considered in the matching process as follows. For each slice, a sub-dictionary was selected containing only the dictionary elements with an excitation fraction (B_1^+ fraction between 0.05 and

1.50) equal to that of the corresponding slice. Second, each slice was matched to its own sub-dictionary. B_1^+ variations due to wave interference were ignored, since these have a minimal effect for the solenoid RF coil used at 50 mT. The matching was performed in 3D after normalizing the dictionary entries, using the inner product as similarity measure. A schematic overview of the processing pipeline can be found in Fig. 3C.

Synthetic MRI

The matched T_1 and T_2 maps obtained with MRF were used to compute synthetic MRI images corresponding to the IR and MESE sequences. This was done by substituting the matched T_1 and T_2 maps in the signal equations.

$$S_{IR}(TI) = M_0 \left(1 - 2e^{-\frac{TI}{T_1}} + e^{-\frac{2TI}{T_1}} \right) \text{ and } S_{MESE}(TE) = M_0 e^{-\frac{TE}{T_2}}$$

using the TI, TR and TE settings that were used for the IR and MESE measurements performed in this study. This allowed a comparison between MRF and standard quantitative measures in terms of generated synthetic images. These synthetic images show the effect of over-/underestimation of relaxation times on the MR image contrast and therefore put the accuracy of the matched MRF times into perspective.

Reconstruction and matching of the MRF data were performed in MATLAB (Mathworks Inc, Natick, MA) and run on a Windows 64-bit machine with Intel Xeon CPU E5-1620 v3 @3.5 GHz and 32 GB internal memory. The processing time for the volumetric MRF data was 29 s for matrix completion reconstruction and 6 min for matching.




Fig. 4 Relaxation time maps in a phantom: comparison between MRF and reference techniques. **A** The T_1 , T_2 and M_0 maps obtained with the optimized flip angle pattern contain less noise compared to those obtained with the standard flip angle pattern. **B** The T_1 and T_2 values obtained with the optimized sequence are much closer to the T_1 and T_2 values obtained with the reference techniques (IR and MESE) and to spectroscopic measurements, compared with those

obtained with the standard flip angle pattern. The error bars represent standard deviations computed in each of the ROIs. The standard deviation of the spectroscopic measurements (computed as the standard deviation over the residuals of the fit) was less than 1% for each sample measured individually, which is much lower than that obtained with the imaging sequences (standard and MRF)

Results

Figure 4 shows the T_1 , T_2 and M_0 maps measured with the optimized and the non-optimized flip angle patterns in a phantom compared to the reference techniques (IR and MESE) and compared to spectroscopic measurements. The maps acquired with the optimized flip angle pattern contain much less noise compared to those measured with the non-optimized one, and the relaxation time values are much closer to those measured with the reference techniques. This is confirmed by the numbers plotted in Fig. 4B, showing the relaxation times averaged in an ROI for each of the four phantom compartments, with a maximum T_1/T_2 difference of 19/55 ms. Note that the noise-like pattern in the parameter maps (hereafter referred to as noise) is not noise in the traditional sense since it is a result of many influencing

factors, including the discretized dictionary matching, the low SNR of the underlying MRF signals and the under-sampling effects superimposed in those pixels, that might impair perfect pattern matching. The standard deviation of the spectroscopic measurements (computed as the standard deviation over the residuals of the fit) was less than 1% for each sample measured individually, which is much lower than that obtained with the imaging sequences (standard and MRF). Online Resource 1 shows that the adopted acquisition/reconstruction scheme is robust to undersampling: undersampling with a factor of 7.5 results in negligible relaxation time differences.

Figure 5 shows a comparison between the optimized MRF sequence and the reference techniques in the center slice of one volunteer's calf. The muscle T_1 and T_2 times measured with MRF are longer than those measured with an IR sequence (T_1 : 182 ± 21.5 ms vs 168 ± 9.89 ms) and with




Fig. 5 Relaxation time maps in a healthy volunteer: comparison between MRF and reference techniques. T_1 , T_2 and M_0 maps obtained with MRF are in the same range as those obtained with an IR sequence (182 ± 21.5 ms vs 168 ± 9.89 ms), but the muscle T_2 times measured with MRF are longer than those measured with an MESE sequence (69.8 ± 19.7 ms vs 46.1 ± 9.65 ms). The maps obtained with the reference techniques contain less noise compared to those obtained with MRF due to larger voxel size and TSE-based

sequences. ROIs used to compute mean and standard deviations in the bone marrow and the gastrocnemius muscle are shown in red in the M_0 maps. **B** The T_1 and T_2 values averaged over an ROI in the muscle and in the bone marrow are overestimated compared to the reference techniques (IR and MESE), potentially influenced by the lower SNR in the MRF scans. Note that the field drift during the undersampled in vivo MRF scans was about 200 Hz. The error bars represent standard deviations computed in each of the ROIs

MESE (T_2 : 69.8 ± 19.7 ms vs 46.1 ± 9.65 ms), respectively. In the bone marrow, T_1 and T_2 times are also longer with MRF compared with IR (T_1 : 165 ± 15.1 ms vs 127 ± 8.28 ms) and with MESE (T_2 : 160 ± 15.0 ms vs 124 ± 4.27 ms). The noise in the MRF maps is higher, due to the higher resolution of MRF ($2.5 \times 2.5 \times 3.0$ mm³) compared to reference techniques ($2.5 \times 2.5 \times 8.3$ mm³), the high undersampling factor for MRF ($R=8.6$) compared to reference techniques ($R=1$) and the different type of sequences used (gradient echo vs spin echo).

The volumetric MRF maps along the entire imaging stack for another volunteer are shown in Fig. 6. The estimated T_1 and T_2 values are consistent along the bore in the volume of the RF coil where the B_1^+ is relatively homogeneous, with the noise amplitude increasing toward the edges of the excited volume as expected. Online Resource 2 presents the same maps, but without excitation profile correction, showing increasing apparent T_2 values and decreasing apparent M_0 values away from the center slice, thus showing the importance of excitation profile correction. Table 1 summarizes the MRF relaxation times (mean \pm standard deviation)




Fig. 6 Volumetric relaxation time maps in one healthy volunteer. The T_1 (A), T_2 (B) and M_0 (C) maps after excitation profile correction show consistent values along the slice dimension. The slices toward

the edges of the excited region show more noise due to lower maximum flip angles used as well as a lower receive coil sensitivity in these regions

Table 1 Relaxation times in 6 volunteers

Vol. #	Muscle	Bone marrow
T₁ (ms)		
1	191 ± 27.4	161 ± 26.4
2	157 ± 15.0	168 ± 14.3
3	177 ± 24.6	147 ± 16.9
4	185 ± 19.8	165 ± 18.6
5	182 ± 21.5	165 ± 15.1
6	200 ± 35.3	155 ± 24.8
Mean ± std	182 ± 14.6	160 ± 7.89
T₂ (ms)		
1	69.6 ± 29.4	131 ± 39.7
2	61.9 ± 12.4	159 ± 17.3
3	52.5 ± 15.9	141 ± 17.9
4	70.9 ± 17.5	152 ± 28.0
5	69.8 ± 19.7	160 ± 15.0
6	72.3 ± 33.4	150 ± 25.2
Mean ± std	66.2 ± 7.63	149 ± 11.0

Mean ± standard deviation values (ms) in a region of interest are reported for the muscle and the bone marrow

measured in an ROI in the muscle and in the bone marrow in each of the 6 volunteers, with overall mean T_1 and T_2 values of 182 ms and 66.2 ms in the muscle and 160 ms and 149 ms in the bone marrow.

One way to reduce the total MRF scan time is by reducing the spatial resolution of the MRF data. However, in inhomogeneous B_0 fields this leads to increased intra-voxel dephasing. Figure 7 shows this effect in one volunteer: reconstructing the same data with a twice as coarse a resolution in the readout dimension results in a large reduction in signal in the left part of the calf, where the B_0 inhomogeneities are strongest. This can be observed in the M_0 map and leads to a larger amount of noise in the corresponding regions in the T_1 and the T_2 maps.

Using an alternating TE pattern along the MRF train results in very similar parameter maps compared to using a constant TE, as shown in Online Resource 3. Estimating a ΔB_0 map from the alternating TE pattern and using it in a model-based MRF reconstruction result in matched parameter maps with slightly reduced distortions. This effect would be more pronounced in case of a stronger inhomogeneous ΔB_0 field. This experiment also shows that the estimated MRF ΔB_0 map is in the same range as that estimated from a TSE sequence, with a maximum error of 105 Hz.

Figure 8 shows that the MRF relaxation time maps can be used to generate synthetic MRI images, in which the MRF parameter matching step effectively serves as a noise filtering step. Online Resource 4 shows a comparison with all the source data, also demonstrating the effect of the overestimation of MRF relaxation times on the synthetic data.


Discussion

In this work, we implemented MRF on a custom-built 50 mT permanent magnet system. We used a three-dimensional Cartesian readout with two-dimensional undersampling ($R=7.5-8.6$) and a matrix completion-based reconstruction to obtain volumetric parameter maps. We showed the proof-of-principle for a volumetric multi-parametric scan with a total acquisition time of ~ 13 min. The muscle T_1 times measured in vivo with MRF were slightly longer compared to those measured with an IR sequence (8.3% difference), while the estimated muscle T_2 times were much longer with MRF compared to MESE (51.4% difference). In the bone marrow, both the T_1 and the T_2 values were overestimated compared to IR (30%) and MESE (29%). Estimating a ΔB_0 map from the MRF data using an alternating TE pattern and using it in a model-based reconstruction framework resulted in parameter maps with reduced geometric distortions.

A large overestimation of T_2 values was consistently observed in muscle tissue in all healthy volunteers with respect to reference measurements in this study and compared to the literature [41]. Fully explaining relaxation time differences between MRF and other techniques has traditionally been challenging, and much previous published work has concentrated on this effort. For example, Ref. [34] found an underestimation of white matter T_2 (17.7%) while gray matter T_2 measures were within the literature range in the same subjects. Muscle MRF T_2 values were overestimated compared to reference techniques, which was in part explained by the presence of fat and flow [42, 43], but even after taking these into consideration a considerable discrepancy remained (37%). Correcting for B_1^+ and slice profile effects improved parameter quantification in phantoms, but generally still resulted in underestimation of T_2 values compared to the literature in vivo (gray matter: ~43%, white matter: 41%) [44].

Accurate and reproducible relaxation time measurements using MR fingerprinting are particularly challenging on a custom-built low-field system such as used here. In addition to the intrinsic low SNR, many of the electronic components have been designed with total cost being an important criterion. This means that their specifications in terms of temporal stability (magnitude and phase) and reproducibility are certainly not as well characterized as the clinical systems from the major vendors. In addition, features such as sophisticated eddy current characterization and compensation, feedback control of RF and gradient systems and continuous performance monitoring are either not present or must be designed from scratch. Although the lack of such features may not manifest themselves in standard low-field imaging, they do become critical when data processing algorithms such as in MR fingerprinting, which depend on fitting

Fig. 7 Effect of intra-voxel dephasing on the quality of the MRF parameter maps. The M_0 map matched from the low-resolution ($2.5 \times 5.0 \times 3.0 \text{ mm}^3$) MRF data shows a large area with reduced signal at the left side of the calf due to intra-voxel dephasing. This results in more noise in the corresponding regions in the T_1 and T_2 maps compared to the high-resolution maps ($2.5 \times 2.5 \times 3.0 \text{ mm}^3$)



routines which assume perfect (or perfectly characterized) performance, are used. As an example, in our early work [45] we already found that our T_2 values were significantly longer than those measured by conventional multi-echo fully sampled sequences.

We have performed numerous experiments as an attempt to understand the discrepancy in our study. We determined that one factor was the particular transmit–receive switch used, which in low-field systems is usually a simple lumped element quarter-wavelength equivalent with passive diodes, unlike the more sophisticated designs on conventional clinical systems. This caused the low tip angles used at various points in the fingerprinting sequence actually having a much lower tip angle than programmed. When the transmit/receive switch was replaced, the T_2

values became much closer to the “gold standard.” A second factor, related to the SNR dependence of the fitting routine, was that most fingerprinting trains on clinical systems use a maximum tip angle of 60° due to SAR considerations. However, low field has no such restrictions, and by increasing this angle to 90° we were able to increase the contribution of stimulated echoes generated along the MRF train, which also resulted in improved fitting, as predicted by simulations (see Online Resource 6). However, we have not performed optimizations with respect to this parameter, and so it is not yet clear what effects it would have on the accuracy of T_1 and T_2 quantification. Higher tip angles could potentially further increase the T_2 sensitivity, but this needs to be further investigated. A third factor was to include the effects of the temperature-induced




Fig. 8 Synthetic MRI images of a volunteer's lower leg obtained with MRF and with reference relaxation time maps. MRF data were first averaged over three neighboring slices before matching to the dictionary to obtain relaxation time maps with a similar voxel size ($2.5 \times 2.5 \times 9.0 \text{ mm}^3$) as the reference maps ($2.5 \times 2.5 \times 8.3 \text{ mm}^3$). **A** T_2 maps measured with MESE and with MRF were used to simu-

late TSE images at echo times $TE=20 \text{ ms}$ and $TE=140 \text{ ms}$. **B** T_1 maps measured with IR and with MRF were used to simulate IR TSE images at different inversion times $TI=50 \text{ ms}$ and $TI=300 \text{ ms}$. Note that the parameter matching step served as a noise filtering step in this process. A comparison with the source reference data can be found in Online Resource 4

changes in resonant frequency into the data processing (see Online Resource 7): these are particularly important to account for in vivo scanning where shifts of several hundreds of Hertz may occur during the scan. Using higher tip angles, improved hardware and accounting for field drift have given us at best percentage errors (with respect to spectroscopy values) below 8% for T_1 and below 6% for T_2 in phantoms, with repeatability tests showing a maximum difference of 11% for T_1 and 26% for T_2 . However, the sensitivity to environmental changes and small changes in the parameter settings seems to be high, and therefore, long-term reproducibility needs to be better characterized in the future. This also requires the design of more robust, tailored MRF encoding approaches that better meet existing hardware constraints.

We showed that the effect of undersampling on the matched relaxation times was negligible using MC as reconstruction scheme. This effect could, however, be stronger when the center of k-space is not representative for the outer k-space region in terms of low-rank structure. This could be the case for example when B_0 effects are stronger, since they introduce phase changes for higher frequencies, or when gradient nonlinearities become worse for high gradient strengths. This is also depending on the order used in the sampling scheme. We used an ordered, but incoherent, sampling scheme in which the jumps between successive k-space lines in one MRF train are small, such that also eddy-current effects are minimized.

The MRF sequence used in this work was optimized for differentiating between the relaxation times expected in the phantom at low field. The final sequence shows a relatively

large number of 'peaks' in the train compared to those used at high field [46, 47]. This could be an effect of the shorter relaxation times compared to high field, or of the low flip angle number constraint, which was motivated by these short relaxation times. Since high-field MRF approaches typically use 500 to 1500 flip angles in the MRF train to achieve sufficient T_2 encoding, it is worth investigating whether a longer (optimized) MRF train can reduce the standard deviation and improve the accuracy of the T_2 measurements in this low-SNR regime.

The parameter maps obtained with MRF contained more noise compared to those obtained with reference techniques (IR and MESE), resulting in a larger standard deviation of the relaxation times. This can in part be explained by the fact that MRF data were acquired with an undersampling factor of 8.6, as opposed to fully sampled acquisitions for the reference techniques. The long scan time for the fully sampled reference techniques limited the through-plane thickness of the reference data to 8.3 mm, which is much larger than the 3.0 mm partition thickness for MRF. This could also explain the difference in level of observable structure between the MRF T_1 maps and the reference T_1 maps. An undersampling factor larger than 5 would need to be implemented for the reference scans to make the combined acquisition more time-efficient than the MRF scans, and such large undersampling for a temporal dimension of size 6 (TI) and 10 (TE) would likely lead to too much loss of information in a potential model-based reconstruction process. Online Resource 5 shows that a comparable through plane resolution for MRF and reference data (by retrospective averaging the MRF data in partition direction image-space) would reduce the noise

difference between the parameter maps considerably. The amount of noise in the reference parameter maps is, however, still lower due to the inherent difference in SNR between spin echo and spoiled gradient echo-based acquisitions. The spoiled SSFP sequence distributes the magnetization among the higher-order coherence pathways, which do not directly contribute to the sampled echoes, and is therefore not optimal in terms of SNR. Balanced SSFP sequences would result in significant banding artifacts and complete signal loss in areas of strong B_0 inhomogeneities and would therefore not be suitable in this case. Such effects are already observed to some extent with the spoiled SSFP sequence (see Fig. 7), which is known to be much less sensitive to inhomogeneous B_0 fields. Although the MRF sequence shows a lower SNR, it samples M_0 , T_1 and T_2 maps more efficiently and currently at a higher resolution than its conventional quantitative mapping counterparts. This might make MRF interesting as an efficient way of producing multiple “synthetic contrasts” from a single scan as outlined in Online Resource 4.

The strong B_0 inhomogeneities away from the center of the bore resulted in image distortions in the MRF images as well as in the matched parameter maps. These effects were corrected by incorporating an MRF-estimated ΔB_0 map in a model-based reconstruction [39]. The maximal difference of 105 Hz between the MRF and TSE-based B_0 map was potentially introduced by the difference in acquisition sequences used (gradient-echo vs spin-echo) or a temperature drift during and in between either of the scans. The remaining reduced signal area at the left side of the calf in Online Resource 3 is likely a result of (through-plane) intra-voxel dephasing, as shown also in Fig. 7, for which the current model-based reconstruction algorithm does not correct. Future work could correct for in-plane and through-plane dephasing by incorporating a higher resolution ΔB_0 map into the reconstruction model, either by directly reconstructing the images on a higher resolution, or by including multiple ΔB_0 frequencies for each voxel in the encoding matrix [48, 49]. It should also be further investigated whether linear field drift correction, as performed in Online Resource 7, improves the accuracy of the MRF parameter maps sufficiently in all in vivo scans, where field drift may not always be approximately linear.

In conclusion, volumetric MRF parameter maps using a spoiled SSFP sequence have been acquired on a 50 mT permanent magnet system in a scan time of 13 min. While the effective resolution of the acquired maps (~ 2.5 mm) is reduced compared to conventional data acquisition methods (IR/MESE) at low field and compared to MRF at high field, sufficient resolution to delineate major anatomical structures is still achieved. The MRF parameter maps were slightly overestimated with respect to T_1 , but severely overestimated with respect to T_2 . Improved hardware components (transmit-receive switch), the use of higher

maximum flip angles or other MRF encoding approaches, and field drift correction can be impacting factors. Further work will help to understand and to control the remaining discrepancy and the reproducibility challenge, which are needed for MRF to become a practical operating sequence on very low-field point-of-care MRI systems in the future.

Supplementary Information The online version contains supplementary material available at <https://doi.org/10.1007/s10334-023-01092-0>.

Acknowledgements This work is supported by the following grants: Horizon 2020 ERC FET-OPEN 737180 Histo MRI, Horizon 2020 ERC Advanced NOMA-MRI 670629, Simon Stevin Meester Award and NWO WOTRO Joint SDG Research Programme W 07.303.101. This publication is also part of the project HTSM (with project number 17104) which is financed by the Dutch Research Council (NWO). We would furthermore like to thank Oleh Dzyubachyck, for running t-SNE analysis on the MRF dictionaries.

Author contributions Study conception and design were performed by TO'R, PB, HL, AW, KK acquisition of data by TO'R, KK; analysis and interpretation of data by KK; drafting of manuscript by TO'R, KK; critical revision by PB, AW, HL.

Funding This work is supported by the following grants: Horizon 2020 ERC FET-OPEN 737180 Histo MRI, Horizon 2020 ERC Advanced NOMA-MRI 670629, Simon Stevin Meester Award and NWO WOTRO Joint SDG Research Programme W 07.303.101. This publication is also part of the project HTSM (with project number 17104) which is financed by the Dutch Research Council (NWO).

Data availability The data that support the findings of this study are available on request from the corresponding author. The data are not publicly available due to privacy or ethical restrictions.

Declarations

Conflict of interest Thomas O'Reilly, Hongyan Liu, Andrew Webb and Kirsten Koolstra declare that they have no conflict of interest. Peter Börner is an employee of Philips Research Hamburg.

Ethical approval All procedures performed involving human participants were in accordance with the ethical standards of our institutional research committee and with the 1964 Helsinki Declaration and its later amendments.

Informed consent Informed consent was obtained from the participants included in this study.

Open Access This article is licensed under a Creative Commons Attribution 4.0 International License, which permits use, sharing, adaptation, distribution and reproduction in any medium or format, as long as you give appropriate credit to the original author(s) and the source, provide a link to the Creative Commons licence, and indicate if changes were made. The images or other third party material in this article are included in the article's Creative Commons licence, unless indicated otherwise in a credit line to the material. If material is not included in the article's Creative Commons licence and your intended use is not permitted by statutory regulation or exceeds the permitted use, you will need to obtain permission directly from the copyright holder. To view a copy of this licence, visit <http://creativecommons.org/licenses/by/4.0/>.

References

- Geethanath S, Vaughan JT (2019) Accessible magnetic resonance imaging: A review. *J Magn Reson Imaging* 49(7):e65–e77
- Hoult D, Richards R (1976) The signal-to-noise ratio of the nuclear magnetic resonance experiment. *J Mag Res* 24(1):71–85
- Marques JP, Simonis FFJ, Webb AG (2019) Low-field MRI: An MR physics perspective. *J Magn Reson Imaging* 49(6):1528–1542
- Rooney WD, Johnson G, Li X et al (2007) Magnetic field and tissue dependencies of human brain longitudinal 1H₂O relaxation in vivo. *Magn Reson Med* 57(2):308–318
- Korb J-P, Bryant RG (2002) Magnetic field dependence of proton spin-lattice relaxation times. *Magn Reson Med* 48(1):21–26
- Bottomley PA, Foster TH, Argersinger RE et al (1984) A review of normal tissue hydrogen NMR relaxation times and relaxation mechanisms from 1–100 MHz: dependence on tissue type, NMR frequency, temperature, species, excision, and age. *Med Phys* 11(4):425–448
- Bottomley PA, Edelstein WA (1981) Power deposition in whole-body NMR imaging. *Med Phys* 8(4):510–512
- Van Speybroeck CDE, O'Reilly T, Teeuwisse W et al (2021) Characterization of displacement forces and image artifacts in the presence of passive medical implants in low-field (<100 mT) permanent magnet-based MRI systems, and comparisons with clinical MRI systems. *Physica Med* 84:116–124
- Allen BD, Schiebler ML, Sommer G et al (2020) Cost-effectiveness of lung MRI in lung cancer screening. *Eur Radiol* 30(3):1738–1746
- Biederer J, Ohno Y, Hatabu H et al (2017) Screening for lung cancer: does MRI have a role? *Eur J Radiol* 86:353–360
- Kutaiba N, Ardalan Z (2021) Risk factors and screening intervals are crucial for evaluating the cost effectiveness of abbreviated MRI in HCC screening. *J Hepatol* 75(5):1258–1259
- Shah S, Luby M, Poole K et al (2015) Screening with MRI for accurate and rapid stroke treatment: SMART. *Neurology* 84(24):2438
- Eldred-Evans D, Tam H, Sokhi H et al (2020) Rethinking prostate cancer screening: could MRI be an alternative screening test? *Nat Rev Urol* 17(9):526–539
- Saadatmand S, Tilanus-Linthorst MMA, Rutgers EJT et al (2013) Cost-effectiveness of screening women with familial risk for breast cancer with magnetic resonance imaging. *J Natl Cancer Inst* 105(17):1314–1321
- Slanetz P (2021) MRI screening of women with a personal history of breast cancer. *Radiology* 300:324–325
- Schröder FF, Verdonschot NJJ, ten Haken B et al (2018) Low-field magnetic resonance imaging offers potential for measuring tibial component migration. *Journal of Experimental Orthopaedics* 5(1):1–9
- Van Beek EJR, Kuhl C, Anzai Y et al (2019) Value of MRI in medicine: more than just another test? HHS public access. *J Magn Reson Imaging* 49(7):14–25
- European Society of Radiology (ESR) (2015) Magnetic Resonance Fingerprinting - a promising new approach to obtain standardized imaging biomarkers from MRI. *Insig Imaging*. 6(2):163–165
- Ravi KS, Geethanath S (2020) Autonomous magnetic resonance imaging. *Magn Reson Imaging* 73:177–185
- Kawooya MG (2012) Training for rural radiology and imaging in sub-saharan Africa: addressing the mismatch between services and population. *J Clin Imag Sci* 2:37
- Mazurek MH, Cahn BA, Yuen MM et al (2021) Portable, bedside, low-field magnetic resonance imaging for evaluation of intracerebral hemorrhage. *Nat Commun* 12(1):1–11
- Ma D, Gulani V, Seiberlich N et al (2013) Magnetic resonance fingerprinting. *Nature* 495(7440):187–192
- Ji S, Yang D, Lee J et al (2020) Synthetic MRI: Technologies and Applications in Neuroradiology. *J Magn Res Imaging* 55:1013–1025
- Blystad I, Warntjes JBM, Smedby O et al (2012) Synthetic MRI of the brain in a clinical setting. *Acta Radiol* 53(10):1158–1163
- Sarracanie M (2021) Fast quantitative low-field magnetic resonance imaging with OPTIMUM—optimized magnetic resonance fingerprinting using a stationary steady-state cartesian approach and accelerated acquisition schedules. *Invest Radiol*. <https://doi.org/10.1097/RLI.0000000000000836>
- Sarracanie M, Cohen O, Rosen MS (2003) 3D Balanced-EPI magnetic resonance fingerprinting at 6.5 mT. In: Proceedings of the 23rd Annual Meeting of ISMRM. Toronto, p 3385
- O'Reilly T, Teeuwisse WM, Webb AG (2019) Three-dimensional MRI in a homogenous 27 cm diameter bore Halbach array magnet. *J Magn Reson* 307:106578
- He Y, He W, Tan L et al (2020) Use of 2.1 MHz MRI scanner for brain imaging and its preliminary results in stroke. *J Mag Res* 319:106829
- O'Reilly T, Teeuwisse WM, de Gans D et al (2021) In vivo 3D brain and extremity MRI at 50 mT using a permanent magnet Halbach array. *Magn Reson Med* 85(1):495–505
- Sarracanie M, Salameh N (2020) Low-field MRI: how low can we go? a fresh view on an old debate. *Front Phys* 8:172
- Gilbert KM, Scholl TJ, Chronik BA (2008) RF coil loading measurements between 1 and 50 MHz to guide field-cycled MRI system design. *Conc Magn Res Part B* 33(3):177–191
- de Vos B, Parsa J, Abdulrazaq Z et al (2021) Design, characterization and performance of an improved portable and sustainable low-field MRI system. *Front Phys* 9:413
- Amthor T, Doneva M, Koken P et al (2017) Magnetic resonance fingerprinting with short relaxation intervals. *Magn Reson Imaging* 41:22–28
- Jiang Y, Ma D, Seiberlich N et al (2015) MR fingerprinting using fast imaging with steady state precession (FISP) with spiral readout. *Magn Reson Med* 74(6):1621–1631
- Koolstra K, Börnert P, Lelieveldt BPF et al (2021) Stochastic neighbor embedding as a tool for visualizing the encoding capability of magnetic resonance fingerprinting dictionaries. *Magn Res Mater Phys, Biol Med* 35:223–234
- Weigel M (2015) Extended phase graphs: dephasing, RF pulses, and Echoes - Pure and Simple. *J Magn Reson Imaging* 41(2):266–295
- Scheffler K (1999) A pictorial description of steady-states in rapid magnetic resonance imaging. *Conc Magn Res* 11(5):291–304
- Koolstra K, Beenakker J-WM, Koken P et al (2019) Cartesian MR fingerprinting in the eye at 7T using compressed sensing and matrix completion-based reconstructions. *Magn Res Med* 81(4):2551–2565
- Koolstra K, O'Reilly T, Börnert P et al (2021) Image distortion correction for MRI in low field permanent magnet systems with strong B₀ inhomogeneity and gradient field nonlinearities. *Magn Reson Mater Phys* 34(4):631–642
- Doneva M, Amthor T, Koken P et al (2016) Matrix completion-based reconstruction for undersampled magnetic resonance fingerprinting data. *Magn Reson Imaging* 41:41–52
- O'Reilly T, Webb AG (2022) In vivo T1 and T2 relaxation time maps of brain tissue, skeletal muscle, and lipid measured in healthy volunteers at 50 mT. *Magn Reson Med* 87(2):884–895
- Koolstra K, Webb AG, Veeger TJJ et al (2020) Water–fat separation in spiral magnetic resonance fingerprinting for high temporal resolution tissue relaxation time quantification in muscle. *Magn Reson Med* 84(2):646–662

43. Koolstra K, Webb A, Börnert P (2020) An attempt to understand why we measure longer relaxation times in quantitative muscle MRI using MRF than using conventional methods. In: Proceedings of the 28th Scientific Meeting, International Society for Magnetic Resonance in Medicine, virtual, p 0869
44. Ma D, Coppo S, Chen Y et al (2017) Slice Profile and B1 Corrections in 2D Magnetic Resonance Fingerprinting. *Magn Reson Med* 78(5):1781–1789
45. O'Reilly T, Börnert P, Webb A et al (2021) 3D magnetic resonance fingerprinting at 50 mT with integrated estimation and correction of image distortions due to B0 inhomogeneities. In: Proceedings of the 30th Scientific Meeting, International Society for Magnetic Resonance in Medicine, virtual, p 2516
46. Cohen O, Rosen MS (2017) Algorithm comparison for schedule optimization in MR fingerprinting. *Magn Reson Imaging* 41:15–21
47. Zhao B, Haldar JP, Liao C et al (2019) Optimal Experiment Design for Magnetic Resonance Fingerprinting: Cramér-Rao Bound Meets Spin Dynamics. *IEEE Trans Med Imaging* 38(3):844–861
48. Galiana G, Stockmann JP, Tam L et al (2012) Spin dephasing under nonlinear gradients: Implications for imaging and field mapping. *Magn Reson Med* 67(4):1120–1126
49. Hennig J, Welz AM, Schultz G et al (2008) Parallel imaging in non-bijective, curvilinear magnetic field gradients: a concept study. *Magn Reson Mater Phys, Biol Med* 21(1):5–14

Publisher's Note Springer Nature remains neutral with regard to jurisdictional claims in published maps and institutional affiliations.