



Universiteit
Leiden
The Netherlands

A review on innovative optical devices for the diagnosis of human soil-transmitted helminthiasis and schistosomiasis: from research and development to commercialization

Meulah, B.; Bengtson, M.; Lieshout, L.V.; Hokke, C.H.; Kreidenweiss, A.; Diehl, J.C.; ... ; Agbana, T.E.

Citation

Meulah, B., Bengtson, M., Lieshout, L. V., Hokke, C. H., Kreidenweiss, A., Diehl, J. C., ... Agbana, T. E. (2022). A review on innovative optical devices for the diagnosis of human soil-transmitted helminthiasis and schistosomiasis: from research and development to commercialization. *Parasitology*, 150(2), 137-149. doi:10.1017/S0031182022001664

Version: Publisher's Version

License: [Creative Commons CC BY-NC-ND 4.0 license](https://creativecommons.org/licenses/by-nc-nd/4.0/)

Downloaded from: <https://hdl.handle.net/1887/3721093>

Note: To cite this publication please use the final published version (if applicable).

Review Article

Cite this article: Meulah B, Bengtson M, Lieshout LV, Hokke CH, Kreidenweiss A, Diehl J-C, Adegnika AA, Agbana TE (2023). A review on innovative optical devices for the diagnosis of human soil-transmitted helminthiasis and schistosomiasis: from research and development to commercialization. *Parasitology* **150**, 137–149. <https://doi.org/10.1017/S0031182022001664>

Received: 22 September 2022

Revised: 25 November 2022

Accepted: 29 November 2022

First published online: 9 December 2022

Keywords:

Artificial intelligence; diagnosis; digital microscopes; helminths; innovation; optical devices; schistosomiasis; soil-transmitted helminthiasis

Author for correspondence:

Brice Meulah,

E-mail: b.meulah_tcheubousou@lumc.nl

A review on innovative optical devices for the diagnosis of human soil-transmitted helminthiasis and schistosomiasis: from research and development to commercialization

Brice Meulah^{1,2} , Michel Bengtson¹, Lisette Van Lieshout¹, Cornelis Hendrik Hokke¹, Andrea Kreidenweiss^{3,4}, Jan-Carel Diehl⁵, Ayola Akim Adegnika^{1,2,3,4} and Temitope Ebenezer Agbana⁶

¹Department of Parasitology, Leiden University Medical Center, Leiden, The Netherlands; ²Centre de Recherches Médicales des Lambaréné, CERMEL, Lambaréné, Gabon; ³Institute of Tropical Medicine, University Hospital Tübingen, Tübingen, Germany; ⁴German Center for Infection Research (DZIF), partner site Tübingen, Tübingen, Germany; ⁵Industrial Design Engineering, Delft University of Technology, Delft, The Netherlands and ⁶Delft Center for Systems and Control, Delft University of Technology, Delft, The Netherlands

Abstract

Diagnosis of soil-transmitted helminth (STH) and schistosome infections relies largely on conventional microscopy which has limited sensitivity, requires highly trained personnel and is error-prone. Rapid advances in miniaturization of optical systems, sensors and processors have enhanced research and development of digital and automated microscopes suitable for the detection of these diseases in resource-limited settings. While some studies have reported proof-of-principle results, others have evaluated the performance of working prototypes in field settings. The extensive commercialization of these innovative devices has, however, not yet been achieved. This review provides an overview of recent publications (2010–2022) on innovative field applicable optical devices which can be used for the diagnosis of STH and schistosome infections. Using an adapted technology readiness level (TRL) scale taking into account the WHO target product profile (TPP) for these diseases, the developmental stages of the devices were ranked to determine the readiness for practical applications in field settings. From the reviewed 18 articles, 19 innovative optical devices were identified and ranked. Almost all of the devices (85%) were ranked with a TRL score below 8 indicating that, most of the devices are not ready for commercialization and field use. The potential limitations of these innovative devices were discussed. We believe that the outcome of this review can guide the end-to-end development of automated digital microscopes aligned with the WHO TPP for the diagnosis of STH and schistosome infections in resource-limited settings.

Introduction

Soil-transmitted helminth (STH) and *schistosome* infections are 2 major neglected tropical diseases (Carbonell *et al.*, 2021) that affect approximately 1.5 billion people, predominantly in low- and middle-income countries (LMICs) (WHO, 2021b). These helminth infections have a considerable public health impact when left untreated with opposing effects on both physical and cognitive development in children and adolescents (Ezeamama *et al.*, 2018; Pabalan *et al.*, 2018). The diagnosis of STH and schistosome infections in LMICs so far relies mainly on conventional microscopy. With this diagnostic technique, parasite-derived products (including eggs) in stool and urine can be identified and quantified. This technique is laborious, requires well-trained personnel as well as electricity which is often lacking in resource-limited settings (Ajibola *et al.*, 2018; Khurana *et al.*, 2021). Alternatives to conventional microscopy for schistosomiasis diagnosis include the urine-based point-of-care test to detect circulating cathodic antigens and the up-converting particle lateral flow assay to detect circulating anodic antigens in either serum or urine (Corstjens *et al.*, 2020; Hoekstra *et al.*, 2021). Nucleic acid amplification tests, which involve the detection of parasite's specific nucleic acid sequences, have been used to detect STH and schistosome infections, although mainly in research settings (Obeng *et al.*, 2008; Meurs *et al.*, 2015; Sanprasert *et al.*, 2019). It is important to note that these existing techniques are still being validated and standardized for use in daily clinical care and medical research in resource-limited settings (Casacuberta-Partal *et al.*, 2019; Amoah *et al.*, 2020; Cools *et al.*, 2020; Cools *et al.*, 2021; Corstjens *et al.*, 2020).

To accelerate control and elimination of STH and schistosome infections, focus is placed mainly on mass drug administration (MDA) of praziquantel prioritizing school age children. In addition, approaches such as increased access to clean water and sanitation, snail control, education and behavioural changes are adopted, together with a multi-diagnostic approach to support monitoring and evaluation of these programmes (Raso *et al.*, 2018). However,

© The Author(s), 2022. Published by Cambridge University Press. This is an Open Access article, distributed under the terms of the Creative Commons Attribution-NonCommercial-NoDerivatives licence (<http://creativecommons.org/licenses/by-nc-nd/4.0/>), which permits non-commercial re-use, distribution, and reproduction in any medium, provided that no alterations are made and the original article is properly cited. The written permission of Cambridge University Press must be obtained prior to any commercial use and/or adaptation of the article.

due to the limitations of the existing diagnostic tools needed to monitor, evaluate and guide these programmes, the WHO developed target product profiles (TPPs) to guide development of new diagnostics to reliably measure the effectiveness of these programmes (WHO, 2021a, 2021b). Therefore, new reliable diagnostic tools, designed to match the WHO TPP, are required for effective disease control and elimination. Automated or semi-automated easy-to-use optical devices for the detection of parasite eggs in feces or urine, and based on low-cost platforms may provide a matching solution.

Innovative optical devices, designed to (semi-) automatically detect helminth eggs, have been built based on technical modifications of the optical train of a smartphone. The integrated sensors are commonly used for image acquisitions in such systems (Hernández-Neuta *et al.*, 2019; Vasiman *et al.*, 2019). Another optical technique is based on the direct coupling of the smartphone to the ocular of a standard microscope to digitally register the images of the sample under examination (Yang *et al.*, 2019; Dacal *et al.*, 2021). Increased processing power of the smartphones has equally been exploited for the development of artificial intelligence (AI) algorithms. The AI algorithms are developed for interpretation and analysis of registered data. Such devices have been reported to possess the potential for automated medical diagnosis (Saeed and Jabbar, 2018; Vasiman *et al.*, 2019). Over the last years, several investigators have reported the use of such devices to detect STHs and schistosome infections. In these studies, high-quality images from samples were acquired and analysed for the presence/absence of these parasite eggs. The readiness of these devices in LIMCs can be determined using the technology readiness level (TRL) developed by the National Aeronautics and Space Administration (NASA) and implemented by the European Commission (EC) (Bruno *et al.*, 2020). However, the TRL scale does not consider extensive medical and design outcomes.

To better understand the progress made so far in the development of these optical devices and the barriers that limit transitions to field applications, we examined publications in peer-reviewed journals with a specific focus on the validation of innovative optical devices (hardware with or without AI) used for the detection and/or quantification of eggs of STHs and *Schistosoma* spp. in human urine and stool samples. We also highlight the progress made in the development of optical devices with AI-assisted detection. Furthermore, we adapted the TRL (combined TRL-TPP) scale and used it to qualify readiness of innovative optical devices developed for LIMCs to diagnose soil-transmitted helminthiasis and schistosomiasis. The clinical diagnostic performance, advantages and disadvantages of the use of these devices are also discussed.

Methodology

Bibliography search strategy and article selection

PubMed and Google Scholar were searched following their respective stipulated database guidelines from March 2020 to May 2022 for English language-based literature published in peer-reviewed journals between 2010 and 2022. The focal terms for devices with or without AI (such as, cell phone OR smartphone OR mobile phone OR digital microscopy OR digital imaging OR microscope OR artificial intelligence OR machine learning OR neural network OR deep learning OR computer vision) were combined with diagnostic OR diagnosis using the AND Boolean operator and the results of this combination operated with the disease names (e.g. helminthiasis OR ascariasis OR ancylostomiasis OR trichuriasis OR schistosomiasis OR Bilharzia OR soil-transmitted helminth infection) OR pathogen names (e.g. *A. lumbricoides* OR

T. trichiura OR *S. haematobium* OR *S. mansoni* OR STH-species) using the AND Boolean operator. In total, 849 hits were found and filtered by scanning the titles for terminologies related to optical devices listed above and parasite names of interest. This resulted in 30 selected articles of which the abstracts were read by the first author and assessed for their relevance to the review. Publications deemed relevant were read in full, and thereafter 16 were selected which focused on innovative optical devices and their assessment for the detection of parasites of interest in human stool and/or urine. In addition, a short report and a conference paper cited within these 16 selected publications were also included, resulting in a total of 18 articles (see Table 1). Publications that focus on AI algorithms for the detection of STHs or *Schistosoma* spp. without device prototype description were excluded (Cooke *et al.*, 2015; Jiménez *et al.*, 2016; Alva *et al.*, 2017; Li *et al.*, 2020; Butpoy *et al.*, 2021).

Analysis

To understand the developmental progress and the performance evaluation process in a conceptual frame, the devices in the various studies were grouped as either smartphone-based optical devices or non-smartphone-based optical devices. Furthermore, each device within these 2 groups was ranked for readiness (for any of the disease-causing parasites; *Ascaris lumbricoides*, *Trichuris trichiura*, *Schistosoma haematobium*, *Schistosoma mansoni* and hookworm) using an adapted version of the 9-level TRL scale (NASA) (Fig. 1). The TRL was modified by distributing the WHO TPP on levels 6, 7 and 8 (WHO, 2021a, 2021b). That is the recommended sensitivity and specificity on TRL-6 and the other profiles (such as prototype use summary, configuration and context design not excluding diagnostic performance) on TRL-7 and -8 as a requirement to be met to transition to the next level on the scale. Three of the co-authors applied the adapted scale to 21 devices from the 18 selected publications. All information about each device was extracted from the respective study such as aim of the study; image of prototype(s) evaluated; evaluation setting; sample size; an image captured by the device; the TPP evaluated for each device including the diagnostic performance. Where available, additional information on context-specific design through stakeholder engagement and usability was examined. With this information, each expert ranked the devices based on the adapted scale (see Tables 2 and 3). A particular rank (1–9) allocated to a device implies it meets the criteria of the rank and all below it. For devices of different iterations that were validated in different studies, the rank of the most recent iteration (study) was considered to determine its overall progress.

Evolution of optical device and performance

Innovative optical devices have evolved over the years based on system designs, and rapid advances in the technological development of integrated functional parts such as general processors, image registration sensors, etc. For simplicity, all the optical devices discussed in this study are grouped into 2 main categories: (1) smartphone-based devices and (2) non-smartphone-based devices.

Smartphone-based optical devices

Modification of the functionalities of smartphones allows for the rapid development of smartphone-based optical devices to visualize and quantify parasite-derived morphological products in samples. Their portability makes them suitable for parasitology field surveys. In the past decade, numerous studies have been carried out using smartphone-based optical devices of which some are limited to proof-of-concept while others have been validated

Table 1. Overview of the 18 included publications in order of year of publications

Study	Geographical region/sample population	Sample type	Sample size	Organism detected	Reference test	Device name and (smartphone used)	Pros and cons of each device per study
Bogoch <i>et al.</i> (2013)	Pemba Island Tanzania/ school-aged children	Stool	199	<i>A. lumbricoides</i> <i>T. trichiura</i> Hookworm	Single KK	Mobile phone microscope (iPhone S4)	Robustness: compact Field compatibility: no (no sample stage) WHO TPP Sen: achieved for <i>A. lumbricoides</i> only WHO TPP Spe: achieved for <i>A. lumbricoides</i> only Functional integrated AI software: no
Linder <i>et al.</i> (2013)	–	Stool Urine		<i>T. trichiura</i> <i>S. mansoni</i> <i>S. haematobium</i>	–	Mini-microscope (–)	WHO TPP Sen: not evaluated WHO TPP Spe: not evaluated Functional integrated AI software: for <i>S. haematobium</i> only
Bogoch <i>et al.</i> (2014a)	Pemba Island Tanzania/ school-aged children	Stool	91	<i>A. lumbricoides</i> <i>T. trichiura</i>	Double KK (2 consecutive stool)	Newton Nm1 (–)	Robustness: compact Field compatibility: yes WHO TPP Sen: not evaluated WHO TPP Spe: not evaluated Functional integrated AI software: no
Bogoch <i>et al.</i> (2014b)	Azagué Makougué south Côte d'Ivoire/ school-aged children	Stool Urine	164 180	<i>T. trichiura</i> <i>S. mansoni</i> <i>S. haematobium</i>	Single KK Urine filtration	Mobile phone microscope and Newton Nm1 (iPhone S4)	Mobile phone microscope WHO TPP Sen: not achieved WHO TPP Spe: not achieved Functional integrated AI software: no Newton Nm 1 WHO TPP Sen: achieved WHO TPP Spe: achieved Functional integrated AI software: no
Switz <i>et al.</i> (2014)	–	Stool	–	<i>A. lumbricoides</i>	–	Mobile phone microscope (iPhone S4)	WHO TPP Sen: not evaluated WHO TPP Spe: not evaluated Functional integrated AI software: no
Ephraim <i>et al.</i> (2015)	Central Region of Ghana/ school-aged children (7–13)	Urine	49	<i>S. haematobium</i>	Urine filtration	Foldscope and CellScope (iPhone S5)	For both devices Field compatibility: no (no sample stage) WHO TPP Sen: not achieved WHO TPP Spe: not achieved Functional integrated AI software: no
Coulibaly <i>et al.</i> (2016a, 2016b)	Grand Moutcho Southern Côte d'Ivoire/school-aged children (6–19 years)	Stool Urine	226	<i>S. mansoni</i> <i>S. haematobium</i>	Single KK Urine filtration	CellScope (iPhone S5) and Newton Nm1	CellScope Field compatibility: no (no sample stage) WHO TPP Sen: not achieved WHO TPP Spe: not achieved Newton Nm1 WHO TPP Sen: achieved WHO TPP Spe: achieved Functional integrated AI software: no
Sowerby <i>et al.</i> (2016)	–	Stool	1	<i>A. lumbricoides</i>	–	Phone-based microscope (Nokia Lumia 1020)	Field compatibility: no (no sample stage) WHO TPP Sen: not evaluated WHO TPP Spe: not evaluated Functional integrated AI: no Sample preparation: complex

(Continued)

Table 1. (Continued.)

Study	Geographical region/sample population	Sample type	Sample size	Organism detected	Reference test	Device name and (smartphone used)	Pros and cons of each device per study
Bogoch <i>et al.</i> (2017)	Sorodofa–Abaasa Ghana/school-aged children (7–16)	Urine	60	<i>S. haematobium</i>	Urine centrifugation	Phone-based microscope (Nokia Lumia 1020)	Field compatibility: yes WHO TPP Sen: achieved WHO TPP Spe: achieved Functional integrated AI software: no
Holmstrom <i>et al.</i> (2017)	–	Stool Urine	3 1	<i>A. lumbricoides</i> <i>T. trichiura</i> Hookworm	–	Portable digital microscope (–)	Field compatibility: yes WHO TPP Sen: not achieved WHO TPP Spe: not achieved Functional integrated AI software: yes
Moser <i>et al.</i> (2018)	Pemba Island, Tanzania/adolescents (age 15–18)	Stool	615	<i>A. lumbricoides</i> <i>T. trichiura</i> Hookworm	Single, double and quadruple KK	FECPAK ^{2G} (–)	Robustness: compact Field compatibility: yes WHO TPP Sen: achieved WHO TPP Spe: achieved Functional integrated AI software: no
Koydemir <i>et al.</i> (2019)	Azagué regions of southern Côte d'Ivoire/school-aged children	Stool Urine	–	<i>S. mansoni</i> <i>S. haematobium</i>	–	Phone-based microscope (Nokia Lumia 1020)	Field compatibility: yes WHO TPP Sen: not evaluated WHO TPP Spe: not evaluated Functional integrated AI: no
Yang <i>et al.</i> (2019)	Infanadiana district in Madagascar/subjects of all age groups	Stool	–	<i>A. lumbricoides</i> <i>T. trichiura</i> Hookworm	Composite KK, SSTT, MIF	UVC microscope (smartphone)	WHO TPP Sen: not evaluated WHO TPP Spe: not evaluated Functional integrated AI: yes
Sukas <i>et al.</i> (2019)	–	Stool	7	<i>A. lumbricoides</i> <i>T. trichiura</i> <i>S. mansoni</i> Hookworm	Mini-FLOTAC, McMaster	Lab-on-a-Disk (–)	WHO TPP Sen: not evaluated WHO TPP Spe: not evaluated Functional integrated AI: no Sample preparation: complex, inefficient
Agbana <i>et al.</i> (2020)	–	Urine	–	<i>S. haematobium</i>	–	SODOS (–)	WHO TPP Sen: not achieved WHO TPP Spe: achieved Functional integrated AI software: yes
Decal <i>et al.</i> (2021)	Kwale county, south coast Kenya	Stool	12	<i>A. lumbricoides</i> <i>T. trichiura</i>	Manual detection on sample images	Mobile Microscope (Xiaomi PocophoneF1, Bq Aquaris X2)	WHO TPP Sen: not evaluated WHO TPP Spe: not evaluated Functional integrated AI: for <i>T. trichiura</i> only
Oyibo <i>et al.</i> (2022)	–	Stool Urine	–	<i>S. mansoni</i> Hookworm <i>S. haematobium</i>	Manual detection on sample images	SchistoScope (–)	WHO TPP Sen: not evaluated WHO TPP Spe: not evaluated Functional integrated AI software: for <i>S. haematobium</i> only
Armstrong <i>et al.</i> (2022)	Sorodofa–Abaasa town, north of Cape Coast, Ghana and Azagué region of southern Côte d'Ivoire	Urine	205	<i>S. haematobium</i>	Urine filtration microscopy	SchistoScope (Iphone 8)	WHO TPP Sen: not achieved WHO TPP Spe: achieved Functional integrated AI software: yes Sample preparation: inefficient

KK, Kato–Katz; (–), not mentioned; Sen, sensitivity; Spe, specificity; TPP, target product profile.




Fig. 1. The 9-scale TRL classification chart. Image adapted from Bruno *et al.* (2020). TPP, target product profile.

through a field evaluation (Bogoch *et al.*, 2013; Bogoch *et al.*, 2014b; Switz *et al.*, 2014; Ephraim *et al.*, 2015; Coulibaly *et al.*, 2016b; Sowerby *et al.*, 2016; Bogoch *et al.*, 2017; Koydemir *et al.*, 2019; Yang *et al.*, 2019; Armstrong *et al.*, 2022). Table 2 recapitulates the diagnostic characteristics of smartphone-based optical devices for the detection of STHs and *Schistosoma* spp. and the corresponding TRL assigned by the authors of this review.

In 4 published studies, 5 smartphone-based optical devices for the detection of helminth eggs in urine and/or stool were reported as experimental proof-of-concept studies (Bogoch *et al.*, 2013; Switz *et al.*, 2014; Sowerby *et al.*, 2016; Yang *et al.*, 2019). First, Bogoch *et al.* carried out a study within an ongoing clinical trial from which 199 Kato–Katz (KK) thick smears were analysed for the presence of STH eggs using a smartphone-based microscope with conventional light microscopy as a reference (Bogoch *et al.*, 2013). The device used for this study consisted of a 3 mm ball lens that was mounted in alignment with the lens of an iPhone 4S using a double-sided tape. The light source was a separate handheld incandescent lamp powered by 1 AA battery. For this design, a moderate sensitivity, specificity, positive predictive value (PPV) and negative predictive value (NPV) for the detection of STH eggs were observed (Table 2). Compared to conventional light microscopy, the authors observed a higher sensitivity of 93.3% (66.0–99.7) for moderate and high infection intensities of *A. lumbricoides*. According to the authors, this device is a promising point-of-care diagnostic tool for helminthic infections. However, this design lacks a functional sample stage which makes slide scanning unfeasible in practice.

Second, Switz *et al.* carried out a proof-of-concept study using a smartphone microscope to image the eggs of *A. lumbricoides* in stool samples over a wide field of view (FOV) (Switz *et al.*, 2014). The smartphone microscope consisted of an iPhone 4S coupled to a reversed iPhone 4S lens and embodied by a frame with a screw-based focusing system, an *x–y* stage and a light-emitting diode (LED) light source. The system set-up provided high-resolution images with significantly large FOV. While the authors conclude that the availability and use of this device in resource-limited settings will increase, its robustness and diagnostic performance at patient level are still unknown and require more testing.

Third, Sowerby *et al.* carried out a proof-of-concept study that uses the flotation technique to accumulate *A. lumbricoides* ova in a single FOV for imaging using a smartphone-based microscope (Sowerby *et al.*, 2016). Stool was collected from a person known to be positive for ascariasis. The smartphone-based microscope consisted of a Nokia Lumia 1020 smartphone, an external double convex objective lens, a cassette containing a sample well and a cassette rack with a light source – all coupled together such that the sample well, external lens and the lens of the smartphone are aligned. Several images of a single FOV at different focal lengths were captured for image reconstruction and visualization of *A. lumbricoides* eggs. According to the authors this device has potential as a point-of-care device for the detection of helminth species upon further development. The sample preparation method and the processing of images through reconstruction clearly do not support the design as a point-of-care test (Peeling and Mabey, 2010; Osredkar, 2017), especially with

Table 2. Diagnostic characteristics of 11 smartphone-based optical devices for the diagnosis of *Schistosoma* and STH infections

Device name (study)	Organism	Sen (95% CI)	Spe (95% CI)	PPV (95% CI)	NPV (95% CI)	TRL	TRL comment (*)
Proof-of-concept studies							
Mobile phone microscope (Bogoch <i>et al.</i> , 2013)	<i>A. lumbricoides</i>	81.0 (65.4–90.9)	87.3 (80.7–91.9)	63.0 (48.7–75.4)	94.5 (89.1–97.4)	6	Large-scale testing ≥50 samples in a field laboratory setting. WHO Sen and Spe achieved for <i>A. lumbricoides</i>
	<i>T. trichiura</i>	54.4 (46.3–62.3)	63.4 (46.9–77.4)	85.1 (76.4–91.2)	26.5 (18.4–36.6)	5	
	Hookworm	14.3 (8.3–23.1)	89.1 (81.0–94.2)	56.0 (35.3–75.0)	51.7 (44.1–59.3)	5	
Mobile phone microscope (Switz <i>et al.</i> , 2014)	<i>A. lumbricoides</i>	–	–	–	–	3	Principle demonstration in technical laboratory
Phone-based microscope (Sowerby <i>et al.</i> , 2016)	<i>A. lumbricoides</i>	–	–	–	–	3	Principle demonstration only
Name not mentioned (Yang <i>et al.</i> , 2019)	<i>A. lumbricoides</i> ^{SSTT}	82.9 (74.4–91.4)	97.1 (91.5–10.3)	–	–	4	Small-scale testing <50 samples prototype in a laboratory
	<i>T. trichiura</i> ^{SSTT}	22.4 (14.1–30.7)	91.7 (76.1–10.7)	–	–	4	
	<i>Hookworm</i> ^{SSTT}	31.8 (12.3–51.3)	96.6 (92.8–100.0)	92.5 (–)	10.5 (–)	4	
	<i>A. lumbricoides</i> ^{KK}	57.9 (46.8–69.0)	97.1 (91.5–10.3)	50.0 (–)	73.8 (–)	4	
	<i>T. trichiura</i> ^{KK}	23.5 (15.1–31.9)	100.0 (100–100.0)	100.0 (–)	100.0 (–)	4	
	<i>Hookworm</i> ^{KK}	38.1 (17.3–58.9)	71.9 (63.7–82.1)			4	
	<i>A. lumbricoides</i> ^{AI}	68.5 (–)	40.0 (–)			4	
	<i>T. trichiura</i> ^{AI}	8.3 (–)	96.9 (–)			4	
	<i>Hookworm</i> ^{AI}	100.0 (–)	100.0 (–)			4	
Validation studies							
Mobile phone microscope (Bogoch <i>et al.</i> , 2014b)	<i>T. trichiura</i>	30.8 (19.9–43.4)	71.0 (61.1–79.6)	40.8 (27.0–55.8)	61.2 (51.7–70.1)	5	Large-scale testing in a field laboratory WHO Sen and Spe achieved for <i>S. mansoni</i>
	<i>S. mansoni</i>	68.2 (60.1–75.5)	64.3 (35.1–87.2)	95.4 (89.5–98.5)	15.8 (7.5–27.9)	6	
Foldscope (Ephraim <i>et al.</i> , 2015)	<i>S. haematobium</i>	55.9 (38.1–72.4)	93.3 (66.0–99.7)	–	–	4	Foldscope commercialized, BUT not together with smartphone small-scale testing (<50 samples) in a laboratory
CellScope (Ephraim <i>et al.</i> , 2015)	<i>S. haematobium</i>	67.6 (49.4–82.0)	100.0 (74.7–100.0)	–	–	4	Small-scale testing (<50 samples) in a laboratory
CellScope (Coulibaly <i>et al.</i> , 2016a, 2016b)	<i>S. mansoni</i>	50.0 (25.4–74.6)	99.5 (97.0–100)	–	–	5	Iteration from Ephraim <i>et al.</i> (2015) with large-scale testing (≥50 samples) in the field laboratory and WHO Sen not achieved
	<i>S. haematobium</i>	35.6 (25.9–46.4)	100 (86.7–100)			5	
Phone-based microscope (Bogoch <i>et al.</i> , 2017)	<i>S. haematobium</i>	72.1 (56.1–84.2)	100.0 (75.9–100)	100.0 (86.3–100)	100.0 (86.3–100)	5	Large-scale testing (≥50 samples) in a field laboratory. WHO Sen not achieved
Phone-based microscope (Koydemir <i>et al.</i> , 2019)	<i>S. mansoni</i>	–	–	–	–	4	Small-scale testing (<50 samples) in a field laboratory
	<i>S. haematobium</i>					4	
SchistoScope (Armstrong <i>et al.</i> , 2022)	<i>S. haematobium</i> ^{UF}	90.9 (–)	91.1(–)	–	–	6	Large-scale testing (≥50 samples) in a field laboratory
	<i>S. haematobium</i> ^{AI}	32.0-91.0 (–)	65.0-90.0 (–)			6	

TRL, technology readiness level scale; (–), not stated; CI, confidence interval; Sen, sensitivity; Spe, specificity; PPV, positive predictive value; NPV, negative predictive value. (*) See Fig. 1 for TRL specification. Superscript KK, SSTT, UF=Kato-Katz, spontaneous sedimentation and filtration technique slides read by a microscopist. Superscript AI=Kato-Katz, spontaneous sedimentation and filtration technique slides analysed by artificial intelligence (AI) algorithm.

Table 3. Diagnostic characteristics of 10 non-smartphone-based optical devices for diagnosis of *Schistosoma* and STH infections

Device name (study)	Organism	Sen (95% CI)	Spe (95% CI)	PPV (95% CI)	NPV (95% CI)	TRL	Scale comment (*)
Proof-of-concept studies							
Mini-microscope with AI (Linder <i>et al.</i> , 2013)	<i>T. trichiura</i>	–	–	–	–	3	Principle demonstration
	<i>S. mansoni</i>					3	
	<i>S. haematobium</i>					3	
Digital microscope with AI (Holmstrom <i>et al.</i> , 2017)	<i>A. lumbricoides</i>	100 (–)	–	93.7 (–)	–	4	Small-scale testing in a laboratory setting
	<i>T. trichiura</i>	83.3 (–)		100.0 (–)		4	
	Hookworm	93.8 (–)		69.7 (–)		4	
Lab-on-a-disk (Sukas <i>et al.</i> , 2019)	<i>A. lumbricoides</i>	–	–	–	–	4	Principle demonstration
	<i>T. trichiura</i> Hookworm					4	
	<i>S. mansoni</i>					4	
SODOS with AI (Agbana <i>et al.</i> , 2020)	<i>S. haematobium</i>	50.6 (–)	98.6 (–)	–	–	4	Small-scale testing in a technical laboratory
Mobile microscope with AI (Decal <i>et al.</i> , 2021)	<i>T. trichiura</i>	80.94 (–)	>99 (–)	–	–	4	Small-scale testing of conventional light microscope attached to smartphone separately commercialized
	<i>A. lumbricoides</i>	93.08 (–)	–			4	
SchistoScope with AI (Oyibo <i>et al.</i> , 2022)	Hookworm	–	–	–	–	4	Small-scale testing (<50 samples)
	<i>S. mansoni</i>						
	<i>S. haematobium</i>						
Validation studies							
Newton Nm1 (Bogoch <i>et al.</i> , 2014a)	<i>A. lumbricoides</i>	–	–	–	–	5	Large-scale testing ≥50 samples. WHO Sen and Spe not reported. Commercially available
	<i>T. trichuria</i>					5	
Newton Nm1 (Bogoch <i>et al.</i> , 2014b)	<i>T. trichiura</i>	81.5 (70.0–90.1)	93.0 (86.1–97.1)	–	–	9	Commercially available. Large-scale testing ≥50 samples. WHO Sen, Spe and other TPPs achieved
	<i>S. mansoni</i>					9	
	<i>S. haematobium</i>	84.8 (78.0–90.1)	85.7 (57.2–98.2)			9	
		78.6 (49.2–95.3)	91.0 (85.5–94.9)				
Newton Nm1 (Coulibaly <i>et al.</i> , 2016a, 2016b)	<i>S. mansoni</i>	91.7 (59.8–99.6)	99.5 (97.0–100.0)	–	–	9	Commercially available for end users. Large-scale testing ≥50 samples. WHO Sen, Spe and other TPPs achieved
	<i>S. haematobium</i>	81.1 (71.2–88.3)	97.1 (92.2–99.1)			9	
FECPAK ^{2G} (Moser <i>et al.</i> , 2018)	<i>A. lumbricoides</i>	75.6 (72.0–77.7)	96.9 (94.8–98.9)	–	–	9	Commercially available for end users. Large-scale testing ≥50 samples. WHO Sen, Spe and other TPPs achieved
	<i>T. trichiura</i>	65.8 (64.9–66.2)	95.3 (91.8–97.6)			9	
	Hookworm	71.5 (67.4–95.3)	91.3 (89.3–93.1)			9	

TRL, technology readiness level scale; (–), not stated; CI, confidence interval; Sen, sensitivity; Spe, specificity; PPV, positive predictive value; NPV, negative predictive value. (*) See Fig. 1 for TRL specification.

stool material. Furthermore, functionalities of point-of-care devices were not examined for this prototype.

The fourth proof-of-concept by Yang *et al.* focused on the development and evaluation of a smartphone-based microscope with and without applying the Kankanet AI algorithm, for the detection of *A. lumbricoides*, *T. trichiura* and hookworm eggs. KK, spontaneous sedimentation technique in tube, and merthiolate-iodine-formaldehyde slides were analysed with standard microscopy as a reference (Yang *et al.*, 2019). This study was embedded in an ongoing epidemiology survey. The smartphone based-microscope setup consisted of a smartphone mounted on a tripod and connected to a USB video class microscope mounted on a translational stage that moves in the *x-y* direction. The number of samples included in the study was not stated and evaluation of the device prototype with Kankanet was based on the detection of eggs at an image level. Sensitivity and specificity were determined based on 124 images captured with the device setup as shown in Table 2. The Kankanet AI yielded a lower sensitivity and specificity for the detection of these parasite eggs compared to the microscopist (see Table 2). According to the authors, the setup with the Kankanet AI, if further developed, can achieve a performance comparable to microscopy, and the device could serve as a point-of-care diagnostic procedure. To support this claim, a more thorough performance evaluation of this tool should be carried out at a sample level.

In addition to the aforementioned proof-of-concept studies, 5 different smartphone-based optical devices were field validated for the detection of helminth eggs in urine and/or stool with 2 of the devices evaluated twice in different settings (Bogoch *et al.*, 2014b; Ephraim *et al.*, 2015; Coulibaly *et al.*, 2016b; Bogoch *et al.*, 2017; Koydemir *et al.*, 2019; Armstrong *et al.*, 2022). First, a cohort validation study by Bogoch *et al.* was integrated into an ongoing study in South Côte d'Ivoire. In this study, a single KK thick smear per sample from a total of 164 stool samples were examined with a smartphone-based microscope to detect the eggs of *T. trichiura* and *S. mansoni*, with conventional light microscopy as the reference (Bogoch *et al.*, 2014b). The smartphone-based microscope consisted of a 3 mm glass ball lens mounted on an S4 iPhone with the support of 1/16 thick foam tape. The sensitivity, specificity, PPV and NPV were reported as shown in Table 2. The authors found this device not ready for use in clinical or epidemiological settings due to its low sensitivity.

The second validation study, carried out by Ephraim *et al.*, was embedded within an ongoing epidemiology survey of schistosomiasis and soil-transmitted helminthiasis in the Central Region of Ghana. A total of 49 participants were randomly selected to provide urine samples for examination of *S. haematobium* eggs using 2 devices, a smartphone-mounted Foldscope and a reversed-lens CellScope, with conventional microscopy as the reference (Ephraim *et al.*, 2015). Both devices were iPhone S5-based. However, the Foldscope consisted of an LED light source attached to the phone lens with the help of a paper-based attachment, a magnet and tape, whereas the CellScope consisted of a 3D-printed plastic embodiment with an integrated lens, fixed to the iPhone's lens. The urine samples (10 mL) were processed by centrifugation and examined by both devices and conventional light microscopy. The sensitivity and specificity of the smartphone-mounted Foldscope and the reversed-lens CellScope are shown in Table 2. The authors concluded that these devices could be of use in epidemiological and clinical settings upon further technical adjustments. The design prototype reported in this study did not have a sample stage without which its application in epidemiological surveys for mass disease screening is not feasible. Coulibaly *et al.* further evaluated the performance of the CellScope in a cross-sectional exploratory study for the diagnosis

of schistosomiasis, with conventional light microscopy as a reference (Coulibaly *et al.*, 2016b). This study examined filtered urine slides and single KK thick smears from 226 participants with sensitivity and specificity reported in Table 2. The sensitivity of the CellScope was lower in low infection intensity samples (<40 eggs 10 mL⁻¹) and strongly correlated with light microscopy in terms of *S. haematobium* egg count estimated with a Pearson's correlation coefficient of 0.92. Based on the outcome, the authors concluded that this device may be of use in areas with moderate and high intensity of infection. However, the lack of a sample stage makes diagnosis with this prototype impractical.

The fourth validation study also by Bogoch *et al.* explored an ongoing epidemiological survey, of which 60 urine specimens were randomly selected and examined for *S. haematobium* eggs using a smartphone-based microscope with conventional microscopy-based centrifugation as the reference standard (Bogoch *et al.*, 2017). The device was a custom-made 3D-printed optomechanical attachment with an inbuilt lens mounted in alignment with the lens of a Nokia Lumia 1020 smartphone. It also contained 2 LED light sources, 2 polymer diffusers and an *x-y* stage for manual adjustment of sample slides. The sensitivity, specificity, PPV and NPV for the detection of *S. haematobium* are shown in Table 2. The sensitivity of the device when separately analysed for low- and high-intensity infections was 65.7% (95% CI 47.7–80.3) and 100.0% (95% CI 59.8–100.0), respectively. Based on these findings, the authors suggested this prototype for use in high infection intensity settings. Similarly, Koydemir *et al.* evaluated the smartphone-based microscope prototype previously validated by Bogoch *et al.* (2017) in an ongoing epidemiology study in a different setting. Urine and KK thick smear slides were imaged for manual identification of *S. mansoni* and *S. haematobium* eggs and detection was at an image level (Koydemir *et al.*, 2019). The number of urine and stool samples used in this study was not stated. The authors concluded that, this device may be a valuable diagnostic tool in endemic settings only after carrying out rigorous field validation. However, a pilot study to show promising diagnostic performance of this prototype at patient level would concretize this conclusion.

The sixth study by Armstrong *et al.* evaluated the SchistoScope; a modified version of the LoaScope [developed previously to detect *Loa loa* microfilaria (Kamgno *et al.*, 2017)] as a point-of-care device to automate quantitative detection of *S. haematobium* eggs with microscopy as a reference (Armstrong *et al.*, 2022). The prototype consists of an iPhone 8 with a reverse lens technology (previously described in this section) and a specialized cartridge to concentrate *S. haematobium* eggs in a single FOV. A total of 205 urine samples were collected to determine the sensitivity and specificity of the SchistoScope with AI and manual detection (Table 3). The authors concluded that this tool has a potential to provide quantitative diagnosis of *S. haematobium*. However, the cartridge used in this study for sample preparation resulted in approximately 79% of eggs lost during the proposed sample preparation method and therefore requires improvements.

Non-smartphone-based optical devices

Of the 8 non-smartphone-based optical devices reviewed here, six were limited to proof-of-concept (Linder *et al.*, 2013; Holmstrom *et al.*, 2017; Sukas *et al.*, 2019; Agbana *et al.*, 2020; Dacal *et al.*, 2021; Oyibo *et al.*, 2022) as reported by the authors. The first study by Linder *et al.* demonstrated the use of a mini-microscope (On-Chip) to image and detect the eggs of *S. haematobium* with AI (Linder *et al.*, 2013). In this study, pooled urine samples from a pre-existing quality and assurance survey were used. The mini-microscope consisted of a webcam image sensor chip and an LED as a light source, all connected to computer software for image capture. Two AI algorithms were deployed for the

detection of *S. haematobium* eggs in urine-captured images (see Table 3). Stool samples were also imaged, and eggs of *T. trichiura* and *S. mansoni* were morphologically distinguished by a technician. Based on the outcome, the authors concluded that this setup could be exploited for building simpler imaging low-cost devices for diagnosis of schistosomiasis and STHs. However, in this study, automated detection was applied for urine parasite detection only, and not for stool parasites.

The second study by Holstrom *et al.* reported the imaging performance of a digital microscope and a deep learning-based image analysis algorithm, with human experts as a comparator, for the identification of helminth eggs in urine and stool samples (Holmstrom *et al.*, 2017). The design of the digital microscope included a white LED with a bandpass filter and an ultraviolet LED to support both bright field and fluorescent imaging. This digital microscope was constructed from mobile phone-based camera components, such as an external motor unit attached for precise manual adjustment of the glass slide, and it was connected to a laptop with the AI algorithm for image processing. Sixteen positive stool slides and 5 positive urine slides were prepared, and 7385 images were captured with the portable microscope. Of this total number of images captured, 410 were labelled as positive for helminth eggs by human experts and were used for subsequent training and evaluation of the AI algorithm. For detection on an image level for eggs of *A. lumbricoides*, *T. trichiura* and hookworm by the algorithm, the sensitivity and PPV were computed (see Table 3). Based on the authors, the device produced sufficient image quality for manual and automatic detection of STH and *S. haematobium* eggs using a deep learning algorithm. Further validation of this setup at patient level is required before implementation in field settings can be considered.

The third study developed and tested a lab-on-a-disk (LOD) digital platform, to image and manually count STHs and *S. mansoni* eggs in a single FOV with Mini-FLOTAC and McMaster as reference tests (Sukas *et al.*, 2019). Stool samples were processed by floatation and eggs collected in a single FOV. The complete design included a setup for imaging, a minicentrifuge and a sample disk where the sample is loaded. The imaging setup consisted of a macro lens, an adapter, a 10 × objective lens and a halogen light source. These parts were mounted in that order to a Sony α5100 camera and the imaging zone of the sample disk, in a vertical axis. The Sony camera was connected to a Samsung tablet for the camera control and image capture. Seven positive stool samples were examined using the 3 methods. The Mini-FLOTAC counts correlated with the LOD counts from the whole chamber of the disk (correlation coefficient 0.91). In addition, samples with low infection intensity of STHs (~30–100 egg per gram) determined with Mini-FLOTAC and McMaster were also detected as egg positive by the LOD device but with a lower egg count. According to the authors, this platform produced quality images to properly detect and quantify eggs of parasites in stool. Nonetheless, the sample disk needs further improvement since eggs were lost along the channel path leading to the collection zone where quantification in a single FOV was aimed.

The fourth study by Agbana *et al.* reported the use of a novel lens-free imaging system in combination with flow cytometry and AI (SODOS) to image and detect *S. haematobium* eggs in comparison with conventional microscopy (Agbana *et al.*, 2020). The SODOS device consists of an imaging sensor, a micro-channel flow cell and a monochromatic light source aligned on the same optical axis. It also contains an automatic urine injection system with a calibrated syringe for pumping the sample into the flow cell and a cup for collecting the outlet of the flow cell. To evaluate the device performance, urine samples collected from uninfected volunteers were spiked with *S. haematobium* eggs

(17 eggs mL⁻¹). Per run, 10.7 mL urine was imaged by the device, followed by image analysis using the AI software programme. An accuracy of 96.8% and a sensitivity and specificity for the AI software programme were determined (see Table 3). According to the authors, the SODOS may serve as a future tool to complement existing techniques for the diagnosis of urinary schistosomiasis if the sensitivity is further improved. Furthermore, this study was limited to spiked urine samples and validation using clinical samples is required.

The study by Dacal *et al.* reported the use of a mobile microscope assisted by AI for the detection of *T. trichiura* and *A. lumbricoides* in stool samples using microscopy as a reference (Dacal *et al.*, 2021). The design of the digital mobile microscope consisted of a smartphone coupled to a conventional microscope (Leica DM-200) with the help of an adapter such that the smartphone's camera aligns with the optics of the microscope. For this study, 51 KK slides were prepared from 12 samples, imaged and uploaded into a telemedicine platform for remote image annotation, training and validation of AI algorithm. The sensitivity and specificity of their setup at image level were computed (see Table 3). According to the authors, the proposed design could serve a great deal to reduce time, distance and expertise needed for coproscopic analysis of STHs detection thereby making accurate diagnosis feasible. Nevertheless, this design still requires an expert to manually operate the attached microscope for sample imaging thereby sharing similar challenges as conventional microscopy.

The sixth proof-of-concept study by Oyibo *et al.* reported the SchistoScope; an automated microscope builds from off-the-shelf components for use in resource-limited settings for the detection of *S. haematobium* eggs with AI (Oyibo *et al.*, 2022). The prototype consists of an optical system with an automated 3-axis scanner (XYZ) to automate sample scanning process and an onboard computer connected to a Raspberry Pi HQ camera. Both urine samples from uninfected persons spiked with eggs from gut tissue of hamsters, clinical urine and stool samples were used to test the prototype. This tool captured quality images with distinguishable features of *S. haematobium*, *S. mansoni* and hookworm eggs. A dice similarity coefficient (DSQ) of 0.44 was obtained when evaluating the egg count outcome from the automated SchistoScope compared to manual counts. Based on the authors, the SchistoScope could be a suitable tool for use in the monitoring and evaluation of schistosomiasis control programmes in endemic settings. However, we think that the SchistoScope's performance in automated egg quantification based on the reported DSQ limits its potential application in monitoring and evaluation of schistosomiasis control programmes.

In 2 non-smartphone-based validation studies, a commercially available research-grade handheld mobile light microscope, Newton Nm1, was evaluated for the diagnosis of STH infections with conventional light microscopy as the reference standard (Bogoch *et al.*, 2014a). In this study, 91 KK slides were randomly selected from a pack already analysed with conventional microscopy. These slides were re-examined using the Newton Nm1 microscope of which, 39.6 and 70.3% were positive for *A. lumbricoides* and *T. trichiura*, respectively, whereas 38.5% and 76.9% were positive for *A. lumbricoides* and *T. trichiura*, respectively, using conventional microscopy. No other known and highly sensitive reference test was used to determine whether samples were truly or falsely positives. Based on these findings, the authors concluded that the Newton Nm1 microscope may serve as a useful tool to detect and quantify STH eggs in endemic settings. Interestingly, Bogoch *et al.* (2014b), and Coulibaly *et al.* (2016b), again evaluated the Newton Nm1 microscope for the detection of helminth eggs with conventional light microscopy (single KK and urine filtration) as a reference standard. From

164 prepared stool slides and 180 urine slides, Bogoch *et al.* determined the sensitivity and specificity of the device for the detection of *Schistosoma* and *T. trichiura* (Bogoch *et al.*, 2014b). A sufficient diagnostic yield was obtained (see Table 3). Furthermore, Coulibaly *et al.* conducted a cross-sectional study and used urine and KK slides prepared from urine and stool samples collected from 226 individuals (Coulibaly *et al.*, 2016b). The Newton Nm1 handheld microscope produced a good sensitivity and an excellent specificity (see Table 3). The authors of these studies concluded that the Newton Nm1 microscope may serve as a valuable tool in clinical and public health settings to detect these infections. Nonetheless, both studies used conventional microscopy as the reference with limited sensitivity. Therefore, thorough evaluation with more sensitive diagnostics is needed to validate this tool. Furthermore, the Newton Nm1 microscope for the detection of these parasites is not fully automated.

The last study evaluated, for the first time, a new and remote diagnostic technique used in veterinary medicine called FECPAK compared to a single, double and quadruplicate KK method for analysing STHs eggs in human stool (Moser *et al.*, 2018). The FECPAK system is made up of a filtration unit, a micro-I cassette with sample loading wells that fit into the micro-I cassette, all connected to a computer for imaging of samples in a single FOV. For this study, 590 positive and 25 negative samples for any of the STHs were collected and analysed by the 4 methods. Images captured with the FECPAK were analysed by a microscopist. The sensitivity and specificity of the FECPAK device were determined for the detection of *A. lumbricoides*, *T. trichiura* and hookworm (Table 3). The sensitivity of FECPAK was found to be lower than single, double and quadruple KK method for the detection of *A. lumbricoides* and hookworm (see Table 3). The FECPAK underestimated the prevalence of these helminth infections compared to the single, double and quadruplicate KK method. Although FECPAK was originally developed for animal helminths diagnosis, authors concluded that it might be an interesting tool for both epidemiological and clinical human studies. Further evaluation studies of FECPAK for diagnosis of human STHs would help to further develop the tool.

Moving beyond test performance and readiness for LMIC

A systematic assessment of innovative optical diagnostic devices for parasitic infection diagnosis is important to understand the progress towards field implementation. In this review, we propose the use of a modified TRL scale as an appropriate standardized approach towards this goal. The FECPAK and the Newton Nm1 microscope met the criteria of the rank TRL-9 on the adapted TRL scale. In addition, of the 3 studies that validated the Nm1 microscope, only Coulibaly *et al.* evaluated the device for egg quantification which limits its application in MDA impact assessment programmes (Wiegand *et al.*, 2021). The Foldscope, although commercially available, was integrated with an iPhone S5 and validated in a field laboratory setting. This combined optical set-up is not yet in the market as an integrated diagnostic system, so we ranked it at TRL-4. Based on our scaling, 21% of the devices reviewed here are simply a demonstration of proof-of-concept (TRL-3), 55% in a stage of validation in the technical/field laboratory facility or community endemic setting (TRL-4, 5), 8% validated on clinical samples and the WHO recommended sensitivity and specificity met (TRL-6), and 15% are already commercialized (TRL-8, 9) (see Tables 2 and 3). Overall, validation in field settings (TRL-5, 6, 7) was revealed to be a bottle neck for 65% of these devices. While further research and development is needed to transition these devices to a commercial level, thorough validation before and after commercialization is required in field settings. Features beyond its performance,

such as robustness, portability, technical simplicity and feasibility of repair and costs, are also crucial parameters to consider for readiness in LMICs.

Non-smartphone-based microscopes developed so far have different cost ranges which may depend on their designs and the functional parts used. The Newton Nm1 microscope, for example, has a price range from US\$465 to US\$930 depending on the series. The WHO TPP for STH and schistosome infection does not mention the minimum capital cost per tool required for control and elimination of these infections; however, a diagnostic test should cost less than US\$3 per test. Aside from Armstrong *et al.*, the reviewed articles did not report material cost analysis per test.

Conclusion and future perspective of optical devices with AI for soil-transmitted helminth and *Schistosoma* spp. detection

Simple, inexpensive, portable and robust diagnostic tools are urgently needed in resource-limited settings for the diagnosis of human helminthiasis using urine and stool samples. Important advancements have been made in the clinical sector with optical devices for automated malaria parasite detection (Frean, 2008; Kaewkamnerd *et al.*, 2012; Coulibaly *et al.*, 2016a; Hashimoto *et al.*, 2017; Agbana *et al.*, 2018; Kumar *et al.*, 2020; de Melo *et al.*, 2021). Comparatively, the automated optical detection of human stool and urine parasites such as STHs and schistosomes is still in its early stages. With advancing technology, optical devices could be developed to have a fully functioning computing system, with the potential to accommodate specific features such as imaging sensors, global positioning system (GPS) navigation, internet access, data storage and high processing power, each with added values for different stakeholders in the health system. This would facilitate data generation and distribution making it useful for mapping of STHs, *Schistosoma* spp. infections and impact assessment of MDA programmes.

Automation of such devices makes them comparatively easy to use and facilitates multitasking during sample analysis. In this review, we summarized 18 publications presenting 21 different innovative optical devices validated within the last decade for the diagnosis of soil-transmitted helminthiasis and schistosomiasis. Some of these devices require infrastructure for sample processing which poses a limitation to device application in resource-limited settings. Most device prototypes reported in this review require human expertise for image interpretation before conclusive decisions could be reached. Fully automated devices with integrated reliable AI models could potentially mitigate the chances of human error and reduce the need for human training. Currently, only a few studies have reported the use of optical devices with AI to detect STH and *Schistosoma* spp. infections in human stool and urine (Linder *et al.*, 2013; Holmstrom *et al.*, 2017; Agbana *et al.*, 2020; Dacal *et al.*, 2021; Armstrong *et al.*, 2022; Oyibo *et al.*, 2022). Much less similar reports have been written for stool samples (Holmstrom *et al.*, 2017; Dacal *et al.*, 2021). The focus on automated detection of parasites in urine rather than stool illustrates the challenges with stool sample analysis and re-affirms the need for more research and development exploring ways to overcome challenges of stool parasite detection with AI. In this regard, a study is currently ongoing with the aim to integrate automated algorithms into different low-cost innovative optical devices for the detection and quantification of STH and *Schistosoma* infections. The performance evaluation of the device is being evaluated in field settings in Nigeria and Gabon through the INSPIRED WP-3 (ClinicalTrials.gov identifier: NCT04505046) project (INSPIRED-project; <http://inspired-diagnostics.info>).

The final rank from the 19 different devices classified in this review reveals that 85% are not ready for commercialization (TRL<8). This is indicative of the need for more research and development on optical devices to detect STH and *Schistosoma* infections. Furthermore, stakeholder engagement is recommended (Levecke *et al.*, 2021) as it provides further information on stakeholder desired products and context-specific needs that in turn accelerate its readiness and enable local productivity. In addition, before such devices become commercially available, validation of their diagnostic performance and operational characteristics by healthcare professionals and public health teams is required, in laboratory, clinical and field settings (Banoo *et al.*, 2006). A proper diagnostic performance evaluation requires an appropriate and justifiable sample size to ensure a meaningful value of sensitivity and specificity with a sufficient level of power and low type-1 errors (Bujang and Adnan, 2016). For all the diagnostic performance evaluation studies reviewed here, no justifiable sample sizes were used which decreases the power and increases the likelihood for type-1 error. Also, the choice of reference tests used to evaluate these diagnostic devices is crucial. In an ideal situation, a reference test should be 100% sensitive and specific which is usually not the case for any of the soil-transmitted helminthiasis and schistosomiasis tests reported. In such a case, a composite of 2 or more reference tests to match-up this specification is encouraged. All the studies reviewed here used conventional microscopy as a reference standard whose sensitivity may vary enormously based on the settings and experience of the microscopists. Thus, the performance of the diagnostic tool is likely to be underestimated. While evaluating such tools aiming for a certain diagnostic performance, the analysis time needed to develop a test result per device is crucial. Only Sukas *et al.* reported the time of analysis of the LOD (15 min).

Of all the parasites of interest from the studies included in this review, hookworm was the least represented. Only 2 out of the 18 studies reported on diagnosis of hookworm infection. We envisage that this might be due to the rapid disintegration of hookworm eggs in a KK smear, which hampers proper comparison between diagnostic procedures. Several diagnostic parameters were used to evaluate these optical devices depending on the goal for which it was developed. PPV and NPV, for example, which are important parameters for clinician's decision making about diagnosis, were not evaluated for most of the studies. This is because most of the optical devices were developed for epidemiology/research surveys and not for point of care. We conclude that there is still much to be done to get automated microscopes ready for application in areas with limited resources. Such settings will require reliable and simple novel diagnostics that are equally good or superior to conventional microscopy which also aligns with the WHO TPP.

Data availability. The authors confirm that the data supporting the findings of this review are available within the article.

Author's contribution. B. M. T., C. H. H., L. V. L., T. E. A. and A. A. A. conceived the review. B. M. T. searched the literature and drafted the manuscript. M. B. designed the figures. B. M. T. and T. E. A. developed the adapted scale; M. B., J.-C. D. and L. V. L. and reviewed it. M. B., L. V. L., A. A. A., A. K., J.-C. D. and T. E. A. revised and edited the manuscript. All authors edited and revised the final manuscript.

Financial support. This work was supported by the Netherland organization for scientific research NWO-WOTRO Science for Global Development programme, grant number W 07.30318.009 (INSPIRED – INclusive diagnoStics for Poverty RElated parasitic Diseases in Nigeria and Gabon).

Conflict of interest. None.

Ethical standards. Not applicable.

References

- Agbana TE, Diehl JC, Van Pul F, Khan SM, Patlan V, Verhaegen M and Vdovin G (2018) Imaging & identification of malaria parasites using cell-phone microscope with a ball lens. *PLoS ONE* **13**, e0205020.
- Agbana TE, Nijman P, Hoerber M, Van Grootheest D, Van Diepen A, Van Lieshout L, Diehl J-C, Verhaegen M, Vdovine G and Coté GL (2020) Detection of *Schistosoma haematobium* using lensless imaging and flow cytometry, a proof of principle study. *Optical Diagnostics and Sensing XX: Toward Point-of-Care Diagnostics*.
- Ajibola O, Gulumbe BH, Eze AA and Obishakin E (2018) Tools for detection of schistosomiasis in resource limited settings. *Medical Sciences* **6**, 39.
- Alva A, Cangalaya C, Quiliano M, Krebs C, Gilman RH, Sheen P and Zimic M (2017) Mathematical algorithm for the automatic recognition of intestinal parasites. *PLoS ONE* **12**, e0175646.
- Amoah AS, Hoekstra PT, Casacuberta-Partal M, Coffeng LE, Corstjens PLAM, Greco B, Van Lieshout L, Lim MD, Markwalter CF, Odiere MR, Reinhard-Rupp J, Roestenberg M, Stothard R, Tchuem Tchuente L-A, De Vlas SJ and Van Dam GJ (2020) Sensitive diagnostic tools and targeted drug administration strategies are needed to eliminate schistosomiasis. *The Lancet Infectious Diseases* **20**, e165–e172.
- Armstrong M, Harris AR, D'ambrosio MV, Coulibaly JT, Essien-Baidoo S, Ephraim RKD, Andrews JR, Bogoch I and Fletcher DA (2022) Point-of-care sample preparation and automated quantitative detection of *Schistosoma haematobium* using mobile phone microscopy. *American Journal of Tropical Medicine and Hygiene* **106**, 1442–1449.
- Banoo S, Bell D, Bossuyt P, Herring A, Mabey D, Poole F, Smith PG, Sriram N, Wongsrichanalai C, Linke R and O'Brien R (2006) Evaluation of diagnostic tests for infectious diseases: general principles. *Nature Reviews Microbiology* **4**, S20–S32.
- Bogoch I, Andrews JR, Speich B, Utzinger J, Ame SM, Ali SM and Keiser J (2013) Mobile phone microscopy for the diagnosis of soil-transmitted helminth infections: a proof-of-concept study. *American Journal of Tropical Medicine and Hygiene* **88**, 626–629.
- Bogoch I, Andrews JR, Speich B, Ame SM, Ali SM, Stothard JR, Utzinger J and Keiser J (2014a) Quantitative evaluation of a handheld light microscope for field diagnosis of soil-transmitted helminth infection. *American Journal of Tropical Medicine and Hygiene* **91**, 1138–1141.
- Bogoch I, Coulibaly JT, Andrews JR, Speich B, Keiser J, Stothard JR, N'goran EK and Utzinger J (2014b) Evaluation of portable microscopic devices for the diagnosis of *Schistosoma* and soil-transmitted helminth infection. *Parasitology* **141**, 1811–1818.
- Bogoch I, Koydemir HC, Tseng D, Ephraim RKD, Duah E, Tee J, Andrews JR and Ozcan A (2017) Evaluation of a mobile phone-based microscope for screening of *Schistosoma haematobium* infection in rural Ghana. *American Journal of Tropical Medicine and Hygiene* **96**, 1468–1471.
- Bruno I, Lobo G, Covino BV, Donarelli A, Marchetti V, Panni AS and Molinari F (2020) Technology readiness revisited: a proposal for extending the scope of impact assessment of European public services. *Proceedings of the 13th International Conference on Theory and Practice of Electronic Governance*, pp. 369–380.
- Bujang MA and Adnan TH (2016) Requirements for minimum sample size for sensitivity and specificity analysis. *Journal of Clinical and Diagnostic Research* **10**, YE01–YE06.
- Butploy N, Kanarkard W and Maleewong Intapan P (2021) Deep learning approach for *Ascaris lumbricoides* parasite egg classification. *Journal of Parasitology Research* **2021**, 6648038.
- Carbonell C, Rodríguez-Alonso B, López-Bernús A, Almeida H, Galindo-Pérez I, Velasco-Tirado V, Marcos M, Pardo-Lledias J and Belhassen-García M (2021) Clinical spectrum of schistosomiasis: an update. *Journal of Clinical Medicine* **10**, 5521.
- Casacuberta-Partal M, Hoekstra PT, Kornelis D, Van Lieshout L and Van Dam GJ (2019) An innovative and user-friendly scoring system for standardised quantitative interpretation of the urine-based point-of-care strip test (POC-CCA) for the diagnosis of intestinal schistosomiasis: a proof-of-concept study. *Acta Tropica* **199**, 105150.
- Cooke IR, Laing CJ, White LV, Wakes SJ and Sowerby SJ (2015) Analysis of menisci formed on cones for single field of view parasite egg microscopy. *Journal of Microscopy* **257**, 133–141.
- Cools P, Van Lieshout L, Koelewijn R, Addiss D, Ajampur SSR, Ayana M, Bradbury RS, Cantera JL, Dana D, Fischer K, Imtiaz R, Kabagenyi J, Lok J, McCarthy J, Mejia R, Mekonnen Z, Njenga SM, Othman N, Shao H,

- Traub R, Van Esbroeck M, Vercruyse J, Vlaminc J, Williams SA, Verweij JJ, Van Hellemond JJ and Levecke B (2020) First international external quality assessment scheme of nucleic acid amplification tests for the detection of *Schistosoma* and soil-transmitted helminths, including *Strongyloides*: a pilot study. *PLoS Neglected Tropical Diseases* **14**, e0008231.
- Cools P, Vlaminc J, Verweij JJ and Levecke B (2021) Quantitative PCR in soil-transmitted helminth epidemiology and control programs: toward a universal standard. *PLoS Neglected Tropical Diseases* **15**, e0009134.
- Corstjens P, De Dood CJ, Knopp S, Clements MN, Ortu G, Umulisa I, Ruberanziza E, Wittmann U, Kariuki T, Loverde P, Secor WE, Atkins L, Kinung'hi S, Binder S, Campbell CH, Colley DG and Van Dam GJ (2020) Circulating anodic antigen (CAA): a highly sensitive diagnostic biomarker to detect active *Schistosoma* infections-improvement and use during SCORE. *American Journal of Tropical Medicine and Hygiene* **103**, 50–57.
- Coulibaly JT, N'goran EK, Keiser J, Ouattara M, Bogoch II, Andrews JR and Bonfoh B (2016a) Evaluation of malaria diagnoses using a handheld light microscope in a community-based setting in rural Côte d'Ivoire. *The American Journal of Tropical Medicine and Hygiene* **95**, 831–834.
- Coulibaly JT, Ouattara M, D'ambrosio MV, Fletcher DA, Keiser J, Utzinger J, N'goran EK, Andrews JR and Bogoch I (2016b) Accuracy of mobile phone and handheld light microscopy for the diagnosis of schistosomiasis and intestinal protozoa infections in Cote d'Ivoire. *PLoS Neglected Tropical Diseases* **10**, e0004768.
- Dacal E, Bermejo-Peláez D, Lin L, Álamo E, Cuadrado D, Martínez Á, Mousa A, Postigo M, Soto A, Sukosd E, Vladimirov A, Mwandawiro C, Gichuki P, Williams NA, Muñoz J, Kepha S and Luengo-Oroz M (2021) Mobile microscopy and telemedicine platform assisted by deep learning for the quantification of *Trichuris trichiura* infection. *PLoS Neglected Tropical Diseases* **15**, e0009677.
- De Melo GC, Netto RLA, Mwangi VI, Salazar YEAR, De Souza Sampaio V, Monteiro WM, De Almeida e Val FF, Rocheleau A, Thota P and Lacerda MVG (2021) Performance of a sensitive haemozoin-based malaria diagnostic test validated for vivax malaria diagnosis in Brazilian Amazon. *Malaria Journal* **20**, 146.
- Ephraim RK, Duah E, Cybulski JS, Prakash M, D'ambrosio MV, Fletcher DA, Keiser J, Andrews JR and Bogoch I (2015) Diagnosis of *Schistosoma haematobium* infection with a mobile phone-mounted Foldscope and a reversed-lens CellScope in Ghana. *American Journal of Tropical Medicine and Hygiene* **92**, 1253–1256.
- Ezeamama AE, Bustinduy AL, Nkwata AK, Martinez L, Pabalan N, Boivin MJ and King CH (2018) Cognitive deficits and educational loss in children with schistosome infection – a systematic review and meta-analysis. *PLoS Neglected Tropical Diseases* **12**, e0005524.
- Frean J (2008) Improving quantitation of malaria parasite burden with digital image analysis. *Transactions of the Royal Society of Tropical Medicine and Hygiene* **102**, 1062–1063.
- Hashimoto M, Yatsushiro S, Yamamura S, Tanaka M, Sakamoto H, Ido Y, Kajimoto K, Bando M, Kido J-I and Kataoka M (2017) Hydrophilic-treated plastic plates for wide-range analysis of Giemsa-stained red blood cells and automated *Plasmodium* infection rate counting. *Malaria Journal* **16**, 321.
- Hernández-Neuta I, Neumann F, Brightmeyer J, Ba Tis T, Madaboosi N, Wei Q, Ozcan A and Nilsson M (2019) Smartphone-based clinical diagnostics: towards democratization of evidence-based health care. *Journal of Internal Medicine* **285**, 19–39.
- Hoekstra PT, Van Dam GJ and Van Lieshout L (2021) Context-specific procedures for the diagnosis of human schistosomiasis – a mini review. *Frontiers in Tropical Diseases* **2**, 722438.
- Holmstrom O, Linder N, Ngasala B, Martensson A, Linder E, Lundin M, Moilanen H, Suutala A, Diwan V and Lundin J (2017) Point-of-care mobile digital microscopy and deep learning for the detection of soil-transmitted helminths and *Schistosoma haematobium*. *Global Health Action* **10**, 1337325.
- Jiménez B, Maya C, Velásquez G, Torner F, Arambula F, Barrios JA and Velasco M (2016) Identification and quantification of pathogenic helminth eggs using a digital image system. *Experimental Parasitology* **166**, 164–172.
- Kaewkamnerd S, Uthaiyibull C, Intarapanich A, Pannarut M, Chaotieing S and Tongsimma S (2012) An automatic device for detection and classification of malaria parasite species in thick blood film. *BMC Bioinformatics* **13**, S18.
- Kamgno J, Pion SD, Chesnais CB, Bakalar MH, D'ambrosio MV, Mackenzie CD, Nana-Djeunga HC, Gounoue-Kamkumo R, Njitchouang G-R, Nwane P, Tchatchueng-Mbouga JB, Wanji S, Stolk WA, Fletcher DA, Klion AD, Nutman TB and Boussinesq M (2017) A test-and-not-treat strategy for onchocerciasis in *Loa loa*-endemic areas. *The New England Journal of Medicine* **377**, 2044–2052.
- Khurana S, Singh S and Mewara A (2021) Diagnostic techniques for soil-transmitted helminths – recent advances. *Research and Reports in Tropical Medicine* **12**, 181–196.
- Koydemir HC, Coulibaly JT, Tseng D, Bogoch I and Ozcan A (2019) Design and validation of a wide-field mobile phone microscope for the diagnosis of schistosomiasis. *Travel Medicine and Infectious Disease* **30**, 128–129.
- Kumar R, Verma AK, Shrivastava S, Thota P, Singh MP, Rajasubramaniam S, Das A and Bharti PK (2020) First successful field evaluation of new, one-minute haemozoin-based malaria diagnostic device. *EClinicalMedicine* **22**, 100347.
- Levecke B, Coffeng LE, Hanna C, Pullan RL and Gass KM (2021) Assessment of the required performance and the development of corresponding program decision rules for neglected tropical diseases diagnostic tests: monitoring and evaluation of soil-transmitted helminthiasis control programs as a case study. *PLoS Neglected Tropical Diseases* **15**, e0009740.
- Li Q, Li S, Liu X, He Z, Wang T, Xu Y, Guan H, Chen R, Qi S and Wang F (2020) FecalNet: automated detection of visible components in human feces using deep learning. *Medical Physics* **47**, 4212–4222.
- Linder E, Grote A, Varjo S, Linder N, Lebbad M, Lundin M, Diwan V, Hannuksela J and Lundin J (2013) On-chip imaging of *Schistosoma haematobium* eggs in urine for diagnosis by computer vision. *PLoS Neglected Tropical Diseases* **7**, e2547.
- Meurs L, Brien E, Mbow M, Ochola EA, Mboup S, Karanja DMS, Secor WE, Polman K and Van Lieshout L (2015) Is PCR the next reference standard for the diagnosis of *Schistosoma* in stool? A comparison with microscopy in Senegal and Kenya. *PLoS Neglected Tropical Diseases* **9**, e0003959.
- Moser W, Barenbold O, Mirams GJ, Cools P, Vlaminc J, Ali SM, Ame SM, Hattendorf J, Vounatsou P, Levecke B and Keiser J (2018) Diagnostic comparison between FEC-PAK2 and the Kato-Katz method for analyzing soil-transmitted helminth eggs in stool. *PLoS Neglected Tropical Diseases* **12**, e0006562.
- Obeng BB, Aryeetey YA, De Dood CJ, Amoah AS, Larbi IA, Deelder AM, Yazdanbakhsh M, Hartgers FC, Boakye DA, Verweij JJ, Van Dam GJ and Van Lieshout L (2008) Application of a circulating-cathodic-antigen (CCA) strip test and real-time PCR, in comparison with microscopy, for the detection of *Schistosoma haematobium* in urine samples from Ghana. *Annals of Tropical Medicine and Parasitology* **102**, 625–633.
- Osredkar J (2017) Point-of-care testing in laboratory medicine. *University Medical Centre Ljubljana, Clinical Institute of Clinical Chemistry and Biochemistry, Zaloška cesta 002, 1000 Ljubljana, Slovenia*.
- Oyibo P, Jujuvarapu S, Meulah B, Agbana T, Braakman I, Van Diepen A, Bengtson M, Van Lieshout L, Oyibo W, Vdovine G and Diehl J-C (2022) Schistoscope: an automated microscope with artificial intelligence for detection of *Schistosoma haematobium* eggs in resource-limited settings. *Micromachines* **13**, 643.
- Pabalan N, Singian E, Tabangay L, Jarjanazi H, Boivin MJ and Ezeamama AE (2018) Soil-transmitted helminth infection, loss of education and cognitive impairment in school-aged children: a systematic review and meta-analysis. *PLoS Neglected Tropical Diseases* **12**, e0005523.
- Peeling RW and Mabey D (2010) Point-of-care tests for diagnosing infections in the developing world. *Clinical Microbiology and Infection* **16**, 1062–1069.
- Raso G, Essé C, Dongo K, Ouattara M, Zouzou F, Hürlimann E, Koffi VA, Coulibaly G, Mahan V, Yapi RB, Koné S, Coulibaly JT, Meité A, Guéhi-Kabran MC, Bonfoh B, N'goran EK and Utzinger J (2018) An integrated approach to control soil-transmitted helminthiasis, schistosomiasis, intestinal protozoa infection, and diarrhea: protocol for a cluster randomized trial. *Jmir Research Protocols* **7**, e145.
- Saeed MA and Jabbar A (2018) 'Smart diagnosis' of parasitic diseases by use of smartphones. *Journal of Clinical Microbiology* **56**, e01469-17.
- Sanprasert V, Kerdkaew R, Srirunguang S, Charuchaibovorn S, Phadungsaksawasdi K and Nuchprayoon S (2019) Development of conventional multiplex PCR: a rapid technique for simultaneous detection of soil-transmitted helminths. *Pathogens* **8**, 152.
- Sowerby SJ, Crump JA, Johnstone MC, Krause KL and Hill PC (2016) Smartphone microscopy of parasite eggs accumulated into a single field of view. *American Journal of Tropical Medicine and Hygiene* **94**, 227–230.

- Sukas S, Van Dorst B, Kryj A, Lagatie O, De Malsche W and Stuyver LJ (2019) Development of a lab-on-a-disk platform with digital imaging for identification and counting of parasite eggs in human and animal stool. *Micromachines* **10**, 852.
- Switz NA, D'ambrosio MV and Fletcher DA (2014) Low-cost mobile phone microscopy with a reversed mobile phone camera lens. *PLoS ONE* **9**, e95330.
- Vasiman A, Stothard JR and Bogoch I (2019) Mobile phone devices and hand-held microscopes as diagnostic platforms for malaria and neglected tropical diseases (NTDs) in low-resource settings: a systematic review, historical perspective and future outlook. *Advances in Parasitology* **103**, 151–173.
- Wiegand RE, Secor WE, Fleming FM, French MD, King CH, Deol AK, Montgomery SP, Evans D, Utzinger J, Vounatsou P and De Vlas SJ (2021) Associations between infection intensity categories and morbidity prevalence in school-age children are much stronger for *Schistosoma haematobium* than for *S. mansoni*. *PLoS Neglected Tropical Diseases* **15**, e0009444.
- World Health Organisation (2021a) Diagnostic target product profiles for monitoring, evaluation and surveillance of schistosomiasis control programmes. Licence: CC BY-NC-SA 3.0 IGO. Geneva: World Health Organization. Available at <https://www.who.int/publications/i/item/9789240031104> [cited 2022 17/03].
- World Health Organisation (2021b) Public consultation: target product profiles for diagnostic tests to meet schistosomiasis and soil-transmitted helminth programme needs. Available at <https://www.who.int/news-room/articles-detail/public-consultation-target-product-profiles-for-diagnostic-tests-to-meet-schistosomiasis-and-soil-transmitted-helminth-programme-needs> [cited 2022 17/03].
- Yang A, Bakhtari N, Langdon-Embry L, Redwood E, Lapierre SG, Rakotomanga P, Rafalimanantsoa A, Santos JDD, Vigan-Womas I and Knoblauch AM (2019) Kankanet: an artificial neural network-based object detection smartphone application and mobile microscope as a point-of-care diagnostic aid for soil-transmitted helminthiasis. *PLoS Neglected Tropical Diseases* **13**, e0007577.