



Universiteit
Leiden
The Netherlands

Quantifying the impact of signal-to-background ratios on surgical discrimination of fluorescent lesions

Azargoshab, S.; Boekestijn, I.; Roestenberg, M.; KleinJan, G.H.; Hage, J.A. van der; Poel, H.G. van der; ... ; Leeuwen, F.W.B. van

Citation

Azargoshab, S., Boekestijn, I., Roestenberg, M., KleinJan, G. H., Hage, J. A. van der, Poel, H. G. van der, ... Leeuwen, F. W. B. van. (2022). Quantifying the impact of signal-to-background ratios on surgical discrimination of fluorescent lesions. *Molecular Imaging And Biology*, 180-189. doi:10.1007/s11307-022-01736-y

Version: Publisher's Version

License: [Creative Commons CC BY 4.0 license](https://creativecommons.org/licenses/by/4.0/)

Downloaded from: <https://hdl.handle.net/1887/3491508>

Note: To cite this publication please use the final published version (if applicable).



Quantifying the Impact of Signal-to-background Ratios on Surgical Discrimination of Fluorescent Lesions

Samaneh Azargoshasb^{1,2} · Imke Boekestijn^{1,3} · Meta Roestenberg^{4,5} · Gijs H. KleinJan⁶ · Jos A. van der Hage⁷ · Henk G. van der Poel² · Daphne D. D. Rietbergen^{1,3} · Matthias N. van Oosterom^{1,2} · Fijs W. B. van Leeuwen^{1,2}

Received: 14 February 2022 / Revised: 28 March 2022 / Accepted: 21 April 2022 / Published online: 16 June 2022
© The Author(s) 2022

Abstract

Purpose Surgical fluorescence guidance has gained popularity in various settings, e.g., minimally invasive robot-assisted laparoscopic surgery. In pursuit of novel receptor-targeted tracers, the field of fluorescence-guided surgery is currently moving toward increasingly lower signal intensities. This highlights the importance of understanding the impact of low fluorescence intensities on clinical decision making. This study uses kinematics to investigate the impact of signal-to-background ratios (SBR) on surgical performance.

Methods Using a custom grid exercise containing hidden fluorescent targets, a da Vinci Xi robot with Firefly fluorescence endoscope and ProGrasp and Maryland forceps instruments, we studied how the participants' (N = 16) actions were influenced by the fluorescent SBR. To monitor the surgeon's actions, the surgical instrument tip was tracked using a custom video-based tracking framework. The digitized instrument tracks were then subjected to multi-parametric kinematic analysis, allowing for the isolation of various metrics (e.g., velocity, jerkiness, tortuosity). These were incorporated in scores for dexterity (*Dx*), decision making (*DM*), overall performance (*PS*) and proficiency. All were related to the SBR values.

Results Multi-parametric analysis showed that task completion time, time spent in fluorescence-imaging mode and total pathlength are metrics that are directly related to the SBR. Below SBR 1.5, these values substantially increased, and handling errors became more frequent. The difference in *Dx* and *DM* between the targets that gave SBR < 1.50 and SBR > 1.50, indicates that the latter group generally yields a 2.5-fold higher *Dx* value and a threefold higher *DM* value. As these values provide the basis for the *PS* score, proficiency could only be achieved at SBR > 1.55.

Conclusion By tracking the surgical instruments we were able to, for the first time, quantitatively and objectively assess how the instrument positioning is impacted by fluorescent SBR. Our findings suggest that in ideal situations a minimum SBR of 1.5 is required to discriminate fluorescent lesions, a substantially lower value than the SBR 2 often reported in literature.

Key words Image-guided surgery · Surgical robotics · Fluorescence imaging · Indocyanine green (ICG) · Surgical performance

✉ Fijs W. B. van Leeuwen
F.W.B.van_Leeuwen@lumc.nl

¹ Interventional Molecular Imaging Laboratory, Department of Radiology, Leiden University Medical Center, Leiden, the Netherlands

² Department of Urology, Netherlands Cancer Institute-Antoni Van Leeuwenhoek Hospital, Amsterdam, the Netherlands

³ Section of Nuclear Medicine, Department of Radiology, Leiden University Medical Center, Leiden, the Netherlands

⁴ Department of Parasitology, Leiden University Medical Center, Leiden, the Netherlands

⁵ Department of Infectious Diseases, Leiden University Medical Center, Leiden, the Netherlands

⁶ Department of Urology, Leiden University Medical Center, Leiden, The Netherlands

⁷ Department of Surgery, Leiden University Medical Center, Leiden, the Netherlands

Introduction

In recent years, fluorescence-guided surgery has rapidly gained popularity. While many different fluorescent dyes have seen implementation in patients [1], surgical use of fluorescence is most often related to use of the near-infrared dye indocyanine green (ICG). This dye has been applied for diagnostic purposes in, for example, cardiology and ophthalmology since the late 1950s [2]. During the last decade surgeons have used ICG in applications such as lymph node (LN) mapping [3, 4], angiography (e.g., anastomosis) [5–8] and real-time identification of lesions (e.g., hepatobiliary lesions) [9, 10]. ICG is most extensively used during plastic surgery [11], urology [12] and gynecology [13, 14], but applications are rapidly expanding into general oncologic surgery [15], and head and neck surgery [16]. To facilitate ICG-based fluorescence guidance, established minimally invasive surgery platforms, such as laparoscopes and the da Vinci surgical robot, now include a near-infrared ICG imaging option [17, 18].

Generally, “free” or non-bound ICG is used for physiological imaging in doses of 5–25 mg/patient [19], thus providing relatively high local fluorescent concentrations. Fueled by the constant flow of nuclear medicine-based molecular imaging successes obtained with small molecules, peptides and antibodies, the field has moved toward the use of receptor-targeted fluorescent tracers. There are, however, some fundamental differences in the application of a receptor-targeted radiotracer at quantities below the micro-dosing level ($< 100 \mu\text{g}/\text{patient}$) [20] and use of a fluorescent-based receptor-targeted tracer at, e.g., 0.18 mg/kg [21]. For one nuclear medicine has a superior sensitivity and is capable of accurately detecting very low tracer quantities, hence the general compatibility with microdosing. The detection sensitivity for fluorescence is a.o. limited by light scattering and tissue attenuation. This means it not only is limited to superficial targets but also its sensitivity is inferior to that of nuclear medicine [21, 22]. Second, while the signal intensity at physiological imaging is directly related to the amount of tracer administered (in relation to the biological clearance half-life), the number of cell-surface receptors that can be targeted in a tumor volume is dictated by biology, at least in patients. Despite the positive effects that receptor internalization has on signal intensity, this still means there is a limit to the degree of tracer that can be accumulated in a tumor. One may even suggest that the accumulation of tracer in a tumor is fixed and can be calculated with, for example, positron emission tomography (PET) standard uptake volume (SUV) values. If that is the case it means that above certain doses the tracer uptake is no longer directly related

to the quantity of tracer administered, while the increase in background signals in non-target tissues is. Combined this means that key challenges are the prevention of false negatives (e.g., lesions missed due to ‘underdosing’ or low camera sensitivity [22]) or false positives (e.g., due to mistaking the reflectance of excitation light as signal or background signals in non-target tissue as the result of ‘overdosing’ [23]). Knowing this, physics indicates that, unless the brightness of fluorescent dyes and the sensitivity of cameras improve radically, most receptor-targeted fluorescence-guided surgery applications will come with relatively low signal intensities and relatively high background signals. This will ultimately reflect on the signal-to-background ratio (SBR) and the ability of fluorescence imaging to guide surgical decision making.

It has been posed that image-guided surgery relies upon obtaining good SBRs [24]. As long-term outcome data and randomized clinical studies are still rare, often visualization of targets, with a $\text{SBR} \geq 2$, is considered a surrogate endpoint for success [25, 26]. The significance of SBR values is further underlined by the rigorous pursuit of ways to increase the SBR. Approaches vary from using ‘therapeutic’ quantities of fluorescent tracers [21], to extended intervals between tracer administration and surgical imaging [27], and tuning of tracer pharmacokinetics to reduce local background [28]. More exotic approaches include the use of activatable dyes [29] or fluorescence lifetime imaging [29]. Despite all these efforts, as far as we know, there are no studies that describe how SBR values, and with that, fluorescence imaging, reflect on the surgical procedure itself. Meaning it is still unclear how a specific SBR value alters the way surgeons approach a target.

In robotic surgery most of the proficiency scoring is still qualitative (i.e., expert (video) assessment). Recent literature, however, suggests that objective and quantitative performance scoring can be realized by tracking mechanical movements of the robotic instruments (e.g., dVLogger Intuitive). A concept has even allowed first steps to be made to relate surgical performance to outcome [30–32]. Kinematic analysis has even become standard during virtual simulation training [33–35]. Recently, we reported that kinematics can also support the comparison between image guidance modalities [36]. We argue such quantitative assessments need to be inclusive for all laparoscopic and robotic surgery. With that in mind video-based isolation of kinematic metrics provides a universal and interchangeable solution that can easily be disseminated across different platforms.

Given the theoretical impact that fluorescence SBR has on surgical movement, we performed multi-dimensional kinematic analysis during a fluorescence guided surgery exercise. Following automated video assessments

surgical instrument movements were digitized and analyzed in detail.

Methods

Phantom Design

To evaluate the sufficient fluorescent SBR needed for robotic surgery, we created a custom silicone (4 × 7) grid-phantom setup (Dragon Skin FXPro silicone, mixed with coloring pigment, FormX, Amsterdam, the Netherlands) that contained 28 possible target locations that could be sealed off with silicone lids. As targets we used 33.33 μL fluorescent beads [37], containing different concentrations of ICG (1–0.0625 mg/mL dissolved in methanol) incorporated in epoxy resin (ratio: 1:6:1 for epoxy resin, epoxy hardener and ICG solution, respectively).

Instrument Tracking Framework

The surgical system used during these experiments was a da Vinci Xi robot with Firefly fluorescence chip-on-a-tip endoscope and ProGrasp™ and Maryland forceps instruments (Intuitive Surgical, Sunnyvale, CA; see Fig. 1). During the exercise as explained below, the ProGrasp served as the dominant instrument. To acquire the coordinate data of the ProGrasp™ instrument tip in 3-dimensional (3D) space, we made use of a previously described custom marker-based tracking framework [36]. In this setup the marker-based tracking accuracy proved to be 1.10 ± 0.74 mm, 0.50 ± 0.53 mm and 0.88 ± 0.99 mm in the x, y and z directions, respectively

[36]. As marker a yellow-colored and rectangular-shaped marker was placed on one side of the ProGrasp (Fig. 2A). Matching computer-vision software was created to segment the markers (based on shape and color) in the endoscopic-video output. The obtained 3D instrument paths were then digitized using custom algorithms (constructed in MATLAB®, the MathWorks, Inc.). From the data we extracted multi-dimensional kinematic metrics such as spatial features (e.g., total pathlength, straightness index) and temporal features (e.g., time of completion, speed, acceleration, jerkiness). For better interpretation of the movement data, the percentage of time spent in each square centimeter of the Firefly image plane (%s/cm²) was calculated and used to calculate a color-coded density plot [36].

Fluorescence Intensity Determination

The SBR values of the individual targets in the grid phantom were determined using a color-based image segmentation algorithm created in MATLAB® on video material recorded from the Firefly camera at around 20 cm distance. In these algorithms, non-fluorescent but saturated parts (i.e., presented as white reflections) of the images were first removed using HSV (hue saturation value) color segmentation. Afterward, green fluorescent intensity was determined of each pixel in the image using the green color channel of the filtered RGB (red green blue) images. The target's fluorescent intensity, and therefore signal value, is then determined as the mean pixel intensity value within a manual defined rectangle-shaped region of interest around the target's location. For background calculation the average of three locations adjacent to the targets was used.





Fig. 1 Schematic overview of the phantom-based study setup. Surgical instruments are tracked to be digitized for kinematic feature extraction. These features are then used to determine performance scores in relation to fluorescence signal-to-background ratios

Fig. 2 A-D) Experiment process consisting: A), start in white light image, B) target identification in Firefly fluorescence mode, C) target resection, D) confirm the right target location in Firefly fluorescence mode. E) X, Y, Z component of instrument movement resulting from customized tracking program, F) digitized traveled path of the ProGrasp instrument tip in 3D with position density (%s/cm²) plot on XY plane overlaid on phantom image and G) cluster analysis of Pearson correlation applied on the kinematic metrics extracted from the traveled paths by the ProGrasp instrument tip, handling errors occurring during the exercises and SBR's from all participants. The color bar implies the correlation strength between the features (red positive and green negative correlation). (SBR: signal-to-background ratio, J_{extr} : number of extreme peeks in Jerkiness, HE: handling error, Cr: curvature, AD: angular dispersion, ST: straightness index, J = jerkiness, a = acceleration, v = speed, t_{task} : task time, $t_{Firefly}$: Firefly fluorescent time, S: pathlength)



Exercise

Participants (n = 16; MD's (n = 4), researchers (n = 6), and engineers (n = 6) aged between 20 and 50) were asked to perform the exercise in which they had to locate and remove three randomly placed fluorescent targets from the grid phantom using the da Vinci Xi surgical system. Each participant was asked to perform one to five of such exercises, and the exercises did not end until all targets (three per exercise) were removed. This resulted in a total of three to fifteen fluorescent targets. The Firefly fluorescent imaging mode could be used to identify the target locations, but in

line with clinical practice removal of the fluorescent beads had to take place in 'white light' mode. Non-visualization in fluorescent imaging mode meant the participant had to randomly check potential target locations in the 4 × 7 grid, thus substantially altering his/her movement kinematics. During the exercise, opening of locations not containing a fluorescent target, was considered a handling error. Exercises were timed, as was the use of fluorescence mode. To limit occlusions in instrument tracking, the participants were instructed to keep the instrument marker visible as much as possible during the exercises.

Surgical Performance Scoring

Surgical performance is a sum of dexterity (Dx) and decision making (DM) [38]. In the exercise, the overall dexterity is represented by the normalized jerkiness (derivative of acceleration) in 3D over the entire procedural time from start (t_1) to finish (t_2) corrected for the pathlength (s). The Dx index can be calculated using Eq. 1 wherein a low dexterity index indicates a more focused performance [39].

$$Dx = \left(\int_{t_1}^{t_2} \left(\frac{\delta^3 x}{\delta t^3} \right)^2 + \left(\frac{\delta^3 y}{\delta t^3} \right)^2 + \left(\frac{\delta^3 z}{\delta t^3} \right)^2 dt \cdot \frac{\Delta t^5}{s^2} \right) \quad (1)$$

The DM index can be described by a correlation between the intentional movements; sudden changes in dexterity (ΔDx_{extr}), handling errors (HE) and procedural fluency (F), where a low DM index indicates a more certain participant. The changes in dexterity are described by the number of extreme peaks within the jerkiness ($\#J_{extr}$; extreme peaks are above 100,000 m/s³). Procedural fluency F is defined as a combination of the time spent in Firefly fluorescence mode ($t_{Firefly}$) and straightness index (ST), as shown below.

$$DM = wf_1 \cdot \Delta Dx_{extr} + wf_2 \cdot HE + wf_3 \cdot F \quad (2)$$

$$\Delta Dx_{extr} = \#J_{extr}$$

$$F = wf_3 \cdot e^{-\log(ST)} + wf_4 \cdot t_{Firefly}$$

The weight factors (i.e., wf_1 , wf_2 , wf_3 and wf_4) in which each of the features contribute to the DM were determined using a maximization on the linear fit between the features and the total pathlength, based on the assumption that lesser movements resemble a more proficient procedure [40]. This was realized using a MATLAB® optimization program. Here the sum of the weight factors equals to 1; $wf_1 + wf_2 + wf_3 + wf_4 = 1$ and wf range between [0, 1] with step size 0.02.

Furthermore, an overall performance score (PS) can be created using the Dx and DM indices. This PS , ranging from 0 to 1, is depicted in Eq. 3, where a low score resembles a poor performance and is constructed in the following manner based on the method published by Ganni et al. [41]. To establish this PS range [0–1], the Dx and DM indices are linearly transformed into performance components P_{Dx} and P_{DM} . This transformation was constructed in such manner that the median Dx value of the exercises with handling errors is resembled by a P_{Dx} of 0.6 and the median Dx value of the exercises without handling errors is resembled by a P_{Dx} of 0.9. The DM indices were transformed in an equal manner. The weightings in which the

P_{Dx} and P_{DM} contribute to the PS were determined using a principal component analysis (PCA) wherein the sum of the weight factors equals to 1: $wf_{Dx} + wf_{DM} = 1$.

$$PS = wf_{Dx} \cdot P_{Dx} + wf_{DM} \cdot P_{DM} \quad (3)$$

To determine what PS cutoff is considered to be proficient and therefore establish a proficiency level, the Z-score method can be used [42]. To avoid an underestimation of proficiency, only exercises showing no handling errors have been included in the Z-score calculation. Here the performance was rated proficient when the individual PS values are within a Z-score interval of [-2, 2] as is most common.

Statistics


Statistical significance between the performance was established via a two sampled t-test with the SPSS statistical software (IBM SPSS Statistics for Windows, Version 25.0), using a confidence interval of 95%.

Results

Kinematic Metrics Extraction of the Traveled Instrument Path and the Intra-feature Correlation

Figure 2A and B shows typical examples of white light and fluorescence imaging visualizations of the grid phantom, respectively. Figure 2C and D illustrates how opening of a lid can reveal a fluorescent target. During the exercises, SBRs of the targets as detected by the Firefly camera varied between 1.0027 and 3.4638 (total 120 targets). Analysis of the path traveled by the ProGrasp instrument helped to objectively evaluate the effect of SBR on the surgical proficiency (Fig. 2E and F). The position density in 2D (%/cm²) could be visualized through color coding. To get an indication in which location instrument time was largest, this value could be overlaid onto the phantom, see Fig. 2F x-, y- plane. Following the direct relation between SBR values and kinematic metrics extracted from the digitized instrument trajectories (Pearson correlation coefficient; Fig. 2G), we observed that the SBR is negatively correlated with pathlength, time to complete the task, time spent in Firefly fluorescent mode and number of handling errors. Positive correlations, in which the two variables move in tandem, are found for SBR and straightness index. Also, pathlength, task time, time spent in Firefly fluorescent mode are positively correlated with the number of handling errors and extremes in jerkiness.

Fig. 3 A) The relation between SBR and handling errors and the sufficient SBR of 1.52. A similar setup for the relation between the SBR and procedural time (B), total pathlength (C) and the time spend in Firefly fluorescence mode (D) with sufficient SBR of 1.42, 1.52 and 1.55, respectively



In general, minimizing handling errors is considered a critical aspect of refining surgical procedures [43, 44]. Hence, we grouped the exercises into sub-classes with and without handling errors. Relating these handling errors to the SBR values, helped identify what SBR support selective target removal and how it affects dexterity and decision making. Figure 3 shows that handling errors, total procedural time, total pathlength, as well as time spent in Firefly fluorescence mode were SBR dependent. The cutoff SBRs were defined where sudden change is observed in these features and results in SBR = 1.52, SBR

1.42, SBR 1.52 and SBR 1.55, respectively. Averaging these values indicates that the ‘sweet spot’ for target identification means SBR > 1.50 is desirable. At lower SBR values participants increased the number of handling errors which also converted to increased pathlength ($S_{avg,SBR<1.50} = 1256$ vs $S_{avg,SBR>1.50} = 568.3\text{mm}$) and task time ($t_{Task_{avg},SBR<1.50} = 99.57$ vs $t_{Task_{avg},SBR>1.50} = 19.77\text{s}$). The uncertainty of the target location also meant participants spent more ‘search’ time in Firefly fluorescence mode ($t_{Firefly_{avg},SBR<1.50} = 10.23$ vs $t_{Firefly_{avg},SBR>1.50} = 3.12\text{s}$). In




Fig. 4 A)The dexterity index, B) decision-making index and C) the overall performance score plotted against the SBR. This clearly shows with SBR > 1.55 the performance score is mostly above proficiency level with no handling errors (green)

some cases, however, the increased time in fluorescence mode did mean the targets could still be identified (see Fig. 3D).

The Relation Between SBR and Surgical Performance Scoring

From the kinematic metrics and handling errors we were able to determine the Dx and DM indices, which are positively correlated with the total pathlength. Note: Pathlength and completion time are the most common measure for performance indicated in the literature [45–47]. Figure 4A and B suggests that lowering of the SBR converts to a decreased performance as indicated by a higher Dx and DM . A comparison between $SBR < 1.50$ and $SBR > 1.50$ groups suggests that the latter group generally yields a 2.5-fold higher Dx value and a threefold higher DM value ($Dx_{avg,SBR>1.50} = 3.75$ vs $Dx_{avg,SBR<1.50} = 9.20$ and $DM_{avg,SBR>1.50} = 2.34$ vs $DM_{avg,SBR<1.50} = 8.09$). To score the performance (PS), these indices were weighted according to the PCA results, yielding $wf_{Dx} = 0.49$ and $wf_{DM} = 0.51$ (see Fig. 4C). Participants were considered proficient at $PS \geq 0.816$. Relating proficiency to the SBR, shows a SBR cutoff at 1.55 (Fig. 4C, pink line) above which all participants performed proficient. Note fully, some individuals still performed proficient even when showing handling errors. However, none of the participants showing multiple handling errors reached the proficiency level. There is a difference between the average PS of the exercises showing no handling errors (Fig. 4C; green) and the exercises with handling errors (Fig. 4C; red) ($PS_{avg} = 0.888$ vs. $PS_{avg} = 0.509$; $p = 0.06$). Furthermore, no handling errors occurred above a SBR of 1.53.

In the current study it was not possible to relate the backgrounds of the individuals to the performance, MD's ($PS_{median} = 0.7920$), researchers ($PS_{median} = 0.8245$) and engineers ($PS_{median} = 0.8641$; $p = 0.27$). If anything, the trend suggests the technical expertise/insight of the participant reflects positively on the score.

Discussion

By analyzing the kinematics of surgical instrument movement, it has become possible to objectively and quantitatively define below which SBR the value of fluorescence guidance starts to deteriorate. Our findings indicate that the cutoff lies around $SBR > 1.5$ (average value calculated from cutoffs indicated in Fig. 3 and 4). This value is markedly lower than the artificial SBR cutoff value of 2 that is often used in the literature [48, 49]. We argue that studies such as the one presented here help gain fundamental insight in the fluorescence signal intensities that are needed to guide surgeons in their actions.

We observed that lower concentration of fluorescent contrast showed impact on a variety of kinematic metrics, namely, time in Firefly fluorescence mode, time of task completion and total pathlength of the movement of the participant. Not only do these metrics relate to each other, feature correlation analysis also indicates they relate to: SBR, handling errors, extremes in jerkiness and straightness index (Fig. 2G). Below the SBR 1.5 cutoff value the decision-making process is negatively impacted by the operators inability to discriminate between the reflection of excitation light and low-intensity fluorescent emissions. Interestingly, this means that at low signal intensities also more non-targets are pursued (so-called false positives) and handling errors become more frequent. In actual surgery, false positives can of course also be caused by accumulation of a fluorescent tracer in non-target tissue.

Surgical training literature relies heavily on handling errors for performance scoring [50]. Our findings indicate that participants could show a high proficiency ($PS \geq 0.816$), despite having a single handling error (Fig. 4C). Since lower fluorescence intensities in targets automatically convert to a drop in SBR, there will be a direct link between our phantom data and clinical findings that indicate that the accuracy of fluorescence guidance deteriorates with decreasing fluorescence concentrations (for ICG in SN surgery, the concentration threshold has been defined as < 0.003 nmol/cm³) [21]. It should however be noted that the exercises that we used present a quite ideal situation viz. no bleeding. Therefore *in vivo* SBR values could be slightly different. Hence, implementing the SBR analysis and kinematic scoring in clinical trials is required to validate if the current SBR threshold directly translates to the *in vivo* situation. This is something that is currently under investigation. Hereby we should realize that the kinematic scoring of instrument movements only relates to the SBR values encountered in situ. Values that in general are much lower than values determined during back table analysis. For example, in their study Moore et al. indicated in situ fluorescent SBR values in the range 1.5–2.5 converted to much higher SBR values *ex vivo* (range 3.0–4.0) [51]. Recently, de Barros et al. reported a similar trend for radioguidance [52].

We based our current study on the commonly used fluorescent dye ICG and the da Vinci platform equipped with a Firefly fluorescence camera that is optimized to identify this particular dye. As such our findings reflect on a situation wherein all parameters are fully optimized. As mentioned before in the clinic multiple fluorescent dyes have been used [1], and there have been reports on using the Firefly camera for other dyes such as fluorescein [53] and IRDye800CW [54]. In these situations, the fluorescent emission may not be optimally matched to the camera. As such the SBR values may be structurally lower. However, assuming the fluorescence is displayed to the surgeon in the same (artificial)

green coloration with black and white background, the interpretation of the images and this SBR values will remain identical. That said, representation of the fluorescent emission in another coloration, e.g., white [55] or blue [56] or heatmap [48, 57], may change the interpretation of the surgical field, as will the display of white light images in conjunction with fluorescence [58].

Almost half the study participants could identify targets and show a proficient performance below the SBR 1.5 cutoff (Fig. 4C), but their accuracy dropped from 100% to 78.95%. This underscores that individual vision can play an important role in overall performance. A difference that did not relate to the background of the individuals. It still needs to be defined if training in fluorescence imaging could help enhance a surgeon's ability to identify low-intensity fluorescent hotspots. This requirement may, however, become obsolete in the future when neural networks take over such tasks and automatically analyze fluorescence signal intensities at a per pixel-level basis and a video-rate speed [59–61]. A similar effect has been reported for the surgeon's ability to perceive objects in 3D [62–65]. For the latter training and selection programs have already been proposed [62].

Our findings show some limitations. For one, we performed our initial evaluations in a phantom setup, rather than in clinical trials. The reason for this is the ease of instrument tracking [36, 66] and the ability to have controlled SBR values. Here it should be noted that the setup used (epoxy fluorescent beads in a silicon phantom) does not allow us to translate the SBR values detected by the camera to fluorescence concentrations in the targets or the attenuating effect of 'tissues' covering a target, for that we refer to previous studies [21, 53]. Furthermore, in this study, the efficiency of the task is based on the movements of the dominant hand only. Future studies should determine whether the same trends extend to both hands. Obviously, handling errors in phantom exercises do not per se translate to a potential surgical complication and thus may appear somewhat artificial. During actual surgery on patients, however, resecting healthy tissue may lead to complications. In these instances, the anatomical location of the target will strongly influence the impact that handling errors yield. Hence the number of potential handling errors should be expanded with a score of severity, meaning the *DM* and *PS* scores should be tailored specifically to these applications.

Conclusion

Through multi-parametric kinematic analysis and performance scoring we have gained practical insight in how fluorescence SBRs impact on the ability to surgically resect a target, indicating that a $SBR > 1.5$ is required. Given that the

quest for receptor-targeted fluorescent agents approaches the use of signal intensities at the boundaries of what is technically feasible, such insights may help refine the implementation of fluorescence guidance in clinical trials and during routine care. Ultimately, they may also help guide the further development of fluorescent tracers and fluorescence imaging modalities.

Supplementary Information The online version contains supplementary material available at <https://doi.org/10.1007/s11307-022-01736-y>.

Acknowledgements We thank all participants for their time and effort in participating, and Dr. Anne van der Eijk for facilitating use of the da Vinci Xi robotic system.

Funding The research is funded by an NWO-TTW-VICI grant (#16141).

Declarations

Conflict of Interest The authors declare that they have no further conflicts of interest.

Open Access This article is licensed under a Creative Commons Attribution 4.0 International License, which permits use, sharing, adaptation, distribution and reproduction in any medium or format, as long as you give appropriate credit to the original author(s) and the source, provide a link to the Creative Commons licence, and indicate if changes were made. The images or other third party material in this article are included in the article's Creative Commons licence, unless indicated otherwise in a credit line to the material. If material is not included in the article's Creative Commons licence and your intended use is not permitted by statutory regulation or exceeds the permitted use, you will need to obtain permission directly from the copyright holder. To view a copy of this licence, visit <http://creativecommons.org/licenses/by/4.0/>.

References

1. van Beurden F, van Willigen DM, Vojnovic B et al (2020) Multi-wavelength fluorescence in image-guided surgery, clinical feasibility and future perspectives. *Mol Imaging* 19:1536012120962333
2. Boni L, David G, Mangano A et al (2015) Clinical applications of indocyanine green (ICG) enhanced fluorescence in laparoscopic surgery. *Surg Endosc* 29:2046–2055
3. Rossi EC, Ivanova A, Boggess JF (2012) Robotically assisted fluorescence-guided lymph node mapping with ICG for gynecologic malignancies: a feasibility study. *Gynecol Oncol* 124:78–82
4. van der Vorst JR, Schaafsma BE, Verbeek FP et al (2013) Near-infrared fluorescence sentinel lymph node mapping of the oral cavity in head and neck cancer patients. *Oral Oncol* 49:15–19
5. Lavazza M, Liu X, Wu C et al (2016) Indocyanine green-enhanced fluorescence for assessing parathyroid perfusion during thyroidectomy. *Gland Surg* 5:512
6. De Nardi P, Elmore U, Maggi G et al (2020) Intraoperative angiography with indocyanine green to assess anastomosis perfusion in patients undergoing laparoscopic colorectal resection: results of a multicenter randomized controlled trial. *Surg Endosc* 34:53–60
7. Aslim EJ, Lee FJ, Gan VHL (2018) The utility of intraoperative near infrared fluorescence (NIR) imaging with indocyanine green

- (ICG) for the assessment of kidney allograft perfusion. *Journal of Transplantation* 2018. <https://doi.org/10.1155/2018/6703056>
8. Hellan M, Spinoglio G, Pigazzi A, Lagares-Garcia JA (2014) The influence of fluorescence imaging on the location of bowel transection during robotic left-sided colorectal surgery. *Surg Endosc* 28:1695–1702
 9. Achterberg FB, Mulder BGS, Meijer RP et al (2020) Real-time surgical margin assessment using ICG-fluorescence during laparoscopic and robot-assisted resections of colorectal liver metastases. *Annals of Translational Medicine* 8(21). <https://doi.org/10.21037/atm-20-1999>
 10. Ishizawa T, Fukushima N, Shibahara J et al (2009) Real-time identification of liver cancers by using indocyanine green fluorescent imaging. *Cancer* 115:2491–2504
 11. Burnier P, Niddam J, Bosc R, Hersant B, Meningaud J-P (2017) Indocyanine green applications in plastic surgery: a review of the literature. *J Plast Reconstr Aesthet Surg* 70:814–827
 12. Cacciamani GE, Shakir A, Tafuri A et al (2020) Best practices in near-infrared fluorescence imaging with indocyanine green (NIRF/ICG)-guided robotic urologic surgery: a systematic review-based expert consensus. *World J Urol* 38:883–896
 13. Zapardiel I, Alvarez J, Barahona M et al (2021) Utility of intraoperative fluorescence imaging in gynecologic surgery: systematic review and consensus statement. *Ann Surg Oncol* 28:3266–3278
 14. Handgraaf HJ, Verbeek FP, Tummers QR et al (2014) Real-time near-infrared fluorescence guided surgery in gynecologic oncology: a review of the current state of the art. *Gynecol Oncol* 135:606–613
 15. Marano A, Priora F, Lenti LM, Ravazzoni F, Quarati R, Spinoglio G (2013) Application of fluorescence in robotic general surgery: review of the literature and state of the art. *World J Surg* 37:2800–2811
 16. van Keulen S, Nishio N, Fakurnejad S et al (2019) The clinical application of fluorescence-guided surgery in head and neck cancer. *J Nucl Med* 60:758–763
 17. Daskalaki D, Aguilera F, Patton K, Giulianotti PC (2015) Fluorescence in robotic surgery. *J Surg Oncol* 112:250–256
 18. Pathak RA, Hemal AK (2019) Intraoperative ICG-fluorescence imaging for robotic-assisted urologic surgery: current status and review of literature. *Int Urol Nephrol* 51:765–771
 19. van Manen L, Handgraaf HJ, Diana M et al (2018) A practical guide for the use of indocyanine green and methylene blue in fluorescence-guided abdominal surgery. *J Surg Oncol* 118:283–300
 20. Tewari T, Mukherjee S (2010) Microdosing: concept, application and relevance. *Perspect Clin Res* 1:61
 21. KleinJan GH, Bunschoten A, van den Berg NS et al (2016) Fluorescence guided surgery and tracer-dose, fact or fiction? *Eur J Nucl Med Mol Imaging* 43:1857–1867. <https://doi.org/10.1007/s00259-016-3372-y>
 22. Meershoek P, Buckle T, van Oosterom MN, KleinJan GH, van der Poel HG, van Leeuwen FW (2020) Can intraoperative fluorescence imaging identify all lesions while the road map created by preoperative nuclear imaging is masked? *J Nucl Med* 61:834–841
 23. Hernot S, van Manen L, Debie P, Mieog JSD, Vahrmeijer AL (2019) Latest developments in molecular tracers for fluorescence image-guided cancer surgery. *Lancet Oncol* 20:e354–e367
 24. Keereweer S, Kerrebijn JD, Van Driel PB et al (2011) Optical image-guided surgery—where do we stand? *Mol Imag Biol* 13:199–207
 25. Jiang JX, Keating JJ, De Jesus EM et al (2015) Optimization of the enhanced permeability and retention effect for near-infrared imaging of solid tumors with indocyanine green. *Am J Nucl Med Mol Imaging* 5:390
 26. Cho SS, Jeon J, Buch L et al (2018) Intraoperative near-infrared imaging with receptor-specific versus passive delivery of fluorescent agents in pituitary adenomas. *J Neurosurg* 131:1974–1984
 27. Xia X, Gai Y, Feng H et al (2020) Fluorescence Imaging Lung Cancer with a Small Molecule MHI-148. *J Fluoresc* 30:1523–1530
 28. Hensbergen AW, Buckle T, van Willigen DM et al (2020) Hybrid tracers based on cyanine backbones targeting prostate-specific membrane antigen: tuning pharmacokinetic properties and exploring dye–protein interaction. *J Nucl Med* 61:234–241
 29. Rood MT, Oikonomou M, Buckle T et al (2014) An activatable, polarity dependent, dual-luminescent imaging agent with a long luminescence lifetime. *Chem Commun* 50:9733–9736
 30. Hung AJ, Chen J, Jarc A, Hatcher D, Djaladat H, Gill IS (2018) Development and validation of objective performance metrics for robot-assisted radical prostatectomy: a pilot study. *J Urol* 199:296–304
 31. Judkins TN, Oleynikov D, Stergiou N (2009) Objective evaluation of expert and novice performance during robotic surgical training tasks. *Surg Endosc* 23:590–597
 32. Hung AJ, Chen J, Che Z et al (2018) Utilizing machine learning and automated performance metrics to evaluate robot-assisted radical prostatectomy performance and predict outcomes. *J Endourol* 32:438–444
 33. Perrenot C, Perez M, Tran N et al (2012) The virtual reality simulator dV-Trainer® is a valid assessment tool for robotic surgical skills. *Surg Endosc* 26:2587–2593
 34. Bric JD, Lumbard DC, Frelich MJ, Gould JC (2016) Current state of virtual reality simulation in robotic surgery training: a review. *Surg Endosc* 30:2169–2178
 35. Chan J, Pangal DJ, Cardinal T et al (2021) A systematic review of virtual reality for the assessment of technical skills in neurosurgery. *Neurosurg Focus* 51:E15
 36. Azargoshab S, van Alphen S, Slof LJ et al (2021) The Click-On gamma probe, a second-generation tethered robotic gamma probe that improves dexterity and surgical decision-making. *Eur J Nucl Med Mol Imaging* 48(13):4142–4151. <https://doi.org/10.1007/s00259-021-05387-z>
 37. van Oosterom MN, den Houting DA, van de Velde CJ, van Leeuwen FW (2018) Navigating surgical fluorescence cameras using near-infrared optical tracking. *J Biomed Opt* 23:056003
 38. Spencer F (1978) Teaching and measuring surgical techniques: the technical evaluation of competence. *Bull Am Coll Surg* 63:9–12
 39. Ghasemloonia A, Maddahi Y, Zareinia K, Lama S, Dort JC, Sutherland GR (2017) Surgical skill assessment using motion quality and smoothness. *J Surg Educ* 74:295–305
 40. Brunyé TT, Gardony AL, Holmes A, Taylor HA (2018) Spatial decision dynamics during wayfinding: Intersections prompt the decision-making process. *Cogn Res: Princ Implications* 3:1–19
 41. Ganni S, Botden SM, Chmarra M, Li M, Goossens RH, Jakimowicz JJ (2020) Validation of Motion Tracking Software for Evaluation of Surgical Performance in Laparoscopic Cholecystectomy. *J Med Syst* 44:1–5
 42. No A, Committee AM (2016) z-Scores and other scores in chemical proficiency testing—their meanings, and some common misconceptions. *Anal Methods* 8:5553–5555
 43. Conen D (2011) Measures to enhance patient safety. Importance of efficiency evaluation. *Bundesgesundheitsbl Gesundheitsforsch Gesundheitsschutz* 54:171–175
 44. Carayon P, Wood KE (2009) Patient safety. *Inf Knowl Syst Manag* 8:23–46
 45. Alaker M, Wynn GR, Arulampalam T (2016) Virtual reality training in laparoscopic surgery: a systematic review & meta-analysis. *Int J Surg* 29:85–94
 46. Mason JD, Ansell J, Warren N, Torkington J (2013) Is motion analysis a valid tool for assessing laparoscopic skill? *Surg Endosc* 27:1468–1477
 47. Farcas MA, Azzie G (2020) Performance assessment-The knowledge, skills and attitudes of surgical performance. *Semin Pediatr*

- Surg 29(2):150903. <https://doi.org/10.1016/j.sempedsurg.2020.150903>
48. de Vries HM, Bekers E, van Oosterom MN et al (2022) c-MET Receptor-Targeted Fluorescence on the Road to Image-Guided Surgery in Penile Squamous Cell Carcinoma Patients. *J Nucl Med* 63:51–56
 49. Korb ML, Huh WK, Boone JD et al (2015) Laparoscopic fluorescent visualization of the ureter with intravenous IRDye800CW. *J Minim Invasive Gynecol* 22:799–806
 50. Gallagher AG (2012) Metric-based simulation training to proficiency in medical education:-what it is and how to do it. *Ulst Med J* 81:107
 51. Moore LS, Rosenthal EL, Chung TK et al (2017) Characterizing the utility and limitations of repurposing an open-field optical imaging device for fluorescence-guided surgery in head and neck cancer patients. *J Nucl Med* 58:246–251
 52. de Barros HA, van Oosterom MN, Donswijk ML et al (2022) Robot-assisted prostate-specific membrane antigen–radioguided salvage surgery in recurrent prostate cancer using a DROP-IN gamma probe: The first prospective feasibility study. *Eur Urol*. <https://doi.org/10.1016/j.eururo.2022.03.002>
 53. Meershoek P, KleinJan GH, van Oosterom MN et al (2018) Multispectral-fluorescence imaging as a tool to separate healthy from disease-related lymphatic anatomy during robot-assisted laparoscopy. *J Nucl Med* 59:1757–1760
 54. Liss MA, Farshchi-Heydari S, Qin Z et al (2014) Preclinical evaluation of robotic-assisted sentinel lymph node fluorescence imaging. *J Nucl Med* 55:1552–1556
 55. van den Berg NS, Miwa M, KleinJan GH et al (2016) (Near-infrared) fluorescence-guided surgery under ambient light conditions: a next step to embedment of the technology in clinical routine. *Ann Surg Oncol* 23:2586–2595
 56. KleinJan GH, van den Berg NS, Brouwer OR et al (2014) Optimisation of fluorescence guidance during robot-assisted laparoscopic sentinel node biopsy for prostate cancer. *Eur Urol* 66:991–998
 57. Krishnan G, van den Berg NS, Nishio N et al (2022) Fluorescent molecular imaging can improve intraoperative sentinel margin detection in oral squamous cell carcinoma. *Journal of Nuclear Medicine*. <https://doi.org/10.2967/jnumed.121.262235>
 58. KleinJan GH, van den Berg NS, de Jong J et al (2016) Multimodal hybrid imaging agents for sentinel node mapping as a means to (re) connect nuclear medicine to advances made in robot-assisted surgery. *Eur J Nucl Med Mol Imaging* 43:1278–1287
 59. Buckle T, van Alphen M, van Oosterom MN et al (2021) Translation of c-Met Targeted Image-Guided Surgery Solutions in Oral Cavity Cancer—Initial Proof of Concept Data. *Cancers* 13:2674
 60. Cahill R, O’shea D, Khan M et al (2021) Artificial intelligence indocyanine green (ICG) perfusion for colorectal cancer intraoperative tissue classification. *Br J Surg* 108:5–9
 61. Zhuk S, Epperlein JP, Nair R et al (2020) Perfusion quantification from endoscopic videos: learning to read tumor signatures. In: *International Conference on Medical Image Computing and Computer-Assisted Intervention MICCAI 2020. Lecture Notes in Computer Science*, vol 12263. Springer, Cham
 62. Bogomolova K, Hierck BP, van der Hage JA, Hovius SE (2020) Anatomy dissection course improves the initially lower levels of visual-spatial abilities of medical undergraduates. *Anat Sci Educ* 13:333–342
 63. Bogomolova K, Vorstenbosch MA, El Messaoudi I et al (2021) Effect of binocular disparity on learning anatomy with stereoscopic augmented reality visualization: A double center randomized controlled trial. *Anat Sci Educ* 00:1–12. <https://doi.org/10.1002/ase.216>
 64. Bogomolova K, Hierck BP, Looijen AE et al (2021) Stereoscopic three-dimensional visualisation technology in anatomy learning: A meta-analysis. *Med Educ* 55:317–327
 65. Bogomolova K, van der Ham IJ, Dankbaar ME et al (2020) The effect of stereoscopic Augmented Reality visualization on learning anatomy and the modifying effect of visual-spatial abilities: a double-center randomized controlled trial. *Anat Sci Educ* 13:558–567
 66. Azargoshasb S, Houwing KH, Roos PR et al (2021) Optical navigation of a DROP-IN gamma probe as a means to strengthen the connection between robot-assisted and radioguided surgery. *Journal of Nuclear Medicine* 62(9):1314–1317. <https://doi.org/10.2967/jnumed.120.259796>

Publisher’s Note Springer Nature remains neutral with regard to jurisdictional claims in published maps and institutional affiliations.