



Universiteit
Leiden
The Netherlands

Little discrepancy between one-stage and chromogenic factor VIII (FVIII)/IX assays in a large international cohort of persons with nonsevere hemophilia A and B

Zwagemaker, A.F.; Kloosterman, F.R.; Gouw, S.C.; Boyce, S.; Brons, P.; Cnossen, M.H.; ... ;
DYNAMO Study Grp

Citation

Zwagemaker, A. F., Kloosterman, F. R., Gouw, S. C., Boyce, S., Brons, P., Cnossen, M. H., ... Cnossen, M. (2023). Little discrepancy between one-stage and chromogenic factor VIII (FVIII)/IX assays in a large international cohort of persons with nonsevere hemophilia A and B. *Journal Of Thrombosis And Haemostasis*, 21(4), 850-861.
doi:10.1016/j.jtha.2022.11.040

Version: Publisher's Version
License: [Creative Commons CC BY 4.0 license](#)
Downloaded from: <https://hdl.handle.net/1887/3761773>

Note: To cite this publication please use the final published version (if applicable).

ORIGINAL ARTICLE

Little discrepancy between one-stage and chromogenic factor VIII (FVIII)/IX assays in a large international cohort of persons with nonsevere hemophilia A and B

Anne-Fleur Zwagemaker¹ | Fabienne R. Kloosterman¹ | Samantha C. Gouw¹ | Sara Boyce² | Paul Brons³ | Marjon H. Cnossen⁴ | Peter W. Collins⁵ | Jeroen Eikenboom⁶ | Charles Hay⁷ | Rutger C. C. Hengeveld⁸ | Shannon Jackson⁹ | Caroline A. M. Kloppe-Tol⁸ | Marieke J. H. A. Kruip¹⁰ | Britta Laros-van Gorkom¹¹ | Christoph Male¹² | Laurens Nieuwenhuizen¹³ | Susan Shapiro^{14,15} | Karin Fijnvandraat^{1,16} | Michiel Coppens^{17,18} | on behalf of the DYNAMO study group

¹Amsterdam UMC, University of Amsterdam, Emma Children's Hospital, Pediatric Hematology, Meibergdreef, Amsterdam, The Netherlands

²Department of Haematology, University Hospital Southampton, Southampton, United Kingdom

³Department of Pediatric Hemato-Oncology, Radboud University Medical Center, Nijmegen, The Netherlands

⁴Department of Pediatric Hematology, Erasmus MC Sophia Children's Hospital, Erasmus University Medical Center, Rotterdam, The Netherlands

⁵Cardiff Haemophilia Centre, School of Medicine, Cardiff University, Cardiff, United Kingdom

⁶Department of Internal Medicine, Division of Thrombosis and Hemostasis, Leiden University Medical Center, Leiden, The Netherlands

⁷Manchester University Department of Haematology, Manchester Royal Infirmary, Manchester, United Kingdom

⁸Amsterdam UMC, University of Amsterdam, Clinical Chemistry, Meibergdreef, Amsterdam, The Netherlands

⁹Adult Bleeding Disorders Program of BC - Adult Division St. Paul's Hospital, Vancouver, British Columbia, Canada

¹⁰Department of Hematology, Erasmus MC, Erasmus University Medical Center, Rotterdam, The Netherlands

¹¹Department of Hematology, Radboud University Medical Center, Nijmegen, The Netherlands

¹²Department of Pediatrics, Medical University of Vienna, Vienna, Austria

¹³Department of Hematology, Maxima Medical Center, Veldhoven, The Netherlands

¹⁴Department of Haematology, Oxford University Hospitals NHS Foundation, Oxford NIHR Biomedical Research Centre, Oxford, United Kingdom

¹⁵Radcliffe Department of Medicine, Oxford University, Oxford, United Kingdom

¹⁶Department of Molecular Cellular Hemostasis, Sanquin Research and Landsteiner Laboratory, Amsterdam, The Netherlands

¹⁷Amsterdam UMC, University of Amsterdam, Vascular Medicine, Meibergdreef, Amsterdam, The Netherlands

¹⁸Amsterdam Cardiovascular Sciences, Pulmonary Hypertension & Thrombosis, Amsterdam, The Netherlands

Correspondence

Michiel Coppens, Department of Vascular Medicine, D3-329, Amsterdam University Medical Centers, Meibergdreef 9, Amsterdam, the Netherlands.
E-mail: m.coppens@amsterdamumc.nl

Abstract

Background: Accurate measurements of coagulation factor activity form an essential part of hemophilia management and are performed by the one-stage or chromogenic assay. Current literature suggests that approximately one-third of persons with non-severe hemophilia A exhibit assay discrepancy, albeit with a high variability between studies. Such data are scarce in nonsevere hemophilia B.

Manuscript handled by: R Gosselin

Final decision: R Gosselin, 28 November 2022

© 2022 Published by Elsevier Inc. on behalf of International Society on Thrombosis and Haemostasis.

Funding information

This work was funded by a grant from Novo Nordisk and supported by the Parelsnoer clinical biobank Hemophilia at the Health-RI (funded by the Ministry of Health, Welfare and Sport).

Objectives: To investigate the extent of factor VIII/IX one-stage and chromogenic assay discrepancy in moderate and mild hemophilia A and B.

Methods: Persons with previously diagnosed nonsevere hemophilia A and B with a factor level of 2 to 35 IU/dL were included from the international DYNAMO cohort study. Central measurements of the factor VIII and IX activity levels were performed by the one-stage and chromogenic assay. Relative and absolute discrepancy definitions were used, with the International Society on Thrombosis and Haemostasis-Scientific and Standardization Committee proposed ratio of >2.0 or <0.5 being the primary outcome. Discrepancy was also evaluated in a subgroup of 13 persons with mutations previously associated with discrepancy (≥ 3 cases reported in literature).

Results: A total of 220 persons were included, of whom 3 (1%) showed assay discrepancy: 2/175 hemophilia A and 1/45 hemophilia B. Six persons (3%) exhibited an absolute difference >10 IU/dL between the assay results. In addition, with more lenient definitions, over 90% of participants ($n = 197$) had no discrepant results. Only 1 out of 13 persons with a mutation previously associated with discrepancy had significant assay discrepancy.

Conclusion: Little assay discrepancy was observed despite the presence of mutations previously associated with discrepancy, suggesting that the presence and magnitude of assay discrepancy are largely determined by laboratory variables.

KEYWORDS

hemophilia A, hemophilia B, factor VIII, factor IX, blood coagulation tests

1 | INTRODUCTION

Hemophilia A and B are rare inherited coagulation disorders that result from a deficiency in clotting factor VIII (FVIII) or factor IX (FIX), respectively. The diagnosis and management of these bleeding disorders are primarily based on the residual endogenous factor activity level that determines their classification as severe (<1 IU/dL), moderate (1–5 IU/dL), or mild (>5 –40 IU/d) [1].

Different laboratory assays are used to measure FVIII and FIX activity. The one-stage assay (OSA) has been most widely adopted, with a reported use in 90% of laboratories [2]. In this assay, patient plasma, plasma deficient of the coagulation factor of interest and reagents are mixed and the resulting activated partial thromboplastin time (APTT) is compared with a calibration curve from serial dilutions of reference plasma [3]. The chromogenic assay (CA) has been less frequently available, as it is generally perceived as more expensive and technically complex [4,5] with a reported use in 68% of laboratories [2]. This assay measures the ability of FVIII or FIX to generate activated FX (FXa) and encompasses 2 stages. In the first stage, reagents are added to the patient plasma, which leads to activation of FX. In the second stage, the amount of FXa produced is measured through its action on a chromogenic substrate. The color intensity generated is proportional to the FVIII or FIX level that was present [3] as compared with a calibration curve from serial dilutions of reference plasma.

Essentials

- Factor VIII and IX activity can be measured by the one-stage and chromogenic assay.
- Central measurements with both assays were performed in a large nonsevere hemophilia cohort.
- Over 90% of 220 persons had no discrepancy irrespective of the definition used.
- Little discrepancy was found in persons with mutations previously associated with discrepancy.

Current literature suggests that around one-third of persons with nonsevere hemophilia A show discrepant results between the OSA and CA, although a high heterogeneity between studies is observed [6–8]. The Scientific and Standardization Committee (SSC) of the International Society on Thrombosis and Haemostasis (ISTH) defined a ratio of >2.0 or <0.5 as discrepant [9], although other studies have also used other definitions (ie, based on absolute difference >10 IU/dL or ratio >1.5 or <0.6). Classically, the CA measures lower FVIII/IX activity levels than the OSA, but reverse discrepancy, where the results of the OSA are lower than those of CA, have also been reported [6,10]. Previous studies demonstrated a correlation between the presence of assay discrepancy and specific mutations in the *F8* gene,

supporting the hypothesis for a genetic basis underlying assay discrepancy [7,11]. In general, lower levels in the CA compared to OSA were associated with missense mutations that result in a reduced stability of the FVIII protein [5,7]. In persons with lower levels observed by the OSA compared to CA, mutations are frequently located at functional sites of the FVIII protein related to thrombin activation, thereby binding to activated FIX (FIXa) or von Willebrand factor (VWF) [5,7]. In nonsevere hemophilia B, knowledge on assay discrepancy is limited with a single study demonstrating discrepant findings in 25% of people, all having lower FIX activity levels measured by OSA compared to CA [10].

Accurate measurement of FVIII and FIX activity levels form a critical part of hemophilia management, as the diagnosis and disease severity are established based on the residual endogenous factor level. In addition, factor levels are measured to guide treatment decisions and monitor therapy. To avoid potential misclassification, it has been recommended to perform both OSA and CA in the diagnostic work-up of nonsevere hemophilia A [7,9,12]. Knowledge on the extent of assay discrepancy and potential factors causing the heterogeneity observed is essential. As studies have reported varying results for hemophilia A and are scarce for hemophilia B, a large international study with centralized measurements may add more robust data to the existing body of work. Therefore, the aim of this study was to investigate the extent of assay discrepancy in persons with nonsevere hemophilia A and B in a multicenter international setting. Additionally, assay discrepancy was assessed in a subgroup of people with mutations associated with discrepancy in previous literature.

2 | METHODS

2.1 | Setting and design

The DYNAMO study was an observational cohort study that recruited persons with nonsevere hemophilia among 15 hemophilia treatment centers from January 2018 to May 2021. Participating centers were located in the Netherlands, United Kingdom, Italy, Austria, and Canada. Additional details on the participating sites can be found in the Supplementary material. The study received approval from all institutional review boards and was registered at [ClinicalTrials.gov](https://clinicaltrials.gov/ct2/show/study/NCT0362395) (NCT0362395). Persons with nonsevere hemophilia A or B with a baseline FVIII or FIX level between 2 and 35 IU/dL and aged 12 to 55 years were eligible for inclusion. As a consequence, the majority of this study cohort has been historically diagnosed with hemophilia based on the results from a OSA. Exclusion criteria were hemophilia B Leyden, history of a clinically relevant inhibitor, participation in a trial with an investigational product, use of anticoagulants, and the presence of a comorbidity affecting the musculoskeletal system. All participants provided written informed consent for this study, which included retrospective clinical data collection and a blood withdrawal. Data collected from medical files included demographics, F8 or F9 mutation according to the Human Genome Variation Society numbering, and treatment regimen. All historical baseline factor levels

measured locally were collected, including assay type (one-stage, chromogenic). Definitions used for data collection are listed in the Supplementary material.

2.2 | Study outcome

The study outcome was a discrepancy between baseline FVIII or FIX levels measured centrally by OSA and CA. The presence of assay discrepancy was defined in 2 ways: first, a ratio of >2.00 or <0.50 was defined as significant assay discrepancy as proposed by the SSC of the ISTH [9], and the second was defined as an absolute difference between assays of at least 10 IU/dL in accordance with a previous work by Schutte et al. [13]. The direction of assay results was reported as standard (higher OSA results) or reverse (higher CA results). For further detailed exploration, lenient assay discrepancy was defined as a relative difference between the assay results with ratios of >1.50 or <0.67 and an absolute difference between the assay results of at least 5 IU/dL. First, assay discrepancy was investigated in the full study population according to different definitions of assay discrepancy. Then, a subanalysis was performed in a subgroup of people carrying mutations associated with assay discrepancy in previous literature. Furthermore, centrally and locally measured levels by OSA and CA were compared in a subgroup of people in whom factor activity levels were measured locally with both assays on the same day at one or multiple occasions.

2.3 | Blood sampling and laboratory analysis

Blood withdrawal was performed during a clinic visit at the local hemophilia treatment center after a washout period of 3 days (hemophilia A) or 5 days (hemophilia B) for standard half-life products. For extended half-life products, the required washout period was determined by the treating physician. Blood was collected in citrated tubes and processed by local laboratories. Samples were double-spun for 15 minutes at 3000 g and 18 °C. The obtained plasma was stored frozen at -80 °C within 4 hours, and shipped on dry ice to the coagulation laboratory of the Amsterdam UMC. After the transport, all samples were thawed once after which the measurements were performed. Measurements of FVIII and FIX activity levels were performed centrally by both OSA and CA. The FVIII and FIX one-stage assays were measured with Actin FS reagents from Siemens and the FVIII and FIX CAs were measured with kits from Siemens and Rossix, respectively. FVIII- and FIX-deficient plasmas were obtained from Siemens in which other remaining coagulation factors, including VWF, were present in levels ≥ 40 IU/dL. The FVIII and FIX one-stage assays and FVIII CA were calibrated using standard human plasma from Siemens. The FIX CA was calibrated using a Biomed Hyphen Plasma Calibrator. Lot numbers are listed in the Supplementary material. The CAs were based on end-point methodology. Different reference curves were used. In the CA, FVIII samples >10 IU/dL and FIX samples >15 IU/dL were analyzed against a high reference standard curve (FVIII range, 136.5–1.4 IU/dL; FIX range,

194.0-12.1 IU/dL). FVIII samples <10 IU/dL and FIX samples <15 IU/dL were analyzed against a low reference standard curve (FVIII range, 11.4-0.4 IU/dL; FIX range, 24.3-0.4 IU/dL). In the one-stage assays, samples were analyzed against one standard reference curve in which specific low-range dilutions were used for samples <10 IU/dL. The incubation times were 260 seconds in the one-stage assays, 210 seconds in the FVIII CA, and 520 seconds in the FIX CA. All assays were validated for clinical patient care. Siemens control N and control P were used for internal quality control. The methods were also evaluated by participation in the ECAT external quality program. In all quality controls, our coagulation laboratory has performed well (at or around the average) in the period covering the central study measurements. All measurements were performed on a CS-2500 analyzer and in bulk. The same citrated plasma sample was used for both assays. The final results obtained centrally were compared to the most recent baseline factor level that was measured at the local laboratory. In case of marked differences, the local center was contacted to double check whether the sample was drawn after a washout period.

2.4 | Literature search

To identify all mutations associated with assay discrepancy, we performed a literature search. PubMed was searched using the key terms "hemophilia" and "assay discrepancy" on January 6, 2022. Studies from peer-reviewed journals were included when information was provided on specific *F8* or *F9* mutations associated with assay discrepancy and the full text was available. Catchments areas were evaluated to ensure that similar patient cohorts in different studies were not included twice. The reference list of a recent literature review on assay discrepancy in hemophilia A [6] was checked to retrieve any additional studies. Data was collected by a single reviewer (A.Z.) on study setting, discrepancy definition, and mutations resulting in assay discrepancy with direction (standard or reverse), including the total number of persons with this mutation in the total cohort if available.

2.5 | Statistical analysis

Data are presented descriptively as medians with interquartile ranges (IQRs) for continuous variables and as absolute numbers with proportions (%) for categorical variables. Scatterplots were constructed to visualize relationships between factor activity levels as measured with both assays. In addition, Bland-Altman plots were created to visualize the agreement [14]. The first plot was presented as ratio (OSA over CA) against the mean of the 2 assay results. The second plot was presented as the absolute difference (OSA – CA) against the mean of the 2 assay results. No limits of agreement were calculated as ratio, and difference did not follow a normal distribution. Instead, upper and lower bounds of the primary assay discrepancy definitions were shown in the plots. The analyses were performed using SPSS version 28 (IBM SPSS Statistics).

2.6 | DATA SHARING STATEMENT

For original data, please contact m.coppens@amsterdamumc.nl.

3 | RESULTS

3.1 | Population characteristics

A total of 304 persons participated in the DYNAMO study [15], and for 227 participants, central measurements were performed. Seven individuals were excluded as assay results showed differences in comparison to the expected values and were confirmed as not baseline or unknown after double checking with the local center. As a result, the present analysis included 220 persons, comprising 175 participants (80%) with hemophilia A and 45 participants (20%) with hemophilia B. The median lowest historical factor level measured at the local laboratories was 9 IU/dL (IQR 4-16). This resulted in 32% and 68% classified as moderate and mild hemophilia, respectively. Most participants received on demand treatment (94%). Table 1 summarizes the characteristics of the cohort.

3.2 | Extent of assay discrepancy

In hemophilia A, the median FVIII activity level was 10.6 IU/dL (IQR 4.8-18.3) and 9.7 IU/dL (IQR: 4.1-19.3), as measured with OSA and CA, respectively. In hemophilia B, the median FIX activity level was 10.7 IU/dL (IQR: 4.6-17.5) and 10.1 IU/dL (IQR 4.3-15.5) as measured with OSA and CA, respectively. An assay discrepancy defined according to the ISTH definition (ratio of >2.00 or <0.50) was detected in 3 people (1%): 2 with hemophilia A and 1 with hemophilia B. In 6 individuals (3%), an absolute difference of >10 IU/dL between assay measurements was found, including 4 individuals with hemophilia A and 2 individuals with hemophilia B. With more lenient discrepancy definitions, 8% and 5% of this cohort exhibited relative (ratios >1.50 or <0.67) and absolute (difference >5 IU/dL) discrepancies, respectively (Table 2). Detailed information on these participants is presented in Table 3. The correlation between assay results is demonstrated for the total cohort (Figure 1) and hemophilia A and B separately (Figure 2). The agreement between assay results is illustrated in Figure 3. The median difference between the assays was generally low with 1.1 IU/dL (IQR: 0.5-2.1) for the total cohort, 1.1 IU/dL (IQR: 0.5-2.0) in the case of higher one-stage results, and 1.3 IU/dL (IQR: 0.7-2.4) in case of higher chromogenic results. Larger ratios were observed within the lower factor activity range and larger absolute differences with increasing factor levels.

3.3 | Mutations with assay discrepancy

The *F8/F9* gene mutation was available in 152 of 220 (69%) participants of our study population. A total of 16 mutations were associated with assay discrepancy in our study according to any of the lenient

TABLE 1 Patient characteristics of the study cohort.

	Total cohort n = 220	Hemophilia A n = 175	Hemophilia B n = 45
Age (years)	39 (26-49)	40 (26-49)	36 (27-47)
Hemophilia severity			
Moderate	71 (32)	54 (31)	17 (38)
Mild	149 (68)	121 (69)	28 (62)
Lifetime lowest factor activity [IU/dL]	9 (4-16)	9 (5-16)	9 (3-15)
Treatment regimen			
Full prophylaxis	8 (4)	6 (3)	2 (4)
Intermittent prophylaxis	5 (2)	4 (2)	1 (2)
On demand	207 (94)	165 (94)	42 (93)

Characteristics are presented as medians with the corresponding IQR or as absolute numbers with proportions (%).

IQR, interquartile range.

definitions (Table 3). Five of these mutations were also identified in 12 other study participants (Arg717Trp, n = 4; Arg1960Gln, n = 2; Pro149, n = 4; Arg550His, n = 1; Ser2030Asn, n = 1). None of them exhibited assay discrepancy (Supplementary Table S1).

The literature search revealed 90 mutations that have been associated with assay discrepancy at least once (Supplementary Table S2). In our cohort, 60 participants (27% of cohort) carried any of these mutations.

When restricted to mutations previously reported in at least 3 people (Supplementary Table S3 [7,8,10,11,16–27]), a total of 13 study participants were identified carrying any of these mutations as presented in Table 4 (Arg1985Gln, n = 2; Arg550Cys, n = 1; Arg550His, n = 2; Arg717Trp, n = 5; Leu1951Phe, n = 1; Thr314Ala, n = 1; Arg19His, n = 1). In only 2 of them, an assay discrepancy was measured according to our strict definitions, and in a third participant, an assay discrepancy was measured according to the more lenient definition (absolute difference

>5 IU/dL). In the other 10 persons with mutations that were frequently reported to demonstrate assay discrepancy, this could not be reproduced in our study. Interestingly, in one participant with discrepancy carrying the Arg717Trp mutation, the CA results were >10 IU/dL higher than the OSA results. This is opposite to the direction of discrepancy typically found for this mutation, as other participants with Arg717Trp in this study and in previous literature had higher one-stage results in comparison to chromogenic results. A similar contradictory finding was also observed in 2 participants that carried the Arg1985Gln mutation. This mutation was previously reported to result in standard assay discrepancy (OSA > CA) but demonstrated slightly higher CA results in our study, although this was not large enough to be classified as assay discrepancy.

3.4 | Comparison with local measurements

In 9 participants from 5 different centers, assay discrepancy was previously identified with historical measurements at the local center. Strikingly, the results in all these persons exhibited no assay discrepancy with the central measurements (Supplementary Table S4). In 4 of 5 participants with multiple measurements over time, the results were also inconsistently discrepant at these different time points.

4 | DISCUSSION

In this international cohort of 220 persons with nonsevere hemophilia A and B, we found very little discrepancy between central OSA and CA measurements. Even with more lenient discrepancy definitions, the vast majority (90%) of this cohort had no discrepant results. This is in contrast to previous studies that reported assay discrepancy in around one-third of persons with nonsevere hemophilia. In addition, the majority of people with a F8/F9 gene mutation previously associated with assay discrepancy exhibited no discrepant results in our study. This suggests that assay discrepancy seems to be largely determined by laboratory variables.

4.1 | Prevalence of assay discrepancy

In nonsevere hemophilia A, assay discrepancy has been frequently addressed by previous studies. A recent literature review including 18 articles reported an estimated pooled prevalence of 36% (95% CI 28–44) assay discrepancy (range 12–84% in the individual studies) [6]. In nonsevere hemophilia B, data on assay discrepancy is scarce with only one article currently published [10]. This study examined 32 persons with nonsevere hemophilia B and found assay discrepancy in 25% of the study population. In our study, a strikingly lower prevalence of significant assay discrepancy was found for both hemophilia A (2/175 people—1%) and hemophilia B (1/45 people—2%). Some differences in prevalence between studies may be related to the criteria used to define assay discrepancy. While we used the definition of the ISTH-SSC for our primary outcome, some other studies applied less stringent definitions. However, even when we applied these more lenient

TABLE 2 Frequencies of assay discrepancy, including primary and secondary outcomes.

	Total cohort n = 220	Hemophilia A n = 175	Hemophilia B n = 45
Primary outcomes			
Ratio >2.00 or <0.50	3 (1)	2 (1)	1 (2)
Absolute difference >10 IU/dL	6 (3)	4 (2)	2 (4)
Secondary outcomes			
Ratio >1.50 or <0.67	17 (8)	13 (7)	4 (9)
Absolute difference >5 IU/dL	11 (5)	8 (5)	3 (7)

Data presented as absolute numbers with proportions (%).

TABLE 3 Characteristics of persons with assay discrepancy.

ID	Hemophilia type	Age (years)	OSA (IU/dL)	CA (IU/dL)	Ratio OSA/CA	Absolute difference (IU/dL)	Mutation protein change
1	A	23	20.0	6.0	3.33 ^a	14.0 ^a	-
2	A	28	7.8	3.5	2.23 ^a	4.3	Arg550Cys ^b
3	B	37	23.5	10.1	2.33 ^a	13.4 ^a	Thr84Ile
4	A	43	29.1	40.3	0.72	11.2 ^a	Arg717Trp ^b
5	A	51	22.8	37.1	0.61 ^c	14.3 ^a	-
6	A	43	29.4	49.2	0.60 ^c	19.8 ^a	Phe698Ser
7	B	51	44.3	23.1	1.92 ^c	21.2 ^a	-
8	A	53	0.8	0.4	2.00 ^c	0.4	-
9	A	45	6.3	3.2	1.97 ^c	3.1	Arg1960Gln
10	A	31	6.0	3.1	1.94 ^c	2.9	Gly498Arg
11	A	25	4.8	2.5	1.92 ^c	2.3	-
12	A	39	1.3	0.7	1.86 ^c	0.6	Arg2182His
13	A	14	3.5	2.1	1.67 ^c	1.4	-
14	A	29	4.3	2.7	1.59 ^c	1.6	Pro149Arg
15	A	54	5.3	8.9	0.60 ^c	3.6	Asp588Glu
16	A	44	1.7	3.2	0.53 ^c	1.5	Arg391His
17	B	40	7.8	4.6	1.70 ^c	3.2	Arg379Pro
18	B	52	2.2	1.3	1.69 ^c	0.9	Arg379Gln
19	A	21	28.7	22.2	1.29	6.5 ^c	Arg550His ^b
20	A	31	29.3	23.0	1.27	6.3 ^c	Ile1901Thr
21	A	14	31.8	37.2	0.85	5.4 ^c	Ser2030Asn
22	A	22	17.0	22.3	0.76	5.3 ^c	Arg2169Cys
23	B	56	23.5	16.1	1.46	7.4 ^c	-

CA, chromogenic assay; OSA, one-stage assay.

^a Results meeting the primary discrepancy definitions.

^b Mutation previously reported to be associated with assay discrepancy in at least 3 people.

^c Results meeting the more lenient discrepancy definitions.

definitions (ratio >1.50 or <0.67 and absolute difference >5 IU/dL), we still found less than 10% discrepancy. The impact of baseline factor level on discrepancy is also reflected by our Bland-Altman plots showing larger ratios with lower factor levels and larger absolute differences with higher factor levels. As such, assay discrepancy based on ratio difference is more likely to be observed in people with moderate hemophilia and that based on absolute difference is more likely to be observed in those with higher factor activities. Furthermore, heterogeneity may arise from differences in design, study population, and factors causing laboratory measurement variation.

4.2 | Mutations and assay discrepancy

In previous studies, assay discrepancy has been linked to several specific mutations, supporting the genetic background of this phenomenon [7,11]. In our study, we did not confirm these observations. The

mutations that resulted in assay discrepancy in our study were also identified in 12 other study participants, of whom none had discrepant results. Furthermore, 27% of our study cohort carried a mutation that was reported to display assay discrepancy in previous literature, but that could not be replicated in our study. Even when we further restricted our analysis to frequently reported mutations associated with discrepancy (reported in at least 3 patients in previous literature), in only one participant, a significant assay discrepancy was present. While the direction of assay results was as expected for most persons, we also observed inconsistencies. For the Arg717Trp mutation, we found one participant with an assay discrepancy (difference > 10 IU/dL) with higher CA results, while 4 other participants with this mutation had higher OSA results. Conflicting results have also been reported for the Arg546Trp mutation (not present in our study), resulting in both standard and reverse assay discrepancy among 3 different studies [7,11,21]. To the same extent, the Arg191His mutation caused assay discrepancy in 6/6 people with hemophilia B in the study by Kihlberg et al. [10]. In our




FIGURE 1 Scatterplot of measured factor activity with one-stage vs chromogenic assay for the total cohort. The colored dots represent people with significant assay discrepancy (red dots), people with an absolute difference >10 IU/dL between the assay results (yellow dots), and people meeting both these criteria (orange dots). The line reflects $x = y$.

study, we included one participant with this mutation in whom similar results by both assays were found. Our results suggest that the impact of specific mutations varies depending on assay conditions, further supporting the notion that the issue of assay discrepancy is mainly driven by laboratory-related factors.

4.3 | Measurement variation

Many different factors may contribute to variation in measured factor activity. First, preanalytical variables can influence the assay measurements, including inappropriate blood collection, processing, storage, or transport [28]. Second, analytical factors may affect measured clotting factor activity. Interassay and interlaboratory variability are frequently observed with estimated variation coefficients around 5 to 20% and 15 to 25%, respectively [28]. Additionally, in our own dataset we observed interlaboratory variability as reflected by differences between central and local measurements. These findings are in line with previous work investigating the analytical variation in FVIII activity in the ECAT external quality assessment program [29]. In that study, OSA results varied between setups from several manufacturers, which was partially explained by the calibrator value [29]. Other studies assessing different one-stage reagents reported varying factor activity results for FVIII [22] and FIX levels [30]. As a consequence, the presence of assay discrepancy and in some cases the direction may also depend on the reagents and setup used. The incubation time in the CA also has a potential impact as longer incubation times can result in lower measured factor activity levels in some patients [31]. Furthermore, many previously reported assay

discrepancies have been based on one-stage and two-stage legacy assays. While CAs are considered two-stage assays, historically, clot-based two-stage assays differ from the modern CAs. Chromogenic assays have only been available on a larger scale since the last decade, and show a considerable variability in the duration of the incubation phase, which may affect different mutations differentially. In addition, the varying use of kinetic or end-point methodology in the chromogenic assay may add to any differences observed. It is likely that all such factors including reagents, setups, and duration of incubation largely determined the magnitude of assay discrepancy. In our study, it is plausible that the little discrepancy observed is influenced by the particular single one-stage and chromogenic assay used. Third, in general, intraindividual variation in baseline factor levels may contribute to measurement variability. Previous work demonstrated that intraindividual variation explained 45% of the variability in FVIII levels adjusted for age, mutation, and hemophilia treatment center [32]. This raises the question of to what extent laboratory issues or other determinants such as stress account for the observed variance. Hence, multiple variables can influence measured factor activity and may contribute to the marked heterogeneity in results that are reported in this area of work.

4.4 | Strengths and limitations

The DYNAMO study included a relatively large cohort of persons with nonsevere hemophilia within a multicenter international setting. All assays were performed centrally in bulk and using the same plasma sample to reduce potential analytical variability as much as possible. A




FIGURE 2 Scatterplots of measured factor activity with one-stage vs chromogenic assay for hemophilia A and hemophilia B. The colored dots represent people with significant assay discrepancy (red dots), people with an absolute difference >10 IU/dL between the assay results (yellow dots), and people meeting both these criteria (orange dots). The line reflects $x = y$. FVIII, factor VIII; FIX, factor IX.

strength of this study was that we included patients with hemophilia B, as only one previous smaller study has focused on assay discrepancies in this population [10]. The local laboratories were provided a protocol with specifications to ensure similar processing and storage of samples. However, remaining variation arising from other pre-analytical factors such as errors in blood collection or inadequate temperature control at local sites cannot be completely ruled out. Another strength was the inclusion of individuals with mutations previously associated with assay discrepancy to put our observations

in perspective. A limitations of the present work is that mutation data was missing for 31% of our cohort. Furthermore, it may be that a longer incubation time for the chromogenic assay could have resulted in more discrepancy, nonetheless demonstrating that laboratory factors have critical impact on measured activity. As we only used one type of OSA and CA, it remains unknown whether the use of alternative kits would have resulted in different findings. Unfortunately, our data collection did not include details of the methodology of local factor level measurements (ie, reagent and manufacturer), which




FIGURE 3 Bland-Altman plots of measured factor activity with one-stage vs chromogenic assay as (A) ratio against mean and (B) absolute difference against mean. The colored dots represent people with significant assay discrepancy (red dots), people with an absolute difference >10 IU/dL between the assay results (yellow dots), and people meeting both these criteria (orange dots). (A) The bold line represents the mean ratio of 1.1 of one-stage over chromogenic results. The dashed lines represent the upper and lower ratios to define significant assay discrepancy (ratio >2.00 or <0.50 , respectively). (B) The bold line represents the mean absolute difference of 0.1 of one-stage minus chromogenic results. The dashed lines represent the upper and lower differences to define an absolute difference >10 IU/dL. CA, chromogenic assay; OSA, one-stage assay.

would allow a more direct comparison between central and local measurements. Finally, it needs to be emphasized that the majority of this study cohort has been historically diagnosed with hemophilia based on the results from a OSA as outlined below.

4.5 | Study cohort

In the DYNAMO study, persons with nonsevere hemophilia aged 12 to 55 years were eligible for inclusion. The median age at diagnosis was 3 years (IQR 0-12) [15] and consequently the majority of this population has been diagnosed based on OSA results. This may have led to a

potential selection bias in which persons with normal OSA results but a factor activity of 35 IU/dL or lower with the CA have not been included. As such, the study population may lack persons with (very) mild hemophilia that have not been accurately identified as hemophilia patient in the past. However, most historical hemophilia cohorts reporting assay discrepancy are prone to the same selection mechanism, and therefore, this does not explain why the assay discrepancy in our study is smaller than that previously reported in other studies of persons diagnosed with hemophilia with a (OSA) factor activity of 2 to 35 IU/dL. Nevertheless, considering the potential selection bias, our findings do not obviate the need to perform both assays in the work-up of a bleeding patient.

TABLE 4 Mutations described in literature associated with assay discrepancy and results in our cohort.

Literature		Central study measurements					
Mutation protein change	Direction of assay results ^a	ID	OSA (IU/dL)	CA (IU/dL)	Ratio OSA/CA	Absolute difference (IU/dL)	Direction of assay results ^a
Hemophilia A							
Thr314Ala	Reverse	1	17.7	22.0	0.80	4.3	Reverse
Arg550Cys	Standard	2	7.8	3.5	2.23 ^b	4.3	Standard
Arg550His	Standard	3	28.7	22.2	1.29	6.5 ^c	Standard
		4	31.0	27.7	1.12	3.3	Standard
Arg717Trp	Standard	5	29.1	40.3	0.72	11.2 ^b	Reverse
		6	20.5	18.7	1.10	1.8	Standard
		7	30.4	29.2	1.04	1.2	Standard
		8	29.7	28.7	1.03	1.0	Standard
		9	23.2	23.5	0.99	0.3	Standard
Leu1951Phe	Standard	10	6.9	5.4	1.28	1.5	Standard
Arg1985Gln	Standard	11	21.5	22.6	0.95	1.1	Reverse
		12	27.5	30.7	0.90	3.2	Reverse
Hemophilia B							
Arg191His	Reverse	13	6.6	7.1	0.93	0.5	Reverse

CA, chromogenic assay; OSA, one-stage assay.

^a Standard defined as higher one-stage compared to chromogenic results. Reverse defined as higher chromogenic compared to one-stage results.

^b Results meeting the primary discrepancy definitions are highlighted in dark orange.

^c Results meeting the more lenient discrepancy definitions are highlighted in light orange.

4.6 | Clinical implications

Accurate measurement of FVIII and FIX activity levels is vital to ensure adequate disease classification and management in hemophilia. The awareness that factor activity levels vary and depend on a range of laboratory- and patient-related variables is required. A low prevalence of assay discrepancy was observed in this cohort, which may be attributed to the particular measurement conditions used in our central measurements. This strengthens our hypothesis that assay discrepancy is largely driven by laboratory variation. Fortunately, in most patients, the factor activity will lead to a correct diagnosis whether it is measured by the OSA or CA. However, there is a subgroup of 5 to 10% of patients with mild hemophilia A that may have normal results by either OSA or CA [6,34] and in whom diagnosis will be missed when the diagnostic work-up is restricted to one assay. Thus, we support recommendations to use both the OSA and CA in the diagnostic work-up of nonsevere hemophilia [6,9,12,33]. For monitoring and routine care, one assay may be sufficiently informative for most clinically relevant decision-making when analytical variability is minimalized. When assay discrepancy does occur, uncertainties on which assay is most reflective of the “true” level, consistent with the clinical bleeding phenotype, will still exist [34]. It has been suggested that the lowest factor level by any assay correlates best with the phenotype [5], while some studies suggest that CA better reflects bleeding tendency [20,34–36]. Interestingly, few previous studies have reported on persons diagnosed with hemophilia based on low

OSA results that were identified following investigation of a prolonged APTT, rather than that of participants based on a clinical suspicion of hemophilia [26,37]. This raises the question of whether the low OSA values found in that study represented clinically relevant results. Within our study, we had a good agreement between assay results, which precludes any exploration on which assay best reflects the bleeding phenotype. Awaiting harmonization of factor level measurements, we suggest that laboratory results should be primarily evaluated in the context of the clinical profile of people. Considering the important influence of laboratory conditions on the presence and magnitude of assay discrepancy, we suggest repeated testing in case of discrepancy. As this study illustrates, some people labeled as discrepant may exhibit concordant activity levels under different circumstances. Further standardization of FVIII and FIX measurements is required to improve adequate diagnosis, clinical management, and future research on assay discrepancy in nonsevere hemophilia A and B.

5 | CONCLUSION

In conclusion, we found little assay discrepancy in this international cohort of 220 people with nonsevere hemophilia A and B, even in those people with mutations previously associated with discrepancy. This suggests that laboratory-related factors contribute largely to the presence and magnitude of assay discrepancy.

ACKNOWLEDGMENTS

This work was funded by a grant from Novo Nordisk and supported by the Parelsnoer clinical biobank Hemophilia at the Health-RI (funded by the Ministry of Health, Welfare and Sport). The funding sponsor had no role in the design and conduct of the study; data collection, management, analysis, and interpretation; and preparation, review, and approval of the manuscript. The authors thank Bianca Bakker, Debby Jongelings, and Caroline Kloppe-Tol (coagulation laboratory, Amsterdam UMC) for their support at the laboratory and for performing the assays. In addition, the authors thank all local research staff for their help in setting up and conducting the study.

AUTHOR CONTRIBUTIONS

K.F. designed the study. A.Z. analyzed the data and wrote the manuscript. A.Z., F.R.K., S.C.G., and M.C. interpreted the data and drafted the manuscript. S.C.G., S.B., P.B., M.H.C., P.W.C., J.E., C.H., S.J., M.J.H.A.K., B.L.G., C.M., L.N., S.S., K.F., and M.C. collected data or supervised data collection. R.C.C.H. and C.A.M.K. performed or supervised the assay measurements and interpreted the data from a laboratory perspective. All authors reviewed this work and approved the final version of the manuscript.

DECLARATION OF COMPETING INTEREST

This work was supported by a grant from Novo Nordisk. S.C.G. has received an unrestricted research grant from Sobi. S.B. received research support from Sangamo Therapeutics and honoraria from CSL Behring. M.H.C. has received investigator-initiated research and travel grants as well as speaker fees over the years from the Netherlands Organisation for Scientific Research (NWO), the Netherlands Organization for Health Research and Development (ZonMw), the Dutch "Innovatiefonds Zorgverzekeraars", Baxter/Baxalta/Shire/Takeda, Pfizer, Bayer Schering Pharma, CSL Behring, Sobi Biogen, Novo Nordisk, Novartis, and Nordic Pharma, and has served as a steering board member for Roche, Bayer, and Novartis. All grants, awards, and fees go to the Erasmus MC. P.W.C. had received research support from CSL Behring, and consultancy fees from CSL Behring, Novo Nordisk, Sobi, and Roche. J.E. received research support from CSL Behring and he has been a teacher on educational activities of Roche. S.J. has received financial support for research from Sanofi Genzyme and honoraria from Bayer, Takeda, Thrombosis Canada, Roche, Octapharma, and Hema-lytic. The institution of M.J.H.A.K. has received unrestricted research grants from Sobi, and her institution received speakers' fees from Sobi, Roche, and BMS. C.M. was a consultant or speaker for Bayer, Biotest, CSL Behring, Grifols, Novo Nordisk, Roche, and Takeda. He has received travel support from Bayer, Biotest, CSL Behring, and Novo Nordisk; his institution has received research support from Bayer, Takeda, Biotest, CSL Behring, Novo Nordisk, Roche, and Sobi. S.S. has received conference support, educational speaker fees, and/or advisory board fees from Pfizer, Bayer, Shire, Takeda, Sobi, Roche, and CSL Behring. K.F.'s institution has received unrestricted research grants from Sobi, Pfizer, CSL Behring, and Novo Nordisk, and also consultancy fees from Grifols, Takeda, Novo Nordisk, and Roche. M.C. reports research funding from

Bayer, CSL Behring, Daiichi Sankyo, Novo Nordisk, Portola/Alexion, Roche, Sanquin Blood Supply, and UniQure and consultancy or lecturing fees from Bayer, CSL Behring, Novo Nordisk, Pfizer, and Sobi. A.Z., F.R.K., P.B., C.H., R.C.C.H., C.A.M.K.T., B.L.G., and L.N. have no potential conflicts of interest.

REFERENCES

- [1] Blanchette VS, Key NS, Ljung LR, Manco-Johnson MJ, van den Berg HM, Srivastava A. Definitions in hemophilia: communication from the SSC of the ISTH. *J Thromb Haemost*. 2014;12:1935–9.
- [2] Kitchen S, Signer-Romero K, Key NS. Current laboratory practices in the diagnosis and management of haemophilia: a global assessment. *Haemophilia*. 2015;21:550–7.
- [3] Marlar RA, Strandberg K, Shima M, Adcock DM. Clinical utility and impact of the use of the chromogenic vs one-stage factor activity assays in haemophilia A and B. *Eur J Haematol*. 2020;104:3–14.
- [4] Kitchen S, Blakemore J, Friedman KD, Hart DP, Ko RH, Perry D, Platton S, Tan-Castillo D, Young G, Luddington RJ. A computer-based model to assess costs associated with the use of factor VIII and factor IX one-stage and chromogenic activity assays. *J Thromb Haemost*. 2016;14:757–64.
- [5] Peyvandi F, Oldenburg J, Friedman KD. A critical appraisal of one-stage and chromogenic assays of factor VIII activity. *J Thromb Haemost*. 2016;14:248–61.
- [6] Al-Huniti A, Sharathkumar A, Krantz M, Watkinson K, Bhagavathi S. Discrepant hemophilia A: an underdiagnosed disease entity. *Am J Clin Pathol*. 2020;154:78–87.
- [7] Pavlova A, Delev D, Pezeshkpoor B, Muller J, Oldenburg J. Haemophilia A mutations in patients with non-severe phenotype associated with a discrepancy between one-stage and chromogenic factor VIII activity assays. *Thromb Haemost*. 2014;111:851–61.
- [8] Bowyer AE, Van Veen JJ, Goodeve AC, Kitchen S, Makris M. Specific and global coagulation assays in the diagnosis of discrepant mild hemophilia A. *Haematologica*. 2013;98:1980–7.
- [9] Makris M, Oldenburg J, Mauser-Bunschoten EP, Peerlinck K, Castaman G, Fijnvandraat K. The definition, diagnosis and management of mild hemophilia A: communication from the SSC of the ISTH. *J Thromb Haemost*. 2018;16:2530–3.
- [10] Kihlberg K, Strandberg K, Rosen S, Ljung R, Astermark J. Discrepancies between the one-stage clotting assay and the chromogenic assay in haemophilia B. *Haemophilia*. 2017;23:620–7.
- [11] Cid AR, Calabuig M, Cortina V, Casana P, Haya S, Moret A, Cabrera N, Aznar JA. One-stage and chromogenic FVIII:C assay discrepancy in mild haemophilia A and the relationship with the mutation and bleeding phenotype. *Haemophilia*. 2008;14:1049–54.
- [12] Srivastava A, Santagostino E, Dougall A, Kitchen S, Sutherland M, Pipe SW, Carcao M, Mahlangu J, Ragni MV, Windyga J, Llinas A, Goddard NJ, Mohan R, Poonnoose PM, Feldman BM, Lewis SZ, van den Berg HM, Pierce GF, WFH Guidelines for the Management of Hemophilia panelists and co-authors. WFH guidelines for the management of hemophilia. *Haemophilia*. 2020;26:1–158, 3rd edition.
- [13] Schutte LM, Hodes LS, van Moort I, Stoof SCM, Leebeek FWG, Cnossen MH, de Maat MPM, Kruip MJHA. The one-stage assay or chromogenic assay to monitor baseline factor VIII levels and desmopressin effect in non-severe haemophilia A: superiority or non-inferiority? *Haemophilia*. 2020;26:916–22.
- [14] Bland JM, Altman DG. Statistical methods for assessing agreement between two methods of clinical measurement. *Lancet*. 1986;1:307–10.
- [15] Kloosterman FR, Zwagemaker A-F, Bagot CN, Beckers EAM, Castaman G, Cnossen MH, Collins PW, Hay C, Hof M, Laros-van Gorkom B, Leebeek FWG, Male C, Meijer K, Pabinger I, Shapiro S, Coppens M, Fijnvandraat K, Gouw SC. The bleeding phenotype in people with non-severe hemophilia. *Blood Adv*. 2022;6:4256–65.

- [16] Trudesson A, Schmidt D, Stralfors A, Soutari N, Norberg EM, Letelier A, Holmstrom M, Ranta S, Ljung R, Antovic J, Bruzelius M. One-stage versus chromogenic factor IX activity in haemophilia B [abstract]. *Res Pract Thromb Haemost*. 2020;4.
- [17] Duncan EM, Rodgers SE, McRae SJ. Diagnostic testing for mild hemophilia A in patients with discrepant one-stage, two-stage, and chromogenic factor VIII:C assays. *Semin Thromb Hemost*. 2013;39:272–82.
- [18] Stralfors A, Mikovic D, Schmidt D, Onelov L, Soutari NMH, Berndtson M, Chaireti R, Holmstrom M, Antovic JP, Bruzelius M. Genetics and hemostatic potential in persons with mild to moderate hemophilia A with a discrepancy between one-stage and chromogenic FVIII assays. *Thromb Haemost*. 2021;121:27–35.
- [19] Trossaert M, Boisseau P, Quemener A, Sigaud M, Fouassier M, Ternisien C, Lefrancois-Bettembourg A, Tesson C, Thomas C, Bezieau S. Prevalence, biological phenotype and genotype in moderate/mild hemophilia A with discrepancy between one-stage and chromogenic factor VIII activity. *J Thromb Haemost*. 2011;9:524–30.
- [20] Trossaert M, Lienhart A, Nougier C, Fretigny M, Sigaud M, Meunier S, Fouassier M, Ternisien C, Negrier C, Dargaud Y. Diagnosis and management challenges in patients with mild haemophilia A and discrepant FVIII measurements. *Haemophilia*. 2014;20:550–8.
- [21] Valikhani A, Mirakhorly M, Namvar A, Rastegarlar G, Toogeh G, Shirayeh FV, Ahmadijead M. Genetic analysis of non-severe hemophilia A phenotype with A discrepancy between one-stage and chromogenic factor VIII activity assays. *Transfus Apher Sci*. 2021;60:103194.
- [22] Suzuki A, Suzuki N, Kanematsu T, Okamoto S, Tamura S, Kikuchi R, Katsumi A, Kiyoi H, Kojima T, Matsushita T. Impact of variation in reagent combinations for one-stage clotting assay on assay discrepancy in nonsevere haemophilia A. *Int J Lab Hematol*. 2021;43:131–8.
- [23] Schwaab R, Oldenburg J, Kemball-Cook G, Albert T, Juhler C, Hanfland P, Ingerslev J. Assay discrepancy in mild haemophilia A due to a factor VIII missense mutation (Asn694Ile) in a large Danish family. *Br J Haematol*. 2000;109:523–8.
- [24] Keeling DM, Sukhu K, Kemball-Cook G, Waseem N, Bagnall R, Lloyd JV. Diagnostic importance of the two-stage factor VIII:C assay demonstrated by a case of mild haemophilia associated with His1954→Leu substitution in the factor VIII A3 domain. *Br J Haematol*. 1999;105:1123–6.
- [25] Hill M, Deam S, Gordon B, Dolan G. Mutation analysis in 51 patients with haemophilia A: report of 10 novel mutations and correlations between genotype and clinical phenotype. *Haemophilia*. 2005;11:133–41.
- [26] Lyall H, Hill M, Westby J, Grimley C, Dolan G. Tyr346→Cys mutation results in factor VIII:C assay discrepancy and a normal bleeding phenotype - is this mild haemophilia A? *Haemophilia*. 2008;14:78–80.
- [27] Mumford AD, Laffan M, O'Donnell J, McVey JH, Johnson DJ, Manning RA, Kemball-Cook G. A Tyr346→Cys substitution in the interdomain acidic region a1 of factor VIII in an individual with factor VIII:C assay discrepancy. *Br J Haematol*. 2002;118:589–94.
- [28] Favaloro EJ, Meijer P, Jennings I, Sioufi J, Bonar RA, Kitchen DP, Kershaw G, Lippi G. Problems and solutions in laboratory testing for hemophilia. *Semin Thromb Hemost*. 2013;39:816–33.
- [29] van Moort I, Meijer P, Priem-Visser D, van Gammeren AJ, Pequeriaux NCV, Leebeek FWG, Crossen MH, de Maat MPM. Analytical variation in factor VIII one-stage and chromogenic assays: experiences from the ECAT external quality assessment programme. *Haemophilia*. 2019;25:162–9.
- [30] Pouplard C, Trossaert M, Querrec ALE, Delahousse B, Giraudeau B, Gruel Y. Influence of source of phospholipids for APTT-based factor IX assays and potential consequences for the diagnosis of mild haemophilia B. *Haemophilia*. 2009;15:365–8.
- [31] Rodgers SE, Duncan EM, Barbulescu DM, Quinn DM, Lloyd JV. In vitro kinetics of factor VIII activity in patients with mild haemophilia A and a discrepancy between one-stage and two-stage factor VIII assay results. *Br J Haematol*. 2007;136:138–45.
- [32] Loomans JI, van Velzen AS, Eckhardt CL, Peters M, Makiperna A, Holmstrom M, Brons PP, Dors N, Haya S, Voorberg J, van der Bom JG, Fijnvandraat K. Variation in baseline factor VIII concentration in a retrospective cohort of mild/moderate hemophilia A patients carrying identical F8 mutations. *J Thromb Haemost*. 2017;15:246–54.
- [33] Adcock DM, Strandberg K, Shima M, Marlar RA. Advantages, disadvantages and optimization of one-stage and chromogenic factor activity assays in haemophilia A and B. *Int J Lab Hematol*. 2018;40:621–9.
- [34] Kitchen S, Hayward C, Negrier C, Dargaud Y. New developments in laboratory diagnosis and monitoring. *Haemophilia*. 2010;16:61–6.
- [35] Potgieter JJ, Damgaard M, Hillarp A. One-stage vs. chromogenic assays in haemophilia A. *Eur J Haematol*. 2015;94:38–44.
- [36] Bowyer AE, Duncan EM, Antovic JP. Role of chromogenic assays in haemophilia A and B diagnosis. *Haemophilia*. 2018;24:578–83.
- [37] Bowyer AE, Goodeve A, Liesner R, Mumford AD, Kitchen S, Makris M. p.Tyr365Cys change in factor VIII: haemophilia A, but not as we know it. *Br J Haematol*. 2011;154:618–25.

SUPPLEMENTARY MATERIAL

All members of the DYNAMO study group are listed in the supplementary material. The online version contains supplementary material available at <https://doi.org/10.1016/j.jth.2022.11.040>