



Universiteit
Leiden

The Netherlands

Touched by technology: automated tactile stimulation in the treatment of apnoea of prematurity

Cramer, S.J.E.

Citation


Cramer, S. J. E. (2025, September 30). *Touched by technology: automated tactile stimulation in the treatment of apnoea of prematurity*. Retrieved from <https://hdl.handle.net/1887/4262038>

Version: Publisher's Version

License: [Licence agreement concerning inclusion of doctoral thesis in the Institutional Repository of the University of Leiden](#)

Downloaded from: <https://hdl.handle.net/1887/4262038>

Note: To cite this publication please use the final published version (if applicable).



PART

5



CHAPTER 7

Technology in the delivery room
supporting the neonatal healthcare
provider's task

**SJE Cramer, KLAM Kuypers, T Martherus,
J Dekker & AB te Pas**

Seminars in Fetal & Neonatal Medicine 2022; 27(5):101333

ABSTRACT

Very preterm infants are a unique and highly vulnerable group of patients that have a narrow physiological margin within which interventions are safe and effective. The increased understanding of the foetal to neonatal transition marks the intricacy of the rapid and major physiological changes that take place, making delivery room stabilisation and resuscitation an increasingly complex and sophisticated activity for caregivers to perform. While modern, automated technologies are progressively implemented in the neonatal intensive care unit (NICU) to enhance the caregivers in providing the right care for these patients, the technology in the delivery room still lags far behind. Diligent translation of well-known and promising technological solutions from the NICU to the delivery room will allow for better support of the caregivers in performing their tasks. In this review we will discuss the current technology used for stabilisation of preterm infants in the delivery room and how this could be optimised in order to further improve care and outcomes of preterm infants in the near future.

INTRODUCTION

Because of their immaturity, very preterm infants are a unique and highly vulnerable group of patients that have a narrow physiological margin within which interventions are safe and effective. A large amount of scientific research along with technological innovations have improved care for preterm infants in the neonatal intensive care unit (NICU). In the NICU, infants are monitored meticulously and devices are used to provide treatment in a safe and sophisticated manner. The arrival of automated technologies, whether or not in combination with artificial intelligence (servo-controlled incubator, algorithm driven ventilators, automated oxygen titration, predictive monitoring) has decreased the manual work in the NICU and improved care and outcome in the last twenty years.

While these automated technologies are increasingly being used in the NICU, technology in the delivery room still lags far behind. Preterm infants can be difficult to manage in the intensive care unit, but this task is considerably more complex at birth due to the infant's rapidly changing physiology. Particularly at birth, automated technologies and/or artificial intelligence could be highly relevant, since the infant's physiology is undergoing large and rapid changes. We now start to understand that the transition to life after birth is an extremely critical phase of life which greatly impacts an individual's risk of death, injury [1, 2] or life-long disability [3, 4], particularly infants born very preterm.

Stabilisation of preterm infants in the delivery room is usually brief, but many interventions need to be performed in order to stabilise the infant's temperature, (spontaneous) ventilation and oxygenation in a time sensitive manner. It has been shown that some interventions are not as effective as caregivers assumed and that the provision of an optimal and safe treatment during this stressful moment is a major challenge for caregivers. [5-8] It also has been demonstrated that caregivers have difficulty in assimilating the complex and rapidly changing physiological information that is required to make accurate strategic decisions with regard to assisting preterm infants as they transition to newborn life [6-10].

In this review we will discuss the current technology used for stabilisation of preterm infants in the delivery room and how this could be optimised by the provision of purpose-built devices and technology that assimilates all of the physiology data and supports decision making processes.



TEMPERATURE MANAGEMENT

The first step of neonatal stabilisation is the prevention of heat loss, which easily occurs in the exposed and wet infants through convection, conduction, radiation and evaporation, resulting in a decreased body temperature. Hypothermia after birth has been recognised as a significant contributor to neonatal morbidity and mortality. [11, 12] Although less is known about the acute and long-term impact of hyperthermia after birth, the potential risks for both hypothermia and hyperthermia are currently recognised in the international resuscitation guidelines with the advice to keep the body temperature of the infant between 36.5 and 37.5°C. [13, 14] While measures to prevent hypothermia – such as increased room temperature and the use of a head cap, a wrap, a radiant heater, a thermal mattress and heated and humidified gases - are commonly performed, keeping the body temperature within the normal range during stabilisation at birth proves to be challenging. [15-18]

Currently, the temperature is often only measured at NICU admission, which does not allow us to take correcting measures until that moment. Although standardised thermoregulation protocols, training, and audits have shown to improve our temperature management [19-21], frequent or continuous measurement of temperature, in combination with a temperature dependent protocol, can further improve this. [22, 23] However, temperature management based on continuous measurements requires constant attention and is more labour intensive. Technology could assist in this process by providing visual or audible cues when the recommended ranges are exceeded in order to capture the attention of the caregiver or by providing decision support on the timing and type of heat loss measures to take. Technology could even further assist caregivers by enabling automated regulation via servo-controlled mattresses and radiant warmers, which are commonly used in the NICU.

Although a recent multi-centre study reported that the use of servo-controlled radiant warmers on the delivery room showed no benefits over the use of radiant warmers on maximal output [24], studies implementing servo-controlled radiant warmers combined with a temperature dependent protocol for additional measures show the highest overall scores of normothermia at NICU admission, ranging from 74% to 100%. [25-28] As infants are much more exposed in the delivery room as compared to the NICU, full automation of thermoregulation in the delivery room probably asks for completely different closed-loop solutions minimising the effects of the environment on their temperature.

TACTILE STIMULATION

As of 2005, local and international resuscitation guidelines recommend tactile stimulation in the form of warming, drying and rubbing the back or soles of the feet to evoke spontaneous breathing in newborn infants. [13, 14, 29] While experimental studies demonstrated tactile stimulation to increase respiratory effort [30, 31], the clinical guidelines are still largely based on many years of experience and expert opinion as there is lack of data on this topic in human infants.

Several retrospective studies recently evaluated current practice, showing a wide variation between caregivers and between centres concerning timing, duration and method of stimulation. [32-36] In addition, stimulation turned out to be often omitted, in particular in preterm infants placed in a polyethylene bag. [34-36] A recent randomised trial showed that repetitive tactile stimulation in preterm infants increased oxygenation, while less oxygen was needed, and improved respiratory effort. [37] However, the trial also led to a high incidence of stimulation in the standard group. This effect could be attributed to the Hawthorne effect and/or the increased focus on tactile stimulation during the study, which in turn implies that omission of stimulation happens because it is simply forgotten.

Albeit the most optimal way of stimulation remains unclear, automated mechanical stimulation could ensure tactile stimulation to be provided, in a more consistent way. [38] Several closed-loop vibratory stimulation devices to treat apnoea's of preterm infants admitted to the NICU have been described in literature, but currently none of these are commercially available. [39-42] No studies have been performed in the delivery room, but mechanical vibratory stimulation in preterm infants in the NICU proved to be as effective as manual stimulation in aborting apnoeic episodes in two preliminary studies, [42, 43] and two other observational studies reported that their closed-loop pulsating and vibrating devices were able to terminate 90% of all apnoea's. [44, 45] Applying this technique in the delivery room has the potential to replace manual intervention, eliminating the chance that stimulation will be forgotten.

OXYGENATION

Currently, oxygen administration is guided by predefined oxygen saturation (SpO₂) target ranges. [46] Caregivers manually titrate the fraction of inspired oxygen (FiO₂) accordingly to avoid hypoxia and hyperoxia. At birth, hypoxia can lead to suppression of spontaneous breathing, and hypoxia that persists for more than 5 minutes after birth is associated with an increased risk of mortality and the development of intraventricular haemorrhages. [47-50] On the other hand, hyperoxia needs to be avoided as this increases the production of

free radicals, but also can inhibit the respiratory centre. [51, 52] As such, it is critical to adequately control oxygenation during this period. However, this is an incredibly difficult and complex task given the fact that immediately after birth, the lung is constantly and rapidly changing.

It has been shown that maintaining SpO₂ values within a predefined target range with manual titration is extremely difficult in the delivery room as well as in the NICU. [53-59] Randomised trials demonstrated the potential of closed-loop titration of FiO₂ in the NICU, increasing the time spent within the SpO₂ target range with a decrease in extreme deviations in oxygenation, including both the duration and the number of episodes. [60-69] The use of a closed-loop oxygen controller in the delivery room has so far only been studied once in a preterm lamb model. [70] In this study, the effect of a closed-loop oxygen controller with timeout restrictions of 30s after each titration step was compared to manual titration of oxygen after evaluation of SpO₂ to be performed every 30s. Results show similar time within the SpO₂ target range and below the target range, while time above the target range was significantly shorter in the automatic titration group. [70]

However, this technique cannot just simply be extrapolated to the delivery room, as there are considerable differences with regard to target ranges, physiology and devices used. In the NICU, the SpO₂ target range is static, while this is dynamic in the first minutes after birth. Oxygen exchange in the lungs is largely determined by the surface area available for gas exchange and the oxygen concentration gradient between the alveoli and adjacent capillaries. At birth, when the airways are mostly liquid-filled, the surface area available for gas exchange is small and a high oxygen concentration is required for adequate exchange (Figure 1A). As the lungs aerate, the surface area available for gas exchange increases exponentially and as such a much lower oxygen concentration is needed for adequate oxygenation (Figure 1B). The oxygen concentration administered after birth should thus be adjusted according to the degree of lung aeration. This would require the closed-loop titration mechanism to adjust the SpO₂ target range continuously based on the time after birth. In addition, the algorithm of the closed-loop oxygen controller should also be calibrated based on the factors present at birth which influence the position of the oxygen-haemoglobin dissociation curve.

Furthermore, titration of oxygen using a T-piece ventilator, which is commonly used for respiratory support at birth [14], can result in a delay between the moment of titration and the delivery of the corresponding FiO₂ at the face mask of the infant. [71] The algorithm used by the closed-loop oxygen controller that is used with the T-piece resuscitator should therefore reckon with this delay.




Figure 1. Overview of the alveoli, surrounded by capillaries. A: Directly at birth, lung liquid needs to be replaced with air. The movement of liquid into the interstitial tissue causes a high airway resistance and the partially liquid-filled alveoli reduces the surface area available for gas exchange. B: As the liquid moves into the interstitial tissue surrounding the alveoli, the airway resistance decrease while the interstitial pressure and lung recoil increase. This causes alveolar collapse and liquid re-entry at end-expiration. Nevertheless, the surface area available for gas exchange increases.

CONTINUOUS POSITIVE AIRWAY PRESSURE

Although most preterm infants breathe at birth, the breathing effort is often insufficient to ensure the large pulmonary physiological changes that are needed to survive the foetal to neonatal transition. While continuous positive airway pressure (CPAP) is often used to support the infant's breathing, there is no data on the optimal pressure level. The CPAP level of 4-8 cmH₂O that is currently used is predominantly extrapolated from data from CPAP later in the NICU, while the underlying physiology during the neonatal transition is strikingly different. [13, 72]

Considering the physiological changes that need to occur during transition, it would be more logical to use a dynamic CPAP strategy wherein the CPAP levels suit the different phases of the transition. In the first phase of the transition (Figure 1A), the role of CPAP is to promote lung aeration and assist movement of lung liquid across the distal airway wall into the interstitial tissue. As a result, the resistance in the airways is high due to the high viscosity of liquid (compared with air) moving across the airway epithelium requiring higher CPAP levels to overcome this. [73-77] Once the lungs become more aerated and liquid is accumulated in the interstitial tissue, the lung characteristics change quickly and the role of CPAP converts to maintaining lung aeration. During this phase of the transition (Figure

1B), airway resistance is considerably lower (~100 fold), but lung recoil and interstitial tissue pressure increase which promote alveolar collapse and liquid re-entry at end-expiration. [73-81] Lower CPAP levels are likely sufficient to maintain aeration and support breathing, while decreasing the risk for lung overexpansion and/or adverse effect on pulmonary blood flow. This dynamic CPAP approach, following the pulmonary physiological changes during transition, has been called physiological based (PB)-CPAP.

Experimental studies in spontaneously breathing preterm animal models demonstrated that PB-CPAP should ideally start with CPAP of 15 cmH₂O which is stepwise decreased to 8 cmH₂O. These studies also showed that PB-CPAP promotes lung aeration (functional residual capacity; FRC), breathing effort and pulmonary blood flow, without causing bulging of the lungs or pneumothoraxes. [82, 83] These results were translated into a small randomised controlled trial wherein PB-CPAP was compared to 5-8 cmH₂O CPAP. This feasibility study demonstrated that PB-CPAP led to a quicker restoration of heart rate and shorter duration of mask ventilation, likely reflecting lung aeration. Nevertheless, post-trial evaluations indicated that caregivers found it difficult to combine standard care with a CPAP protocol that requires constant evaluations and changes in CPAP levels. [84]

This is where technological innovation could help the caregiver in assimilating complex physiological changes and fine-tuning and optimising the respiratory support. Mathematical modelling with currently available physiological data could be used to create algorithms, which will allow us to develop a decision or even automated pressure support system in the delivery room.

POSITIVE PRESSURE VENTILATION

If preterm infants fail to clear their lung liquid, establish FRC and initiate spontaneous breathing to facilitate gas exchange [85], manual non-invasive positive pressure ventilation (NIPPV) is provided by occluding the aperture of a T-piece resuscitator with a thumb or finger. The sufficiency of the provided tidal volumes is confirmed by adequate chest rise, auscultation or, indirectly, by an increase in heart rate. [86] However, due to rapidly changing pulmonary physiology and inconsistent respiratory drive of infants at birth, variable tidal volume are administered that might be inadequate or excessive. Large tidal volumes could overstretch the delicate alveoli and airways (volutrauma), while small tidal volumes could lead to loss of lung volume or cycling between collapse and recruitment (atelectotrauma) thereby injuring the lungs. [87, 88] A recent multicentre trial evaluating tidal volume monitoring during manual ventilation reported that, despite using a respiratory function monitor (RFM)(Figure 2), ineffective ventilation <4 mL/kg and potentially harmful ventilation >8 mL/kg was provided 40.7% and 20.0% of the time, respectively. [89]

The high percentage of ineffective manual ventilation could be caused by pharyngeal ventilation as the glottis is predominantly closed after birth and only opens when a spontaneous breath is taken. [90] When ventilation is provided to a closed glottis, no air is able to enter into the lungs. [90] Providing inflations which coincides with spontaneous breaths would be more effective, but also increases the risks of high tidal volumes and thus the risk of lung and/or cerebral injury. [91] As it is difficult for caregivers to evaluate the presence and quality of spontaneous breathing at birth [8], especially during manual ventilation [7], this hampers safe and effective ventilation at birth.

Again, automation can offer a solution. In this case, several solutions already exist and are being applied as features of a neonatal ventilator. Replacing the T-piece resuscitator for a regular neonatal ventilator in the delivery room therefore brings several opportunities to prevent inappropriate ventilation. The first solution is automated synchronised NIPPV (sNIPPV). Caregivers can only detect breathing after a breath has been taken, while a ventilator can detect the start of a breath. This enables ventilators to synchronise their ventilation. In addition, caregivers have to keep overview of the clinical condition of the infant and are, therefore, not able to continuously focus on the infant's breathing while a ventilator can. Although there is no evidence for the effectiveness of synchronised ventilation in the delivery room, it has shortened the duration and improved the effectiveness of ventilation in the NICU. [92, 93]




Figure 2. Respiratory Function Monitor Display

Even with sNIPPV, it remains difficult to provide tidal volumes in a safe target range. To date, caregivers are only able to apply pressure-limited manual ventilation in the delivery room due to the lack of appropriate technology. However, a recent neonatal resuscitation simulation study showed it is feasible to use a ventilator with RFM in the delivery room as it increased the proportion of tidal volumes within the target range and reduced the number of large tidal volumes during different simulated scenarios of changing pulmonary mechanics commonly encountered at birth. [86] Also, the delivery of consistent tidal volumes during changing pulmonary mechanics could be improved by implementing volume-targeted ventilation. While, this ventilation mode showed to improve outcome and is a widely accepted in the NICU, there is no data on using this mode in preterm infants at birth. [87, 94, 95] This effect might even be increased when using a ventilator with synchronised ventilation and/or volume-targeted ventilation.

MONITORING

Regular feedback on the patient's physiological state is a pivotal element of neonatal stabilisation after birth, guiding corrective actions and clinical decision making of the caregivers. Despite its importance, monitoring in this critical period is still relatively basic compared to the continuous and extensive monitoring techniques used in the NICU.

The current guidelines recommend the use of pulse oximetry and/or ECG for physiological feedback instead of rudimentary methods such as auscultation, palpation of the umbilical cord and assessment of skin colour, as these methods proved to be prone to subjectivity [96, 97]. The same applies to the assessment of administered tidal volumes by observing chest excursions [98] but the evidence for using a RFM instead remains conflicting. Although manikin studies demonstrated that providing continuous feedback on ventilation pressures, tidal volumes, mask leak, SpO₂, heart rate and FiO₂ via a RFM improved the performance of the caregiver during PPV [86, 99-101], a recent multicentre randomized controlled trial showed no difference between neonatal resuscitation with or without integrated feedback by RFM. [89] This result might be explained by previous findings that the use and interpretation of a RFM in the delivery room is experienced as challenging and therefore not helpful to all caregivers in critical decision making. [7], [102]

Although continuous, objective and accurate data acquisition is necessary to further implement modern technological innovations such as closed-loop interventions and prediction models, the question is whether presenting all this data directly to the caregivers is always useful. Future research should also be focused on which data to present, and in particular in what manner, to facilitate quick assimilation and easy interpretation by caregivers so that they can recognize and act upon abnormalities or changes in physiology.

In other words, in the design or development of monitoring methods, one should consider carefully whether the data is processed by algorithms or a human brain.

MAN AND MACHINES

Our understanding of the foetal to neonatal transition and the underlying physiological changes has evidently increased in the recent decades, facilitating clear opportunities aiming to improve clinical outcome. However, these insights also underscore the eminent complexity of the transition process, especially in ill or preterm infants who cannot meet the required physiological challenges on their own.

Whilst the expansion of monitoring solutions and intervention strategies and the finetuning of protocols and target ranges can definitely aid caregivers in providing the right support, it makes the resuscitation process increasingly sophisticated. As of today, caregivers continuously have to assimilate and interpret many physiological parameters from different devices in order to decide if, when and which intervention is required, in just a small-time window. The more difficult, dynamic and versatile the process, the more prone it becomes to human errors such as forgetfulness and lack of continued focus.

Over the last decades the development and adoption of automated medical technology has tremendously increased and accordingly revolutionised medical practice, but not yet in the delivery room. We argue that the development and implementation of automation, closed-loop systems and artificial intelligence could serve as a next iteration in improving resuscitation management by reducing human error and unwanted variability in human behaviour. However, this can only be achieved if we critically validate the added value using a holistic approach; not only taking into account the patient but also the caregivers. This means that we should not blindly use existing solutions for new problems but find new ones fitting the entire context. We should not use or implement innovations because it is technically feasible, but because it is desirable and we should not endlessly extend and expand existing solutions but come up with solutions that replace a bundle of existing ones.

Although some might dream, and others fear, a completely automated transition support system, it is more likely that technology will take on an integral part of resuscitation management, resulting in an increased caregiver-machine interaction. Given the growing complexity of automated systems, the poor explainability of artificial intelligence and the consequences of possible erroneous automated interventions, a paradigm shift is necessary. Caregivers should not only be clinically aware, understanding the status of the patient with regard to the interlinked physiological changes, but be situational aware, also understanding the status of all automated devices, systems and software during the transition process.

Shaping this role is however not the sole responsibility of the caregivers. To make the most out of it, designers and developers should indeed focus on the explainability and interpretability of automated systems and error prone interfaces including clear user feedback. Managers and medical engineers should moreover ensure that caregivers are trained like pilots; focussing on the capabilities to identify and respond to system errors or failure. As it is utopian to think that capitalising some strength of computers will fully replace human weaknesses, caregivers have to accept that improvement of care will always remain an iterative process.

CONCLUSION

Although the complexity of stabilisation after birth increases by our growing understanding of the complex physiology, the development and implementation of technology to assist in this process lags behind. Implementing state-of-the-art technology during the neonatal stabilisation would enable us to i) prevent hypo- and hyperthermia through closed-loop temperature management, ii) stimulate spontaneous breathing by providing automatic repetitive tactile stimulation to all infants , iii) control oxygenation in relation to neonatal transition through closed-loop oxygenation, iv) support spontaneous breathing during




Figure 3. Visualisation of evolving technologies for use in the delivery room. From providing manual heat loss measures, manual tactile stimulation, manual supplemental oxygen control and manual CPAP and NIPPV with a T-piece resuscitator (A) to automated thermoregulation, automated tactile stimulation, automated oxygen control and automated ventilation using a ventilator (B).

neonatal transition by automated PB-CPAP algorithms and v) provide safe and effective ventilation by using synchronised volume targeted ventilation (Figure 3). By using technology to assist caregivers to provide the optimal care, caregivers would be able to comprehend an overview of the infant's clinical condition more easily and finetune the stabilisation where appropriate.

Although most of the technology discussed in this review is already used in the NICU, it cannot simply be extrapolated to the delivery room because of the difference in physiology, environment and situation. The adoption of automation has great potential to improve the care we provide in the delivery room, as long as we put humans, not technology, first. Above all, we must realise that technology does not make man superfluous: the clinical view remains necessary.



REFERENCES

1. Sobotka, K.S., et al., Circulatory responses to asphyxia differ if the asphyxia occurs in utero or ex utero in near-term lambs. *PLoS One*, 2014. 9(11): p. e112264.
2. Polglase, G.R., et al., Respiratory support for premature neonates in the delivery room: effects on cardiovascular function and the development of brain injury. *Pediatr. Res*, 2014. 75(6): p. 682-688.
3. Davidson, L.M. and S.K. Berkelhamer, Bronchopulmonary Dysplasia: Chronic Lung Disease of Infancy and Long-Term Pulmonary Outcomes. *J Clin Med*, 2017. 6(1).
4. O'Shea, T.M., Cerebral palsy in very preterm infants: new epidemiological insights. *Ment Retard Dev Disabil Res Rev*, 2002. 8(3): p. 135-45.
5. van Vonderen, J.J., et al., Effects of a sustained inflation in preterm infants at birth. *J Pediatr*, 2014. 165(5): p. 903-8 e1.
6. Huberts, T.J.P., et al., The Breathing Effort of Very Preterm Infants at Birth. *J Pediatr*, 2018. 194: p. 54-59.
7. Schilleman, K., et al., Evaluating manual inflations and breathing during mask ventilation in preterm infants at birth. *J Pediatr*, 2013. 162(3): p. 457-63.
8. Schilleman, K., et al., Auditing resuscitation of preterm infants at birth by recording video and physiological parameters. *Resuscitation*, 2012. 83(9): p. 1135-9.
9. van Vonderen, J.J., et al., Cardiorespiratory Monitoring during Neonatal Resuscitation for Direct Feedback and Audit. *Front Pediatr*, 2016. 4: p. 38.
10. Root, L., et al., Improving Guideline Compliance and Documentation Through Auditing Neonatal Resuscitation. *Front Pediatr*, 2019. 7: p. 294.
11. Chitty, H. and J. Wyllie, Importance of maintaining the newly born temperature in the normal range from delivery to admission. *Semin Fetal Neonatal Med*, 2013. 18(6): p. 362-8.
12. Laptook, A.R., et al., Admission temperature of low birth weight infants: predictors and associated morbidities. *Pediatrics*, 2007. 119(3): p. e643-9.
13. Wyllie, J., et al., European Resuscitation Council Guidelines for Resuscitation 2015: Section 7. Resuscitation and support of transition of babies at birth. *Resuscitation*, 2015. 95: p. 249-63.

14. Madar, J., et al., European Resuscitation Council Guidelines 2021: Newborn resuscitation and support of transition of infants at birth. *Resuscitation*, 2021. 161: p. 291-326.
15. Wilson, E., et al., Admission Hypothermia in Very Preterm Infants and Neonatal Mortality and Morbidity. *J Pediatr*, 2016. 175: p. 61-67 e4.
16. Mank, A., et al., Hypothermia in Preterm Infants in the First Hours after Birth: Occurrence, Course and Risk Factors. *PLoS One*, 2016. 11(11): p. e0164817.
17. Lyu, Y., et al., Association between admission temperature and mortality and major morbidity in preterm infants born at fewer than 33 weeks' gestation. *JAMA Pediatr*, 2015. 169(4): p. e150277.
18. de Almeida, M.F., et al., Hypothermia and early neonatal mortality in preterm infants. *J Pediatr*, 2014. 164(2): p. 271-5 e1.
19. Choi, H.S., et al., The impact of a quality improvement effort in reducing admission hypothermia in preterm infants following delivery. *Korean J Pediatr*, 2018. 61(8): p. 239-244.
20. Yip, W.Y., et al., A quality improvement project to reduce hypothermia in preterm infants on admission to the neonatal intensive care unit. *Int J Qual Health Care*, 2017. 29(7): p. 922-928.
21. Billimoria, Z., et al., Improving admission temperature in extremely low birth weight infants: a hospital-based multi-intervention quality improvement project. *J Perinat Med*, 2013. 41(4): p. 455-60.
22. Russo, A., et al., Reducing hypothermia in preterm infants following delivery. *Pediatrics*, 2014. 133(4): p. e1055-62.
23. Harer, M.W., et al., Implementation of a multidisciplinary guideline improves preterm infant admission temperatures. *J Perinatol*, 2017. 37(11): p. 1242-1247.
24. Cavallin, F., et al., Thermal management with and without servo-controlled system in preterm infants immediately after birth: a multicentre, randomised controlled study. *Arch Dis Child Fetal Neonatal Ed*, 2021. 106(6): p. 572-577.
25. Young, A., et al., A multimodal quality improvement approach to promote normothermia in very preterm infants. *Acta Paediatr*, 2021. 110(10): p. 2745-2752.
26. Pinheiro, J.M., et al., Decreasing hypothermia during delivery room stabilization of preterm neonates. *Pediatrics*, 2014. 133(1): p. e218-26.
27. Manani, M., et al., Elimination of admission hypothermia in preterm very low-birth-weight infants by standardization of delivery room management. *Perm J*, 2013. 17(3): p. 8-13.

28. DeMauro, S.B., et al., Improving delivery room management for very preterm infants. *Pediatrics*, 2013. 132(4): p. e1018-25.
29. Wyckoff, M.H., et al., Part 13: Neonatal Resuscitation: 2015 American Heart Association Guidelines Update for Cardiopulmonary Resuscitation and Emergency Cardiovascular Care. *Circulation*, 2015. 132(18 Suppl 2): p. S543-60.
30. Scarpelli, E., S. Condorelli, and E. Cosmi, Cutaneous stimulation and generation of breathing in the fetus. *Pediat Res*, 1977. 11: p. 24-28.
31. Ronca, A.E. and J.R. Alberts, Cutaneous induction of breathing in perinatal rats. *Psychobiology*, 1995. 23(4): p. 261-269.
32. Dekker, J., et al., Tactile Stimulation to Stimulate Spontaneous Breathing during Stabilization of Preterm Infants at Birth: A Retrospective Analysis. *Front Pediatr*, 2017. 5: p. 61.
33. Pietravalle, A., et al., Neonatal tactile stimulation at birth in a low-resource setting. *BMC Pediatrics*, 2018. 18(1): p. 306.
34. van Henten, T.M.A., et al., Tactile stimulation in the delivery room: do we practice what we preach? *Arch Dis Child Fetal Neonatal Ed*, 2019. 104(6): p. F661-F662.
35. Gaertner, V.D., et al., Physical stimulation of newborn infants in the delivery room. *Arch Dis Child Fetal Neonatal Ed*, 2018. 103(2): p. F132-F136.
36. Baik-Schneditz, N., et al., Tactile stimulation during neonatal transition and its effect on vital parameters in neonates during neonatal transition. *Acta Paediatr*, 2018. 107(6): p. 952-957.
37. Dekker, J., et al., Repetitive versus standard tactile stimulation of preterm infants at birth - A randomized controlled trial. *Resuscitation*, 2018. 127: p. 37-43.
38. Cramer, S.J.E., et al., Effect of Tactile Stimulation on Termination and Prevention of Apnea of Prematurity: A Systematic Review. *Front Pediatr*, 2018. 6: p. 45.
39. Marcotte, A.L., et al., Development of Apnea Interruption System by Vibratory Stimulus. *Proceedings of the IEEE 22nd Annual Northeast Bioengineering Conference*, 1996: p. 28-29.
40. Faille, E.O., A. Setya, and L. Eisenfeld, A Computerized System to Diagnose and Treat Neonatal Apnea Using Vibrotactile Stimulation. *Connecticut Medicine*, 2013. 77(9): p. 517-22.
41. Marayong, P. and M.S. Mostoufi, Foot Vibrotactile Device for Central Apnea Interruption in Premature Infants. *Medicine Meets Virtual Reality*, 2009. 17: p. 180-182.

42. Pichardo, R., et al., Validation of a vibrotactile stimulation system. Proceedings of the IEEE 27nd Annual Northeast Bioengineering Conference, 2001: p. 13-14.
43. Lovell, J.R., et al., Vibrotactile stimulation for treatment of neonatal apnea: a preliminary study. Connecticut Medicine, 1999. 63(6): p. 323-325.
44. Frank UA, et al., Treatment of apnea in neonates with an automated monitor-actuated apnea arrestor. Pediatrics, 1973. 51(5): p. 878-83.
45. Camargo, V.C., et al., Instrumentation for the detection and interruption of apnea. Conf Proc IEEE Eng Med Biol Soc, 2014: p. 2127-2130.
46. Dawson, J.A., et al., Defining the reference range for oxygen saturation for infants after birth. Pediatrics, 2010. 125(6): p. e1340-7.
47. Dekker, J., et al., Stimulating and maintaining spontaneous breathing during transition of preterm infants. Pediatr Res, 2019.
48. Davey, M.G., et al., Prematurity alters hypoxic and hypercapnic ventilatory responses in developing lambs. Respir Physiol, 1996. 105(1-2): p. 57-67.
49. Dekker, J., et al., Increasing Respiratory Effort With 100% Oxygen During Resuscitation of Preterm Rabbits at Birth. Front Pediatr, 2019. 7: p. 427.
50. Oei, J.L., et al., Outcomes of oxygen saturation targeting during delivery room stabilisation of preterm infants. Arch Dis Child Fetal Neonatal Ed, 2018. 103(5): p. F446-f454.
51. Clyman, R.I., O.D. Saugstad, and F. Mauray, Reactive oxygen metabolites relax the lamb ductus arteriosus by stimulating prostaglandin production. Circ Res, 1989. 64(1): p. 1-8.
52. Saugstad, O.D., Oxidative stress in the newborn--a 30-year perspective. Biol Neonate, 2005. 88(3): p. 228-36.
53. White, L.N., et al., Achievement of saturation targets in preterm infants <32 weeks' gestational age in the delivery room. Arch Dis Child Fetal Neonatal Ed, 2017. 102(5): p. F423-f427.
54. Goos, T.G., et al., Observing the resuscitation of very preterm infants: are we able to follow the oxygen saturation targets? Resuscitation, 2013. 84(8): p. 1108-13.
55. Phillipos, E., et al., Oxygen Saturation and Heart Rate Ranges in Very Preterm Infants Requiring Respiratory Support at Birth. J Pediatr, 2017. 182: p. 41-46. e2.
56. Hagadorn, J.I., et al., Achieved versus intended pulse oximeter saturation in infants born less than 28 weeks' gestation: the AVIOx study. Pediatrics, 2006. 118(4): p. 1574-82.

57. Lupton, A.R., et al., Pulse oximetry in very low birth weight infants: can oxygen saturation be maintained in the desired range? *J Perinatol*, 2006. 26(6): p. 337-41.
58. Lim, K., et al., Oxygen saturation targeting in preterm infants receiving continuous positive airway pressure. *J Pediatr*, 2014. 164(4): p. 730-736.e1.
59. Schmidt, B., et al., Effects of targeting higher vs lower arterial oxygen saturations on death or disability in extremely preterm infants: a randomized clinical trial. *Jama*, 2013. 309(20): p. 2111-20.
60. Claure, N., C. D'Ugard, and E. Bancalari, Automated adjustment of inspired oxygen in preterm infants with frequent fluctuations in oxygenation: a pilot clinical trial. *J Pediatr*, 2009. 155(5): p. 640-5.e1-2.
61. Plattner, G.K., et al., Clinical evaluation of a novel adaptive algorithm for automated control of oxygen therapy in preterm infants on non-invasive respiratory support. *Arch Dis Child Fetal Neonatal Ed*, 2017. 102(1): p. F37-F43.
62. Lal, M., W. Tin, and S. Sinha, Automated control of inspired oxygen in ventilated preterm infants: crossover physiological study. *Acta Paediatr*, 2015. 104(11): p. 1084-9.
63. van Kaam, A.H., et al., Automated versus Manual Oxygen Control with Different Saturation Targets and Modes of Respiratory Support in Preterm Infants. *J Pediatr*, 2015. 167(3): p. 545-50.e1-2.
64. Urschitz, M.S., et al., Automatic control of the inspired oxygen fraction in preterm infants: a randomized crossover trial. *Am J Respir Crit Care Med*, 2004. 170(10): p. 1095-100.
65. Hallenberger, A., et al., Closed-loop automatic oxygen control (CLAC) in preterm infants: a randomized controlled trial. *Pediatrics*, 2014. 133(2): p. e379-85.
66. Waitz, M., et al., Effects of automated adjustment of the inspired oxygen on fluctuations of arterial and regional cerebral tissue oxygenation in preterm infants with frequent desaturations. *J Pediatr*, 2015. 166(2): p. 240-4.e1.
67. Claure, N., et al., Multicenter crossover study of automated control of inspired oxygen in ventilated preterm infants. *Pediatrics*, 2011. 127(1): p. e76-83.
68. Zapata, J., et al., A randomised controlled trial of an automated oxygen delivery algorithm for preterm neonates receiving supplemental oxygen without mechanical ventilation. *Acta Paediatr*, 2014. 103(9): p. 928-33.

69. Van Zanten, H.A., et al., The effect of implementing an automated oxygen control on oxygen saturation in preterm infants. *Arch Dis Child Fetal Neonatal Ed*, 2017. 102(5): p. F395-F399.
70. Hütten, M.C., et al., Fully automated predictive intelligent control of oxygenation (PRICO) in resuscitation and ventilation of preterm lambs. *Pediatr Res*, 2015. 78(6): p. 657-63.
71. Dekker, J., et al., Time to achieve desired fraction of inspired oxygen using a T-piece ventilator during resuscitation of preterm infants at birth. *Resuscitation*, 2019. 136: p. 100-104.
72. Wyckoff, M.H., et al., Neonatal Life Support 2020 International Consensus on Cardiopulmonary Resuscitation and Emergency Cardiovascular Care Science With Treatment Recommendations. *Resuscitation*, 2020. 156: p. A156-A187.
73. Hooper, S.B., et al., Imaging lung aeration and lung liquid clearance at birth. *FASEB J*, 2007. 21(12): p. 3329-37.
74. Siew, M.L., et al., Inspiration regulates the rate and temporal pattern of lung liquid clearance and lung aeration at birth. *J Appl Physiol (1985)*, 2009. 106(6): p. 1888-95.
75. Siew, M.L., et al., The role of lung inflation and sodium transport in airway liquid clearance during lung aeration in newborn rabbits. *Pediatr Res*, 2013. 73(4 Pt 1): p. 443-9.
76. te Pas, A.B., et al., Effect of sustained inflation length on establishing functional residual capacity at birth in ventilated premature rabbits. *Pediatr Res*, 2009. 66(3): p. 295-300.
77. te Pas, A.B., et al., Establishing functional residual capacity at birth: the effect of sustained inflation and positive end-expiratory pressure in a preterm rabbit model. *Pediatr Res*, 2009. 65(5): p. 537-41.
78. Siew, M.L., et al., Surfactant increases the uniformity of lung aeration at birth in ventilated preterm rabbits. *Pediatr Res*, 2011. 70(1): p. 50-5.
79. Miserocchi, G., B.H. Poskurica, and M. Del Fabbro, Pulmonary interstitial pressure in anesthetized paralyzed newborn rabbits. *J Appl Physiol (1985)*, 1994. 77(5): p. 2260-8.
80. Bland, R.D., et al., Clearance of liquid from lungs of newborn rabbits. *J Appl Physiol Respir Environ Exerc Physiol*, 1980. 49(2): p. 171-7.
81. Hooper, S.B., A.B. Te Pas, and M.J. Kitchen, Respiratory transition in the newborn: a three-phase process. *Arch Dis Child Fetal Neonatal Ed*, 2016. 101(3): p. F266-71.
82. Martherus, T., et al., High-CPAP Does Not Impede Cardiovascular Changes at Birth in Preterm Sheep. *Front Pediatr*, 2020. 8: p. 584138.

83. Martherus, T., et al., Higher CPAP levels improve functional residual capacity at birth in preterm rabbits. *Pediatr Res*, 2021.
84. Martherus, T., et al., Feasibility and Effect of Physiological-Based CPAP in Preterm Infants at Birth. *Frontiers in Pediatrics*, 2021. 9.
85. Mian, Q.N., et al., Tidal volumes in spontaneously breathing preterm infants supported with continuous positive airway pressure. *J Pediatr*, 2014. 165(4): p. 702-6 e1.
86. Jain, D., et al., Use of a Mechanical Ventilator with Respiratory Function Monitoring Provides More Consistent Ventilation during Simulated Neonatal Resuscitation. *Neonatology*, 2019: p. 1-8.
87. Morley, C.J., Volume-limited and volume-targeted ventilation. *Clin Perinatol*, 2012. 39(3): p. 513-23.
88. Singh, J., et al., Long term follow-up of very low birthweight infants from a neonatal volume versus pressure mechanical ventilation trial. *Arch Dis Child Fetal Neonatal Ed*, 2009. 94(5): p. F360-2.
89. van Zanten, H.A., et al., A multi-centre randomised controlled trial of respiratory function monitoring during stabilisation of very preterm infants at birth. *Resuscitation*, 2021.
90. Crawshaw, J.R., et al., Laryngeal closure impedes non-invasive ventilation at birth. *Archives of Disease in Childhood - Fetal and Neonatal Edition*, 2018. 103(2): p. F112-F119.
91. Morley, C.J., Monitoring Neonatal Resuscitation: Why is it needed? *Neonatology*, 2018. 113: p. 387-392.
92. Greenough, A., et al., Synchronized mechanical ventilation for respiratory support in newborn infants. *Cochrane Database Syst Rev*, 2016. 9: p. CD000456.
93. Claure, N. and E. Bancalari, New modes of mechanical ventilation in the preterm newborn: evidence of benefit. *Arch Dis Child Fetal Neonatal Ed*, 2007. 92(6): p. F508-12.
94. Keszler, M., Volume-targeted ventilation: one size does not fit all. Evidence-based recommendations for successful use. *Arch Dis Child Fetal Neonatal Ed*, 2019. 104(1): p. F108-F112.
95. Klingenberg, C., et al., Volume-targeted versus pressure-limited ventilation in neonates. *Cochrane Database Syst Rev*, 2017. 10: p. CD003666.
96. O'Donnell, C.P., et al., Clinical assessment of infant colour at delivery. *Arch Dis Child Fetal Neonatal Ed*, 2007. 92(6): p. F465-7.

97. Schmolzer, G.M., et al., Respiratory function monitor guidance of mask ventilation in the delivery room: a feasibility study. *J Pediatr*, 2012. 160(3): p. 377-381 e2.
98. Schmolzer, G.M., et al., Assessment of tidal volume and gas leak during mask ventilation of preterm infants in the delivery room. *Arch Dis Child Fetal Neonatal Ed*, 2010. 95(6): p. F393-7.
99. Wood, F.E., et al., A respiratory function monitor improves mask ventilation. *Arch Dis Child Fetal Neonatal Ed*, 2008. 93(5): p. F380-1.
100. Binder, C., et al., Human or monitor feedback to improve mask ventilation during simulated neonatal cardiopulmonary resuscitation. *Arch Dis Child Fetal Neonatal Ed*, 2014. 99(2): p. F120-3.
101. Kelm, M., et al., Manual neonatal ventilation training: a respiratory function monitor helps to reduce peak inspiratory pressures and tidal volumes during resuscitation. *J Perinat Med*, 2012. 40(5): p. 583-6.
102. Milner, A., et al., Evaluation of respiratory function monitoring at the resuscitation of prematurely born infants. *Eur J Pediatr*, 2015. 174(2): p. 205-8.

