Universiteit
Leiden
The Netherlands

Safety and performance of high-risk medical devices: the role of real-world data

Hoogervorst, L.A.

Citation
Hoogervorst, L. A. (2025, November 18). *Safety and performance of high-risk medical devices: the role of real-world data*. Retrieved from https://hdl.handle.net/1887/4283146

Version:	Publisher's Version
License:	Licence agreement concerning inclusion of doctoral thesis in the Institutional Repository of the University of Leiden
Downloaded from:	https://hdl.handle.net/1887/4283146

Note: To cite this publication please use the final published version (if applicable).

Chapter 1 – General introduction and outline of the thesis



Historical perspective of arthroplasty implants

One of the most frequently used medical devices in orthopedics are traumatology implants (e.g. nails and plates) and arthroplasty implants. The latter referred to as joint replacement implants, joint prosthesis, or artificial joints.[1] The earliest written evidence of the use of a prosthetic implant refers to the "Rig Veda", a sacred Indian poem (3500 – 1800 Before Christ).[2,3] This poem recounts the story of a warrior-queen, who lost her leg while fighting a battle. To enable her to return to the battle, her leg was replaced by an iron prosthetic implant.

Whereas the prosthetic amputee implant used in Rig Veda is written evidence of the use of a prosthesis, the oldest arthroplasty implant found by archaeologists originates to the ancient Egyptians. A female mummy was found at the necropolis of Thebes-West (built 1500 – 1300 Before Christ) with a missing big toe on her right foot.[2] The missing toe was replaced by a wooden prosthesis including a joint (i.e. an external arthroplasty implant) consisting of three separate well-manufactured components; a "body" (replacing the toe) and two smaller wooden plates, used for keeping the "body" in the correct position. This construction was expected to provide sufficient stability and to allow her to move without major restrictions.

Next to these external amputee prosthesis used for amputated limbs, the first "in-body" arthroplasty implant has been described in the 19th century by Professor Themistocles Glück.[4,5] Glück designed and implanted arthroplasty knee and hip implants made from ivory. Although Glück's arthroplasty implants were deemed successful in the short term (i.e. the first two years), longer follow-up results were less favourable, with all implants ultimately failing due to chronic infection.[4] Around the same time, the first total shoulder arthroplasty implant was implanted in 1893 by the French surgeon Péan.[6]

Innovations in both arthroplasty materials, designs and surgical techniques have greatly improved since the 19th and early 20th century with the development of better implant materials. Also, since the 1970s -1980s with better, more reproducible surgical tools. Since the 1990s, total hip and total knee arthroplasty implants are considered the most successful treatments for several hip- and knee-related disorders and diseases, while also one of the most commonly performed elective surgeries worldwide.[7-9] Unfortunately, despite these improvements in both implants and surgical technique, failures related to the safety and performance of arthroplasty implants still occur.[10-12] The latter and the fact that the number of arthroplasty surgeries performed worldwide is still increasing , demands for a more rigorous system for evaluating clinical evidence of arthroplasty implants.

The most commonly used clinical outcome to indicate the safety and performance of orthopedic arthroplasty implants is "revision surgery". Revision surgery can be done with or without component exchange.[13] Revision without the exchange of components is usually performed in case of implant loosening. Revision with component exchange (i.e. replacing the old component for a new component) is the most commonly used clinical outcome, since it has significant consequences for the patient (e.g. undergoing a re-operation, having a higher risk of infection when compared to primary arthroplasty, and worse patient reported outcome measurements (PROMs)).[14-17] Next to the impact for patients, revision surgery also leads to incurring personal and societal costs.[18]

A well-known safety incident in the field of arthroplasty implants relates to the metal-on-metal total hip arthroplasty implants.[12,19,20] Originally developed as a more durable alternative to total hip arthroplasty implants that incorporated ceramic or polyethylene components, mid-term follow-up registry data of patients with metal-on-metal total hip arthroplasty implants showed far higher revision rates when compared with other total hip arthroplasty implants.[12] In addition, literature showed the occurrence of pseudo-tumours and adverse reactions to metal debris in patients with metal-on-metal total hip arthroplasty implants.[12,20] As a consequence, metal-on-metal total hip arthroplasty implants were taken off the market ("recalled") by their manufacturers to prevent further patient harm. However, at the time of recall, many metal-on-metal total hip arthroplasty implants were already implanted in patients, affecting over one million patients worldwide.[12,21] The case of metal-on-metal total hip arthroplasty implants demonstrates that to prevent patient harm it is important to use orthopedic medical devices with high safety and good performance[22] as well as the need for monitoring the performance of devices after they have entered the market, to detect any devices with safety concerns as soon as possible.[23,24]

Bringing a medical device on the European market

The regulation regarding medical devices as set by the European Union defines requirements of medical device design and fabrication to ensure quality assurance and safe use of medical devices.[25] The first European directive on medical devices (MDD) came into force in 1990.[25-28] Since then, several amendments to this directive have been made, resulting in stricter and more explicit requirements on the safety and performance of medical devices. In 2017, the European Union has significantly revised its regulatory framework by the Medical Device Regulation.[29] The Medical Device Regulation has a four-year transition period, which was extended to 2028. The aim of the new regulatory framework is to provide clinical evidence and to improve patient safety while ensuring that innovative medical devices remain available for patients.[30]

Before new medical devices can be introduced to the European market, they must first obtain a "conformité européenne" (CE) marking.[31] The CE marking for medical devices implies (but does not guarantee) that the medical device meets the general safety and performance requirements.[31] These general safety and performance requirements are set by European authorities to ensure that new medical devices will be safe and perform as intended throughout their entire life cycle. Once a medical device has attained a CE marking, it is approved to be sold in the European member states and thus allowed to be used in patients. To obtain a CE marking, clinical evidence has to be collected and reviewed by manufacturers (i.e. pre-market clinical investigation) to show that their medical device is safe, has good clinical performance and an acceptable benefit/risk ratio.[24,31,32]

The pre-market clinical investigation must be objective, which means that both favourable (e.g. high implant survival percentages, low number of adverse events, and good PROMs) and unfavourable (e.g. low implant survival percentages, high number of adverse events, and low PROMs) clinical data have to be included. If the conformity of the medical devices are deemed to be sufficient, the device will be certified with a CE marking. However, CE marking does not indicate that a medical device is effective, only that it complies with the legal requirements.[31] Moreover, pre-market clinical evaluation is an assessment of clinical data from any relevant source, including literature on similar medical devices.[31] Thus, the pre-market clinical evaluation may not include an extensive analysis of real-world data of the actual medical device used in all patients. In addition, pre-market clinical investigation is often based on studies with a relatively short study duration when compared with their (expected) long duration of use in patients, thereby not capturing failures at longer follow-up.[31,33,34] Therefore, infrequent adverse events may not be captured due to the relatively short follow-up period or because of a small patient population (e.g. not all patients receiving the medical device might be analysed during pre-market clinical evaluation).[33,34] Given these pre-market clinical investigation shortcomings, manufactures have to show high safety and good performance of their medical device in daily clinical practice after having entered the market to ensure patients' safety, using real-world data.

Real-world data sources to assess the safety and performance of medical devices

For decades, data from randomised controlled trials were considered as the highest-quality evidence in clinical research as they provide the ideal study design to minimise confounding and thereby can demonstrate causality between specific interventions and outcomes.[34-39] However, randomised controlled trials do have limitations such as: i) mostly including highly selective patient populations, often not representative of the general patient population; ii) having a limited period of follow-up and thereby a small chance to capture infrequent adverse events or events

that occur at late follow-up; iii) it may pose ethical challenges, and iv) not always the most suitable study design (e.g. when comparing the safety and performance of a specific arthroplasty implant between different countries).[37,38] Consequently, the safety and performance as demonstrated in randomised controlled trials will frequently differ from the safety and performance of medical devices in daily clinical practice.[38,40-42] Hence, outcomes from high quality clinical data collection systems, such as registries where real-world data is collected, are needed.

Different real-world data sources can be used for the detection of any safety and performance concerns for specific medical devices (e.g. electronic health records, administrative data, patient generated data, safety notices, and registries), all having their own strengths and weaknesses.

Data obtained from electronic health records (i.e. an electronic system maintained by healthcare providers to collect and store patients' medical information[43]) may enable clinicians and researchers, after informed consent has been obtained by patients, to exchange information between different hospitals in order to improve the ability to conduct research and (in)directly patient care.[44] On the other hand, the data may be of poor data quality such as incomplete or missing records because they are gathered as part of routine patient care.[45] Administrative data are relatively easily accessible and cost-efficient as well as covering a large patient population. Administrative data however is collected as part of routine care and thus not intended to be used for research purposes, which may lower the validity and reliability of administrative databases for clinical research.[46,47] Patient generated data (i.e. data generated by and from patients) may capture other outcomes, such as PROMs, when compared to the other four data sources. But it may be challenging to get sufficient response rates which limits the generalisability of patient generated data.[48] Another data source which can be used to assess the safety and performance of medical devices are safety notices. Safety notices can be viewed as incident reporting data, issued when an adverse event related to their medical devices triggers action from the manufacturer.[49] Incident reporting is often voluntary, where known disadvantages include that it is heavily influenced by reporting behaviour and therefore primarily reflects the safety culture of e.g. an organisation rather than the underlying safety epidemiology.[50] However, safety notices are legally mandatory to be reported and therefore less influenced by reporting behaviour. They are typically publicly reported on the websites of national competent authorities or regulatory agencies. Safety notices can be issued for a wide variety of issues (e.g. from packaging and labelling to material integrity). Thus, these safety notices do not always indicate a problem with the mechanical or clinical performance of a particular medical device. Although reporting safety notices is legally mandatory, safety notices only provide the numerator of a possible safety concern "the number of devices with safety concerns" and not the denominator "how

many patients are at risk to develop a safety concern" (i.e. in which the devices were implanted). Medical device registries have both the numerator as well as the denominator, if high implant-level completeness is present, which makes them well suited to provide clinical evidence of the safety and performance of medical devices used on the market.

A registry is defined by the International Medical Device Regulators Forum as an "organized system with a primary aim to increase the knowledge on medical devices contributing to improve the quality of patient care that continuously collects relevant data, evaluates meaningful outcomes and comprehensively covers the population defined by exposure to particular device(s) at a reasonably generalizable scale (e.g. international, national, regional, and health system)".[51] A medical device registry is thus an unselected population-based health information system collecting large numbers of real-world data regarding the safety and performance of specific medical devices over time, while having a longer follow-up period than in randomised controlled trials. The main purpose of arthroplasty registries is to collect information on all patients receiving a specific arthroplasty implant to monitor and improve clinical outcomes.[52] The value of registry data is internationally acknowledged, which has resulted in a strong rise of hospital-based, regional, and national arthroplasty registries in the last century (Figure 1).



Figure 1. *World map of countries (coloured in green, n=29) with a hospital-based, regional or national arthroplasty registry (Figure adapted from the International Society of Arthroplasty Registries template "Arthroplasty registries at a glance"[53])*

Registries may however vary in terms of: i) design; ii) organisation; iii) methods used for data collection, and iv) collected outcomes (e.g. different definitions of outcomes and follow-up durations).[54,55] Such variation in structural and methodological characteristics may influence the quality of data collected. In addition, individual registries may not have sufficient sample size to analyse the safety and performance for less frequently used medical devices, requiring

coordination and collaboration of registries across countries. Pooling data across registries has several benefits, such as: i) a larger set of data, which will better reflect daily clinical practice while also allowing to analyse less frequently occurring adverse events, and ii) potentially resulting in earlier detection of safety and performance issues for specific medical devices. Before pooling data across registries could be achieved, consensus on a common dataset of outcomes (including similar definitions) that all medical device registries would collect is needed. Several principles have been proposed by orthopedic associations (e.g. the Network of Orthopaedic Registries of Europe (NORE)[56]) and regulators (e.g. the International Medical Device Regulators Forum (IMDRF)[57]) to evaluate whether the quality of data used to assess the safety and performance of medical devices on the market meets the scientific standards including; coverage ("the extent of participation in data collection"), completeness ("data used in analyses are consistently captured"), accuracy ("data recorded is an accurate reflection of the healthcare event"), consistency ("uniformity in following the same procedures for data capture"), integrity ("consistent recording of unique identification of medical devices"), and reliability ("reproducibility of data elements")[51]. Specific criteria have however not been proposed, and it is therefore unknown if existing medical device registries meet the medical device regulation requirements to an acceptable standard and can be reliably used to assess the safety and performance of medical devices on the market.

Methods used to evaluate the safety and performance of medical devices

In orthopedic practice, different methods to assess the safety and performance of arthroplasty implants are currently used to guarantee that clinicians and hospitals select the best-performing implants, in which implants are compared to: i) the best-performing implant, based on implant survival analysis (i.e. revision as end-point); ii) the average performance (again, based on implant survival analysis) of other comparable implants, and iii) absolute thresholds by using objective-performance-criteria.[58-68] Arthroplasty registries have different approaches in place for identification of outlier implants (i.e. implants having a significantly worse performance (i.e. revision) than expected), mainly using the first two methods; comparing specific implants to the best-performing implants or to other comparable implants. Outlier identification in registries has resulted in multiple outlier medical devices being taken off the market.[58]

Using an absolute threshold (i.e. objective-performance criteria) has the benefit of relatively easy interpretation and thereby making the assessment of implant safety and performance more straightforward. The limitation of absolute thresholds is that they may have to be updated over time (given that implants performances may change over time), so it has to be considered whether the absolute threshold is still acceptable. The most frequently used objective-performance-criterium in orthopedic practice is the Orthopaedic Data Evaluation

Panel (ODEP) rating.[62] The ODEP criterium is designed and used to promote evidence-based selection of orthopedic implants by assigning a rating (the ODEP-rating) to orthopedic implants based on implant survival analysis criteria (i.e. revision rates). ODEP ratings are available for TH components (cups and stems), TK implants (tibiofemoral combinations), unicondylar knee implants, shoulder components (glenoids and stems), reverse shoulder implants, total elbow implants, and spine implants (cervical discs). As different data sources can be used by manufacturers to submit their application for an ODEP-rating (e.g. single-center studies, manufacturers' in-house sources, and registry data), this data may not be representative of daily clinical practice. Therefore, before submission, manufacturers must declare that "the clinical data submitted is representative of all studies that have been conducted in relation to it." Although originally focused on the United Kingdom, the ODEP-rating is increasingly used in other countries for the quality assessment of arthroplasty implants.[69-71] Despite being increasingly used, external validation of ODEP-ratings across multiple registries has never been undertaken. Therefore, it is unclear whether the performance of a particular implant in another country than the United Kingdom is consistent with the assigned ODEP-rating and thereby if surgeons or hospitals outside the United Kingdom can use the ODEP-rating to guide implant selection.

The safety and performance of a specific medical device may however differ between hospitals within a single registry, but also between registries, as research has shown that besides factors related to the medical device itself, there are many more factors influencing the safety and performance of medical devices.[72-74] Focusing on arthroplasty implants, patient related factors (e.g. body mass index (BMI), comorbidity classification, gender, and indication for surgery) may differ between specific implants and are known to influence revision rates and clinical outcomes.[72,73,75-77] Thus, before differences in revision risks between registries for specific implants can be interpretated, a better understanding of any differences in patient characteristics between registries for specific implants is needed, which may also point to differences in patient selection for the same implant across registries. Only a few studies have assessed differences in patient characteristics across registries.[78,81] However, the majority of these studies only focused on variations in pre-operative pain and function levels, even more important, all these studies analysed all arthroplasty implants combined (e.g. all cemented total hip implants combined) instead of analysing differences for specific medical devices from specific manufacturers. Hence, a more comprehensive analysis of the similarities and differences in patients receiving these specific arthroplasty implants in different countries is required to better understand differences in implants' safety and performances across registries.

Besides these medical device- and patient-related factors, surgeon- and hospital-factors are known to effect performance of medical devices.[74,82-86] Therefore, an essential part of clinical practice should involve reducing complication risks (e.g. revisions) by continuously improving the quality of care delivered by surgeons and hospitals thereby minimising patient harm and unnecessary costs. Hence, an increased number of arthroplasty registries are providing feedback to orthopedic surgeons and hospitals, with the aim to minimise patients' morbidity and costs.[87-90] This feedback is intended to serve two goals: i) individual hospitals can monitor their performance over time to examine if outcomes are improving or deteriorating, and ii) hospitals performing worse than other hospitals while treating comparable patients, might compare their practices to better-performing hospitals to identify possibilities for improvement. Literature has shown that if this feedback is provided in an active approach by also discussing these results with clinicians, outcomes for patients will improve (e.g. revision of implants decreases, less complications, and a reduced length of hospital stay).[91] However, variation in overall revision risks does not give sufficient information to guide where to improve care as revisions may be performed for different indications (e.g. infection or surgical technique failure) each aligned with different quality improvement initiatives.

In the Netherlands, large between-hospital variation in one- and three-year revision rates following primary total hip and total knee arthroplasty has been found.[74] For most of the hospitals performing worse than expected, a specific indication for revision could be identified as contributing to the overall worse performance. The latter can guide quality improvement initiatives. However, whether this also applies to other arthroplasty implants, such as shoulder arthroplasty implants, is currently unknown. As shoulder arthroplasty, both primarily and revisions, are less commonly performed than hip- and knee-arthroplasty, between-hospital variation in revision rates following primary shoulder arthroplasty might not be reliable given that large numbers (both for primary and revision surgery) are needed to reliably assess between-hospital performances.

Outline of this thesis

This thesis aims to investigate how real-world data of implantable high-risk medical devices – particularly from arthroplasty registries and safety notices – can be used to i) ensure high-quality data and evidence regarding safety and performance of high-risk medical devices, and ii) develop methods for the evaluation of safety and performance of high-risk medical devices.

In **Chapter 2,** European orthopedic registries (hip and knee arthroplasty implants) and cardiovascular registries (coronary stents and valve repair/replacement) will be identified and reviewed on the extent to which 33 structural and methodological items are publicly reported that influence the quality of registry data, as well as the definitions used and reported outcomes. By assessing this, the utility of current orthopedic and cardiovascular medical device registries for regulatory purposes will be examined. Based on these results, a three-round Delphi study will then be conducted, including all stakeholders in the regulatory evaluation of medical devices. This Delphi study aims to achieve consensus on a minimum dataset needed to assess the quality of registry data and analysis of medical device safety and performance (**Chapter 3**).

Combining registry data with other real-world data sources may highlight safety problems that otherwise would remain unnoticed, and therefore **Chapter 4** will assess the extent to which safety notices issued by manufacturers signal the same total knee arthroplasty implants as those identified by registries as outlier implants (i.e. implants having a significantly worse performance (i.e. revision) than expected).

Chapter 5 will extend the results from Chapter 4 by not only analysing registry data and safety notices, but also safety problems identified in the peer-reviewed literature and from manufacturers' websites. The frequency of safety problems will be assessed for a random sample set of ten hip arthroplasty stems, ten hip arthroplasty cups, and ten knee arthroplasty implants to also obtain an estimate of the percentage of implants without safety concerns in any of these data sources.

In **Chapter 6**, multi-registry data will be used to validate the ODEP-rating, using revision risks of total hip and total knee arthroplasty implants across nine European and non-European registries. These results will show if the ODEP-rating could be used across countries to provide evidence on the safety and performance of total hip and total knee arthroplasty implants even if based originally on data obtained from another country.

Understanding and identifying differences in patient characteristics is important to ensure fair comparison of implant performances, given that these characteristics are known to influence the risk of revision. Hence, in **Chapter 7**,

multi-registry data will be used to assess differences in patient characteristics of patients receiving the same total knee arthroplasty implants in different registries. If patient-mix differs across countries, this should be taken into account for a fair comparison of safety and performance regarding specific implants across registries.

In **Chapter 8**, national Dutch registry data will be used to assess whether the occurrence of revision surgery can be used for quality assessment to judge hospital performance following primary shoulder arthroplasty. In addition, the extent of between-hospital variation in revision following primary shoulder arthroplasty in the Netherlands will be assessed, both overall and for specific revision indications to guide quality improvement initiatives and to improve shoulder arthroplasty performances.

Chapter 9 includes a summary, general discussion, and offers some future perspectives.

References

1. Haglin JM, Eltorai AE, Gil JA, et al. Patient-Specific Orthopaedic Implants. Orthop Surg. 2016;8(4):417-24
2. Thurston AJ. Paré and prosthetics: the early history of artificial limbs. ANZ J Surg. 2007;77(12):1114-9
3. Kaul H, Ventikos Y. On the genealogy of tissue engineering and regenerative medicine. Tissue Eng Part B Rev. 2015;21(2):203-17
4. Hernigou P. Earliest times before hip arthroplasty: from John Rhea Barton to Themistocles Glück. Int Orthop. 2013;37(11):2313-8
5. Glück T, Wessinghage D. Referat über die durch das moderne chirurgische Experiment gewonnenen positiven Resultate, betreffend die Naht und den Ersatz von Defecten höherer Gewebe, sowie über die Verwerthung resorbirbarer und lebendiger Tampons in der Chirurgie: Schattauer; 1988
6. Willems WJ. History of shoulder arthroplasty. Journal of Arthroscopy and Joint Surgery. 2021;8(1):2-6
7. Kremers HH, Larson DR, Crowson CS, et al. Prevalence of Total Hip and Knee Replacement in the United States. J Bone Joint Surg Am. 2015;97(17):1386-97
8. Pivec R, Johnson AJ, Mears SC, et al. Hip arthroplasty. Lancet. 2012;380(9855):1768-77
9. Learmonth ID, Young C, Rorabeck C. The operation of the century: total hip replacement. Lancet. 2007;370(9597):1508-19
10. Wilson RA, Gwynne-Jones DP, Sullivan TA, et al. Total Hip and Knee Arthroplasties Are Highly Cost-Effective Procedures: The Importance of Duration of Follow-Up. J Arthroplasty. 2021;36(6):1864-72.e10
11. Malahias MA, Iacobelli DA, Torres L, et al. Early Aseptic Loosening with Increased Presence and Severity of Backside Burnishing in the Optetrak Logic Posterior-Stabilized Total Knee Arthroplasty Femoral Component. J Knee Surg. 2022;35(14):1595-603
12. Pijls BG, Meessen JMTA, Tucker K, et al. MoM total hip replacements in Europe: a NORE report. EFORT Open Rev. 2019;4(6):423-9
13. van Schie P, Hasan S, van Bodegom-Vos L, et al. International comparison of variation in performance between hospitals for THA and TKA: Is it even possible? A systematic review including 33 studies and 8 arthroplasty register reports. EFORT Open Rev. 2022;7(4):247-63
14. Heo K, Karzon A, Shah J, et al. Trends in Costs and Professional Reimbursements for Revision Total Hip and Knee Arthroplasty. J Arthroplasty. 2024;39(3):612-8.e1
15. Fang CJ, Shaker JM, Ward DM, et al. Financial Burden of Revision Hip and Knee Arthroplasty at an Orthopedic Specialty Hospital: Higher Costs and Unequal Reimbursements. J Arthroplasty. 2021;36(8):2680-4
16. Harada S, Hamai S, Shiomoto K, et al. Patient-reported outcomes after primary or revision total hip arthroplasty: A propensity score-matched Asian cohort study. PLoS One. 2021;16(5):e0252112
17. Stirling P, Middleton SD, Brenkel IJ, et al. Revision total knee arthroplasty versus primary total knee arthroplasty: a matched cohort study. Bone Jt Open. 2020;1(3):29-34
18. Roof MA, Levine BR, Schwarzkopf R. The Hidden Cost of Revision Hip and Knee Arthroplasty. Arthroplast Today. 2022;16:167-8
19. Graves SE, Rothwell A, Tucker K, et al. A Multinational Assessment of Metal-on-Metal Bearings in Hip Replacement. The Journal of bone and joint surgery American volume. 2011;93 Suppl 3:43-7
20. Smeekes C, Cleven AHG, van der Wal BCH, et al. Current Pathologic Scoring Systems for Metal-on-metal THA Revisions are not Reproducible. Clin Orthop Relat Res. 2017;475(12):3005-11
21. Berber R, Skinner JA, Hart AJ. Management of metal-on-metal hip implant patients: Who, when and how to revise? World J Orthop. 2016;7(5):272-9
22. Faro LM, Huiskes R. Quality assurance of joint replacement Legal regulation and medical

judgement. Acta Orthopaedica Scandinavica. 1992;63(sup250):1-33
23. Nelissen RGHH, Pijls BG, Kärrholm J, et al. RSA and registries: the quest for phased introduction of new implants. J Bone Joint Surg Am. 2011;93 Suppl 3:62-5
24. Overgaard S, Grupp TM, Nelissen RGHH, et al. Introduction of innovations in joint arthroplasty: Recommendations from the 'EFORT implant and patient safety initiative'. EFORT Open Rev. 2023;8(7):509-21
25. MedTech Europe. The transition to a new regulatory framework for in vitro diagnostic medical devices in the European Union. Available: https://www.medtecheurope.org/wp-content/uploads/2018/07/180505_MTE_Information-on-the-Transition-to-the-IVD-Regulation_0.pdf (last accessed 11 June, 2024)
26. European Commission. Active implantable medical devices. Available: https://single-market-economy.ec.europa.eu/single-market/european-standards/harmonised-standards/implantable-medical-devices_en (last accessed 11 June, 2024)
27. Melvin T. The European Medical Device Regulation-What Biomedical Engineers Need to Know. IEEE J Transl Eng Health Med. 2022;10:4800105
28. McAllister P, Jeswiet J. Medical device regulation for manufacturers. Proc Inst Mech Eng H. 2003;217(6):459-67
29. Melvin T, Torre M. New medical device regulations: the regulator's view. EFORT Open Rev. 2019;4(6):351-6
30. Health and Youth Care Inspectorate - Ministry of Health, Welfare and Sport. European regulations for medical devices and IVDs. Available: https://english.igj.nl/medical-technology/new-european-regulations-mdr-and-ivdr (last accessed 11 June, 2024)
31. Hulstaert F, Pouppez C, Primus-de Jong C, et al. Gaps in the evidence underpinning high-risk medical devices in Europe at market entry, and potential solutions. Orphanet J Rare Dis. 2023;18(1):212
32. Nieuwenhuijse MJ, Randsborg PH, Hyde JH, et al. Evidence-based objective performance criteria for the evaluation of hip and knee replacement devices and technologies. Int J Surg. 2023;109(5):1125-35
33. Nieuwenhuijse MJ, Nelissen RGHH, Schoones JW, et al. Appraisal of evidence base for introduction of new implants in hip and knee replacement: a systematic review of five widely used device technologies. Bmj. 2014;349:g5133
34. Lübbeke A, Combescure C, Barea C, et al. Clinical investigations to evaluate high-risk orthopaedic devices: a systematic review of the peer-reviewed medical literature. EFORT Open Rev. 2023;8(11):781-91
35. Makady A, de Boer A, Hillege H, et al. What Is Real-World Data? A Review of Definitions Based on Literature and Stakeholder Interviews. Value Health. 2017;20(7):858-65
36. Hariton E, Locascio JJ. Randomised controlled trials - the gold standard for effectiveness research: Study design: randomised controlled trials. Bjog. 2018;125(13):1716
37. Booth CM, Tannock IF. Randomised controlled trials and population-based observational research: partners in the evolution of medical evidence. Br J Cancer. 2014;110(3):551-5
38. Kim HS, Lee S, Kim JH. Real-world Evidence versus Randomized Controlled Trial: Clinical Research Based on Electronic Medical Records. J Korean Med Sci. 2018;33(34):e213
39. Spieth PM, Kubasch AS, Penzlin AI, et al. Randomized controlled trials - a matter of design. Neuropsychiatr Dis Treat. 2016;12:1341-9
40. Revicki DA, Frank L. Pharmacoeconomic evaluation in the real world. Effectiveness versus efficacy studies. Pharmacoeconomics. 1999;15(5):423-34
41. Dekkers OM, Groenwold RHH. Study design: what's in a name? Eur J Endocrinol. 2020;183(6):E11-e3
42. Ankarfeldt MZ, Adalsteinsson E, Groenwold RHH, et al. A systematic literature review on the

efficacy-effectiveness gap: comparison of randomized controlled trials and observational studies of glucose-lowering drugs. Clin Epidemiol. 2017;9:41-51
43. Gliklich RE, Leavy MB, Dreyer NA. Tools and Technologies for Registry Interoperability, Registries for Evaluating Patient Outcomes: A User's Guide, 3rd Edition, Addendum 2. Rockville (MD): Agency for Healthcare Research and Quality (US); 2019 Oct. Report No.: 19(20)-EHC017-EF
44. Menachemi N, Collum TH. Benefits and drawbacks of electronic health record systems. Risk Manag Healthc Policy. 2011;4:47-55
45. Honeyford K, Expert P, Mendelsohn EE, et al. Challenges and recommendations for high quality research using electronic health records. Front Digit Health. 2022;4:940330
46. Timofte D, Pantea Stoian A, Razvan H, et al. A Review on the Advantages and Disadvantages of Using Administrative Data in Surgery Outcome Studies. The Journal of Surgery. 2018;14
47. Johnson EK, Nelson CP. Values and pitfalls of the use of administrative databases for outcomes assessment. J Urol. 2013;190(1):17-8
48. Rolfson O, Bohm E, Franklin P, et al. Patient-reported outcome measures in arthroplasty registries Report of the Patient-Reported Outcome Measures Working Group of the International Society of Arthroplasty Registries Part II. Recommendations for selection, administration, and analysis. Acta Orthop. 2016;87 Suppl 1(Suppl 1):9-23
49. Health and Youth Care Inspectorate - Ministry of Health, Welfare and Sport. Safety notices. Available: https://english.igj.nl/medical-technology/safety-notices (last accessed 11 June, 2024)
50. Macrae C. The problem with incident reporting. BMJ Quality Saf. 2016;25(2):71-5
51. International Medical Device Regulators Forum (IMDRF). Principles of International System of Registries Linked to Other Data Sources and Tools. Available: https://www.imdrf.org/sites/default/files/docs/imdrf/final/technical/imdrf-tech-160930-principles-system-registries.pdf (last accessed 11 June, 2024)
52. Lübbeke A, Carr AJ, Hoffmeyer P. Registry stakeholders. EFORT Open Rev. 2019;4(6):330-6
53. Lübbeke A, Hoogervorst LA, Marang-van de Mheen PJ, et al. Arthroplasty registries at a glance: an initiative of the International Society of Arthroplasty Registries (ISAR) to facilitate access, understanding, and reporting of registry data from an international perspective. Acta Orthop. 2025;96:116-26
54. Niederländer C, Wahlster P, Kriza C, et al. Registries of implantable medical devices in Europe. Health Policy. 2013;113(1-2):20-37
55. Faxon DP, Burgess A. Cardiovascular Registries: Too Much of Good Thing? Circ Cardiovasc Interv. 2016;9(4):e003866
56. Network of Orthopaedic Registries of Europe (NORE). MINIMAL DATASETS. Available: https://nore.efort.org/minimal-datasets (last accessed 4 April, 2025)
57. International Medical Device Regulators Forum (IMDRF). Methodological Principles in the Use of International Medical Device Registry Data. Available: https://www.imdrf.org/sites/default/files/docs/imdrf/final/technical/imdrf-tech-170316-methodological-principles.pdf (last accessed 29 April, 2025)
58. de Steiger RN, Miller LN, Davidson DC, et al. Joint registry approach for identification of outlier prostheses. Acta Orthop. 2013;84(4):348-52
59. Australian Government - Department of Health. Prostheses List - Guide to listing and setting benefits for prostheses. Available: https://www.health.gov.au/sites/default/files/documents/2020/06/prostheses-list-guide.pdf (last accessed 11 June, 2024)
60. Poolman RW, Verhaar JAN, Schreurs BW, et al. Finding the right hip implant for patient and surgeon: the Dutch strategy--empowering patients. Hip Int. 2015;25(2):131-7
61. Chubb HA, Cornish ER, Hallstrom BR, et al. Early Benchmarking Total Hip Arthroplasty Implants Using Data from the Michigan Arthroplasty Registry Collaborative Quality Initiative

(MARCQI). Orthop Res Rev. 2021;13:215-28
62. Orthopaedic Data Evaluation Panel (ODEP). What is ODEP? Available: https://www.odep.org.uk/ (last accessed 14 June, 2024)
63. Keurentjes JC, Pijls BG, Van Tol FR, et al. Which implant should we use for primary total hip replacement? A systematic review and meta-analysis. J Bone Joint Surg Am. 2014;96 Suppl 1:79-97
64. Deere KC, Whitehouse MR, Porter M, et al. Assessing the non-inferiority of prosthesis constructs used in hip replacement using data from the National Joint Registry of England, Wales, Northern Ireland and the Isle of Man: a benchmarking study. BMJ Open. 2019;9(4):e026685
65. Deere KC, Whitehouse MR, Porter M, et al. Assessing the non-inferiority of prosthesis constructs used in total and unicondylar knee replacements using data from the National Joint Registry of England, Wales, Northern Ireland and the Isle of Man: a benchmarking study. BMJ Open. 2019;9(4):e026736
66. Wyatt M, Frampton C, Whitehouse M, et al. Benchmarking total hip replacement constructs using noninferiority analysis: the New Zealand joint registry study. BMC Musculoskelet Disord. 2021;22(1):719
67. Nederlandse Orthopaedische Vereniging (NOV). Procedure Uitkomstanalyse. Available: https://www.orthopeden.org/kwaliteitsbeleid/procedure-uitkomstanalyse/ (last accessed 11 June, 2024)
68. Nelissen RG, Brand R, Rozing PM. Survivorship analysis in total condylar knee arthroplasty. A statistical review. J Bone Joint Surg Am. 1992;74(3):383-9
69. Orthopaedic Data Evaluation Panel (ODEP). Impact of ODEP. Available: https://www.odep.org.uk/about/impact-of-odep/ (last accessed 11 June, 2024)
70. Dutch Arthroplasty Register (LROI). Online LROI annual report 2022. Available: https://www.lroi-report.nl/app/uploads/2022/11/PDF-LROI-annual-report-2022.pdf (last accessed 11 June, 2024)
71. Gundersen T, Furnes O, Gjertsen JE, et al. Annual report 2021 Norwegian National Advisory Unit on Arthroplasty and Hip Fractures Norwegian Arthroplasty Register Norwegian Cruciate Ligament Register Norwegian Hip Fracture Register Norwegian Paediatric Hip Register
72. Sayed-Noor AS, Mukka S, Mohaddes M, et al. Body mass index is associated with risk of reoperation and revision after primary total hip arthroplasty: a study of the Swedish Hip Arthroplasty Register including 83,146 patients. Acta Orthop. 2019;90(3):220-5
73. Ferguson RJ, Silman AJ, Combescure C, et al. ASA class is associated with early revision and reoperation after total hip arthroplasty: an analysis of the Geneva and Swedish Hip Arthroplasty Registries. Acta Orthop. 2019;90(4):324-30
74. van Schie P, van Steenbergen LN, van Bodegom-Vos L, et al. Between-Hospital Variation in Revision Rates After Total Hip and Knee Arthroplasty in the Netherlands: Directing Quality-Improvement Initiatives. J Bone Joint Surg Am. 2020;102(4):315-24
75. Santaguida PL, Hawker GA, Hudak PL, et al. Patient characteristics affecting the prognosis of total hip and knee joint arthroplasty: a systematic review. Can J Surg. 2008;51(6):428-36
76. Peters RM, van Steenbergen LN, Stewart RE, et al. Which patients improve most after total hip arthroplasty? Influence of patient characteristics on patient-reported outcome measures of 22,357 total hip arthroplasties in the Dutch Arthroplasty Register. Hip Int. 2021;31(5):593-602
77. Peters RM, van Steenbergen LN, Stewart RE, et al. Patient Characteristics Influence Revision Rate of Total Hip Arthroplasty: American Society of Anesthesiologists Score and Body Mass Index Were the Strongest Predictors for Short-Term Revision After Primary Total Hip Arthroplasty. J Arthroplasty. 2020;35(1):188-92.e2
78. Lingard EA, Katz JN, Wright EA, et al. Predicting the outcome of total knee arthroplasty. J Bone Joint Surg Am. 2004;86(10):2179-86
79. Ackerman IN, Dieppe PA, March LM, et al. Variation in age and physical status prior to total

knee and hip replacement surgery: a comparison of centers in Australia and Europe. Arthritis Rheum. 2009;61(2):166-73
80. Franklin PD, Miozzari H, Christofilopoulos P, et al. Important patient characteristics differ prior to total knee arthroplasty and total hip arthroplasty between Switzerland and the United States. BMC Musculoskelet Disord. 2017;18(1):14
81. Paxton EW, Cafri G, Nemes S, et al. An international comparison of THA patients, implants, techniques, and survivorship in Sweden, Australia, and the United States. Acta Orthop. 2019;90(2):148-52
82. Ravi B, Jenkinson R, Austin PC, et al. Relation between surgeon volume and risk of complications after total hip arthroplasty: propensity score matched cohort study. Bmj. 2014;348:g3284
83. Hoskins W, Rainbird S, Lorimer M, et al. What Can We Learn From Surgeons Who Perform THA and TKA and Have the Lowest Revision Rates? A Study from the Australian Orthopaedic Association National Joint Replacement Registry. Clin Orthop Relat Res. 2022;480(3):464-81
84. Roof MA, Sharan M, Merkow D, et al. High-volume revision surgeons have better outcomes following revision total knee arthroplasty. Bone Joint J. 2021;103-b(6 Supple A):131-6
85. Okoro T, Tomescu S, Paterson JM, et al. Analysis of the relationship between surgeon procedure volume and complications after total knee arthroplasty using a propensity-matched cohort study. BMJ Surg Interv Health Technol. 2021;3(1):e000072
86. Van Gestel R, Broekman N, Müller T. Surgeon supply and healthcare quality: Are revision rates for hip and knee replacements lower in hospitals that employ more surgeons? Health Econ. 2023;32(10):2298-321
87. Van Schie P, Van Bodegom-Vos L, Zijdeman TM, et al. Awareness of performance on outcomes after total hip and knee arthroplasty among Dutch orthopedic surgeons: how to improve feedback from arthroplasty registries. Acta Orthop. 2021;92(1):54-61
88. National Joint Registry (NJR). NJR Surgeon and Hospital Profile. Available: https://surgeonprofile.njrcentre.org.uk/ (last accessed 4 April, 2025)
89. Swedish Hip Arthroplasty Register. Annual Report 2017. Available: https://registercentrum.blob.core.windows.net/sar/r/Swedish-Hip-Arthroplasty-Register-Annual-report-2017-BylrAoP8ro.pdf (last accessed 11 June, 2024)
90. American Academy of Orthopaedic Surgeons (AAOS). AJRR Year In Review. Available: https://www.aaos.org/globalassets/registries/3.5.20_ajrr-a-year-in-review.pdf (last accessed 11 June, 2024)
91. van Schie P, van Bodegom-Vos L, Zijdeman TM, et al. Effectiveness of a multifaceted quality improvement intervention to improve patient outcomes after total hip and knee arthroplasty: a registry nested cluster randomised controlled trial. BMJ Qual Saf. 2023;32(1):34-46