



Universiteit  
Leiden

The Netherlands

## Safety and performance of high-risk medical devices: the role of real-world data

Hoogervorst, L.A.

### Citation

Hoogervorst, L. A. (2025, November 18). *Safety and performance of high-risk medical devices: the role of real-world data*. Retrieved from <https://hdl.handle.net/1887/4283146>


Version: Publisher's Version

License: [Licence agreement concerning inclusion of doctoral thesis in the Institutional Repository of the University of Leiden](#)

Downloaded from: <https://hdl.handle.net/1887/4283146>

**Note:** To cite this publication please use the final published version (if applicable).

**Chapter 4 – Safety notices and registry outlier data measure different aspects of safety and performance of total knee implants: a comparative study of safety notices and register outliers**



Lotje A. Hoogervorst<sup>\*1</sup>, Yijun Ren<sup>\*2</sup>, Tom Melvin<sup>3</sup>, Ashley A. Stratton-Powell<sup>3</sup>, Anne Lübbeke<sup>4,5</sup>, Robert E. Geertsma<sup>6</sup>, Alan G. Fraser<sup>7</sup>, Rob G.H.H. Nelissen<sup>1</sup>, Enrico G. Caiani<sup>‡2,8</sup>, Perla J. Marang-van de Mheen<sup>‡9</sup>

<sup>1</sup> Department of Orthopaedics, Leiden University Medical Center, Leiden, the Netherlands

<sup>2</sup> Department of Electronics, Information and Biomedical Engineering, Politecnico di Milano, Milan, Italy

<sup>3</sup> School of Medicine, Trinity College Dublin, Dublin, Ireland

<sup>4</sup> Division of Orthopaedic Surgery and Traumatology, Geneva University Hospitals and University of Geneva, Geneva, Switzerland

<sup>5</sup> Nuffield Department of Orthopaedics, Rheumatology and Musculoskeletal Sciences, University of Oxford, Oxford, the United Kingdom

<sup>6</sup> Centre for Health Protection, National Institute for Public Health and the Environment, Bilthoven, the Netherlands

<sup>7</sup> Department of Cardiology, University Hospital of Wales, Heath Park Way, Cardiff, the United Kingdom

<sup>8</sup> Istituto Auxologico Italiano IRCCS, Milan, Italy

<sup>9</sup> Safety & Security Science and Centre for Safety in Healthcare, Delft University of Technology, Delft, the Netherlands

\* = shared first authorship; ‡ = shared last authorship

*Acta Orthop. 2024 Nov 25;95:667-676*

## **Abstract**

*Background:* Safety notices for medical devices such as total knee (TK) implants may indicate problems in their design or performance that require corrective action to prevent patient harm. Safety notices are often published on national Ministries of Health or regulatory agencies websites. It is unknown whether problems triggering safety notices identify the same implants as those identified by registries as “outlier.” We aimed to assess the extent to which safety notices and outlier identification in registries signal the same or different TK implants.

*Methods:* The CORE-MD tool, an automated web scraper tool, was used to collect safety notices related to TK implants on 13 national Ministries of Health websites and regulatory agencies. Safety notices were defined according to the Medical Device Regulation (MDR) as “a communication sent by a manufacturer to users or customers in relation to a field safety corrective action.” Identified TK outliers, defined as having a significantly higher revision risk than other comparable TK implants, were extracted from registry reports.

*Results:* 787 safety notices for 38 TK implants and 35 TK outliers were identified, together identifying 47 unique TK implants. 26 (55%) TK implants had safety notices and were also outliers, 12 (26%) TK implants had only safety notices, and 9 (19%) were outliers only. TK implants with safety notices only had similar types of problems to TK outliers with safety notices, with “Manufacturing/Packaging/Shipping” problems being most frequent (44%). Cumulative revision risks (1/5/10-years) were lower for TK implants with safety notices only than for TK outliers with safety notices.

*Conclusion:* 55% of the TK with a safety notice were identified as outliers in the registry, whereas around 25% of TK outliers are not the subject of publicly released safety notices, with safety notices pointing to TK implants not identified by registries as potentially

having a higher risk of failure. This suggests that safety notices and registry outlier data measure different aspects of safety and performance.

## **Background**

Medical devices are subject to post-market surveillance (PMS) where manufacturers have to collect and review data on experience with their devices.<sup>1</sup> Once collected, these data must be analysed by the manufacturer to evaluate if any corrective or preventive actions are needed. If action is required to prevent patient harm, a safety notice must be issued.<sup>2</sup> Safety notices can be published on the websites of manufacturers, Ministries of Health, and regulatory agencies. Safety notices may include a recall, amended instructions for use, adverse events, or additional information concerning the device. From a safety and performance perspective, total knee (TK) implants are of interest, as together with total hip implants they are the most used arthroplasty implants.

Safety notices are relevant for clinicians and hospitals as they may guide implant selection. Safety notices can be issued for a wide variety of implant-related issues (e.g. packaging and labelling), which are not always associated with the safety or performance of a TK implant. On the other hand, several arthroplasty registries have procedures in place to identify TK implants with outlier performance (i.e. a significantly higher revision risk than other comparable implants)<sup>3</sup>, defined solely based on revision risk.<sup>4,5</sup> Safety notices, however, may also refer to signals based on other outcomes (e.g. poor patient satisfaction scores), meaning that safety notices and outlier identification may reflect different aspects of patient safety. Furthermore, safety notices may be issued based on other data sources such as peer-reviewed publications. Hence, it is unknown whether problems triggering safety notices identify the same TK implants as those identified by registries as outliers.

We aimed to assess the extent to which safety notices and outliers identified by registries signal the same or different TK implants, and to explore possible reasons for any discrepancies.

## **Methods**

### *Study design*

This study focused on the agreement between two real-world data sources that are intended to signal problems related to medical devices, and more specifically to assess whether TK implants for which safety notices were published on the websites of Ministries of Health and regulatory agencies were the same as the TK implants identified and publicly reported by registries as outliers. Only TK implants currently used on the market were included. The study was conducted according to the STARD guidelines.<sup>6</sup>

### *Data collection of safety notices reporting TK implants*

The Coordinating Research and Evidence for Medical Devices (CORE-MD) PMS tool<sup>7</sup>, an automated web scraper tool, was used to identify TK implants with safety notices on the websites of Ministries of Health and regulatory agencies. 13 countries were included in the CORE-MD tool and were therefore assessed in the current study: Australia, Czechia, Denmark, France, Germany, Greece, Ireland, Italy, Portugal, Spain, Sweden, the United States of America (USA), and the Netherlands. Note that all historical and publicly available safety notices were retrieved for each country with their respective last update (Supplementary Table 1 – online available).

Details of the applied methodology in the CORE-MD PMS tool have been published.<sup>7</sup> Briefly, the tool screens the website of each Ministry of Health and regulatory agency to collect all safety information, including safety notices, alerts, and recalls. We refer to safety notices to indicate the collective safety information found on these websites.

To include only safety notices for TK implants currently on the market, a list of all TK implants from the latest annual reports from the following national and regional registries was constructed: American Joint Replacement Registry (AJRR)<sup>8</sup>, Australian Orthopaedic Association National Joint Replacement Registry

(AOANJRR)<sup>9</sup>, Dutch Arthroplasty Register (LROI)<sup>10</sup>, Emilia-Romagna Register (RIPO)<sup>11</sup>, German Arthroplasty Registry (EPRD)<sup>12</sup>, Swiss National Hip & Knee Joint Registry (SIRIS)<sup>13</sup>, and the National Joint Registry for England, Wales, Northern Ireland, the Isle of Man and Guernsey (NJR)<sup>14</sup>, and up-to-date registry website data from the Finnish Arthroplasty Register (FAR)<sup>15</sup>. Note that some countries included in the CORE-MD PMS tool to identify safety notices are not used to construct the list of TK implants currently on the market as they lack a(n) (active) regional or national arthroplasty registry capturing data on TK implants<sup>3</sup>. We assumed that safety notices would identify problems that relate to the implant itself rather than reflecting, e.g. limited experience by surgeons or patient case-mix, and thereby that the problems highlighted in these countries would reflect problems elsewhere.

The brand name of each TK implant on this list was used as input for the CORE-MD PMS tool, to extract all associated safety notices for further analysis. Based on the extended safety notice text, the described adverse event was linked to an International Medical Device Regulators Forum (IMDRF) medical device problem code.<sup>16</sup> These IMDRF codes have a hierarchical alphanumerical coding structure, including a letter (i.e. referring to the Annex A in our case) followed by numerical codes at different levels of detail.<sup>16,17</sup> Level one terms are represented by the first two digits, referring to 27 different medical device problems (Table 1). Level two and three terms are described by the digits three to four and five to six respectively, representing a more detailed description of the problem under one of the overarching 27 groups. In this study only the Level one terms were used, as they are already detailed enough to distinguish different device problems. All safety notices related to TK implants were independently classified to an IMDRF code by two researchers (LH and YR); possible discrepancies in coding were resolved by discussion. To determine interobserver variability, the Cohen's kappa ( $\kappa$ ) was calculated. Kappa values were categorised into six levels: i)  $\kappa \leq 0$  (no agreement); ii)  $\kappa = 0.01-0.20$  (none to

slight); iii)  $\kappa=0.21-0.40$  (fair); iv)  $\kappa=0.41-0.60$  (moderate); v)  $\kappa=0.61-0.80$  (substantial), and vi)  $\kappa=0.81-1.00$  (almost perfect).<sup>18</sup> Analysis was performed using Python (version 3.11.5; <https://www.python.org/downloads/release/python-3115/>).

#### *Data collection of registries reporting TK outliers*

Outlier TK implants currently on the market were identified by European registries publicly reporting on TK outliers, as found in a recent systematic review<sup>3</sup> and non-European registries as listed on the website of the AOANJRR.<sup>19</sup> All available registries' annual reports and websites were screened, and any reported TK outlier was extracted. For all extracted TK outliers, it was assessed whether they were reported in the latest annual reports and up-to-date website, representing TK implants currently on the market in these registries. If the TK outlier was not reported in the latest available registry data (i.e. not implanted in the past year in the included registries), the outlier was considered an off-market implant and excluded from further analysis.

TK outlier definitions differed between these registries (AOANJRR: "The revision rate (per 100 component years) exceeds twice that for the group and the Poisson probability of observing that number of revisions, given the rate of the group is significant ( $p\text{-value}<0.05$ )"; NJR: "having a more than twice the prosthesis time incident rate when compared to the group, allowing for confidence intervals"; SIRIS: "Revision rates of more than twice compared to the relevant group"; and the definition of an outlier was not reported for the SAR).<sup>20</sup>

For all TK outliers, the year of first identification and its cumulative revision risks (1/5/10-years), including standard errors (SE) and/or 95% confidence intervals (CI), were extracted. If only the 95% CI was provided, the SE was calculated by subtracting the upper and lower bound of the 95% CI and dividing it by 3.92.<sup>21</sup>

### *Statistics*

First, the overlap between TK implants with safety notices and TK outliers was determined by comparing the brand name reported in both safety notices and registry data. Three groups were distinguished: i) TK implants with safety notices but not identified as an outlier (“safety notices only”); ii) TK implants with both safety notices and identified as an outlier (“both”); iii) TK implants without safety notices but identified as an outlier (“outlier only”). The percentage of TK implants in each of these groups was related to the number of unique TK implants identified by both data sources.

Second, to prevent camouflage (i.e. multiple compatible construct combinations existing within one implant brand name<sup>22</sup>), the overlap between TK implants with safety notices and TK outliers across different variants/subtypes under the same brand name was analysed. We considered possible subtypes with the same brand name by: i) fixation (e.g. cemented *versus* uncemented); ii) stability (e.g. cruciate retaining *versus* hinged), and iii) mobility (e.g. fixed *versus* mobile).

Third, to explore possible reasons for not signalling the same TK implants we examined: i) differences in the frequency of IMDRF codes (Table 1) between the “safety notices only” and “both” groups, and ii) whether the “safety notices only” group had lower cumulative revision risks (and thus seemingly better performance) than the “both” group, which may explain why they were not detected as TK outliers. Random effects models were used to calculate the pooled cumulative revision risks (1/5/10-years) across all registries reporting on the specific TK implant, for the “safety notices only” and “both” groups.

The metafor package in R-statistics (version 4.1.2; R Foundation for Statistical Computing, Vienna, Austria) was used for analyses.

*Ethics, registration, data sharing, use of AI, funding, and disclosures*  
According to Dutch law, no institutional approval was required. This work was supported by the European Union's Horizon 2020 Research and Innovation 41 Programme (grant number 965246) and was part of the CORE-MD project. Complete disclosure of interest forms according to ICMJE are available on the article page, doi: 10.2340/17453674.2024.42361

## **Results**

### *TK implants with safety notices*

The CORE-MD PMS tool retrieved 104,638 safety notices from 13 Ministries of Health and regulatory agencies websites, of which 1,327 safety notices were considered relevant as they matched with a specific TK implant included in the latest annual registry reports. For the selected 1,327 safety notices, 540 safety notices were excluded because they were not related to a TK implant (i.e. associated with surgical protocols) thus resulting in 787 safety notices included for further analysis (Figure and Supplementary Table 1 – online available). These 787 safety notices were relevant to 38 unique TK implant brand names. Most safety notices originated from the USA and were associated with the Nexgen (Zimmer Biomet) (n=243, 31%) (Table 2).

### *Outlier TK implants*

Four national registries (AOANJRR, NJR, Swedish Arthroplasty Register [SAR], and SIRIS) publicly reported TK outliers while others might report them on a secure website.<sup>3</sup> After removing duplicate TK outlier brand names (i.e. the same brand name was mentioned in multiple annual reports) and off-market TK outliers, 35 unique TK outlier brand names were included for further analysis (Table 3). Supplementary Table 2 (online available) gives more detailed information on specific subtypes within a brand that were identified as outlier along with their performance.

#### *Overlap between TK outliers and TK implants with safety notices*

Combining the brand names of the 38 TK implants with a safety notice and the 35 TK outliers resulted in 47 unique TK implant brand names, of which 26 (55%) were in the “both” group, 12 (26%) in the “safety notices only” group, and 9 (19%) in the “outlier only” group (Tables 3 and 4). Thus, safety notices did not signal 9 (26%) of the 35 TK outliers and registries did not signal 12 (32%) of the 38 TK implants that had safety notices.

Considering the 26 TK implants in the “both” group, 7 (27%) TK implants did not have any information in the safety notice regarding fixation, 10 (38%) had no information regarding stability, and 15 (57%) no information regarding mobility, which would be needed to determine whether the exact same TK implant was concerned (white colour, Table 4). Focusing on fixation, 4 out of 26 (15%) TK implants could be matched to the cemented subtype and 6 (23%) to the uncemented subtype (Table 4). With regard to stability, 2 out of 26 (8%) related to the cruciate retaining, 2 (8%) to the hinged, and 8 (31%) to the posterior stabilised subtype. For mobility, one (4%) signalled the fixed, one (4%) the mobile, and five (19%) the rotating subtype. However, 14 (54%) cemented and 3 (12%) uncemented TK implants did not relate to the same fixation subtype (Table 4). Similarly, six (23%) cruciate retaining, two (8%) hinged, and seven (27%) posterior stabilised TK implants did not have the same stability and three (12%) fixed, five (19%) mobile, and two (8%) rotating TK implants did not relate to the same mobility subtype.

#### *Revision rates, timing of safety concerns, and implant problems*

For the “both” group, the median 1/5/10-year cumulative revision risks were 1.6% (range: 0.9–9.5), 6.3 (range: 3.6–23.8), and 8.2% (range: 5.6–23.8), respectively, compared with 0.6% (range: 0.3–1.1), 2.3% (range: 1.4–3.7), and 3.8% (range: 3.1–5.1), for the “safety notices only” group (Table 5).

When comparing the dates of the first issuance of safety notices with the dates when the implant was first identified as outlier by registries, no specific data source consistently published safety signals earlier (Table 5).

For the 26 TK implants in the “both” group, 728 safety notices were issued with the most frequently reported problem being related to “A02–Manufacturing, Packaging or Shipping” (43%), followed by “A23–Use of Device” (16%) (Table 6). The most frequent type of problem found was similar for the 12 TK implants in the “safety notices only” group (n=59 safety notices): “A02–Manufacturing, Packaging or Shipping” (44%) (Table 6). Focusing on differences between the two groups, safety notices related to “A05–Mechanical Problem” (6%) and “A17–Compatibility Problem” (8%), respectively, were reported only for the “both” group (Table 6) but not encountered for the “safety notices only” group (Table 6). The interobserver agreement to classify safety notices according to the IMDRF codes among the two observers was substantial ( $\kappa=0.79$ ; CI 0.76–0.82).

## **Discussion**

Our study is the first to assess the extent of overlap in TK implants for which safety notices were issued and that were identified as outliers in registry data. We aimed to assess the extent to which safety notices and outlier identification in registries signal the same or different TK implants. We found that approximately half (55%) of the TK implants were identified by both safety notices and registries outlier identification procedures, but a quarter of the TK outliers did not have any publicly released safety notices on the websites of Ministries of Health or regulatory agencies. In addition, there were implant problems identified by safety notices that did not manifest in an outlier status. TK implants with both safety notices and an outlier status had higher cumulative revision risks (1/5/10-years) than TK implants with safety notices only.

A recent review that assessed the current state of medical device safety signal detection stated that a global dataset of medical devices should be created using automatic reports from national/regional databases.<sup>23</sup> In the absence of such a global dataset, the CORE-MD PMS tool was developed recently.<sup>7</sup> Our results add that such a global dataset of safety notices may still not identify a quarter of TK implants with statistically significant poor performance (i.e. TK outliers). Additionally, a published safety notice by itself does not constitute a sufficient and necessary condition for being identified as a TK outlier (the “safety notices only” group). We identified that certain IMDRF codes “A05–Mechanical Problem” and “A17–Compatibility Problem” were not encountered in the “safety notices only” group, suggesting that these are more closely related to poorer implant performance. This observation could result in a helpful indication to highlight a higher risk for certain TK implants with such IMDRF codes identified in safety notices to become an outlier, thus warranting closer scrutiny of these TK implants.

This multi-registry analysis examined the content of safety notice text, which does not typically include information needed to identify specific variants/subtypes of TK implants, characterised by fixation, stability, and mobility. Such a lack of information causes camouflage (i.e. multiple implant subtypes exist under the same implant brand name)<sup>22</sup> making it difficult or even impossible to link the correct TK implants with safety notices to registry data, or to combine data from different real-world data sources. This information is, however, important for action to be taken, as illustrated by a recent study showing good performance for the Nexgen system but significantly higher revision risks for specific combinations with the Nexgen LPS Flex (Supplementary Table 2 – online available).<sup>24</sup> In addition, registries often publicly report only TK implants’ brand names without listing more detailed information (e.g. fixation, stabilization, and mobility) to identify which specific subtype of an implant is concerned. Product codes and unique device identifiers (UDIs), which would be needed to deal with such

camouflage, were also not reported in safety notices or publicly by registries, except for the American medical device recall database. Accordingly, we highly recommend minimal reporting requirements for manufacturers with respect to safety notices and also for registries when reporting outliers, including: full brand name, fixation, mobility, stability, and product codes or UDIs.

Arthroplasty registries currently only identify TK outliers based on revision risks, which may take several years (at least 1) before sufficient numbers are available to detect performance problems.<sup>3,4</sup> Using revision risk may seem a relatively straightforward endpoint (the occurrence of revision surgery), but surgeon, implant, and patient factors determine whether an implant is revised. Moreover, between-registry variation exists regarding definitions and reasons for revision<sup>3,25</sup> although all included registries identifying TK outliers defined revision as “the replacement/removal/addition of one or more prosthetic components”. But, for instance, revisions due to infection are excluded from the all-cause revision risk in the Swedish registry.<sup>26,27</sup> In contrast, the NJR also includes revision due to infection if no prosthetic component was exchanged, which can result in specific TK implants being identified as an outlier in the NJR but not in other registries. Interestingly, the number of TK outliers publicly reported by the AOANJRR is much higher when compared with other registries publicly reporting on outliers. Part of the explanation may be related to the definition, such as the minimum number of implants required for the publication and analysis of implant-specific revision rates, which is much lower in the AOANJRR (500 procedures compared with 2,500 procedures required in the NJR). These heterogeneities highlight the importance of an international agreement on definitions and outcomes, as well as time-points and methodology used for measuring outcomes within registries.

Some safety notices may be released based on implant related problems causing clinical performance issues relating to a

specific TK implant but also on a case-by-case analysis (i.e. no minimum number of implants at risk is required), meaning that safety notices may provide an earlier signal of a possible performance problem than registries.<sup>28</sup> Accordingly, registries could use this as a signal to analyse specific TK implants with released safety notices to detect potential adverse trends in performance earlier. However, when considering the timing of safety notices and outlier data being published, none of the data sources consistently released safety signals earlier than any other, highlighting the importance of a multifaced approach combining these two data sources. While this provides relevant information and includes all TK implants for which safety concerns were reported in safety notices or reported as an outlier, it does not answer the question as to what percentage of TK implants did not have any safety concerns reported. It would seem rather infeasible to estimate this percentage based on all TK implants currently on the market in all countries examined in the present study. Creating a random sample of TK implants would be a more feasible alternative to provide such information as a next step.

### *Limitations*


First, the CORE-MD PMS tool searched for safety notices published on the websites of Ministries of Health and regulatory agencies, but we may have missed safety notices if these were reported only on manufacturers' websites, which would have underestimated the number of TK outliers with safety notices. Second, both TK outliers and TK implants not identified as a TK outlier had a relatively similar distribution of IMDRF-problem types, suggesting that the IMDRF code may not be sufficient to distinguish between these two groups. However, only the Level one IMDRF codes were used due to the large number of safety notices to be manually classified, so there may be differences in distribution when Level two or three problem terms were used. On the other hand, one could argue that such differences in these more detailed problem-type descriptions would not likely entail clinically relevant differences in problems. Third, other factors

such as surgeon or hospital performance are known to influence revisions, which may skew the revision risks data. Nonetheless, as we used data from four national registries consisting of a large number of TK outliers, the impact on our results is likely to have been small. Fourth, safety notices were collected from websites in more countries than those for which registry outlier identification data were available, which might have underestimated the number of TK outliers and explain part of the “safety notices only” group. On the other hand, assuming that safety notices point to a problem with the implant itself, we would expect any performance issue to be similar across countries and thereby picked up by other registries as well. Finally, our analysis does not exclude possible duplicates of the same safety notices published in different countries or for different models/lots within the country. This is because different countries use diverse formats and criteria to issue safety notices: some countries issue separate safety notices for each model (e.g. the USA, resulting in a high number of safety notices from the USA), while others publish only one safety notice with multiple models. However, the safety notices would still signal the same TK implant, which was used as the unit of analysis in the present study, so excluding duplicate safety notices would not have changed our results.

## **Conclusion**

We found that approximately half (55%) of the TK implants were identified by both safety notices and registries outlier identification procedures, whereas around 25% of TK outliers are not the subject of publicly released safety notices, highlighting the potential of adopting a multifaceted approach, integrating various real-world data sources and methods to combine information to enhance medical device safety signal detection.

## Figures and Tables



**Figure 1.** Flowchart showing the selection process of TK implants with safety notices

| <b>IMDRF code</b>                                  | <b>IMDRF description of medical device problem</b>                                                                                                                                                                                                             |
|----------------------------------------------------|----------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|
| A01 – Patient Device Interaction Problem           | Problem related to the interaction between the patient and the device.                                                                                                                                                                                         |
| A02 – Manufacturing, Packaging or Shipping Problem | Problem associated with any deviations from the documented specifications of the device that relate to nonconformity during manufacture to the design of an item or to <u>specified manufacturing, packaging, or shipping processes (out of box problem)</u> . |
| A03 – Chemical Problem                             | Problem associated with any deviations from the documented specifications of the device that relate to any chemical characterization, i.e. element, compound, or mixture.                                                                                      |
| A04 – Material Integrity Problem                   | Problem associated with any deviations from the documented specifications of the device that relate to the <u>limited durability</u> of all material used to construct device.                                                                                 |
| A05 – Mechanical Problem                           | Problems associated with mechanical actions or defects, including moving parts or subassemblies, etc.                                                                                                                                                          |
| A06 – Optical Problem                              | Problem associated with transmission of visible light affecting the quality of the image transmitted or otherwise affecting the intended application of the visible light path.                                                                                |
| A07 – Electrical /Electronic Property Problem      | Problem associated with the function of the electrical circuitry of the device.                                                                                                                                                                                |
| A08 – Calibration Problem                          | Problem associated with the operation of the device, related to its accuracy, and associated with the calibration of the device.                                                                                                                               |
| A09 – Output Problem                               | Problem associated with any deviation from the documented specifications of the device that relate to the end result, data, or test results provided by the device.                                                                                            |
| A10 – Temperature Problem                          | Problem associated with the device producing unintended temperatures.                                                                                                                                                                                          |
| A11 – Computer Software Problem                    | Problem associated with written programs, codes, and/or software system that affects <u>device performance or communication with another device</u> .                                                                                                          |
| A12 – Connection Problem                           | Problem associated with linking of the device and/or the functional units set up to provide means for a transfer of liquid, gas, electricity, or data.                                                                                                         |

|                                                     |                                                                                                                                                                                                                                                                           |
|-----------------------------------------------------|---------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|
| A13 – Communication or Transmission Problem         | Problem associated with the device sending or receiving signals or data. This includes transmission among internal components of the device to which the device is intended to communicate.                                                                               |
| A14 – Infusion or Flow Problem                      | Problem associated with the device failing to deliver or draw liquids or gases as intended (e.g. delivering drugs at incorrect rate, problems with drawing fluid from a system). This includes vacuum collection devices and manual or mechanical pumps.                  |
| A15 – Activation, Positioning or Separation Problem | Problem associated with any deviations from the documented specifications of the device that relate to the sequence of events for activation, positioning or separation of device. Note: Deployment is synonymous with activation.                                        |
| A16 – Protective Measures Problem                   | Problem associated with any deviations from the documented specifications of the device that relate to the implemented and inherited design features specific to devices used for reducing risks to patient or caregiver or maintaining risks within specified levels.    |
| A17 – Compatibility Problem                         | Problem associated with compatibility between device, patients, or substances (medication, body fluid, etc.).                                                                                                                                                             |
| A18 – Contamination /Decontamination Problem        | Problem associated with the presence of any unexpected foreign substance found in the device, on its surface or in the package materials, which may affect performance or intended use of the device, or problem that compromise effective decontamination of the device. |
| A19 – Environmental Compatibility Problem           | Problem associated with the surrounding conditions in which the device is being used such as temperature, noise, lighting, ventilation, or other external factors such as power supply.                                                                                   |
| A20 – Installation-Related Problem                  | Problem associated with unsatisfactory installation, configuration, and/or setup of a specific device.                                                                                                                                                                    |

|                                                              |                                                                                                                                                                                                                                                                                                                         |
|--------------------------------------------------------------|-------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|
| A21 – Labelling, Instructions for Use or Training Problem    | Problem associated with device markings/labelling, instructions for use, training and maintenance documentation or guidelines.                                                                                                                                                                                          |
| A22 – Human–Device Interface Problem                         | Problem associated with an act or omission of an act that has a different result than that intended by the manufacturer or expected by the operator.                                                                                                                                                                    |
| A23 – Use of Device Problem                                  | Problem associated with failure to process, service, or operate the device according to the manufacturer’s recommendations or recognised best practices.                                                                                                                                                                |
| A24 – Adverse Event Without Identified Device or Use Problem | An adverse event (e.g. patient harm) appears to have occurred, but there does not appear to have been a problem with the device or the way it was used.                                                                                                                                                                 |
| A25 – No Apparent Adverse Event                              | A report has been received but the description provided does not appear to relate to an adverse event. This code allows a report to be recorded for administration purposes, even if it does not meet the requirements for adverse event reporting.                                                                     |
| A26 – Insufficient Information                               | An adverse event appears to have occurred but there is not yet enough information available to classify the device problem.                                                                                                                                                                                             |
| A27 – Appropriate Term/Code Not Available                    | The device problem is not adequately described by any other term. Note: this code must not be used unless there is no other feasible code. The preferred term should be documented when submitting an adverse event report. This information will be used to determine if a new term should be added to the code table. |

**Table 1.** The 27 International Medical Device Regulators Forum (IMDRF) medical device problem codes and relevant description<sup>15</sup>

| Implant      | Au | Cz | Dk | Fr | Ge | Gr | Ir | It | Ne | Sp | Sw | USA |
|--------------|----|----|----|----|----|----|----|----|----|----|----|-----|
| Active Knee  | 1  |    |    |    |    |    |    |    |    |    |    |     |
| Advance      |    |    |    | 1  |    |    |    | 1  |    |    |    | 11  |
| AGC Anatomic | 2  | 1  |    | 1  |    |    |    | 1  |    |    |    | 2   |
| Attune       |    |    | 1  | 1  | 3  |    | 2  | 2  | 1  |    |    | 20  |
| Balansys     | 1  | 1  |    | 1  | 3  |    |    | 2  |    |    |    |     |
| Columbus     | 1  |    |    |    | 1  |    |    |    |    |    |    |     |
| Duracon      | 1  |    | 1  |    |    |    |    | 1  |    |    |    | 3   |
| EFK          |    |    |    |    | 1  |    |    |    |    |    |    |     |
| Endo-Model   |    |    | 1  | 1  | 4  |    | 1  | 2  |    | 2  | 1  | 4   |
| Evolution    |    |    |    |    | 2  |    |    | 1  |    |    |    | 3   |
| Gemini       |    |    |    |    |    |    | 1  |    |    | 1  |    |     |
| Genesis      | 2  | 1  |    | 1  | 11 | 1  |    | 3  | 1  | 2  |    | 16  |
| GMK          |    |    |    |    | 1  |    | 2  | 3  |    |    |    | 3   |
| Innex        |    | 1  |    |    | 7  |    | 1  | 2  | 1  | 1  |    |     |
| iTotal       |    |    |    |    | 1  |    |    |    |    |    |    | 8   |
| Journey      | 3  |    | 1  | 1  | 2  |    |    | 1  | 1  | 1  |    | 4   |
| Kinemax      |    |    |    |    |    |    |    |    |    |    |    | 1   |
| K-mod        |    |    |    |    |    |    |    |    | 1  |    |    |     |
| LCS          |    |    |    |    | 4  |    | 2  | 3  |    |    |    | 32  |
| Legion       |    |    |    |    | 6  |    |    | 1  | 1  |    |    | 23  |
| MET'S Smiles |    |    |    |    |    |    |    |    |    |    |    | 1   |
| MRK          |    |    |    |    |    |    |    |    | 1  |    |    |     |
| Multigen     |    | 1  |    | 1  |    |    |    | 1  |    | 1  |    |     |
| Mutars       | 1  |    |    |    | 1  |    |    |    |    |    |    |     |
| Natural-knee |    |    |    |    |    |    |    |    |    |    |    | 1   |
| Nexgen       | 2  |    | 5  | 1  | 14 |    | 8  | 5  | 1  |    | 1  | 206 |
| Noiles       |    |    | 1  |    | 1  |    | 2  | 1  |    |    |    | 6   |
| Optetrak     | 2  |    |    |    |    | 1  |    |    | 1  |    |    | 47  |
| Persona      | 2  |    | 1  |    | 5  |    |    | 8  |    |    |    | 24  |
| PFC Sigma    |    |    |    |    | 5  |    | 2  | 2  |    |    |    | 21  |
| Physica      |    |    |    |    |    |    |    | 2  |    |    |    |     |
| Saiph        | 1  |    |    |    |    |    |    |    |    |    |    |     |
| Score        | 1  |    |    |    | 1  |    |    |    |    |    |    |     |
| Scorpio      |    | 1  |    |    | 4  |    | 4  | 2  | 1  |    |    | 23  |
| TC-plus      |    |    |    |    | 1  |    |    |    |    |    |    |     |
| Triathlon    | 6  |    |    |    | 5  |    | 2  | 6  |    | 1  |    | 32  |
| Unity        |    |    |    |    | 1  |    |    |    |    |    |    | 4   |
| Vanguard     | 3  | 1  |    |    | 3  | 1  |    | 1  | 3  |    |    | 45  |

**Table 2.** TK implants with the number of safety notices by country

| Outlier TK implant | Outlier reported         | TK implants implanted     | Identified by the CORE-MD PMS tool (number of safety notices) |
|--------------------|--------------------------|---------------------------|---------------------------------------------------------------|
| ACS                | AOANJRR, NJR             | 2,900                     | No                                                            |
| Active Knee        | AOANJRR                  | 7,215                     | Yes (1)                                                       |
| Advance            | AOANJRR                  | 1,009                     | Yes (12)                                                      |
| AGC Anatomic       | SAR                      |                           | Yes (7)                                                       |
| Apex Knee          | AOANJRR                  | 513                       | No                                                            |
| Attune             | AOANJRR                  | 854                       | Yes (30)                                                      |
| Columbus           | AOANJRR                  | 6,334                     | Yes (2)                                                       |
| Duracon            | SAR                      |                           | Yes (6)                                                       |
| E.Motion           | AOANJRR, NJR, SIRIS      | AOANJRR: 1,014; NJR: 339  | No                                                            |
| Endo-Model         | NJR                      | 309                       | Yes (16)                                                      |
| Gemini             | AOANJRR                  | 21                        | Yes (2)                                                       |
| Genesis            | AOANJRR, NJR, SAR        | AOANJRR: 826; NJR 9,190   | Yes (38)                                                      |
| Journey            | AOANJRR, NJR, SAR, SIRIS | AOANJRR: 3,033; NJR 1,714 | Yes (14)                                                      |
| Kinemax            | SAR                      |                           | Yes (1)                                                       |
| LCS                | AOANJRR, NJR             | 5,729                     | Yes (41)                                                      |
| Legion             | AOANJRR, SAR             | AOANJRR: 1,017            | Yes (31)                                                      |
| Maxim              | AOANJRR                  | 413                       | No                                                            |
| METS Smiles        | NJR                      | 954                       | Yes (1)                                                       |
| Miller-Galante     | SAR                      |                           | No                                                            |
| Mutars             | AOANJRR                  | 357                       | Yes (2)                                                       |
| Nexgen             | AOANJRR, SAR             | AOANJRR: 316              | Yes (30)                                                      |
| Noiles             | NJR                      | 594                       | Yes (11)                                                      |
| Optetrak           | AOANJRR, NJR             | 4,098                     | Yes (51)                                                      |
| Persona            | SAR                      |                           | Yes (40)                                                      |
| PFC Sigma          | AOANJRR, SAR             | AOANJRR: 316              | Yes (30)                                                      |
| Physica            | SIRIS                    |                           | Yes (2)                                                       |
| Profix             | AOANJRR, SAR             | AOANJRR: 1,895            | No                                                            |
| Rotaglide Plus     | AOANJRR                  | 631                       | No                                                            |
| Score              | AOANJRR                  | 4,686                     | Yes (2)                                                       |
| Scorpio            | AOANJRR                  | 1,172                     | Yes (35)                                                      |
| TC-plus            | AOANJRR                  | 63                        | Yes (1)                                                       |
| Trekking           | AOANJRR                  | 1,263                     | No                                                            |
| Triathlon          | SAR                      |                           | Yes (52)                                                      |
| Vanguard           | AOANJRR, SAR             | AOANJRR: 6,225            | Yes (57)                                                      |

TK = total knee; CORE-MD = Coordinating Research and Evidence for Medical Devices; PMS = post-market surveillance; AOANJRR = Australian Orthopaedic Association National Joint Replacement Registry, SIRIS = Swiss National Hip & Knee Joint Registry, SAR = Swedish Arthroplasty Register, NJR = National Joint Registry for England, Wales, Northern Ireland, the Isle of Man and Guernsey.

**Table 3.** Outlier TK implants currently used on the market

| Implant name                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                        | Fixation Cemented |     | Stability          |        |                      | Mobility |        |          |
|-----------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|-------------------|-----|--------------------|--------|----------------------|----------|--------|----------|
|                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                     | Yes               | No  | Cruciate retaining | Hinged | Posterior Stabilised | Fixed    | Mobile | Rotating |
| Active Knee                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                         | No                | Yes |                    |        |                      |          |        |          |
| Advance                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                             |                   |     |                    |        |                      |          |        |          |
| AGC Anatomic                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                        | No                |     |                    |        | No                   | No       |        |          |
| Attune                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                              | Yes               | Yes | No                 |        | Yes                  | No       |        | Yes      |
| Columbus                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                            |                   | Yes |                    |        |                      |          |        |          |
| Duracon                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                             | No                |     |                    |        | No                   |          |        |          |
| Endo-Model                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                          | No                | No  |                    | Yes    | No                   |          | No     | Yes      |
| Gemini                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                              |                   |     |                    |        |                      |          |        | Yes      |
| Genesis                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                             | Yes               | Yes | Yes                |        | Yes                  |          | No     |          |
| Journey                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                             | No                |     |                    |        |                      |          |        |          |
| Kinemax                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                             | No                |     |                    |        | No                   |          |        | No       |
| LCS                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                 | No                |     |                    |        | Yes                  | No       | No     | Yes      |
| Legion                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                              | No                |     | Yes                | No     | Yes                  |          |        |          |
| METS Smiles                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                         |                   |     |                    |        |                      |          |        |          |
| Mutars                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                              |                   |     |                    |        |                      |          |        |          |
| Nexgen                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                              | No                | Yes | No                 | No     | Yes                  | Yes      | No     | Yes      |
| Noiles                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                              | No                |     |                    | Yes    |                      |          |        | No       |
| Optetrak                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                            | Yes               |     |                    |        | Yes                  |          |        |          |
| Persona                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                             | No                | No  | No                 |        | No                   |          |        |          |
| PFC Sigma                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                           | Yes               | Yes |                    |        | Yes                  |          |        |          |
| Physica                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                             |                   |     |                    |        | No                   |          |        |          |
| Score                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                               | No                | No  |                    |        |                      |          | No     |          |
| Scorpio                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                             | No                |     | No                 |        | Yes                  |          |        |          |
| TC-plus                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                             |                   |     |                    |        |                      |          |        |          |
| Triathlon                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                           | No                |     | No                 |        | No                   |          |        |          |
| Vanguard                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                            |                   |     |                    |        |                      |          |        |          |
| <i>'' = total knee implants without any information reported in the safety notice on specific fixation/stability/mobility methods. Yes = total knee implants with information in the safety notice about its fixation, stability, or mobility, thus overlapping with an outlier implant based on its implant characteristic (fixation/stability/mobility). No = total knee implants with information in the safety notice about its fixation, stability, or mobility method but without overlapping with an outlier implant based on its implant characteristics (fixation/stability/mobility).</i> |                   |     |                    |        |                      |          |        |          |

**Table 4.** Overlap of total knee implants in the “both” group based on fixation, mobility, and stability

| Implant      | Date of first safety notice | Identified as outlier | Year first identified as outlier and registry reporting | Pooled cumulative revision risk (CI) for specified implant brand name |                               |                              |
|--------------|-----------------------------|-----------------------|---------------------------------------------------------|-----------------------------------------------------------------------|-------------------------------|------------------------------|
|              |                             |                       |                                                         | 1-year                                                                | 5-year                        | 10-year                      |
| Active Knee  | 21/10/2016                  | Yes                   | 2016 (AOANJRR)                                          | 1.1 (0.9–1.4)*                                                        | 5.0 (4.6–5.6)*                | 8.8 (8.1–9.5)*               |
| Advance      | 11/07/2016                  | Yes                   | 2013 (AOANJRR)                                          | 2.0 (1.3–3.1)*                                                        | 6.4 (5.0–8.2)*                | 8.1 (6.4–10.2)*              |
| AGC Anatomic | 21/07/2015                  | Yes                   | 2014 (SAR)                                              |                                                                       |                               |                              |
| Attune       | 29/06/2015                  | Yes                   | 2023 (AOANJRR)                                          | 1.8 (1.0–3.0)*                                                        |                               |                              |
| Balansys     | 29/01/2014                  | No                    |                                                         | 0.9 (0.5–1.2)* <sup>§¶¶</sup>                                         | 3.1 (2.3–3.9)* <sup>§¶¶</sup> | 5.1 (2.2–8.1)* <sup>§</sup>  |
| Columbus     | 17/01/2008                  | Yes                   | 2009 (AOANJRR)                                          | 1.2 (0.9–1.5)*                                                        | 4.4 (3.7–5.3)*                | 7.3 (6.0–8.8)*               |
| Duracon      | 20/09/2007                  | Yes                   | 2004 (SAR)                                              |                                                                       |                               |                              |
| EFK          | 15/04/2014                  | No                    |                                                         | 0.6 (0.1–1.2) <sup>¶</sup>                                            | 1.7 (0.5–3.0) <sup>¶</sup>    |                              |
| Endo-Model   | 16/04/2012                  | Yes                   | 2019 (NJR)                                              | 1.3 (0.8–2.2) <sup>†</sup>                                            | 4.8 (3.7–6.3) <sup>†</sup>    | 7.0 (5.3–9.2) <sup>†</sup>   |
| Evolution    | 17/02/2015                  | No                    |                                                         | 0.7 (0.3–1.1)* <sup>¶¶£</sup>                                         | 2.8 (2.1–3.5)* <sup>¶£</sup>  |                              |
| Gemini       | 07/09/2010                  | Yes                   | 2007 (AOANJRR)                                          | 9.5 (2.5–33.0)*                                                       | 23.8 (10.7–48.1)*             | 23.8 (10.7–48.1)*            |
| Genesis      | 09/05/2006                  | Yes                   | 2004 (AOANJRR), 2018 (SAR), 2021 (NJR)                  | 1.0 (0.7–1.3)* <sup>†</sup>                                           | 3.6 (3.2–4.1)* <sup>†</sup>   | 5.6 (4.8–6.3)* <sup>†</sup>  |
| GMK Sphere   | 03/07/2017                  | No                    |                                                         | 1.1 (0.9–1.4)* <sup>†¶¶£</sup>                                        | 3.7 (2.9–4.5)* <sup>†¶£</sup> | 4.3 (2.4–6.1)*               |
| Innex        | 25/07/2005                  | No                    |                                                         | 0.9 (0.5–1.3) <sup>§¶¶</sup>                                          | 2.8 (2.0–3.6) <sup>§¶¶</sup>  | 3.5 (2.4–4.6) <sup>§</sup>   |
| iTotal       | 23/07/2012                  | No                    |                                                         | 0.4 (0.2–0.9) <sup>¶</sup>                                            | 3.5 (2.5–5.0) <sup>¶</sup>    |                              |
| Journey      | 03/01/2014                  | Yes                   | 2009 (AOANJRR), 2018 (SAR), 2019 (SIRIS), 2014 (NJR)    | 1.6 (0.1–3.1)* <sup>†¶</sup>                                          | 6.3 (1.8–10.8)* <sup>†¶</sup> | 11.0 (9.9–12.2)*             |
| Kinemax      | 14/05/2015                  | Yes                   | 2006 (SAR)                                              |                                                                       |                               |                              |
| K-mod        | 19/05/2014                  | No                    |                                                         |                                                                       |                               |                              |
| LCS          | 02/12/2005                  | Yes                   | 2012 (AOANJRR), 2021 (NJR)                              | 0.9 (0.2–1.6)* <sup>†</sup>                                           | 5.6 (1.8–9.5)* <sup>†</sup>   | 7.7 (2.5–12.8)* <sup>†</sup> |
| Legion       | 22/08/2009                  | Yes                   | 2017 (AOANJRR), 2019 (SAR)                              | 3.3 (2.3–4.6)*                                                        | 6.3 (4.8–8.3)*                | 9.9 (7.5–13.0)*              |
| METS Smiles  | 17/08/2016                  | Yes                   | 2018 (NJR)                                              |                                                                       |                               |                              |
| MRK          | 31/12/2021                  | No                    |                                                         | 0.3 (0.0–0.6)* <sup>†§</sup>                                          | 1.8 (1.2–2.3)* <sup>†§</sup>  | 3.1 (1.6–4.6)* <sup>†</sup>  |
| Multigen     | 12/05/2021                  | No                    |                                                         |                                                                       |                               |                              |
| Mutars       | 03/04/2013                  | Yes                   | 2023 (AOANJRR)                                          | 6.5 (4.2–9.9)*                                                        |                               |                              |
| Natural-knee | 07/11/2019                  | No                    |                                                         | 0.4 (0.2–0.7)* <sup>§¶£</sup>                                         | 1.7 (1.2–2.1)* <sup>§¶£</sup> | 3.2 (2.4–3.9)* <sup>§</sup>  |

|           |            |     |                            |                             |                            |                  |
|-----------|------------|-----|----------------------------|-----------------------------|----------------------------|------------------|
| Nexgen    | 13/09/2004 | Yes | 2018 (AOANJRR), 2015 (SAR) | 2.4 (1.9-3.2)*              | 5.0 (4.2-6.1)*             | 6.9 (5.1-9.2)*   |
| Noiles    | 02/03/2014 | Yes | 2018 (NJR)                 |                             |                            |                  |
| Optetrak  | 01/06/2006 | Yes | 2007 (AOANJRR)             | 1.0 (0.0-2.1)*              | 10.3 (4.1-16.4)*           | 13.7 (7.0-20.4)* |
| Persona   | 21/11/2012 | Yes | 2021 (SAR)                 |                             |                            |                  |
| PFC Sigma | 02/12/2005 | Yes | 2018 (AOANJRR), 2012 (SAR) | 2.2 (1.1-4.6)*              | 7.1 (4.7-10.5)*            | 7.4 (5.0-10.9)*  |
| Physica   | 18/04/2019 | Yes | 2019 (SIRIS)               | 1.7 (1.3-2.3) <sup>y</sup>  | 6.8 (5.9-7.9) <sup>y</sup> |                  |
| Saiph     | 25/03/2022 | No  |                            | 0.6 (0.3-1.0) <sup>†</sup>  | 1.4 (0.9-2.0) <sup>†</sup> |                  |
| Score     | 04/10/2019 | Yes | 2013 (AOANJRR)             | 1.5 (0.8-2.2)*              | 6.5 (5.5-7.6)*             | 11.1 (9.3-12.8)* |
| Scorpio   | 26/08/2005 | Yes | 2014 (AOANJRR)             | 1.2 (0.7-2.0)*              | 6.1 (4.9-7.7)*             | 7.4 (6.0-9.2)*   |
| TC-plus   | 10/06/2008 | Yes | 2008 (AOANJRR)             | 1.6 (0.2-10.7)*             | 8.4 (3.6-19.1)*            | 14.4 (7.4-26.9)* |
| Triathlon | 07/02/2007 | Yes | 2021 (SAR)                 |                             |                            |                  |
| Unity     | 30/09/2021 | No  |                            | 0.4 (0.1-0.9)* <sup>¶</sup> | 1.5 (0.7-2.3) <sup>¶</sup> |                  |
| Vanguard  | 17/11/2016 | Yes | 2012 (AOANJRR), 2009 (SAR) | 1.9 (1.2-2.6)*              | 5.9 (4.7-7.1)*             | 8.2 (6.8-9.5)*   |

\* = based on revision risks (RR) as reported by the AOANJRR; † = based on RR as reported by the NJR; ‡ = based on RR as reported by the RIPO; § = based on RR as reported by the LROI; ¶ = based on RR as reported by the SIRIS; ¶ = based on RR as reported by the EPRD; £ = based on RR as reported by the AJRR. TK = total knee; CI = confidence intervals; AOANJRR = Australian Orthopaedic Association National Joint Replacement Registry; SIRIS = Swiss National Hip & Knee Joint Registry; SAR = Swedish Arthroplasty Register; NJR = National Joint Registry for England; Wales, Northern Ireland, the Isle of Man and Guernsey.

**Table 5.** Total knee implants brand names with at least one safety notice

| IMDRF code | TK implants (n=26) in the "both" group | TK implants (n=12) in the "safety notices only" group |
|------------|----------------------------------------|-------------------------------------------------------|
| A01        |                                        | 4 (6.7)                                               |
| A02        | 313 (43)                               | 26 (44)                                               |
| A04        | 56 (7.7)                               | 7 (12)                                                |
| A05        | 41 (5.6)                               |                                                       |
| A09        | 6 (0.8)                                |                                                       |
| A17        | 59 (8.1)                               |                                                       |
| A18        | 9 (1.2)                                | 1 (1.7)                                               |
| A20        | 2 (0.3)                                |                                                       |
| A21        | 70 (10)                                | 11 (19)                                               |
| A23        | 113 (16)                               | 6 (10)                                                |
| A24        | 34 (4.7)                               | 1 (1.7)                                               |
| A26        | 25 (3.4)                               | 3 (5.1)                                               |
| Total      | 728                                    | 59                                                    |

**Table 6.** IMDRF medical device problem codes described in safety notices.  
Values are count (%)

## References

1. Medical Device Regulation (MDR). MDR – Article 83 – Post-market surveillance system of the manufacturer.
2. Medical Device Regulation (MDR). MDR – Article 87 – Reporting of serious incidents and field safety corrective actions.
3. Hoogervorst LA, Geurkink TH, Lübbecke A, et al. Quality and utility of European cardiovascular and orthopaedic registries for the regulatory evaluation of medical device safety and performance across the implant lifecycle: a systematic review. *Int J Health Policy Manag* 2023; 12: 7648
4. de Steiger RN, Miller LN, Davidson DC, et al. Joint registry approach for identification of outlier prostheses. *Acta Orthop* 2013; 84(4): 348-52
5. Puijk R, Sierevelt IN, Pijls BGCW, et al. Increased risk of aseptic loosening for posterior stabilized compared with posterior cruciate-retaining uncemented total knee replacements: a cohort study of 13,667 knees from the Dutch Arthroplasty Registry. *Acta Orthop* 2023; 94: 600-6
6. Enhancing the QUALity and Transparency Of health Research. STARD 2015: an updated list of essential items for reporting diagnostic accuracy studies. Available: <https://www.equator-network.org/wp-content/uploads/2015/03/STARD-2015-checklist.pdf> (last accessed November 5, 2024)
7. Ren Y, Bertoldi M, Fraser AG, et al. Validation of CORE-MD PMS support tool: a novel strategy for aggregating information from notices of failures to support medical devices' post-market surveillance. *Ther Innov Regul Sci* 2023; 57(3): 589-602
8. American Joint Replacement Registry (AJRR). The American Joint Replacement Registry Annual Report 2023. Available: <https://www.aaos.org/registries/publications/ajrr-annual-report/> (last accessed November 6, 2024)
9. Australian Orthopaedic Association National Joint Replacement Registry (AOANJRR). Other Registries Worldwide. Available: <https://aoanjrr.sahmri.com/registries> (last accessed November 5, 2024)
10. Dutch Arthroplasty Register (LROI). Online LROI annual report 2023. Available: <https://www.lroi-report.nl/app/uploads/2023/10/PDF-LROI-annual-report-2023-1.pdf> (last accessed November 5, 2024)
11. Regional Register of Orthopaedic Prosthetic Implantology (RIPO) overall data hip, knee and shoulder arthroplasty in Emilia-Romagna region (Italy) 2000–2020. Available: <https://ripo.cineca.it/authzssl> (last accessed November 5, 2024)
12. Endoprothesenregister Deutschland (EPRD). Annual Report 2022. Available: <https://www.eprd.de/en/downloads/reports> (last accessed November 5, 2024)
13. Schweizerisches implantatregister registre suisse des implants (SIRIS). Swiss National Hip & Knee Joint Registry Report 2023. Available: [https://www.swiss-medtech.ch/sites/default/files/2023-11/231117\\_SIRIS-Report-2023\\_Final\\_online.pdf](https://www.swiss-medtech.ch/sites/default/files/2023-11/231117_SIRIS-Report-2023_Final_online.pdf) (last accessed November 5, 2024)
14. National Joint Registry (NJR). 20th Annual Report. Available: <https://reports.njrcentre.org.uk/> (last accessed November 5, 2024)
15. Finnish Arthroplasty register (FAR). Available:

- <https://www.thl.fi/far/#index>. (data until March 2021)
16. International Medical Device Regulators Forum (IMDRF). IMDRF terminologies for categorized adverse event reporting (AER): terms, terminology structure and codes. Available: <https://www.imdrf.org/sites/default/files/docs/imdrf/final/technical/imdrf-tech-200318-ae-terminologies-n43.pdf> (last accessed February 23, 2024)
17. International Medical Device Regulators Forum (IMDRF). Annex A: Medical device problem. Available: <https://www.imdrf.org/work-ing-groups/adverse-event-terminology/annex-medical-device-problem> (last accessed February 23, 2024)
18. McHugh ML. Interrater reliability: the kappa statistic. *Biochem Med (Zagreb)* 2012; 22(3): 276-82. PMID: 23092060
19. Australian Orthopaedic Association National Joint Replacement Registry (AOANJRR). Other Registries Worldwide. Available: <https://aoanjrr.sahmri.com/registries> (last accessed February 23, 2024)
20. de Steiger RN, Hallstrom BR, Lübbeke A, et al. Identification of implant outliers in joint replacement registries. *EFORT Open Rev* 2023; 8(1): 11-17
21. Hozo SP, Djulbegovic B, Hozo I. Estimating the mean and variance from the median, range, and the size of a sample. *BMC Med Res Meth-odol* 2005; 5: 13
22. Phillips JRA, Tucker K. Implant brand portfolios, the potential for camouflage of data, and the role of the Orthopaedic Data Evaluation Panel in total knee arthroplasty. *Bone Joint J* 2021; 103-b(10): 1555-60
23. Pane J, Verhamme KMC, Villegas D, et al. Challenges associated with the safety signal detection process for medical devices. *Med Devices (Auckl)* 2021; 14: 43-57
24. Wilton T, Skinner JA, Haddad FS. Camouflage uncovered: what should happen next? *Bone Joint J* 2023; 105-b(3): 221-6
25. van Schie P, Hasan S, van Bodegom-Vos L, et al. International comparison of variation in performance between hospitals for THA and TKA: is it even possible? A systematic review including 33 studies and 8 arthroplasty register reports. *EFORT Open Rev* 2022; 7(4): 247-63
26. Hoogervorst LA, van Tilburg MM, Lübbeke A, et al. Validating Orthopaedic Data Evaluation Panel (ODEP) ratings across 9 orthopaedic registries: total hip implants with an ODEP rating perform better than those without an ODEP rating. *J Bone Joint Surg Am* 2024; 106(17): 1583-93
27. Robertsson O, Lidgren L, Sundberg M, et al. The Swedish Knee Arthroplasty Register: annual report 2020. Available: [https://www.myknee.se/pdf/SVK\\_2020\\_Eng\\_1.0.pdf](https://www.myknee.se/pdf/SVK_2020_Eng_1.0.pdf) (last accessed November 5, 2024)
28. Medicines & Healthcare products Regulatory Agency (MHRA). First generation JOURNEY BCS Knee System: higher than expected risk of revision. Available: <https://www.cas.mhra.gov.uk/ViewandAcknowl-edgment/> (last accessed November 5, 2024)